

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031723\
 Data File : BM039035.D
 Acq On : 17 Mar 2023 15:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 01:45:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 05:25:34 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.00
2	1,4-Dioxane	0.586	0.548	6.5	111	0.00
3	Pyridine	1.624	1.605	1.2	114	0.00
4	n-Nitrosodimethylamine	0.661	0.653	1.2	117	0.00
5 S	2-Fluorophenol	1.316	1.284	2.4	115	0.00
6	Aniline	2.245	2.241	0.2	117	0.00
7 S	Phenol-d6	1.710	1.667	2.5	115	0.00
8	2-Chlorophenol	1.401	1.360	2.9	114	0.00
9	Benzaldehyde	0.814	0.746	8.4	118	0.00
10 C	Phenol	1.816	1.778	2.1	116	0.00
11	bis(2-Chloroethyl)ether	1.476	1.441	2.4	116	0.00
12	1,3-Dichlorobenzene	1.489	1.427	4.2	113	0.00
13 C	1,4-Dichlorobenzene	1.515	1.448	4.4	113	0.00
14	1,2-Dichlorobenzene	1.459	1.387	4.9	112	0.00
15	Benzyl Alcohol	1.226	1.253	-2.2	119	0.00
16	2,2'-oxybis(1-Chloropropane	1.864	1.902	-2.0	122	0.00
17	2-Methylphenol	1.247	1.241	0.5	117	0.00
18	Hexachloroethane	0.559	0.551	1.4	115	0.00
19 P	n-Nitroso-di-n-propylamine	1.077	1.135	-5.4	120	0.00
20	3+4-Methylphenols	1.666	1.655	0.7	116	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
22	Acetophenone	0.561	0.540	3.7	115	0.00
23 S	Nitrobenzene-d5	0.407	0.408	-0.2	118	0.00
24	Nitrobenzene	0.428	0.425	0.7	118	0.00
25	Isophorone	0.749	0.778	-3.9	121	0.00
26 C	2-Nitrophenol	0.169	0.189	-11.8	128	0.00
27	2,4-Dimethylphenol	0.326	0.329	-0.9	118	0.00
28	bis(2-Chloroethoxy)methane	0.481	0.479	0.4	119	0.00
29 C	2,4-Dichlorophenol	0.310	0.309	0.3	116	0.00
30	1,2,4-Trichlorobenzene	0.338	0.317	6.2	113	0.00
31	Naphthalene	1.140	1.086	4.7	114	0.00
32	Benzoic acid	0.216	0.256	-18.5	138	0.00
33	4-Chloroaniline	0.486	0.484	0.4	116	0.00
34 C	Hexachlorobutadiene	0.194	0.181	6.7	112	0.00
35	Caprolactam	0.094	0.100	-6.4	121	0.00
36 C	4-Chloro-3-methylphenol	0.330	0.338	-2.4	119	0.00
37	2-Methylnaphthalene	0.764	0.742	2.9	116	0.00
38	1-Methylnaphthalene	0.716	0.701	2.1	116	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
40	1,2,4,5-Tetrachlorobenzene	0.650	0.593	8.8	112	0.00
41 P	Hexachlorocyclopentadiene	0.357	0.363	-1.7	122	0.00
42 S	2,4,6-Tribromophenol	0.231	0.225	2.6	115	0.00
43 C	2,4,6-Trichlorophenol	0.421	0.417	1.0	117	0.00
44	2,4,5-Trichlorophenol	0.493	0.488	1.0	116	0.00
45 S	2-Fluorobiphenyl	1.536	1.440	6.3	114	0.00
46	1,1'-Biphenyl	1.694	1.589	6.2	115	0.00
47	2-Chloronaphthalene	1.278	1.190	6.9	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031723\
 Data File : BM039035.D
 Acq On : 17 Mar 2023 15:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 01:45:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 05:25:34 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.367	0.411	-12.0	124	0.00
49	Acenaphthylene	2.035	1.970	3.2	115	0.00
50	Dimethylphthalate	1.539	1.503	2.3	116	0.00
51	2,6-Dinitrotoluene	0.313	0.325	-3.8	118	0.00
52 C	Acenaphthene	1.251	1.178	5.8	113	0.00
53	3-Nitroaniline	0.363	0.381	-5.0	116	0.00
54 P	2,4-Dinitrophenol	0.169	0.191	-13.0	131	0.00
55	Dibenzofuran	1.957	1.838	6.1	113	0.00
56 P	4-Nitrophenol	0.306	0.306	0.0	115	0.00
57	2,4-Dinitrotoluene	0.401	0.433	-8.0	117	0.00
58	Fluorene	1.533	1.476	3.7	113	0.00
59	2,3,4,6-Tetrachlorophenol	0.380	0.374	1.6	114	0.00
60	Diethylphthalate	1.497	1.517	-1.3	117	0.00
61	4-Chlorophenyl-phenylether	0.753	0.720	4.4	114	0.00
62	4-Nitroaniline	0.362	0.390	-7.7	115	0.00
63	Azobenzene	1.584	1.626	-2.7	118	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
65	4,6-Dinitro-2-methylphenol	0.120	0.129	-7.5	125	0.00
66 c	n-Nitrosodiphenylamine	0.674	0.645	4.3	114	0.00
67	4-Bromophenyl-phenylether	0.216	0.209	3.2	116	0.00
68	Hexachlorobenzene	0.241	0.223	7.5	112	0.00
69	Atrazine	0.184	0.196	-6.5	119	0.00
70 C	Pentachlorophenol	0.178	0.174	2.2	116	0.00
71	Phenanthrene	1.174	1.109	5.5	113	0.00
72	Anthracene	1.171	1.128	3.7	112	0.00
73	Carbazole	1.068	1.038	2.8	111	0.00
74	Di-n-butylphthalate	1.187	1.347	-13.5	125	0.00
75 C	Fluoranthene	1.239	1.207	2.6	111	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
77	Benzidine	0.417	0.491	-17.7	117	0.00
78	Pyrene	1.502	1.486	1.1	108	0.00
79 S	Terphenyl-d14	1.094	1.088	0.5	106	0.00
80	Butylbenzylphthalate	0.558	0.684	-22.6	122	0.00
81	Benzo(a)anthracene	1.440	1.405	2.4	104	0.00
82	3,3'-Dichlorobenzidine	0.443	0.471	-6.3	113	0.00
83	Chrysene	1.401	1.336	4.6	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.819	1.015	-23.9	121	0.00
85 c	Di-n-octyl phthalate	1.414	1.693	-19.7	123	0.00
86 I	Perylene-d12	1.000	1.000	0.0	112	0.00
87	Indeno(1,2,3-cd)pyrene	1.522	1.456	4.3	99	0.00
88	Benzo(b)fluoranthene	1.278	1.237	3.2	101	0.00
89	Benzo(k)fluoranthene	1.330	1.277	4.0	101	0.00
90 C	Benzo(a)pyrene	1.230	1.208	1.8	101	0.00
91	Dibenzo(a,h)anthracene	1.289	1.235	4.2	99	0.00
92	Benzo(g,h,i)perylene	1.261	1.186	5.9	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031723\
Data File : BM039035.D
Acq On : 17 Mar 2023 15:28
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 18 01:45:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 17 05:25:34 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0