

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM031919\
 Data File : BM019233.D
 Acq On : 19 Mar 2019 12:51
 Operator : JU/SJ
 Sample : K1970-07
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P-7-WEST-WALL-SOUTH

Quant Time: Mar 19 13:42:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM031119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 19 12:12:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.64	152	268570	20.00	ng	0.00
21) Naphthalene-d8	10.43	136	1181871	20.00	ng	0.01
39) Acenaphthene-d10	14.32	164	901627	20.00	ng	0.02
64) Phenanthrene-d10	17.07	188	1766080	20.00	ng	0.02
76) Chrysene-d12	21.26	240	1306608	20.00	ng	0.00
87) Perylene-d12	23.48	264	1225122	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	2060665	124.26	ng	0.00
7) Phenol-d6	6.83	99	2987769	123.17	ng	0.00
23) Nitrobenzene-d5	8.82	82	1953138	78.12	ng	0.00
42) 2,4,6-Tribromophenol	15.83	330	1718042	129.06	ng	0.03
45) 2-Fluorobiphenyl	12.92	172	4571854	65.96	ng	0.01
79) Terphenyl-d14	19.70	244	6169241	93.73	ng	0.01
Target Compounds						
31) Naphthalene	10.48	128	794130	12.748	ng	# 74
37) 2-Methylnaphthalene	12.12	142	3541345	78.232	ng	# 100
38) 1-Methylnaphthalene	12.33	142	3418787	76.623	ng	# 100
50) Dimethylphthalate	13.77	163	524256	6.797	ng	# 89
52) Acenaphthene	14.38	154	583476	10.218	ng	# 86
58) Fluorene	15.37	166	1623034	21.382	ng	# 84
71) Phenanthrene	17.12	178	6908644	66.740	ng	# 96
78) Pyrene	19.50	202	1567092	18.956	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM031919\
Data File : BM019233.D
Acq On : 19 Mar 2019 12:51
Operator : JU/SJ
Sample : K1970-07
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P-7-WEST-WALL-SOUTH

Quant Time: Mar 19 13:42:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM031119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 19 12:12:53 2019
Response via : Initial Calibration

