

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032023\
 Data File : BM039043.D
 Acq On : 20 Mar 2023 14:07
 Operator : CG/JU
 Sample : PB151375BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB151375BL

Quant Time: Mar 21 04:07:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 05:25:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	374879	20.000	ng	0.00
21) Naphthalene-d8	10.786	136	1586691	20.000	ng	0.00
39) Acenaphthene-d10	14.616	164	967590	20.000	ng	0.00
64) Phenanthrene-d10	17.357	188	2115943	20.000	ng	0.00
76) Chrysene-d12	21.533	240	1759503	20.000	ng	0.00
86) Perylene-d12	23.968	264	1727025	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	3849315	156.003	ng	0.00
7) Phenol-d6	7.134	99	4958890	154.724	ng	0.00
23) Nitrobenzene-d5	9.145	82	2953067	91.421	ng	0.00
42) 2,4,6-Tribromophenol	16.104	330	1654700	147.827	ng	0.00
45) 2-Fluorobiphenyl	13.245	172	6270417	84.362	ng	0.00
79) Terphenyl-d14	19.980	244	9861481	102.472	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032023\
Data File : BM039043.D
Acq On : 20 Mar 2023 14:07
Operator : CG/JU
Sample : PB151375BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB151375BL

Quant Time: Mar 21 04:07:41 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 17 05:25:34 2023
Response via : Initial Calibration

