

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032122\
 Data File : BM034284.D
 Acq On : 21 Mar 2022 09:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 21 17:53:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 16 02:38:26 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	40.342	-0.9	106	0.00
3	Pyridine	40.000	40.129	-0.3	105	-0.01
4	n-Nitrosodimethylamine	40.000	39.700	0.7	106	0.00
5 S	2-Fluorophenol	80.000	81.336	-1.7	105	-0.01
6	Aniline	40.000	39.104	2.2	101	0.00
7 S	Phenol-d6	80.000	79.055	1.2	102	0.00
8	2-Chlorophenol	40.000	39.701	0.7	105	0.00
9	Benzaldehyde	40.000	36.941	7.6	101	0.00
10 C	Phenol	40.000	39.457	1.4	102	-0.01
11	bis(2-Chloroethyl)ether	40.000	38.721	3.2	102	0.00
12	1,3-Dichlorobenzene	40.000	39.122	2.2	104	-0.01
13 C	1,4-Dichlorobenzene	40.000	38.780	3.0	104	0.00
14	1,2-Dichlorobenzene	40.000	38.594	3.5	103	0.00
15	Benzyl Alcohol	40.000	39.234	1.9	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.534	3.7	103	-0.01
17	2-Methylphenol	40.000	39.354	1.6	102	0.00
18	Hexachloroethane	40.000	39.385	1.5	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.053	4.9	100	-0.01
20	3+4-Methylphenols	40.000	38.962	2.6	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	38.941	2.6	101	-0.01
23 S	Nitrobenzene-d5	80.000	79.664	0.4	101	0.00
24	Nitrobenzene	40.000	39.723	0.7	102	0.00
25	Isophorone	40.000	38.930	2.7	99	-0.01
26 C	2-Nitrophenol	40.000	40.438	-1.1	99	0.00
27	2,4-Dimethylphenol	40.000	39.786	0.5	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.390	1.5	100	-0.01
29 C	2,4-Dichlorophenol	40.000	40.323	-0.8	101	0.00
30	1,2,4-Trichlorobenzene	40.000	39.149	2.1	102	-0.01
31	Naphthalene	40.000	39.011	2.5	102	0.00
32	Benzoic acid	40.000	37.636	5.9	94	0.00
33	4-Chloroaniline	40.000	40.187	-0.5	100	0.00
34 C	Hexachlorobutadiene	40.000	39.265	1.8	102	-0.01
35	Caprolactam	40.000	39.374	1.6	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.866	0.3	100	0.00
37	2-Methylnaphthalene	40.000	39.330	1.7	102	0.00
38	1-Methylnaphthalene	40.000	39.295	1.8	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.784	3.0	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.594	1.0	99	0.00
42 S	2,4,6-Tribromophenol	80.000	80.500	-0.6	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.556	1.1	100	0.00
44	2,4,5-Trichlorophenol	40.000	39.806	0.5	100	0.00
45 S	2-Fluorobiphenyl	80.000	79.003	1.2	102	0.00
46	1,1'-Biphenyl	40.000	39.161	2.1	102	0.00
47	2-Chloronaphthalene	40.000	38.937	2.7	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032122\
 Data File : BM034284.D
 Acq On : 21 Mar 2022 09:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 21 17:53:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 16 02:38:26 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.874	-2.2	102	0.00
49	Acenaphthylene	40.000	39.883	0.3	102	0.00
50	Dimethylphthalate	40.000	38.960	2.6	101	0.00
51	2,6-Dinitrotoluene	40.000	40.620	-1.5	103	0.00
52 C	Acenaphthene	40.000	39.651	0.9	102	0.00
53	3-Nitroaniline	40.000	41.458	-3.6	102	0.00
54 P	2,4-Dinitrophenol	40.000	41.769	-4.4	101	0.00
55	Dibenzofuran	40.000	39.293	1.8	103	0.00
56 P	4-Nitrophenol	40.000	41.160	-2.9	103	0.00
57	2,4-Dinitrotoluene	40.000	41.189	-3.0	104	0.00
58	Fluorene	40.000	39.897	0.3	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.287	1.8	100	0.00
60	Diethylphthalate	40.000	39.866	0.3	102	0.00
61	4-Chlorophenyl-phenylether	40.000	39.358	1.6	103	0.00
62	4-Nitroaniline	40.000	43.153	-7.9	104	0.00
63	Azobenzene	40.000	40.749	-1.9	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.168	-2.9	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.829	2.9	103	0.00
67	4-Bromophenyl-phenylether	40.000	38.969	2.6	105	0.00
68	Hexachlorobenzene	40.000	38.878	2.8	104	0.00
69	Atrazine	40.000	41.562	-3.9	106	0.00
70 C	Pentachlorophenol	40.000	40.023	-0.1	103	0.00
71	Phenanthrene	40.000	39.963	0.1	108	0.00
72	Anthracene	40.000	40.350	-0.9	107	0.00
73	Carbazole	40.000	41.177	-2.9	109	0.00
74	Di-n-butylphthalate	40.000	41.636	-4.1	108	0.00
75 C	Fluoranthene	40.000	41.462	-3.7	112	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	130	0.00
77	Benzydine	40.000	44.156	-10.4	135	0.00
78	Pyrene	40.000	36.895	7.8	116	0.00
79 S	Terphenyl-d14	80.000	73.895	7.6	114	0.00
80	Butylbenzylphthalate	40.000	39.829	0.4	124	0.00
81	Benzo(a)anthracene	40.000	39.316	1.7	131	0.00
82	3,3'-Dichlorobenzidine	40.000	40.205	-0.5	133	0.00
83	Chrysene	40.000	38.519	3.7	129	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.247	-0.6	129	0.00
85 c	Di-n-octyl phthalate	40.000	42.049	-5.1	142	0.00
86 I	Perylene-d12	20.000	20.000	0.0	146	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.709	5.7	142	0.00
88	Benzo(b)fluoranthene	40.000	39.772	0.6	147	0.00
89	Benzo(k)fluoranthene	40.000	40.429	-1.1	145	0.00
90 C	Benzo(a)pyrene	40.000	39.476	1.3	146	0.00
91	Dibenzo(a,h)anthracene	40.000	38.263	4.3	146	0.00
92	Benzo(g,h,i)perylene	40.000	36.737	8.2	140	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032122\
Data File : BM034284.D
Acq On : 21 Mar 2022 09:56
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 21 17:53:43 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 16 02:38:26 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0