

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM032219\
 Data File : BM019341.D
 Acq On : 23 Mar 2019 08:25
 Operator : JU/SJ
 Sample : K1968-12
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C09C2

Manual Integrations
 APPROVED

Sohil
 3/25/2019 6:36:51 PM

Quant Time: Mar 24 00:02:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Mar 23 23:32:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	138622	20.00	ng/ul	0.00
18) Naphthalene-d8	10.39	136	589681	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.27	164	342463	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.03	188	763932	20.00	ng/ul	0.00
78) Chrysene-d12	21.23	240	831990	20.00	ng/ul	0.00
86) Perylene-d12	23.46	264	958223	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	16927	4.78	ng/uL	0.00
5) Phenol-d5	6.80	99	86505	7.01	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.96	67	285124	34.50	ng/ul	0.00
9) 2-Chlorophenol-d4	7.15	132	271152	28.98	ng/ul	0.00
13) 4-Methylphenol-d8	8.33	113	163885	17.69	ng/ul	0.00
19) Nitrobenzene-d5	8.78	128	162100	38.54	ng/ul	0.00
22) 2-Nitrophenol-d4	9.49	143	182619	39.01	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.03	165	286647	32.45	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul	
44) Dimethylphthalate-d6	13.69	166	1059020	38.32	ng/ul	0.00
47) Acenaphthylene-d8	13.96	160	1355666	39.23	ng/ul	0.00
52) 4-Nitrophenol-d4	14.54	143	16995m	3.92	ng/ul	0.03
58) Fluorene-d10	15.27	176	919889	38.64	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.41	200	153039	30.29	ng/ul	0.00
71) Anthracene-d10	17.13	188	1439019	39.25	ng/ul	0.00
79) Pyrene-d10	19.43	212	1581887	41.16	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.31	264	1845836	36.34	ng/ul	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM032219\
Data File : BM019341.D
Acq On : 23 Mar 2019 08:25
Operator : JU/SJ
Sample : K1968-12
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C09C2

Manual Integrations
APPROVED

Sohil
3/25/2019 6:36:51 PM

Quant Time: Mar 24 00:02:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Mar 23 23:32:22 2019
Response via : Initial Calibration

