

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM032219\
 Data File : BM019317.D
 Acq On : 22 Mar 2019 17:55
 Operator : JU/SJ
 Sample : PB118087BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK87

Quant Time: Mar 22 23:32:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 22 22:31:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	115250	20.00	ng/ul	0.00
18) Naphthalene-d8	10.40	136	467987	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.27	164	259422	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.03	188	568867	20.00	ng/ul	0.00
78) Chrysene-d12	21.24	240	577203	20.00	ng/ul	0.00
86) Perylene-d12	23.46	264	673868	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	16830	5.72	ng/uL	0.00
5) Phenol-d5	6.80	99	264359	25.76	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.97	67	180082	26.21	ng/ul	0.00
9) 2-Chlorophenol-d4	7.15	132	210856	27.11	ng/ul	0.00
13) 4-Methylphenol-d8	8.33	113	209830	27.24	ng/ul	0.00
19) Nitrobenzene-d5	8.79	128	95792	28.69	ng/ul	0.00
22) 2-Nitrophenol-d4	9.50	143	105378	28.36	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.03	165	174033	24.82	ng/ul	0.00
29) 4-Chloroaniline-d4	10.56	131	275341	32.70	ng/ul	0.00
44) Dimethylphthalate-d6	13.69	166	674859	32.24	ng/ul	0.00
47) Acenaphthylene-d8	13.97	160	790383	30.20	ng/ul	0.00
52) 4-Nitrophenol-d4	14.52	143	67380	20.54	ng/ul	0.01
58) Fluorene-d10	15.27	176	567448	31.46	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.42	200	80716	21.46	ng/ul	0.00
71) Anthracene-d10	17.13	188	873494	32.00	ng/ul	0.00
79) Pyrene-d10	19.44	212	941311	35.31	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.32	264	1096730	30.70	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM032219\
Data File : BM019317.D
Acq On : 22 Mar 2019 17:55
Operator : JU/SJ
Sample : PB118087BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK87

Quant Time: Mar 22 23:32:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 22 22:31:25 2019
Response via : Initial Calibration

