

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032320\
 Data File : BM025742.D
 Acq On : 24 Mar 2020 23:08
 Operator : CG/JU
 Sample : L1940-17
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0DC1

Quant Time: Mar 25 00:53:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 24 23:08:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	181888	20.00	ng/ul	0.00
18) Naphthalene-d8	10.33	136	727454	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.21	164	459648	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.96	188	996575	20.00	ng/ul	0.00
78) Chrysene-d12	21.14	240	1092804	20.00	ng/ul	-0.01
86) Perylene-d12	23.30	264	1236950	20.00	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.15	96	4568	1.05	ng/uL	0.00
5) Phenol-d5	6.76	99	87184	6.07	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.92	67	229999	28.07	ng/ul	0.00
9) 2-Chlorophenol-d4	7.11	132	270481	23.22	ng/ul	0.00
13) 4-Methylphenol-d8	8.27	113	136965	11.61	ng/ul	-0.01
19) Nitrobenzene-d5	8.71	128	166018	30.71	ng/ul	0.00
22) 2-Nitrophenol-d4	9.43	143	169218	29.44	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.96	165	311141	26.47	ng/ul	0.00
29) 4-Chloroaniline-d4	10.49	131	10823	0.79	ng/ul	0.00
44) Dimethylphthalate-d6	13.63	166	1109514	31.76	ng/ul	0.00
47) Acenaphthylene-d8	13.90	160	1096467	26.20	ng/ul	0.00
52) 4-Nitrophenol-d4	14.42	143	31860	4.79	ng/ul	0.00
58) Fluorene-d10	15.20	176	980345	31.97	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.33	200	132276	21.75	ng/ul	-0.01
71) Anthracene-d10	17.05	188	1597517	34.32	ng/ul	0.00
79) Pyrene-d10	19.35	212	1790800	34.12	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.16	264	2197990	34.71	ng/ul	-0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032320\
Data File : BM025742.D
Acq On : 24 Mar 2020 23:08
Operator : CG/JU
Sample : L1940-17
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0DC1

Quant Time: Mar 25 00:53:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 24 23:08:23 2020
Response via : Initial Calibration

