

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032320\
 Data File : BM025722.D
 Acq On : 24 Mar 2020 10:29
 Operator : CG/JU
 Sample : PB127564BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK64

Quant Time: Mar 24 11:26:56 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 24 00:45:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	126261	20.00	ng/ul	0.00
18) Naphthalene-d8	10.33	136	500495	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.21	164	309033	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.96	188	707863	20.00	ng/ul	0.00
78) Chrysene-d12	21.14	240	816773	20.00	ng/ul	-0.01
86) Perylene-d12	23.30	264	907025	20.00	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.14	96	21441	7.12	ng/uL	0.00
5) Phenol-d5	6.76	99	325968	32.71	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.92	67	198258	34.86	ng/ul	0.00
9) 2-Chlorophenol-d4	7.11	132	281431	34.80	ng/ul	0.00
13) 4-Methylphenol-d8	8.28	113	260575	31.83	ng/ul	0.00
19) Nitrobenzene-d5	8.72	128	136267	36.63	ng/ul	0.00
22) 2-Nitrophenol-d4	9.43	143	140397	35.50	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.96	165	267934	33.13	ng/ul	0.00
29) 4-Chloroaniline-d4	10.47	131	378314	40.29	ng/ul	0.00
44) Dimethylphthalate-d6	13.63	166	864355	36.80	ng/ul	0.00
47) Acenaphthylene-d8	13.90	160	1020285	36.26	ng/ul	0.00
52) 4-Nitrophenol-d4	14.42	143	135214	30.24	ng/ul	0.00
58) Fluorene-d10	15.21	176	762764	37.00	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.33	200	110492	25.58	ng/ul	0.00
71) Anthracene-d10	17.05	188	1204604	36.43	ng/ul	0.00
79) Pyrene-d10	19.35	212	1333309	33.99	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.16	264	1690437	36.40	ng/ul	-0.01

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032320\
Data File : BM025722.D
Acq On : 24 Mar 2020 10:29
Operator : CG/JU
Sample : PB127564BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK64

Quant Time: Mar 24 11:26:56 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 24 00:45:06 2020
Response via : Initial Calibration

