

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032320\
 Data File : BM025774.D
 Acq On : 25 Mar 2020 18:55
 Operator : CG/JU
 Sample : PB127603BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK03

Quant Time: Mar 26 00:42:14 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 24 00:45:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	195782	20.00	ng/ul	0.00
18) Naphthalene-d8	10.33	136	790071	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.20	164	496501	20.00	ng/ul	-0.01
62) Phenanthrene-d10	16.95	188	1071807	20.00	ng/ul	-0.01
78) Chrysene-d12	21.14	240	1166709	20.00	ng/ul	-0.01
86) Perylene-d12	23.30	264	1300300	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.14	96	21074	4.52	ng/uL	0.00
5) Phenol-d5	6.75	99	363628	23.53	ng/ul	-0.01
7) Bis-(2-Chloroethyl)ether-d	6.91	67	213310	24.19	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.10	132	311906	24.87	ng/ul	-0.01
13) 4-Methylphenol-d8	8.27	113	300887	23.70	ng/ul	-0.01
19) Nitrobenzene-d5	8.71	128	155550	26.49	ng/ul	0.00
22) 2-Nitrophenol-d4	9.43	143	167174	26.78	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.96	165	303784	23.80	ng/ul	0.00
29) 4-Chloroaniline-d4	10.47	131	434461	29.31	ng/ul	-0.01
44) Dimethylphthalate-d6	13.62	166	943593	25.01	ng/ul	-0.01
47) Acenaphthylene-d8	13.89	160	1158882	25.63	ng/ul	-0.01
52) 4-Nitrophenol-d4	14.41	143	150612	20.96	ng/ul	-0.01
58) Fluorene-d10	15.20	176	831764	25.11	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.33	200	122114	18.67	ng/ul	-0.01
71) Anthracene-d10	17.05	188	1303326	26.03	ng/ul	0.00
79) Pyrene-d10	19.35	212	1406497	25.10	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.16	264	1724643	25.91	ng/ul	-0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032320\
Data File : BM025774.D
Acq On : 25 Mar 2020 18:55
Operator : CG/JU
Sample : PB127603BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK03

Quant Time: Mar 26 00:42:14 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 24 00:45:06 2020
Response via : Initial Calibration

