

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032321\
 Data File : BM029185.D
 Acq On : 22 Mar 2021 19:40
 Operator : CG/JU
 Sample : SST00.101
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST00.1001

Quant Time: Mar 23 02:18:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 02:06:52 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	2091	0.40	ng/ul	0.00
4) Naphthalene-d8	10.55	136	6446	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.39	164	3228	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.13	188	6915	0.40	ng/ul	0.00
17) Chrysene-d12	21.29	240	6454	0.40	ng/ul	0.00
23) Perylene-d12	23.53	264	6643	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/ul	
6) 2-Methylnaphthalene-d10	12.14	152	975	0.10	ng/ul	0.00
18) Fluoranthene-d10	19.15	212	1780	0.09	ng/ul	0.00
Target Compounds				Ovalue		
5) Naphthalene	10.60	128	1972	0.110	ng/ul#	94
7) 2-Methylnaphthalene	12.21	142	1293	0.106	ng/ul	97
8) 1-Methylnaphthalene	12.43	142	1227	0.105	ng/ul	98
10) Acenaphthylene	14.11	152	1468	0.102	ng/ul#	91
11) Acenaphthene	14.45	153	1278	0.114	ng/ul	98
12) Fluorene	15.44	166	1410	0.110	ng/ul	97
15) Phenanthrene	17.17	178	2336	0.110	ng/ul	95
16) Anthracene	17.26	178	1927	0.099	ng/ul	94
19) Fluoranthene	19.18	202	2396	0.099	ng/ul#	93
20) Pyrene	19.54	202	2434	0.099	ng/ul#	94
21) Benzo(a)anthracene	21.28	228	2062	0.090	ng/ul	95
22) Chrysene	21.33	228	2373	0.100	ng/ul	94
24) Benzo(b)fluoranthene	22.85	252	2520	0.096	ng/ul	75
25) Benzo(k)fluoranthene	22.90	252	2606	0.101	ng/ul#	73
26) Benzo(a)pyrene	23.43	252	2049	0.091	ng/ul#	65
27) Indeno(1,2,3-cd)pyrene	25.79	276	3136	0.105	ng/ul#	96
28) Dibenzo(a,h)anthracene	25.80	278	2600	0.104	ng/ul#	80
29) Benzo(a,h,i)perylene	26.47	276	2555	0.102	ng/ul#	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032321\
Data File : BM029185.D
Acq On : 22 Mar 2021 19:40
Operator : CG/JU
Sample : SST0.101
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.1001

Quant Time: Mar 23 02:18:06 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 02:06:52 2021
Response via : Initial Calibration

