

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032321\
 Data File : BM029186.D
 Acq On : 22 Mar 2021 20:17
 Operator : CG/JU
 Sample : SST00.202
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST00.2002

Quant Time: Mar 23 02:12:22 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 02:06:52 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	1896	0.40	ng/ul	0.00
4) Naphthalene-d8	10.55	136	5980	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.39	164	3157	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.12	188	6962	0.40	ng/ul	0.00
17) Chrysene-d12	21.29	240	7044	0.40	ng/ul	0.00
23) Perylene-d12	23.52	264	7353	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.30	96	345	0.15	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.14	152	1887	0.21	ng/ul	0.00
18) Fluoranthene-d10	19.15	212	3891	0.19	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.33	88	352	0.159	ng/ul#	91
5) Naphthalene	10.60	128	3759	0.227	ng/ul	98
7) 2-Methylnaphthalene	12.21	142	2469	0.218	ng/ul	99
8) 1-Methylnaphthalene	12.43	142	2363	0.219	ng/ul	98
10) Acenaphthylene	14.11	152	2872	0.205	ng/ul	97
11) Acenaphthene	14.45	153	2548	0.231	ng/ul	99
12) Fluorene	15.44	166	2855	0.228	ng/ul	100
14) Pentachlorophenol	16.78	266	396	0.207	ng/ul	98
15) Phenanthrene	17.17	178	4726	0.222	ng/ul	98
16) Anthracene	17.26	178	4014	0.204	ng/ul	98
19) Fluoranthene	19.18	202	5044	0.192	ng/ul	98
20) Pyrene	19.54	202	5063	0.188	ng/ul	99
21) Benzo(a)anthracene	21.28	228	4471	0.178	ng/ul	98
22) Chrysene	21.33	228	5139	0.198	ng/ul	99
24) Benzo(b)fluoranthene	22.85	252	5618	0.194	ng/ul	93
25) Benzo(k)fluoranthene	22.90	252	5764	0.202	ng/ul#	94
26) Benzo(a)pyrene	23.43	252	4502	0.180	ng/ul#	88
27) Indeno(1,2,3-cd)pyrene	25.79	276	7275	0.221	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.80	278	6122	0.222	ng/ul	95
29) Benzo(g,h,i)perylene	26.47	276	5802	0.209	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032321\
Data File : BM029186.D
Acq On : 22 Mar 2021 20:17
Operator : CG/JU
Sample : SST0.202
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.2002

Quant Time: Mar 23 02:12:22 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 02:06:52 2021
Response via : Initial Calibration

