

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032321\
 Data File : BM029189.D
 Acq On : 22 Mar 2021 22:06
 Operator : CG/JU
 Sample : SSTD1.605
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD1.6005

Quant Time: Mar 23 02:12:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 02:06:52 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	2185	0.40	ng/ul	0.00
4) Naphthalene-d8	10.55	136	6817	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.39	164	3486	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.12	188	7149	0.40	ng/ul	0.00
17) Chrysene-d12	21.29	240	7638	0.40	ng/ul	0.00
23) Perylene-d12	23.53	264	7242	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.30	96	2787	1.07	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.14	152	14732	1.41	ng/ul	0.00
18) Fluoranthene-d10	19.15	212	27732	1.25	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.33	88	3127	1.227	ng/ul#	82
5) Naphthalene	10.60	128	29212	1.548	ng/ul	97
7) 2-Methylnaphthalene	12.21	142	19329	1.496	ng/ul	100
8) 1-Methylnaphthalene	12.43	142	18563	1.506	ng/ul	100
10) Acenaphthylene	14.11	152	22982	1.485	ng/ul	98
11) Acenaphthene	14.46	153	19405	1.596	ng/ul	99
12) Fluorene	15.44	166	21670	1.566	ng/ul	98
14) Pentachlorophenol	16.77	266	2815	1.436	ng/ul	98
15) Phenanthrene	17.17	178	33447	1.529	ng/ul	99
16) Anthracene	17.26	178	29862	1.479	ng/ul	98
19) Fluoranthene	19.18	202	36776	1.289	ng/ul	100
20) Pyrene	19.54	202	36170	1.240	ng/ul	100
21) Benzo(a)anthracene	21.28	228	33351	1.227	ng/ul	99
22) Chrysene	21.33	228	36605	1.304	ng/ul	99
24) Benzo(b)fluoranthene	22.85	252	38740	1.355	ng/ul	93
25) Benzo(k)fluoranthene	22.90	252	40293	1.432	ng/ul#	92
26) Benzo(a)pyrene	23.43	252	32103	1.306	ng/ul#	89
27) Indeno(1,2,3-cd)pyrene	25.78	276	49521	1.525	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.80	278	41961	1.546	ng/ul	97
29) Benzo(a,h,i)perylene	26.47	276	40285	1.473	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032321\
Data File : BM029189.D
Acq On : 22 Mar 2021 22:06
Operator : CG/JU
Sample : SSTD1.605
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD1.6005

Quant Time: Mar 23 02:12:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 02:06:52 2021
Response via : Initial Calibration

