

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032322\
 Data File : BM034340.D
 Acq On : 23 Mar 2022 15:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 02:42:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 02:40:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	133	0.00
2	1,4-Dioxane	40.000	41.029	-2.6	142	0.00
3	Pyridine	40.000	41.364	-3.4	142	0.00
4	n-Nitrosodimethylamine	40.000	40.315	-0.8	142	0.00
5 S	2-Fluorophenol	80.000	85.003	-6.3	144	0.00
6	Aniline	40.000	42.175	-5.4	144	0.00
7 S	Phenol-d6	80.000	85.755	-7.2	145	0.00
8	2-Chlorophenol	40.000	41.474	-3.7	144	0.00
9	Benzaldehyde	40.000	42.346	-5.9	152	0.00
10 C	Phenol	40.000	42.178	-5.4	144	0.00
11	bis(2-Chloroethyl)ether	40.000	41.037	-2.6	142	0.00
12	1,3-Dichlorobenzene	40.000	39.581	1.0	138	0.00
13 C	1,4-Dichlorobenzene	40.000	39.573	1.1	140	0.00
14	1,2-Dichlorobenzene	40.000	39.962	0.1	140	0.00
15	Benzyl Alcohol	40.000	43.100	-7.8	144	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.431	-6.1	150	0.00
17	2-Methylphenol	40.000	42.044	-5.1	144	0.00
18	Hexachloroethane	40.000	40.192	-0.5	140	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.389	-6.0	146	0.00
20	3+4-Methylphenols	40.000	42.252	-5.6	144	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	138	0.00
22	Acetophenone	40.000	40.866	-2.2	145	0.00
23 S	Nitrobenzene-d5	80.000	82.692	-3.4	144	0.00
24	Nitrobenzene	40.000	40.910	-2.3	144	0.00
25	Isophorone	40.000	41.634	-4.1	145	0.00
26 C	2-Nitrophenol	40.000	43.550	-8.9	147	0.00
27	2,4-Dimethylphenol	40.000	41.066	-2.7	144	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.887	-2.2	142	0.00
29 C	2,4-Dichlorophenol	40.000	41.076	-2.7	141	0.00
30	1,2,4-Trichlorobenzene	40.000	39.081	2.3	139	0.00
31	Naphthalene	40.000	39.984	0.0	143	0.00
32	Benzoic acid	40.000	41.519	-3.8	142	0.00
33	4-Chloroaniline	40.000	42.000	-5.0	144	0.00
34 C	Hexachlorobutadiene	40.000	39.480	1.3	141	0.00
35	Caprolactam	40.000	42.765	-6.9	148	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.810	-4.5	144	0.00
37	2-Methylnaphthalene	40.000	40.439	-1.1	144	0.00
38	1-Methylnaphthalene	40.000	40.397	-1.0	144	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	140	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.432	1.4	142	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.166	-5.4	147	0.00
42 S	2,4,6-Tribromophenol	80.000	85.041	-6.3	153	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.695	-1.7	143	0.00
44	2,4,5-Trichlorophenol	40.000	41.060	-2.7	143	0.00
45 S	2-Fluorobiphenyl	80.000	80.171	-0.2	144	0.00
46	1,1'-Biphenyl	40.000	39.838	0.4	145	0.00
47	2-Chloronaphthalene	40.000	39.715	0.7	143	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032322\
 Data File : BM034340.D
 Acq On : 23 Mar 2022 15:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 02:42:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 02:40:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.657	-11.6	154	0.00
49	Acenaphthylene	40.000	40.694	-1.7	145	0.00
50	Dimethylphthalate	40.000	40.322	-0.8	145	0.00
51	2,6-Dinitrotoluene	40.000	41.226	-3.1	145	0.00
52 C	Acenaphthene	40.000	40.784	-2.0	146	0.00
53	3-Nitroaniline	40.000	42.980	-7.4	147	0.00
54 P	2,4-Dinitrophenol	40.000	45.677	-14.2	153	0.00
55	Dibenzofuran	40.000	39.862	0.3	145	0.00
56 P	4-Nitrophenol	40.000	46.585	-16.5	162	0.00
57	2,4-Dinitrotoluene	40.000	42.773	-6.9	150	0.00
58	Fluorene	40.000	40.887	-2.2	147	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.792	-2.0	144	0.00
60	Diethylphthalate	40.000	41.310	-3.3	148	0.00
61	4-Chlorophenyl-phenylether	40.000	40.667	-1.7	148	0.00
62	4-Nitroaniline	40.000	45.982	-15.0	154	0.00
63	Azobenzene	40.000	42.024	-5.1	150	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	144	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.824	-9.6	154	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.154	-0.4	147	0.00
67	4-Bromophenyl-phenylether	40.000	40.755	-1.9	151	0.00
68	Hexachlorobenzene	40.000	39.567	1.1	147	0.00
69	Atrazine	40.000	41.605	-4.0	146	0.00
70 C	Pentachlorophenol	40.000	41.314	-3.3	146	0.00
71	Phenanthrene	40.000	40.396	-1.0	150	0.00
72	Anthracene	40.000	40.472	-1.2	148	0.00
73	Carbazole	40.000	41.409	-3.5	151	0.00
74	Di-n-butylphthalate	40.000	42.443	-6.1	152	0.00
75 C	Fluoranthene	40.000	40.034	-0.1	149	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	148	0.00
77	Benzydine	40.000	48.041	-20.1	168	0.00
78	Pyrene	40.000	41.929	-4.8	151	0.00
79 S	Terphenyl-d14	80.000	84.118	-5.1	148	0.00
80	Butylbenzylphthalate	40.000	45.136	-12.8	161	0.00
81	Benzo(a)anthracene	40.000	40.979	-2.4	156	0.00
82	3,3'-Dichlorobenzidine	40.000	41.593	-4.0	158	0.00
83	Chrysene	40.000	38.991	2.5	150	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.386	-11.0	163	0.00
85 c	Di-n-octyl phthalate	40.000	41.805	-4.5	162	0.00
86 I	Perylene-d12	20.000	20.000	0.0	143	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.606	1.0	146	0.00
88	Benzo(b)fluoranthene	40.000	41.617	-4.0	150	0.00
89	Benzo(k)fluoranthene	40.000	42.155	-5.4	148	0.00
90 C	Benzo(a)pyrene	40.000	40.689	-1.7	147	0.00
91	Dibenzo(a,h)anthracene	40.000	40.299	-0.7	150	0.00
92	Benzo(g,h,i)perylene	40.000	39.123	2.2	145	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032322\
Data File : BM034340.D
Acq On : 23 Mar 2022 15:54
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 24 02:42:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 02:40:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0