

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032519\
 Data File : BM019435.D
 Acq On : 25 Mar 2019 17:25
 Operator : JU/SJ
 Sample : PB118220BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK20

Quant Time: Mar 26 05:40:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 25 03:11:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	141502	20.00	ng/ul	0.00
18) Naphthalene-d8	10.38	136	600769	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.26	164	337895	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.02	188	733289	20.00	ng/ul	0.00
78) Chrysene-d12	21.23	240	740747	20.00	ng/ul	0.00
86) Perylene-d12	23.44	264	807940	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	20347	5.63	ng/uL	0.00
5) Phenol-d5	6.78	99	384264	30.49	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.96	67	241063	28.58	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	307598	32.21	ng/ul	0.00
13) 4-Methylphenol-d8	8.32	113	307908	32.55	ng/ul	0.00
19) Nitrobenzene-d5	8.77	128	140535	32.79	ng/ul	0.00
22) 2-Nitrophenol-d4	9.49	143	167926	35.21	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.02	165	282485	31.39	ng/ul	0.00
29) 4-Chloroaniline-d4	10.55	131	410565	37.99	ng/ul	0.00
44) Dimethylphthalate-d6	13.68	166	991719	36.37	ng/ul	0.00
47) Acenaphthylene-d8	13.95	160	1222330	35.85	ng/ul	0.00
52) 4-Nitrophenol-d4	14.50	143	123523	28.90	ng/ul	0.00
58) Fluorene-d10	15.26	176	853031	36.31	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.40	200	127212	26.23	ng/ul	0.00
71) Anthracene-d10	17.12	188	1332773	37.87	ng/ul	0.00
79) Pyrene-d10	19.43	212	1419966	41.50	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.30	264	1566078	36.56	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032519\
Data File : BM019435.D
Acq On : 25 Mar 2019 17:25
Operator : JU/SJ
Sample : PB118220BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK20

Quant Time: Mar 26 05:40:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 25 03:11:42 2019
Response via : Initial Calibration

