

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032620\
 Data File : BM025828.D
 Acq On : 27 Mar 2020 08:44
 Operator : CG/JU
 Sample : L2003-02
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0EK3

Quant Time: Mar 27 11:45:05 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 26 18:34:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	203722	20.00	ng/ul	0.00
18) Naphthalene-d8	10.33	136	821395	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.20	164	508366	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.95	188	1110255	20.00	ng/ul	0.00
78) Chrysene-d12	21.14	240	1216484	20.00	ng/ul	0.00
86) Perylene-d12	23.29	264	1389258	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.14	96	6044	1.24	ng/uL	0.00
5) Phenol-d5	6.75	99	82482	5.13	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.91	67	228810	24.93	ng/ul	0.00
9) 2-Chlorophenol-d4	7.10	132	266059	20.39	ng/ul	0.00
13) 4-Methylphenol-d8	8.27	113	169738	12.85	ng/ul	0.00
19) Nitrobenzene-d5	8.70	128	170102	27.86	ng/ul	0.00
22) 2-Nitrophenol-d4	9.42	143	179695	27.69	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.96	165	324103	24.42	ng/ul	0.00
29) 4-Chloroaniline-d4	10.47	131	21395	1.39	ng/ul	0.00
44) Dimethylphthalate-d6	13.62	166	1246792	32.27	ng/ul	0.00
47) Acenaphthylene-d8	13.89	160	1415212	30.57	ng/ul	0.00
52) 4-Nitrophenol-d4	14.41	143	39385	5.35	ng/ul	0.00
58) Fluorene-d10	15.20	176	1055738	31.13	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.33	200	186557	27.54	ng/ul	0.00
71) Anthracene-d10	17.04	188	1786892	34.46	ng/ul	0.00
79) Pyrene-d10	19.34	212	2006151	34.33	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.16	264	2540245	35.71	ng/ul	0.00

Target Compounds

					Ovalue	
20) Nitrobenzene	8.75	77	762688	51.279	ng/ul	99
23) 2-Nitrophenol	9.45	139	7711	1.058	ng/ul	91
45) Dimethylphthalate	13.67	163	206756	5.141	ng/ul	99
73) 1,2,3,4-Tetrachlorobenzene	12.99	216	91371	5.371	ng/uL	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032620\
Data File : BM025828.D
Acq On : 27 Mar 2020 08:44
Operator : CG/JU
Sample : L2003-02
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0EK3

Quant Time: Mar 27 11:45:05 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 26 18:34:50 2020
Response via : Initial Calibration

