

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032620\
 Data File : BM025824.D
 Acq On : 27 Mar 2020 06:20
 Operator : CG/JU
 Sample : PB127720BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK20

Quant Time: Mar 27 08:26:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 26 18:34:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	191074	20.00	ng/ul	0.00
18) Naphthalene-d8	10.33	136	773314	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.20	164	493276	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.95	188	1097849	20.00	ng/ul	0.00
78) Chrysene-d12	21.14	240	1227576	20.00	ng/ul	0.00
86) Perylene-d12	23.29	264	1372790	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.13	96	27254	5.98	ng/uL	0.00
5) Phenol-d5	6.75	99	444076	29.45	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.91	67	266817	31.00	ng/ul	0.00
9) 2-Chlorophenol-d4	7.10	132	372558	30.44	ng/ul	0.00
13) 4-Methylphenol-d8	8.27	113	370825	29.93	ng/ul	0.00
19) Nitrobenzene-d5	8.70	128	186004	32.36	ng/ul	0.00
22) 2-Nitrophenol-d4	9.42	143	200838	32.87	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.96	165	383193	30.67	ng/ul	0.00
29) 4-Chloroaniline-d4	10.47	131	546857	37.69	ng/ul	0.00
44) Dimethylphthalate-d6	13.62	166	1234985	32.94	ng/ul	0.00
47) Acenaphthylene-d8	13.89	160	1479298	32.94	ng/ul	0.00
52) 4-Nitrophenol-d4	14.41	143	202213	28.33	ng/ul	0.00
58) Fluorene-d10	15.20	176	1095374	33.29	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.33	200	165149	24.66	ng/ul	0.00
71) Anthracene-d10	17.05	188	1728068	33.70	ng/ul	0.00
79) Pyrene-d10	19.34	212	1905275	32.31	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.16	264	2382415	33.90	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM032620\
Data File : BM025824.D
Acq On : 27 Mar 2020 06:20
Operator : CG/JU
Sample : PB127720BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK20

Quant Time: Mar 27 08:26:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM031420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 26 18:34:50 2020
Response via : Initial Calibration

