

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM032819\
 Data File : BM019576.D
 Acq On : 29 Mar 2019 12:31
 Operator : JU/SJ
 Sample : K2085-02
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BF7W9

Quant Time: Mar 29 14:09:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 29 09:53:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	103487	20.00	ng/ul	0.00
18) Naphthalene-d8	10.36	136	472411	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.24	164	276493	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.00	188	618743	20.00	ng/ul	0.00
78) Chrysene-d12	21.20	240	758866	20.00	ng/ul	0.00
86) Perylene-d12	23.41	264	899677	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.13	96	16510	6.25	ng/uL	0.00
5) Phenol-d5	6.76	99	151434	16.43	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.93	67	108456	17.58	ng/ul	0.00
9) 2-Chlorophenol-d4	7.12	132	124979	17.89	ng/ul	0.00
13) 4-Methylphenol-d8	8.30	113	103314	14.94	ng/ul	0.00
19) Nitrobenzene-d5	8.75	128	57937	17.19	ng/ul	0.00
22) 2-Nitrophenol-d4	9.46	143	65486	17.46	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.00	165	102403	14.47	ng/ul	0.00
29) 4-Chloroaniline-d4	10.54	131	119854	14.10	ng/ul	0.02
44) Dimethylphthalate-d6	13.66	166	427271	19.15	ng/ul	0.00
47) Acenaphthylene-d8	13.93	160	474212	17.00	ng/ul	0.00
52) 4-Nitrophenol-d4	14.50	143	32739	9.36	ng/ul	0.02
58) Fluorene-d10	15.24	176	342223	17.80	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.39	200	46400	11.34	ng/ul	0.00
71) Anthracene-d10	17.10	188	531228	17.89	ng/ul	0.00
79) Pyrene-d10	19.40	212	584380	16.67	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.27	264	750833	15.74	ng/ul	0.00

Target Compounds					Qvalue
45) Dimethylphthalate	13.71	163	70215	3.137 ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM032819\
Data File : BM019576.D
Acq On : 29 Mar 2019 12:31
Operator : JU/SJ
Sample : K2085-02
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleID :
BF7W9

Quant Time: Mar 29 14:09:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 29 09:53:26 2019
Response via : Initial Calibration

