


Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM032819\
Data File : BM019537.D
Acq On : 28 Mar 2019 10:52
Operator : JU/SJ
Sample : DFTPP57
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP57

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
Title : SVOA CALIBRATION
Last Update : Fri Mar 29 16:04:25 2019



AutoFind: Scans 2454, 2455, 2456; Background Corrected with Scan 2449

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.3	43314	PASS
68	69	0.00	2	1.9	1253	PASS
69	198	0.00	100	53.0	66981	PASS
70	69	0.00	2	0.5	360	PASS
127	198	10	80	52.5	66392	PASS
197	198	0.00	2	0.6	744	PASS
198	198	100	100	100.0	126373	PASS
199	198	5	9	6.7	8522	PASS
275	198	10	60	29.5	37250	PASS
365	198	1	100	4.9	6130	PASS
441	198	0.01	100	14.0	17657	PASS
442	198	50	100	88.8	112221	PASS
443	442	15	24	19.3	21687	PASS

m/z	Abundance
37.00	405.0
37.90	707.0
39.00	3852.0
40.00	815.0
49.10	379.0
50.10	11931.0
51.00	42144.0
52.00	2166.0
54.90	432.0
56.00	1412.0
57.00	3087.0
61.00	699.0
62.00	771.0
63.00	1949.0
64.00	335.0
65.10	827.0
68.10	1176.0
69.00	64096.0
69.90	309.0
73.10	1072.0
74.00	6637.0
75.00	9618.0
76.00	2941.0
77.10	68824.0
78.00	5512.0
79.00	4777.0
80.00	3735.0
81.00	4772.0
82.00	1216.0
82.90	1403.0
85.00	1144.0
86.00	1432.0
86.90	592.0
87.90	320.0
90.90	1122.0
91.90	1582.0
93.00	8107.0
94.00	609.0
96.00	370.0
98.00	5970.0
99.00	5264.0
100.10	435.0
101.00	2361.0
103.00	755.0
104.00	1426.0
105.00	1721.0
106.00	644.0
107.00	19016.0
108.00	2767.0
109.10	421.0
110.00	30976.0
111.00	5077.0
111.90	730.0
116.00	1190.0
117.00	17760.0
117.90	1348.0
122.00	1381.0
123.00	2023.0
124.00	951.0
125.00	864.0
127.00	60640.0
128.00	4316.0
129.00	22664.0
130.00	2361.0
131.00	436.0
134.10	506.0
135.00	1898.0
136.00	868.0
137.00	1293.0
140.10	301.0
141.00	3044.0
142.00	983.0
142.90	785.0
143.80	300.0
146.00	615.0
147.00	1784.0
148.00	4067.0
148.90	1195.0
150.80	392.0
152.90	875.0
153.90	751.0
155.00	1444.0
156.00	2864.0
157.00	729.0
157.90	509.0
158.80	655.0
159.90	1139.0
160.90	1402.0
162.00	412.0
165.00	1281.0
165.90	697.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

167.00	7019.0
167.90	4765.0
168.90	805.0
171.00	393.0
171.90	767.0
173.00	726.0
174.00	1289.0
175.00	2544.0
176.00	767.0
177.00	1432.0
177.90	310.0
178.90	4464.0
179.90	3229.0
180.90	1348.0
182.00	372.0
183.80	371.0
185.00	2309.0
186.00	17328.0
187.00	4896.0
188.00	755.0
189.00	1102.0
190.90	847.0
192.00	1393.0
192.90	1822.0
196.00	3702.0
197.00	825.0
197.90	107680.0
198.90	7280.0
199.90	534.0
201.30	481.0
203.00	1111.0
204.00	4030.0
205.00	7573.0
206.00	29104.0
207.00	5362.0
208.00	1460.0
208.90	527.0
210.00	754.0
210.90	1148.0
216.00	750.0
216.90	8348.0
218.00	1070.0
221.00	4510.0
221.80	1384.0
223.00	1908.0
224.00	15701.0
225.00	4150.0
226.00	388.0
227.00	7055.0
227.90	980.0
229.00	1549.0
231.00	681.0
234.10	324.0
234.90	491.0
237.00	1104.0
238.90	381.0
241.00	424.0
241.90	981.0
243.00	928.0
244.00	11928.0
245.00	1748.0
246.00	2570.0
246.90	571.0
248.90	569.0
250.80	301.0
254.10	611.0
255.00	65000.0
256.00	9511.0
256.90	1166.0
258.00	4127.0
258.90	538.0
265.00	2012.0
265.90	344.0
272.90	2118.0
274.00	4805.0
275.00	28776.0
276.00	3790.0
277.00	2645.0
277.90	644.0
281.00	366.0
283.00	374.0
285.00	435.0
292.90	593.0
294.10	338.0
296.00	9171.0
296.90	1339.0
303.00	928.0
314.00	386.0
315.00	749.0
315.90	709.0
323.00	2819.0
323.90	485.0
326.90	601.0
328.00	331.0
334.00	1684.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

334.90	371.0
341.00	362.0
345.90	556.0
351.90	798.0
353.00	630.0
354.00	653.0
355.00	321.0
365.00	4520.0
366.00	723.0
372.00	1473.0
382.90	557.0
389.80	300.0
401.90	569.0
403.00	1089.0
404.00	425.0
421.00	713.0
422.00	628.0
423.00	4443.0
424.10	1814.0
425.00	324.0
441.00	11243.0
442.00	72312.0
443.00	13983.0
444.00	1392.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
38.00	828.0
39.00	4337.0
40.00	883.0
49.00	450.0
50.00	13393.0
51.10	51688.0
52.00	2533.0
54.90	356.0
56.00	1539.0
57.00	4148.0
61.00	842.0
62.00	1090.0
63.00	3105.0
64.00	397.0
65.00	1123.0
68.00	1667.0
69.00	79712.0
69.90	361.0
73.00	1533.0
74.00	9399.0
75.00	13230.0
76.00	3968.0
77.00	87704.0
78.10	6735.0
79.00	6766.0
80.00	4719.0
81.00	6511.0
82.00	1604.0
83.00	1291.0
84.00	348.0
84.90	1352.0
86.00	1920.0
86.90	713.0
88.00	351.0
90.90	1448.0
92.00	1772.0
93.00	9938.0
94.00	868.0
95.80	461.0
98.00	7929.0
99.00	6542.0
100.00	625.0
100.90	2811.0
102.90	959.0
104.00	2197.0
105.00	2106.0
105.90	553.0
107.00	24992.0
108.00	3485.0
109.00	802.0
110.00	41760.0
111.00	6369.0
112.00	1025.0
113.00	474.0
116.00	1223.0
117.00	22760.0
117.90	1845.0
122.00	1874.0
123.00	2328.0
124.00	1062.0
124.90	1049.0
126.10	319.0
127.00	78216.0
128.00	6474.0
129.00	32520.0
130.00	2563.0
130.90	612.0
133.90	1082.0
135.00	3026.0
135.90	1244.0
137.00	1669.0
138.00	348.0
140.00	531.0
140.90	3932.0
142.00	1233.0
142.90	995.0
144.00	444.0
145.90	937.0
146.90	2568.0
148.00	5438.0
149.00	1207.0
151.10	733.0
151.70	346.0
153.00	1310.0
153.90	857.0
155.00	2430.0
156.00	3740.0
156.90	833.0
157.90	950.0
158.80	627.0
159.90	1436.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

161.00	2345.0
161.90	634.0
165.00	1835.0
166.00	871.0
167.00	9148.0
168.00	5865.0
169.00	890.0
170.80	451.0
171.90	856.0
173.00	1043.0
174.00	1875.0
175.00	3759.0
175.90	923.0
177.00	1616.0
178.00	573.0
178.90	6712.0
180.00	4489.0
180.90	1849.0
183.90	635.0
185.00	3001.0
186.00	22904.0
187.00	6825.0
187.90	920.0
188.90	1432.0
190.90	1005.0
192.00	1918.0
192.90	2348.0
194.00	600.0
195.10	337.0
196.00	4434.0
197.00	735.0
197.90	148032.0
199.00	9798.0
200.00	863.0
201.40	741.0
203.00	1387.0
204.00	6154.0
205.00	10616.0
206.00	42760.0
207.00	6417.0
208.00	1554.0
209.00	747.0
210.00	847.0
211.00	1524.0
215.00	422.0
216.00	943.0
216.90	11620.0
218.00	1592.0
221.00	6145.0
221.90	1388.0
223.00	2853.0
224.00	23472.0
225.00	6017.0
226.00	1044.0
227.00	10518.0
228.00	1594.0
228.90	2256.0
229.80	405.0
230.90	951.0
233.90	829.0
235.00	705.0
235.90	631.0
236.90	1284.0
238.90	538.0
240.10	364.0
240.90	657.0
242.00	1460.0
242.90	1334.0
244.00	18904.0
245.00	2504.0
246.00	3963.0
246.90	776.0
249.00	616.0
253.00	515.0
254.20	903.0
255.00	99040.0
256.00	14000.0
256.90	1221.0
258.00	6084.0
259.00	1004.0
265.00	2695.0
265.90	480.0
273.00	3286.0
274.00	7682.0
275.00	42968.0
276.00	6434.0
276.90	4253.0
278.00	759.0
281.10	375.0
283.00	722.0
285.00	860.0
291.90	309.0
292.90	886.0
295.10	368.0
296.00	14006.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

297.00	1971.0
303.00	1385.0
304.00	451.0
308.00	340.0
314.00	624.0
314.90	1568.0
315.90	927.0
320.90	469.0
323.00	4401.0
323.90	620.0
326.90	958.0
328.00	421.0
331.90	449.0
332.90	384.0
334.00	2813.0
334.90	677.0
341.00	628.0
345.90	896.0
352.10	1421.0
353.00	925.0
354.00	1832.0
355.10	371.0
365.00	7299.0
365.90	1332.0
370.90	389.0
372.00	2095.0
372.90	602.0
383.00	716.0
390.00	522.0
402.00	1128.0
403.10	1551.0
404.10	563.0
421.00	1186.0
422.00	1235.0
423.00	7718.0
424.00	2926.0
441.00	21360.0
442.00	128928.0
443.00	25352.0
444.00	2349.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
38.00	594.0
39.00	3106.0
39.90	652.0
48.90	380.0
50.00	9655.0
51.00	36112.0
52.00	1831.0
56.00	1344.0
56.90	3153.0
61.00	660.0
62.00	849.0
63.00	2122.0
65.10	955.0
67.90	916.0
69.00	57136.0
70.00	412.0
73.10	1082.0
74.00	6168.0
75.00	9592.0
76.00	3396.0
77.10	62776.0
78.00	4389.0
79.00	4190.0
80.00	3074.0
81.00	4482.0
81.90	1110.0
83.00	895.0
85.00	899.0
86.00	1283.0
86.90	595.0
90.90	1048.0
91.90	1082.0
93.00	7826.0
93.90	445.0
96.00	526.0
98.00	6331.0
99.00	4632.0
100.00	462.0
101.00	2618.0
103.00	805.0
104.00	1556.0
105.00	1762.0
106.10	483.0
107.00	17704.0
108.00	3125.0
109.10	486.0
110.00	30008.0
111.00	4937.0
111.90	590.0
115.90	1122.0
117.00	18000.0
118.00	1280.0
121.90	1018.0
122.90	1795.0
123.90	1135.0
125.00	654.0
126.20	465.0
127.00	60320.0
128.00	4202.0
129.00	23072.0
130.00	2044.0
130.80	354.0
134.00	897.0
135.00	2003.0
136.00	993.0
137.00	951.0
141.00	3490.0
142.00	996.0
142.90	801.0
146.00	783.0
146.90	2194.0
148.00	4046.0
148.90	707.0
149.80	347.0
150.90	484.0
152.90	1059.0
153.90	735.0
155.00	1580.0
156.00	2533.0
157.10	456.0
157.70	531.0
158.90	430.0
159.90	1084.0
160.90	1549.0
161.90	519.0
165.00	1248.0
166.00	832.0
167.00	6899.0
168.00	4496.0
168.90	709.0
169.80	322.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

170.80	332.0
171.90	663.0
173.00	722.0
174.00	1207.0
175.00	2477.0
176.00	893.0
177.00	1525.0
177.80	301.0
179.00	5359.0
179.90	3752.0
181.00	1478.0
183.90	472.0
185.00	2429.0
186.00	18632.0
187.00	5775.0
187.90	639.0
188.90	1470.0
190.90	689.0
192.00	1572.0
193.00	2134.0
194.00	376.0
196.00	3505.0
197.10	674.0
197.90	123408.0
198.90	8490.0
200.00	721.0
201.30	503.0
203.00	1242.0
204.00	5321.0
205.00	8673.0
206.00	34984.0
207.10	5466.0
208.00	1768.0
209.00	600.0
210.20	750.0
211.00	1821.0
211.70	305.0
216.00	780.0
216.90	9542.0
217.90	1313.0
221.00	5013.0
222.00	1176.0
223.00	2159.0
224.00	20232.0
225.00	4748.0
226.00	707.0
227.00	8870.0
228.00	1205.0
228.90	1520.0
231.00	597.0
233.90	519.0
234.80	575.0
235.90	352.0
236.90	896.0
238.90	329.0
240.00	331.0
241.00	469.0
242.00	1089.0
243.00	1046.0
244.00	15816.0
245.00	2062.0
246.00	3264.0
246.90	631.0
248.90	613.0
252.90	408.0
254.10	743.0
255.00	87176.0
256.00	12535.0
257.00	1399.0
257.90	5306.0
259.00	912.0
265.00	2079.0
266.00	470.0
273.00	2756.0
274.00	7056.0
275.00	40008.0
276.00	5692.0
277.00	3721.0
277.90	612.0
280.90	341.0
283.10	415.0
284.00	346.0
285.00	664.0
288.80	303.0
292.90	937.0
294.90	400.0
296.00	12963.0
297.00	1747.0
302.90	1734.0
303.90	337.0
314.00	682.0
315.00	1582.0
315.90	623.0
320.90	371.0
323.00	4156.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

324.10	720.0
326.90	864.0
327.80	346.0
332.10	394.0
332.90	540.0
334.00	2864.0
335.00	757.0
341.00	476.0
345.90	989.0
352.00	1501.0
353.00	963.0
354.00	1242.0
355.00	317.0
365.00	6573.0
366.10	793.0
371.00	356.0
372.00	2708.0
372.90	477.0
382.90	696.0
391.10	327.0
401.90	1172.0
403.00	1345.0
403.90	820.0
421.00	1156.0
422.00	1375.0
423.00	8255.0
424.00	2457.0
425.00	303.0
441.00	20368.0
442.00	135424.0
443.00	25728.0
444.10	2331.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

Scan 2449 (17.392 min): BM019537.D
DFTPP57

m/z	Abundance
40.00	468.0
73.10	629.0
207.00	1120.0
280.90	342.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57