

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032824\
 Data File : BM044806.D
 Acq On : 28 Mar 2024 23:28
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM032824

Quant Time: Mar 29 01:29:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 01:24:21 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	39.361	1.6	105	0.00
3	Pyridine	40.000	40.960	-2.4	107	0.00
4	n-Nitrosodimethylamine	40.000	40.336	-0.8	109	0.00
5 S	2-Fluorophenol	80.000	78.855	1.4	104	0.00
6	Aniline	40.000	40.357	-0.9	108	0.00
7 S	Phenol-d6	80.000	79.485	0.6	106	0.00
8	2-Chlorophenol	40.000	39.559	1.1	106	0.00
9	Benzaldehyde	40.000	39.644	0.9	107	0.00
10 C	Phenol	40.000	39.968	0.1	107	0.00
11	bis(2-Chloroethyl)ether	40.000	39.487	1.3	104	0.00
12	1,3-Dichlorobenzene	40.000	39.703	0.7	106	0.00
13 C	1,4-Dichlorobenzene	40.000	39.262	1.8	105	0.00
14	1,2-Dichlorobenzene	40.000	39.689	0.8	107	0.00
15	Benzyl Alcohol	40.000	40.037	-0.1	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.464	1.3	107	-0.01
17	2-Methylphenol	40.000	39.666	0.8	107	0.00
18	Hexachloroethane	40.000	39.444	1.4	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.995	0.0	108	0.00
20	3+4-Methylphenols	40.000	39.910	0.2	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	108	0.00
22	Acetophenone	40.000	39.032	2.4	107	0.00
23 S	Nitrobenzene-d5	80.000	79.280	0.9	107	0.00
24	Nitrobenzene	40.000	39.635	0.9	108	0.00
25	Isophorone	40.000	38.840	2.9	108	0.00
26 C	2-Nitrophenol	40.000	40.151	-0.4	107	0.00
27	2,4-Dimethylphenol	40.000	39.363	1.6	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.822	0.4	109	0.00
29 C	2,4-Dichlorophenol	40.000	39.927	0.2	107	0.00
30	1,2,4-Trichlorobenzene	40.000	39.482	1.3	108	0.00
31	Naphthalene	40.000	38.966	2.6	107	0.00
32	Benzoic acid	40.000	38.334	4.2	106	0.00
33	4-Chloroaniline	40.000	39.593	1.0	109	0.00
34 C	Hexachlorobutadiene	40.000	39.253	1.9	109	0.00
35	Caprolactam	40.000	37.455	6.4	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.809	3.0	108	0.00
37	2-Methylnaphthalene	40.000	38.818	3.0	109	0.00
38	1-Methylnaphthalene	40.000	38.531	3.7	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.541	-1.4	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.079	-7.7	109	0.00
42 S	2,4,6-Tribromophenol	80.000	77.453	3.2	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.215	-0.5	108	0.00
44	2,4,5-Trichlorophenol	40.000	40.251	-0.6	108	0.00
45 S	2-Fluorobiphenyl	80.000	80.186	-0.2	108	0.00
46	1,1'-Biphenyl	40.000	39.875	0.3	108	0.00
47	2-Chloronaphthalene	40.000	39.896	0.3	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032824\
 Data File : BM044806.D
 Acq On : 28 Mar 2024 23:28
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM032824

Quant Time: Mar 29 01:29:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 01:24:21 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.334	1.7	104	0.00
49	Acenaphthylene	40.000	39.736	0.7	108	0.00
50	Dimethylphthalate	40.000	38.230	4.4	105	0.00
51	2,6-Dinitrotoluene	40.000	38.634	3.4	103	0.00
52 C	Acenaphthene	40.000	39.538	1.2	108	0.00
53	3-Nitroaniline	40.000	39.830	0.4	104	0.00
54 P	2,4-Dinitrophenol	40.000	37.849	5.4	105	0.00
55	Dibenzofuran	40.000	38.887	2.8	107	0.00
56 P	4-Nitrophenol	40.000	39.472	1.3	106	0.00
57	2,4-Dinitrotoluene	40.000	38.850	2.9	103	0.00
58	Fluorene	40.000	38.647	3.4	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.800	3.0	105	0.00
60	Diethylphthalate	40.000	38.034	4.9	104	0.00
61	4-Chlorophenyl-phenylether	40.000	38.459	3.9	106	0.00
62	4-Nitroaniline	40.000	39.738	0.7	104	0.00
63	Azobenzene	40.000	38.509	3.7	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.466	-1.2	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.230	-0.6	105	0.00
67	4-Bromophenyl-phenylether	40.000	39.286	1.8	102	0.00
68	Hexachlorobenzene	40.000	39.509	1.2	104	0.00
69	Atrazine	40.000	40.621	-1.6	101	0.00
70 C	Pentachlorophenol	40.000	40.830	-2.1	102	0.00
71	Phenanthrene	40.000	39.263	1.8	102	0.00
72	Anthracene	40.000	39.633	0.9	102	0.00
73	Carbazole	40.000	39.904	0.2	101	0.00
74	Di-n-butylphthalate	40.000	39.438	1.4	99	0.00
75 C	Fluoranthene	40.000	39.629	0.9	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	43.306	-8.3	95	0.00
78	Pyrene	40.000	38.500	3.8	101	0.00
79 S	Terphenyl-d14	80.000	76.739	4.1	99	0.00
80	Butylbenzylphthalate	40.000	39.026	2.4	97	0.00
81	Benzo(a)anthracene	40.000	39.464	1.3	97	0.00
82	3,3'-Dichlorobenzidine	40.000	40.008	-0.0	95	0.00
83	Chrysene	40.000	39.238	1.9	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.305	4.2	92	0.00
85 c	Di-n-octyl phthalate	40.000	38.184	4.5	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.235	-0.6	99	0.00
88	Benzo(b)fluoranthene	40.000	40.098	-0.2	101	0.00
89	Benzo(k)fluoranthene	40.000	39.528	1.2	95	0.00
90 C	Benzo(a)pyrene	40.000	40.070	-0.2	98	0.00
91	Dibenzo(a,h)anthracene	40.000	40.013	-0.0	98	0.00
92	Benzo(g,h,i)perylene	40.000	40.142	-0.4	100	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032824\
Data File : BM044806.D
Acq On : 28 Mar 2024 23:28
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM032824

Quant Time: Mar 29 01:29:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 01:24:21 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0