

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
 Data File : BM029250.D
 Acq On : 29 Mar 2021 18:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 29 19:12:48 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 11:14:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.527	0.538	-2.1	88	0.00
3	Pyridine	1.271	1.409	-10.9	86	0.00
4	n-Nitrosodimethylamine	0.571	0.639	-11.9	85	0.00
5 S	2-Fluorophenol	1.160	1.197	-3.2	86	0.00
6	Aniline	1.804	1.866	-3.4	85	0.00
7 S	Phenol-d6	1.665	1.730	-3.9	85	0.00
8	2-Chlorophenol	1.309	1.343	-2.6	86	0.00
9	Benzaldehyde	1.046	1.054	-0.8	85	0.00
10 C	Phenol	1.645	1.712	-4.1	86	0.00
11	bis(2-Chloroethyl)ether	1.177	1.224	-4.0	84	0.00
12	1,3-Dichlorobenzene	1.465	1.462	0.2	85	0.00
13 C	1,4-Dichlorobenzene	1.485	1.477	0.5	84	0.00
14	1,2-Dichlorobenzene	1.427	1.427	0.0	84	0.00
15	Benzyl Alcohol	1.205	1.275	-5.8	85	0.00
16	2,2'-oxybis(1-Chloropropane	1.805	1.956	-8.4	89	0.01
17	2-Methylphenol	1.057	1.092	-3.3	84	0.00
18	Hexachloroethane	0.540	0.564	-4.4	87	0.00
19 P	n-Nitroso-di-n-propylamine	1.074	1.132	-5.4	82	0.00
20	3+4-Methylphenols	1.415	1.459	-3.1	82	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
22	Acetophenone	0.501	0.523	-4.4	84	0.00
23 S	Nitrobenzene-d5	0.412	0.443	-7.5	85	0.00
24	Nitrobenzene	0.430	0.464	-7.9	86	0.00
25	Isophorone	0.708	0.760	-7.3	82	0.00
26 C	2-Nitrophenol	0.143	0.153	-7.0	78	0.00
27	2,4-Dimethylphenol	0.275	0.283	-2.9	81	0.00
28	bis(2-Chloroethoxy)methane	0.374	0.395	-5.6	83	0.00
29 C	2,4-Dichlorophenol	0.299	0.308	-3.0	80	0.00
30	1,2,4-Trichlorobenzene	0.339	0.340	-0.3	82	0.00
31	Naphthalene	1.054	1.080	-2.5	83	0.00
32	Benzoic acid	0.160	0.165	-3.1	74	0.00
33	4-Chloroaniline	0.417	0.438	-5.0	82	0.00
34 C	Hexachlorobutadiene	0.197	0.193	2.0	81	0.00
35	Caprolactam	0.091	0.094	-3.3	78	0.00
36 C	4-Chloro-3-methylphenol	0.319	0.338	-6.0	81	0.00
37	2-Methylnaphthalene	0.737	0.755	-2.4	81	0.00
38	1-Methylnaphthalene	0.701	0.725	-3.4	82	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	0.560	0.538	3.9	80	0.00
41 P	Hexachlorocyclopentadiene	0.308	0.293	4.9	77	0.00
42 S	2,4,6-Tribromophenol	0.185	0.185	0.0	76	0.00
43 C	2,4,6-Trichlorophenol	0.365	0.367	-0.5	79	0.00
44	2,4,5-Trichlorophenol	0.406	0.407	-0.2	79	0.00
45 S	2-Fluorobiphenyl	1.410	1.398	0.9	82	0.00
46	1,1'-Biphenyl	1.520	1.500	1.3	81	0.00
47	2-Chloronaphthalene	1.197	1.177	1.7	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
 Data File : BM029250.D
 Acq On : 29 Mar 2021 18:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 29 19:12:48 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 11:14:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.351	0.392	-11.7	84	0.00
49	Acenaphthylene	1.821	1.858	-2.0	82	0.00
50	Dimethylphthalate	1.398	1.421	-1.6	81	0.00
51	2,6-Dinitrotoluene	0.302	0.314	-4.0	80	0.00
52 C	Acenaphthene	1.157	1.154	0.3	81	0.00
53	3-Nitroaniline	0.306	0.327	-6.9	81	0.00
54 P	2,4-Dinitrophenol	0.151	0.148	2.0	75	0.00
55	Dibenzofuran	1.821	1.830	-0.5	81	0.00
56 P	4-Nitrophenol	0.280	0.285	-1.8	79	0.00
57	2,4-Dinitrotoluene	0.398	0.425	-6.8	80	0.00
58	Fluorene	1.476	1.518	-2.8	82	0.00
59	2,3,4,6-Tetrachlorophenol	0.332	0.336	-1.2	80	0.00
60	Diethylphthalate	1.377	1.433	-4.1	82	0.00
61	4-Chlorophenyl-phenylether	0.681	0.680	0.1	80	0.00
62	4-Nitroaniline	0.322	0.359	-11.5	82	0.00
63	Azobenzene	1.574	1.782	-13.2	88	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
65	4,6-Dinitro-2-methylphenol	0.111	0.108	2.7	75	0.00
66 c	n-Nitrosodiphenylamine	0.640	0.646	-0.9	81	0.00
67	4-Bromophenyl-phenylether	0.183	0.166	9.3	73	0.00
68	Hexachlorobenzene	0.216	0.212	1.9	80	0.00
69	Atrazine	0.202	0.215	-6.4	81	0.00
70 C	Pentachlorophenol	0.133	0.135	-1.5	80	0.00
71	Phenanthrene	1.160	1.189	-2.5	83	0.00
72	Anthracene	1.127	1.166	-3.5	83	0.00
73	Carbazole	1.109	1.168	-5.3	84	0.00
74	Di-n-butylphthalate	1.069	1.194	-11.7	84	0.00
75 C	Fluoranthene	1.281	1.368	-6.8	85	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
77	Benzydine	0.578	0.570	1.4	85	0.00
78	Pyrene	1.354	1.309	3.3	84	0.00
79 S	Terphenyl-d14	0.996	0.983	1.3	85	0.00
80	Butylbenzylphthalate	0.457	0.500	-9.4	86	0.00
81	Benzo(a)anthracene	1.305	1.320	-1.1	88	0.00
82	3,3'-Dichlorobenzidine	0.390	0.397	-1.8	84	0.00
83	Chrysene	1.306	1.276	2.3	86	0.00
84	Bis(2-ethylhexyl)phthalate	0.658	0.775	-17.8	92	0.00
85 c	Di-n-octyl phthalate	1.066	1.275	-19.6	92	0.00
86 I	Perylene-d12	1.000	1.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	1.463	1.511	-3.3	89	0.00
88	Benzo(b)fluoranthene	1.301	1.314	-1.0	86	0.00
89	Benzo(k)fluoranthene	1.264	1.307	-3.4	91	0.00
90 C	Benzo(a)pyrene	1.179	1.211	-2.7	88	0.00
91	Dibenzo(a,h)anthracene	1.252	1.294	-3.4	89	0.00
92	Benzo(g,h,i)perylene	1.227	1.253	-2.1	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
Data File : BM029250.D
Acq On : 29 Mar 2021 18:31
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 29 19:12:48 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 29 11:14:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0