

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
 Data File : BM029247.D
 Acq On : 29 Mar 2021 14:37
 Operator : CG/JU
 Sample : M1458-09
 Misc : MDL-WATER-8PPM
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Manual Integrations
 APPROVED

mohammad
 3/30/2021 12:41:36 PM

Quant Time: Mar 29 15:09:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 12:59:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.751	152	102071	20.00	ng	0.00
21) Naphthalene-d8	10.533	136	395589	20.00	ng	0.00
39) Acenaphthene-d10	14.374	164	233027	20.00	ng	0.00
64) Phenanthrene-d10	17.115	188	475907	20.00	ng	0.00
76) Chrysene-d12	21.280	240	489669	20.00	ng	0.00
86) Perylene-d12	23.515	264	506944	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.351	112	454592	76.77	ng	0.00
7) Phenol-d6	6.916	99	657961	77.43	ng	0.00
23) Nitrobenzene-d5	8.904	82	528206	64.83	ng	0.00
42) 2,4,6-Tribromophenol	15.863	330	158899	73.63	ng	0.00
45) 2-Fluorobiphenyl	13.004	172	1066693	64.93	ng	0.00
79) Terphenyl-d14	19.739	244	1539781	63.17	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.304	88	20862	7.75	ng	97
3) Pyridine	3.704	79	47345	7.30	ng	# 93
4) n-Nitrosodimethylamine	3.622	42	22420	7.70	ng	# 95
6) Aniline	7.087	93	84817	9.21	ng	99
8) 2-Chlorophenol	7.316	128	53644	8.03	ng	97
9) Benzaldehyde	6.904	77	51201	9.60	ng	98
10) Phenol	6.945	94	69896	8.33	ng	98
11) bis(2-Chloroethyl)ether	7.187	93	49794	8.29	ng	97
12) 1,3-Dichlorobenzene	7.645	146	61614	8.24	ng	99
13) 1,4-Dichlorobenzene	7.787	146	62325	8.22	ng	96
14) 1,2-Dichlorobenzene	8.104	146	59317	8.14	ng	99
15) Benzyl Alcohol	7.992	79	48070	7.82	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.281	45	82195	8.92	ng	98
17) 2-Methylphenol	8.187	107	42088	7.80	ng	99
18) Hexachloroethane	8.828	117	22040	8.00	ng	93
19) n-Nitroso-di-n-propyla...	8.557	70	43454	7.93	ng	98
20) 3+4-Methylphenols	8.510	107	56373	7.81	ng	98
22) Acetophenone	8.569	105	79879	8.07	ng	96
24) Nitrobenzene	8.945	77	72487	8.52	ng	99
25) Isophorone	9.469	82	103723	7.41	ng	99
26) 2-Nitrophenol	9.651	139	18783	6.62	ng	97
27) 2,4-Dimethylphenol	9.710	122	45028	8.28	ng	99
28) bis(2-Chloroethoxy)met...	9.951	93	65285	8.82	ng	99
29) 2,4-Dichlorophenol	10.181	162	42152	7.13	ng	97
30) 1,2,4-Trichlorobenzene	10.398	180	52572	7.83	ng	97
31) Naphthalene	10.580	128	165413	7.93	ng	99
32) Benzoic acid	9.769	122	15443m	4.89	ng	
33) 4-Chloroaniline	10.692	127	64846	7.87	ng	99
34) Hexachlorobutadiene	10.875	225	29089	7.46	ng	98
35) Caprolactam	11.445	113	10040	5.60	ng	96
36) 4-Chloro-3-methylphenol	11.810	107	45401	7.19	ng	99
37) 2-Methylnaphthalene	12.198	142	113396	7.78	ng	96
38) 1-Methylnaphthalene	12.416	142	106349	7.67	ng	98
40) 1,2,4,5-Tetrachloroben...	12.563	216	51191	7.84	ng	97
41) Hexachlorocyclopentadiene	12.551	237	28608	7.98	ng	98
43) 2,4,6-Trichlorophenol	12.804	196	29269	6.88	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
 Data File : BM029247.D
 Acq On : 29 Mar 2021 14:37
 Operator : CG/JU
 Sample : M1458-09
 Misc : MDL-WATER-8PPM
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Manual Integrations
 APPROVED

mohammad
 3/30/2021 12:41:36 PM

Quant Time: Mar 29 15:09:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 12:59:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.869	196	34450	7.29	ng	99
46) 1,1'-Biphenyl	13.210	154	146061	8.25	ng	97
47) 2-Chloronaphthalene	13.251	162	111533	8.00	ng	98
48) 2-Nitroaniline	13.451	65	28449	6.96	ng	94
49) Acenaphthylene	14.098	152	162382	7.65	ng	99
50) Dimethylphthalate	13.833	163	132825	8.16	ng	98
51) 2,6-Dinitrotoluene	13.951	165	25275	7.18	ng	89
52) Acenaphthene	14.439	154	107892	8.00	ng	96
53) 3-Nitroaniline	14.280	138	24900	6.98	ng	97
54) 2,4-Dinitrophenol	14.480	184	8714	4.95	ng	91
55) Dibenzofuran	14.774	168	167896	7.91	ng	97
56) 4-Nitrophenol	14.574	139	19335	5.94	ng	91
57) 2,4-Dinitrotoluene	14.733	165	29989	6.48	ng	# 85
58) Fluorene	15.421	166	134481	7.82	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.998	232	27276	7.04	ng	97
60) Diethylphthalate	15.204	149	128555	8.01	ng	98
61) 4-Chlorophenyl-phenyle...	15.421	204	63170	7.96	ng	95
62) 4-Nitroaniline	15.433	138	24620	6.56	ng	99
63) Azobenzene	15.715	77	145363	7.93	ng	98
65) 4,6-Dinitro-2-methylph...	15.492	198	13689	5.16	ng	96
66) n-Nitrosodiphenylamine	15.633	169	114455	7.51	ng	97
67) 4-Bromophenyl-phenylether	16.315	248	29003	6.65	ng	96
68) Hexachlorobenzene	16.427	284	39098	7.62	ng	97
69) Atrazine	16.580	200	31523	6.55	ng	99
70) Pentachlorophenol	16.762	266	18689	5.91	ng	96
71) Phenanthrene	17.157	178	214243	7.76	ng	99
72) Anthracene	17.245	178	204738	7.64	ng	99
73) Carbazole	17.515	167	189628	7.19	ng	99
74) Di-n-butylphthalate	18.086	149	180441	7.09	ng	99
75) Fluoranthene	19.168	202	228780	7.50	ng	99
77) Benzidine	19.351	184	89755	6.34	ng	99
78) Pyrene	19.527	202	247397	7.46	ng	98
80) Butylbenzylphthalate	20.433	149	71838	6.41	ng	97
81) Benzo(a)anthracene	21.268	228	236038	7.39	ng	98
82) 3,3'-Dichlorobenzidine	21.203	252	75408	7.89	ng	96
83) Chrysene	21.315	228	236051	7.38	ng	99
84) Bis(2-ethylhexyl)phtha...	21.209	149	110297	6.85	ng	98
85) Di-n-octyl phthalate	22.086	149	181238	6.95	ng	96
87) Indeno(1,2,3-cd)pyrene	25.768	276	290713	7.84	ng	100
88) Benzo(b)fluoranthene	22.844	252	256277	7.77	ng	100
89) Benzo(k)fluoranthene	22.886	252	237731	7.42	ng	99
90) Benzo(a)pyrene	23.415	252	220441	7.38	ng	97
91) Dibenzo(a,h)anthracene	25.785	278	240558	7.58	ng	96
92) Benzo(g,h,i)perylene	26.456	276	241278	7.76	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
Data File : BM029247.D
Acq On : 29 Mar 2021 14:37
Operator : CG/JU
Sample : M1458-09
Misc : MDL-WATER-8PPM
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021

Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM

Quant Time: Mar 29 15:09:04 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 29 12:59:16 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.751 min Scan# 827
Delta R.T. -0.001 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37





Tgt Ion:152 Resp: 102071
Ion Ratio Lower Upper
152 100
150 154.3 124.3 186.5
115 59.8 46.4 69.6



#2
1,4-Dioxane
Concen: 7.75 ng
RT: 3.304 min Scan# 71
Delta R.T. 0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion: 88 Resp: 20862
Ion Ratio Lower Upper
88 100
58 73.2 59.9 89.9
43 36.6 26.2 39.4



Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021


Manual Integrations
APPROVED

mohammad
3/30/2021 12:41:36 PM



#3
Pyridine
Concen: 7.30 ng
RT: 3.704 min Scan# 139
Delta R.T. 0.011 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion: 79 Resp: 47345
Ion Ratio Lower Upper
79 100
52 70.0 53.2 79.8
51 44.9 29.8 44.8#

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#4
n-Nitrosodimethylamine
Concen: 7.70 ng
RT: 3.622 min Scan# 125
Delta R.T. 0.017 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion: 42 Resp: 22420
Ion Ratio Lower Upper
42 100
74 96.5 80.6 121.0
44 10.1 5.4 8.0#





#5
2-Fluorophenol
Concen: 76.77 ng
RT: 5.351 min Scan# 419
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:	112	Resp:	454592
Ion Ratio	Lower	Upper	
112	100		
64	63.8	51.6	77.4
63	34.5	29.0	43.4

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#6
Aniline
Concen: 9.21 ng
RT: 7.087 min Scan# 714
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion:	93	Resp:	84817
Ion Ratio	Lower	Upper	
93	100		
66	41.1	33.2	49.8
65	21.1	17.8	26.6









#7
Phenol-d6
Concen: 77.43 ng
RT: 6.916 min Scan# 685
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	99	Resp:	657961
Ion	Ratio	Lower	Upper
99	100		
42	24.2	18.5	27.7
71	39.0	31.8	47.6

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#8
2-Chlorophenol
Concen: 8.03 ng
RT: 7.316 min Scan# 753
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion:	128	Resp:	53644
Ion	Ratio	Lower	Upper
128	100		
130	32.2	13.1	53.1
64	54.0	30.9	70.9





#9
Benzaldehyde
Concen: 9.60 ng
RT: 6.904 min Scan# 683
Delta R.T. 0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion: 77 Resp: 51201
Ion Ratio Lower Upper
77 100
105 88.1 71.4 111.4
106 87.7 66.5 106.5

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#10
Phenol
Concen: 8.33 ng
RT: 6.945 min Scan# 690
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion: 94 Resp: 69896
Ion Ratio Lower Upper
94 100
65 32.2 13.0 53.0
66 42.4 24.1 64.1





#11
bis(2-Chloroethyl)ether
Concen: 8.29 ng
RT: 7.187 min Scan# 731
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion: 93 Resp: 49794
Ion Ratio Lower Upper
93 100
63 84.9 63.9 103.9
95 29.2 12.8 52.8

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#12
1,3-Dichlorobenzene
Concen: 8.24 ng
RT: 7.645 min Scan# 809
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion: 146 Resp: 61614
Ion Ratio Lower Upper
146 100
148 63.9 50.5 75.7
75 37.2 29.7 44.5









#13
1,4-Dichlorobenzene
Concen: 8.22 ng
RT: 7.787 min Scan# 833
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:146 Resp: 62325
Ion Ratio Lower Upper
146 100
148 68.2 51.8 77.6
111 42.6 32.8 49.2

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#14
1,2-Dichlorobenzene
Concen: 8.14 ng
RT: 8.104 min Scan# 887
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion:146 Resp: 59317
Ion Ratio Lower Upper
146 100
148 62.5 50.2 75.4
111 43.9 35.8 53.8





#15
Benzyl Alcohol
Concen: 7.82 ng
RT: 7.992 min Scan# 868
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion: 79 Resp: 48070
Ion Ratio Lower Upper
79 100
108 66.6 56.6 84.8
77 63.0 54.9 82.3

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 8.92 ng
RT: 8.281 min Scan# 917
Delta R.T. 0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion: 45 Resp: 82195
Ion Ratio Lower Upper
45 100
77 15.1 0.0 34.5
79 12.0 0.0 30.9





#17
2-Methylphenol
Concen: 7.80 ng
RT: 8.187 min Scan# 901
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021



Tgt Ion:	107	Resp:	42088
Ion	Ratio	Lower	Upper
107	100		
108	111.8	89.8	134.6
77	54.9	42.8	64.2
79	48.9	40.4	60.6




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#18
Hexachloroethane
Concen: 8.00 ng
RT: 8.828 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	117	Resp:	22040
Ion	Ratio	Lower	Upper
117	100		
119	99.0	75.0	112.6
201	85.2	74.6	111.8









#19
n-Nitroso-di-n-propylamine
Concen: 7.93 ng
RT: 8.557 min Scan# 964
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	70	Resp:	43454
Ion	Ratio	Lower	Upper
70	100		
42	67.2	53.3	79.9
101	7.2	6.7	10.1
130	17.6	16.1	24.1

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#20
3+4-Methylphenols
Concen: 7.81 ng
RT: 8.510 min Scan# 956
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion:	107	Resp:	56373
Ion	Ratio	Lower	Upper
107	100		
108	85.9	68.5	108.5
77	38.2	17.0	57.0
79	31.6	12.0	52.0





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.533 min Scan# 1300
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:	136	Resp:	395589
Ion	Ratio	Lower	Upper
136	100		
137	10.4	9.3	13.9
54	9.0	6.9	10.3
68	6.4	5.4	8.0

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#22
Acetophenone
Concen: 8.07 ng
RT: 8.569 min Scan# 966
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion:	105	Resp:	79879
Ion	Ratio	Lower	Upper
105	100		
71	3.3	2.8	4.2
51	32.8	24.7	37.1
120	24.0	17.4	26.2









#23
Nitrobenzene-d5
Concen: 64.83 ng
RT: 8.904 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion: 82 Resp: 528206
Ion Ratio Lower Upper
82 100
128 37.0 28.6 42.8
54 50.3 40.7 61.1

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#24
Nitrobenzene
Concen: 8.52 ng
RT: 8.945 min Scan# 1030
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion: 77 Resp: 72487
Ion Ratio Lower Upper
77 100
123 37.9 31.3 46.9
65 13.6 10.8 16.2





#25
Isophorone
Concen: 7.41 ng
RT: 9.469 min Scan# 1119
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion: 82 Resp: 103723
Ion Ratio Lower Upper
82 100
95 7.8 5.6 8.4
138 16.6 13.3 19.9

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#26
2-Nitrophenol
Concen: 6.62 ng
RT: 9.651 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion: 139 Resp: 18783
Ion Ratio Lower Upper
139 100
109 33.8 26.6 39.8
65 54.5 45.6 68.4









#27
2,4-Dimethylphenol
Concen: 8.28 ng
RT: 9.710 min Scan# 1160
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:122 Resp: 45028
Ion Ratio Lower Upper
122 100
107 118.7 95.0 142.6
121 59.3 45.8 68.6

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#28
bis(2-Chloroethoxy)methane
Concen: 8.82 ng
RT: 9.951 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion: 93 Resp: 65285
Ion Ratio Lower Upper
93 100
95 32.6 26.6 39.8
123 10.5 8.6 13.0









#29
2,4-Dichlorophenol
Concen: 7.13 ng
RT: 10.181 min Scan# 1240
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:162 Resp: 42152
Ion Ratio Lower Upper
162 100
164 62.5 45.7 85.7
98 36.7 16.3 56.3

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#30
1,2,4-Trichlorobenzene
Concen: 7.83 ng
RT: 10.398 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:180 Resp: 52572
Ion Ratio Lower Upper
180 100
182 99.0 77.0 115.4
145 33.7 25.4 38.2









#31
Naphthalene
Concen: 7.93 ng
RT: 10.580 min Scan# 1308
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	128	Resp:	165413
Ion Ratio	Lower	Upper	
128	100		
129	11.0	9.0	13.6
127	12.8	10.7	16.1

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#32
Benzoic acid
Concen: 4.89 ng m
RT: 9.769 min Scan# 1170
Delta R.T. -0.059 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	122	Resp:	15443
Ion Ratio	Lower	Upper	
122	100		
105	141.6	108.2	148.2
77	123.8	86.0	126.0









#33
4-Chloroaniline
Concen: 7.87 ng
RT: 10.692 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	127	Resp:	64846
Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.4	38.0
65	36.1	28.6	42.8
92	19.2	15.8	23.6

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#34
Hexachlorobutadiene
Concen: 7.46 ng
RT: 10.875 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	225	Resp:	29089
Ion	Ratio	Lower	Upper
225	100		
223	64.7	51.8	77.8
227	68.1	52.2	78.4









#35
Caprolactam
Concen: 5.60 ng
RT: 11.445 min Scan# 1455
Delta R.T. -0.018 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:113 Resp: 10040
Ion Ratio Lower Upper
113 100
55 166.0 149.4 189.4
56 124.0 110.8 150.8

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#36
4-Chloro-3-methylphenol
Concen: 7.19 ng
RT: 11.810 min Scan# 1517
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:107 Resp: 45401
Ion Ratio Lower Upper
107 100
144 26.3 22.7 34.1
142 84.2 67.0 100.4









#37
2-Methylnaphthalene
Concen: 7.78 ng
RT: 12.198 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:142 Resp: 113396
Ion Ratio Lower Upper
142 100
141 84.4 72.0 108.0
115 37.7 29.8 44.6

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#38
1-Methylnaphthalene
Concen: 7.67 ng
RT: 12.416 min Scan# 1620
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion:142 Resp: 106349
Ion Ratio Lower Upper
142 100
141 94.2 73.5 110.3
116 4.3 3.2 4.8





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.374 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:164 Resp: 233027
Ion Ratio Lower Upper
164 100
162 100.6 84.4 126.6
160 45.8 36.6 54.8

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 7.84 ng
RT: 12.563 min Scan# 1645
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion:216 Resp: 51191
Ion Ratio Lower Upper
216 100
214 80.8 62.3 93.5
179 24.1 18.6 27.8
108 24.0 18.2 27.4









#41
Hexachlorocyclopentadiene
Concen: 7.98 ng
RT: 12.551 min Scan# 1643
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	237	Resp:	28608
Ion Ratio	Lower	Upper	
237	100		
235	63.3	41.8	81.8
272	11.0	0.0	31.1

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#42
2,4,6-Tribromophenol
Concen: 73.63 ng
RT: 15.863 min Scan# 2206
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	330	Resp:	158899
Ion Ratio	Lower	Upper	
330	100		
332	95.9	77.1	115.7
141	49.4	39.2	58.8









#43
2,4,6-Trichlorophenol
Concen: 6.88 ng
RT: 12.804 min Scan# 1686
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:196 Resp: 29269
Ion Ratio Lower Upper
196 100
198 94.3 75.3 112.9
200 30.4 25.2 37.8

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#44
2,4,5-Trichlorophenol
Concen: 7.29 ng
RT: 12.869 min Scan# 1697
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion:196 Resp: 34450
Ion Ratio Lower Upper
196 100
198 93.5 74.6 112.0
97 57.0 46.4 69.6
132 28.3 21.5 32.3





#45
2-Fluorobiphenyl
Concen: 64.93 ng
RT: 13.004 min Scan# 1720
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:172 Resp: 1066693
Ion Ratio Lower Upper
172 100
171 35.7 28.7 43.1
170 23.7 18.8 28.2

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#46
1,1'-Biphenyl
Concen: 8.25 ng
RT: 13.210 min Scan# 1755
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion:154 Resp: 146061
Ion Ratio Lower Upper
154 100
153 40.3 21.5 61.5
76 15.4 0.0 33.6





#47
2-Chloronaphthalene
Concen: 8.00 ng
RT: 13.251 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:162 Resp: 111533
Ion Ratio Lower Upper
162 100
127 40.2 30.4 45.6
164 33.5 26.3 39.5

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#48
2-Nitroaniline
Concen: 6.96 ng
RT: 13.451 min Scan# 1796
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion: 65 Resp: 28449
Ion Ratio Lower Upper
65 100
92 65.9 52.4 78.6
138 83.6 75.0 112.6





#49
Acenaphthylene
Concen: 7.65 ng
RT: 14.098 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:152 Resp: 162382
Ion Ratio Lower Upper
152 100
151 19.6 16.2 24.4
153 12.8 10.2 15.2

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#50
Dimethylphthalate
Concen: 8.16 ng
RT: 13.833 min Scan# 1861
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion:163 Resp: 132825
Ion Ratio Lower Upper
163 100
194 3.7 3.2 4.8
164 10.7 7.9 11.9









#51
2,6-Dinitrotoluene
Concen: 7.18 ng
RT: 13.951 min Scan# 1881
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:165 Resp: 25275
Ion Ratio Lower Upper
165 100
63 74.0 51.4 77.2
89 71.4 51.4 77.2

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#52
Acenaphthene
Concen: 8.00 ng
RT: 14.439 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion:154 Resp: 107892
Ion Ratio Lower Upper
154 100
153 118.3 90.5 135.7
152 55.8 43.4 65.2





#53
3-Nitroaniline
Concen: 6.98 ng
RT: 14.280 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:138 Resp: 24900
Ion Ratio Lower Upper
138 100
108 12.8 9.0 13.4
92 129.4 101.4 152.0

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#54
2,4-Dinitrophenol
Concen: 4.95 ng
RT: 14.480 min Scan# 1971
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion:184 Resp: 8714
Ion Ratio Lower Upper
184 100
63 87.1 61.9 92.9
154 71.1 52.3 78.5









#55
Dibenzenofuran
Concen: 7.91 ng
RT: 14.774 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:168 Resp: 167896
Ion Ratio Lower Upper
168 100
139 40.5 30.6 46.0
169 12.6 10.2 15.4

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#56
4-Nitrophenol
Concen: 5.94 ng
RT: 14.574 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion:139 Resp: 19335
Ion Ratio Lower Upper
139 100
109 92.7 69.5 109.5
65 133.1 98.1 138.1





#57
2,4-Dinitrotoluene
Concen: 6.48 ng
RT: 14.733 min Scan# 2014
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:165 Resp: 29989
Ion Ratio Lower Upper
165 100
63 62.0 42.2 63.4
89 97.2 65.6 98.4
182 3.8 2.2 3.4#

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#58
Fluorene
Concen: 7.82 ng
RT: 15.421 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion:166 Resp: 134481
Ion Ratio Lower Upper
166 100
165 100.6 79.6 119.4
167 13.7 11.0 16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 7.04 ng
RT: 14.998 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37





Tgt Ion:232 Resp: 27276
Ion Ratio Lower Upper
232 100
131 54.4 41.1 61.7
130 3.3 2.6 3.8
166 35.7 27.5 41.3



#60
Diethylphthalate
Concen: 8.01 ng
RT: 15.204 min Scan# 2094
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion:149 Resp: 128555
Ion Ratio Lower Upper
149 100
177 20.1 17.3 25.9
150 11.8 9.8 14.6




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021

Manual Integrations
APPROVED




mohammad
3/30/2021 12:41:36 PM




#61
4-Chlorophenyl-phenylether
Concen: 7.96 ng
RT: 15.421 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37





Tgt Ion:204 Resp: 63170
Ion Ratio Lower Upper
204 100
206 31.4 26.7 40.1
141 64.4 48.4 72.6



#62
4-Nitroaniline
Concen: 6.56 ng
RT: 15.433 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion:138 Resp: 24620
Ion Ratio Lower Upper
138 100
92 56.7 38.2 78.2
108 92.0 71.7 111.7




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021

Manual Integrations
APPROVED




mohammad
3/30/2021 12:41:36 PM




#63
Azobenzene
Concen: 7.93 ng
RT: 15.715 min Scan# 2181
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37





Tgt Ion: 77 Resp: 145363
Ion Ratio Lower Upper
77 100
182 21.5 2.3 42.3
105 15.7 0.0 36.4
51 32.1 12.9 52.9



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.115 min Scan# 2419
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion: 188 Resp: 475907
Ion Ratio Lower Upper
188 100
94 9.6 8.1 12.1
80 11.0 9.2 13.8



Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021


Manual Integrations
APPROVED

mohammad
3/30/2021 12:41:36 PM



#65
4,6-Dinitro-2-methylphenol
Concen: 5.16 ng
RT: 15.492 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:198 Resp: 13689
Ion Ratio Lower Upper
198 100
51 58.0 35.1 75.1
105 47.0 29.4 69.4

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#66
n-Nitrosodiphenylamine
Concen: 7.51 ng
RT: 15.633 min Scan# 2167
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion:169 Resp: 114455
Ion Ratio Lower Upper
169 100
168 71.2 54.6 81.8
167 37.5 29.4 44.0









#67
4-Bromophenyl-phenylether
Concen: 6.65 ng
RT: 16.315 min Scan# 2283
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	248	Resp:	29003
Ion Ratio	Lower	Upper	
248	100		
250	97.3	78.4	117.6
141	76.8	67.1	100.7

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#68
Hexachlorobenzene
Concen: 7.62 ng
RT: 16.427 min Scan# 2302
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:	284	Resp:	39098
Ion Ratio	Lower	Upper	
284	100		
142	40.2	33.4	50.2
249	30.6	25.9	38.9









#69
Atrazine
Concen: 6.55 ng
RT: 16.580 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:200 Resp: 31523
Ion Ratio Lower Upper
200 100
173 25.9 6.1 46.1
215 44.2 23.8 63.8

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#70
Pentachlorophenol
Concen: 5.91 ng
RT: 16.762 min Scan# 2359
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:266 Resp: 18689
Ion Ratio Lower Upper
266 100
268 64.8 52.2 78.2
264 58.0 50.9 76.3









#71
Phenanthrene
Concen: 7.76 ng
RT: 17.157 min Scan# 2426
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:178 Resp: 214243
Ion Ratio Lower Upper
178 100
176 19.1 15.7 23.5
179 15.1 11.9 17.9

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#72
Anthracene
Concen: 7.64 ng
RT: 17.245 min Scan# 2441
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:178 Resp: 204738
Ion Ratio Lower Upper
178 100
176 18.3 14.9 22.3
179 14.8 12.1 18.1









#73
 Carbazole
 Concen: 7.19 ng
 RT: 17.515 min Scan# 2487
 Delta R.T. -0.000 min
 Lab File: BM029247.D
 Acq: 29 Mar 2021 14:37

Tgt Ion:	167	Resp:	189628
Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.5	26.3
139	12.2	10.6	15.8

Instrument :
 BNA_M
 ClientSampleId :
 MDL-WATER-03-QT1-2021




Manual Integrations
 APPROVED


mohammad
 3/30/2021 12:41:36 PM



#74
 Di-n-butylphthalate
 Concen: 7.09 ng
 RT: 18.086 min Scan# 2584
 Delta R.T. -0.000 min
 Lab File: BM029247.D
 Acq: 29 Mar 2021 14:37

Tgt Ion:	149	Resp:	180441
Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.3	10.9
104	6.2	4.6	6.8









#75
Fluoranthene
Concen: 7.50 ng
RT: 19.168 min Scan# 2768
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:202 Resp: 228780
Ion Ratio Lower Upper
202 100
101 12.0 0.0 31.5
203 17.4 0.0 37.7

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.280 min Scan# 3127
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:240 Resp: 489669
Ion Ratio Lower Upper
240 100
120 13.1 9.0 13.6
236 26.2 20.3 30.5









#77
Benzidine
Concen: 6.34 ng
RT: 19.351 min Scan# 2799
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:184 Resp: 89755
Ion Ratio Lower Upper
184 100
185 14.3 11.3 16.9
183 12.0 9.8 14.6

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#78
Pyrene
Concen: 7.46 ng
RT: 19.527 min Scan# 2829
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:202 Resp: 247397
Ion Ratio Lower Upper
202 100
200 19.9 17.0 25.6
203 16.9 14.1 21.1









#79
Terphenyl-d14
Concen: 63.17 ng
RT: 19.739 min Scan# 2865
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:244 Resp: 1539781
Ion Ratio Lower Upper
244 100
212 8.0 6.5 9.7
122 11.6 9.2 13.8

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM




#80
Butylbenzylphthalate
Concen: 6.41 ng
RT: 20.433 min Scan# 2983
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:149 Resp: 71838
Ion Ratio Lower Upper
149 100
91 77.8 60.5 90.7
206 18.4 15.7 23.5







#81
Benzo(a)anthracene
Concen: 7.39 ng
RT: 21.268 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion:228 Resp: 236038
Ion Ratio Lower Upper
228 100
226 28.4 21.9 32.9
229 18.5 15.7 23.5




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#82
3,3'-Dichlorobenzidine
Concen: 7.89 ng
RT: 21.203 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion:252 Resp: 75408
Ion Ratio Lower Upper
252 100
254 62.6 52.6 78.8
126 12.0 10.8 16.2





#83
Chrysene
Concen: 7.38 ng
RT: 21.315 min Scan# 3133
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:228 Resp: 236051
Ion Ratio Lower Upper
228 100
226 30.1 24.7 37.1
229 19.4 15.8 23.6

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#84
Bis(2-ethylhexyl)phthalate
Concen: 6.85 ng
RT: 21.209 min Scan# 3115
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion:149 Resp: 110297
Ion Ratio Lower Upper
149 100
167 24.9 20.6 31.0
279 3.6 3.3 4.9









#85
Di-n-octyl phthalate
Concen: 6.95 ng
RT: 22.086 min Scan# 3264
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:149 Resp: 181238
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 13.8 9.8 14.6

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Manual Integrations
APPROVED


mohammad
3/30/2021 12:41:36 PM



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.515 min Scan# 3507
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37


Tgt Ion:264 Resp: 506944
Ion Ratio Lower Upper
264 100
260 22.9 19.6 29.4
265 21.4 17.8 26.8





#87
Indeno(1,2,3-cd)pyrene
Concen: 7.84 ng
RT: 25.768 min Scan# 3890
Delta R.T. -0.012 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:276 Resp: 290713
Ion Ratio Lower Upper
276 100
138 28.2 22.6 34.0
277 25.2 20.2 30.4

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#88
Benzo(b)fluoranthene
Concen: 7.77 ng
RT: 22.844 min Scan# 3393
Delta R.T. -0.000 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Tgt Ion:252 Resp: 256277
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 11.3 9.0 13.4





#89
Benzo(k)fluoranthene
Concen: 7.42 ng
RT: 22.886 min Scan# 3400
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37




Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:252 Resp: 237731
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 11.5 8.7 13.1

Manual Integrations
APPROVED



mohammad
3/30/2021 12:41:36 PM




#90
Benzo(a)pyrene
Concen: 7.38 ng
RT: 23.415 min Scan# 3490
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37



Tgt Ion:252 Resp: 220441
Ion Ratio Lower Upper
252 100
253 22.3 17.0 25.4
125 13.5 9.5 14.3









#91
Dibenzo(a,h)anthracene
Concen: 7.58 ng
RT: 25.785 min Scan# 3893
Delta R.T. -0.006 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:278 Resp: 240558
Ion Ratio Lower Upper
278 100
139 15.0 14.8 22.2
279 23.4 19.3 28.9

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-03-QT1-2021

Manual Integrations
APPROVED

mohammad
3/30/2021 12:41:36 PM



#92
Benzo(g,h,i)perylene
Concen: 7.76 ng
RT: 26.456 min Scan# 4007
Delta R.T. -0.018 min
Lab File: BM029247.D
Acq: 29 Mar 2021 14:37

Tgt Ion:276 Resp: 241278
Ion Ratio Lower Upper
276 100
277 22.2 18.7 28.1
138 24.2 19.0 28.4

