


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
 Data File : BM029240.D
 Acq On : 29 Mar 2021 10:01
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Mar 29 11:14:33 2021






AutoFind: Scans 2494, 2495, 2496; Background Corrected with Scan 2488

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	56.9	49608	PASS
68	69	0.00	2	2.0	986	PASS
69	198	0.00	100	56.9	49648	PASS
70	69	0.00	2	0.4	214	PASS
127	198	10	80	56.9	49637	PASS
197	198	0.00	2	0.5	468	PASS
198	198	100	100	100.0	87189	PASS
199	198	5	9	7.0	6131	PASS
275	198	10	60	26.4	23025	PASS
365	198	1	100	3.9	3398	PASS
441	198	0.01	100	10.2	8858	PASS
442	442	50	100	100.0	55035	PASS
443	442	15	24	19.4	10663	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
Data File : BM029240.D
Acq On : 29 Mar 2021 10:01
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Mar 29 10:47:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 25 16:05:40 2021
Response via : Initial Calibration



TIC: BM029240.D\data.ms

(77) Benzidine

19.357min (-0.005) 0.00 ng




response 808195

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.10	13.85
183.00	12.20	12.05
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032921\
Data File : BM029240.D
Acq On : 29 Mar 2021 10:01
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Mar 29 10:47:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 25 16:05:40 2021
Response via : Initial Calibration



TIC: BM029240.D\data.ms

(70) Pentachlorophenol (C)

16.763min (-0.011) 0.00 ng

response 102145

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.20	65.30
264.00	63.60	65.93
0.00	0.00	0.00

DDT Breakdown

Instrument :
BNA_M
ClientSampleId :
DFTPP

Date	Instrument Name	DFTPP Data File
03/29/2021	BNA_M	<u>BM029240.D</u>
Compound Name	Response	Retention Time
DDT	345465	20.59
DDD	1571	20.15
DDE	510	19.65
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
2081	347546	0.60

Instrument :
BNA_M
ClientSampleId :
DFTPP