

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032923\
 Data File : BM039215.D
 Acq On : 29 Mar 2023 21:32
 Operator : CG/JU
 Sample : PB151721BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB151721BL

Quant Time: Mar 30 05:56:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 30 05:15:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	196935	20.000	ng	0.00
21) Naphthalene-d8	10.739	136	747140	20.000	ng	0.00
39) Acenaphthene-d10	14.568	164	401169	20.000	ng	0.00
64) Phenanthrene-d10	17.315	188	771840	20.000	ng	0.00
76) Chrysene-d12	21.503	240	643209	20.000	ng	0.00
86) Perylene-d12	23.915	264	715370	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	1906951	145.887	ng	0.00
7) Phenol-d6	7.092	99	2333356	138.012	ng	0.00
23) Nitrobenzene-d5	9.092	82	1423022	92.569	ng	0.00
42) 2,4,6-Tribromophenol	16.062	330	656704	128.956	ng	0.00
45) 2-Fluorobiphenyl	13.198	172	2834264	95.725	ng	0.00
79) Terphenyl-d14	19.945	244	3671588	105.463	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM032923\
Data File : BM039215.D
Acq On : 29 Mar 2023 21:32
Operator : CG/JU
Sample : PB151721BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB151721BL

Quant Time: Mar 30 05:56:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 30 05:15:45 2023
Response via : Initial Calibration

