


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040119\
 Data File : BM019645.D
 Acq On : 01 Apr 2019 08:34
 Operator : JU/SJ
 Sample : DFTPP60
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP60

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Apr 03 07:04:07 2019



AutoFind: Scans 2453, 2454, 2455; Background Corrected with Scan 2447

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.0	57272	PASS
68	69	0.00	2	1.9	1716	PASS
69	198	0.00	100	52.7	88886	PASS
70	69	0.00	2	0.6	535	PASS
127	198	10	80	54.0	91077	PASS
197	198	0.00	2	0.5	869	PASS
198	198	100	100	100.0	168554	PASS
199	198	5	9	6.8	11408	PASS
275	198	10	60	31.0	52304	PASS
365	198	1	100	5.1	8524	PASS
441	198	0.01	100	13.8	23322	PASS
442	198	50	100	94.6	159530	PASS
443	442	15	24	18.8	29922	PASS

Scan 2447 (17.380 min): BM019645.D
DFTPP60

m/z	Abundance
39.90	483.0
43.00	301.0
55.00	305.0
69.00	332.0
73.00	934.0
207.00	1380.0
281.10	302.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

m/z	Abundance
37.00	396.0
38.00	1024.0
39.00	5363.0
40.00	738.0
41.00	455.0
43.90	347.0
49.10	526.0
50.10	16424.0
51.00	65320.0
52.00	3380.0
54.90	616.0
56.00	2482.0
57.00	5137.0
58.00	336.0
61.00	1132.0
62.00	1193.0
63.00	3630.0
64.00	604.0
65.10	1790.0
68.10	1887.0
69.00	99840.0
69.90	576.0
73.10	1675.0
74.00	10247.0
75.00	15853.0
76.00	5296.0
77.10	108400.0
78.00	7945.0
79.00	7874.0
80.00	5398.0
81.00	7217.0
82.00	2080.0
83.00	1985.0
84.00	565.0
85.00	1364.0
86.00	2306.0
87.00	872.0
88.00	374.0
91.00	1627.0
92.00	1948.0
93.00	11809.0
94.00	815.0
95.00	417.0
96.00	827.0
97.10	416.0
98.00	9346.0
99.00	8243.0
99.90	741.0
101.00	3884.0
102.00	391.0
102.90	1484.0
103.90	2143.0
105.00	2517.0
105.90	734.0
107.00	29120.0
108.00	4456.0
108.90	862.0
109.20	840.0
110.00	49584.0
111.00	7445.0
112.00	1050.0
113.00	454.0
116.00	1524.0
117.00	25992.0
118.00	2030.0
119.00	310.0
120.00	440.0
122.00	2154.0
123.00	3053.0
124.00	1459.0
125.00	1226.0
126.20	506.0
127.00	93496.0
128.00	6986.0
129.00	36840.0
130.00	2808.0
131.00	703.0
132.90	433.0
134.00	943.0
135.00	2798.0
136.00	1348.0
137.00	2006.0
140.00	623.0
141.00	5225.0
142.00	1632.0
142.90	1121.0
144.00	353.0
145.00	404.0
146.10	1088.0
147.00	2501.0
148.00	6690.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

149.00	1668.0
150.00	472.0
151.10	543.0
151.80	605.0
153.00	1657.0
154.00	1339.0
155.00	3411.0
156.00	3775.0
157.00	949.0
157.90	942.0
159.00	666.0
160.00	1582.0
161.00	2132.0
162.00	631.0
164.00	321.0
164.90	2017.0
165.90	1439.0
167.00	10563.0
168.00	6978.0
168.90	1158.0
170.00	307.0
170.70	338.0
172.00	1002.0
173.00	1195.0
174.00	2349.0
175.00	3671.0
175.90	991.0
176.90	2195.0
178.00	852.0
179.00	7178.0
180.00	5726.0
181.00	2009.0
181.90	379.0
183.80	540.0
185.00	3748.0
186.00	26080.0
187.00	8002.0
188.00	1343.0
189.00	2177.0
191.00	974.0
192.00	2374.0
193.00	2319.0
194.00	659.0
194.90	548.0
195.90	5063.0
197.10	1049.0
197.90	160384.0
198.90	11249.0
200.00	1117.0
201.40	675.0
203.00	1554.0
204.00	7009.0
205.00	11149.0
206.00	45280.0
207.00	7427.0
207.90	2060.0
209.00	688.0
210.20	1123.0
210.90	1962.0
211.80	413.0
214.90	561.0
216.00	833.0
216.90	12561.0
217.90	1792.0
221.00	6073.0
221.80	1848.0
223.00	2557.0
224.00	22896.0
225.00	6480.0
226.00	729.0
227.00	11013.0
227.90	1710.0
229.00	2516.0
231.00	906.0
232.80	321.0
234.00	652.0
234.90	856.0
236.00	540.0
237.00	1548.0
239.10	431.0
240.00	414.0
240.90	694.0
242.00	1577.0
243.00	1442.0
244.00	18472.0
245.00	3283.0
246.00	4111.0
246.90	754.0
248.90	570.0
253.00	474.0
254.20	974.0
255.00	102632.0
256.00	14818.0
256.90	1092.0
258.00	6074.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

259.00	1264.0
265.00	2697.0
266.00	481.0
271.90	350.0
273.00	3435.0
274.00	7680.0
275.00	47296.0
275.90	6521.0
277.00	3944.0
277.90	1144.0
281.00	398.0
282.90	551.0
284.10	374.0
284.90	767.0
292.90	913.0
294.90	408.0
296.00	15003.0
296.90	1673.0
301.00	304.0
303.00	1715.0
304.00	517.0
313.90	718.0
315.00	1672.0
316.00	752.0
321.00	452.0
323.00	4013.0
323.90	862.0
327.00	677.0
328.00	634.0
331.90	323.0
333.00	431.0
334.00	2899.0
335.00	847.0
341.00	547.0
345.90	916.0
352.00	1162.0
353.10	914.0
354.10	1417.0
355.00	439.0
365.00	6936.0
365.90	1026.0
372.00	2434.0
373.00	438.0
383.10	640.0
401.90	970.0
403.00	1351.0
404.10	500.0
421.00	956.0
421.90	841.0
423.00	7207.0
424.00	2734.0
425.00	571.0
441.00	16992.0
442.00	112640.0
443.00	21928.0
444.00	2314.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

m/z	Abundance
36.80	374.0
38.00	1137.0
39.00	5859.0
40.00	506.0
41.00	454.0
43.00	359.0
43.90	305.0
49.00	678.0
50.00	18816.0
51.10	66224.0
52.00	3705.0
55.00	672.0
55.90	2389.0
57.00	5044.0
61.10	1095.0
62.00	1400.0
63.00	3651.0
64.00	499.0
65.10	1872.0
68.10	2006.0
69.00	105776.0
70.00	680.0
71.10	302.0
73.10	1722.0
74.00	11506.0
75.00	16528.0
76.00	5904.0
77.00	117304.0
78.00	8108.0
79.00	7640.0
80.00	5635.0
81.00	8505.0
82.00	1924.0
83.00	1849.0
83.90	381.0
85.00	1790.0
86.00	2050.0
87.00	1089.0
87.90	374.0
91.00	2218.0
92.00	2345.0
93.00	14112.0
94.00	893.0
96.00	941.0
97.00	415.0
98.00	10297.0
99.00	7833.0
99.90	805.0
101.00	4198.0
103.00	1384.0
104.00	3070.0
105.00	2765.0
106.00	782.0
107.00	33488.0
108.00	5368.0
108.90	867.0
110.00	56224.0
111.00	8497.0
111.90	886.0
115.00	307.0
116.00	2195.0
117.00	28152.0
118.00	1969.0
119.00	315.0
120.00	301.0
121.00	303.0
122.00	2189.0
123.00	3181.0
124.00	1400.0
124.90	1383.0
127.00	110888.0
128.00	8911.0
129.00	43624.0
130.00	3767.0
131.00	941.0
131.90	388.0
134.00	1258.0
135.00	3602.0
136.00	1460.0
136.90	2022.0
137.90	562.0
139.00	411.0
140.00	571.0
141.00	6822.0
142.00	1966.0
143.00	1246.0
143.90	509.0
144.90	451.0
146.00	1197.0
147.00	3465.0
148.00	7856.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

148.90	1483.0
150.00	530.0
151.00	744.0
151.70	585.0
153.00	1913.0
154.00	1339.0
155.00	3144.0
156.00	4830.0
157.10	1001.0
157.90	1182.0
158.90	826.0
159.90	1773.0
161.00	2912.0
162.00	754.0
163.00	336.0
165.00	2247.0
166.00	2100.0
167.00	12186.0
168.00	7556.0
169.00	1180.0
170.00	345.0
171.00	866.0
171.90	1197.0
172.90	1574.0
174.00	2130.0
175.00	4965.0
175.90	1101.0
176.90	2439.0
177.90	1037.0
179.00	9781.0
180.00	6350.0
181.00	2869.0
182.00	433.0
183.90	708.0
185.00	4195.0
186.00	32816.0
187.00	9831.0
187.90	1049.0
189.00	2288.0
190.00	367.0
191.00	1128.0
191.90	2955.0
193.00	3288.0
194.00	948.0
194.90	392.0
196.00	6728.0
197.10	1559.0
197.90	205120.0
198.90	13570.0
200.00	1522.0
201.50	937.0
202.90	1616.0
204.00	8943.0
205.00	13755.0
206.00	60856.0
207.00	8752.0
208.00	2442.0
209.00	740.0
210.00	947.0
211.00	2144.0
215.00	696.0
215.90	1062.0
217.00	16218.0
218.00	2026.0
221.00	9024.0
222.00	1788.0
223.00	3509.0
224.00	32136.0
225.00	7854.0
226.00	1154.0
227.00	15275.0
228.00	2524.0
229.00	2855.0
229.90	464.0
230.90	1071.0
233.90	1030.0
235.00	1317.0
236.00	636.0
236.90	2004.0
238.90	633.0
239.90	657.0
241.00	782.0
241.90	2028.0
243.00	1638.0
244.00	26096.0
245.00	3545.0
246.00	6319.0
246.90	1242.0
248.90	909.0
253.00	501.0
254.10	1192.0
255.00	141952.0
256.00	20432.0
257.00	1713.0
258.00	8429.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

259.00	1530.0
265.00	3716.0
266.00	456.0
272.00	433.0
273.00	4375.0
274.00	10354.0
275.00	62912.0
276.00	8809.0
276.90	5464.0
278.00	1322.0
281.10	473.0
282.90	642.0
284.00	378.0
285.10	926.0
293.00	1489.0
293.90	554.0
294.90	519.0
296.00	20072.0
297.00	2709.0
303.00	2435.0
304.00	515.0
307.90	342.0
314.00	770.0
315.00	2399.0
316.00	1421.0
320.90	739.0
323.00	6793.0
324.00	1142.0
326.90	1394.0
327.90	695.0
331.90	472.0
332.90	533.0
334.00	4473.0
335.00	1053.0
341.10	886.0
345.90	1458.0
347.00	354.0
352.00	2322.0
353.00	1425.0
354.00	2251.0
354.90	487.0
365.00	9944.0
365.90	1303.0
371.00	576.0
372.00	3172.0
373.00	856.0
382.90	946.0
389.90	471.0
391.90	342.0
402.00	1319.0
403.00	2195.0
404.10	730.0
421.00	1848.0
422.00	1716.0
423.00	12058.0
424.00	3563.0
425.10	408.0
441.00	27928.0
442.00	195776.0
443.00	36896.0
444.00	3575.0

Instrument :

BNA_M

ClientSampleId :

DFTPP60

m/z	Abundance
38.10	534.0
39.00	3222.0
40.00	519.0
40.90	306.0
49.00	436.0
50.00	10864.0
51.00	40272.0
52.00	2286.0
55.00	535.0
56.00	1574.0
57.00	3191.0
61.00	794.0
62.00	998.0
63.00	1955.0
63.90	338.0
65.00	1025.0
66.90	349.0
68.00	1255.0
69.00	62040.0
70.00	351.0
73.00	1484.0
74.00	6762.0
75.00	10088.0
76.10	3195.0
77.10	69672.0
78.00	4465.0
79.00	5198.0
80.00	3730.0
81.00	5191.0
82.00	1271.0
82.90	1185.0
84.90	1176.0
86.00	1468.0
87.00	637.0
90.90	1290.0
92.00	1000.0
93.00	7643.0
93.90	507.0
95.90	322.0
98.00	6257.0
99.00	5155.0
99.90	471.0
101.00	2623.0
103.00	991.0
104.00	1785.0
104.90	1887.0
106.00	579.0
107.00	19392.0
108.00	3363.0
109.00	612.0
110.00	34304.0
111.00	5450.0
112.00	825.0
113.00	304.0
116.00	1171.0
117.00	18040.0
117.90	1240.0
120.00	338.0
121.90	1237.0
123.00	2418.0
123.90	1155.0
125.00	1055.0
127.00	68848.0
128.00	4992.0
129.00	27512.0
130.00	2224.0
130.90	562.0
131.90	315.0
134.00	819.0
135.00	2304.0
136.00	907.0
137.10	1448.0
139.90	463.0
140.90	3619.0
142.00	1358.0
143.00	800.0
144.10	322.0
145.90	689.0
147.00	1971.0
148.00	5050.0
149.00	1402.0
151.00	662.0
151.80	474.0
152.90	1288.0
154.00	749.0
155.00	2290.0
156.00	3093.0
157.00	782.0
157.90	520.0
159.00	735.0
160.00	997.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

161.00	1997.0
161.80	651.0
165.00	1701.0
166.00	1374.0
167.00	8517.0
168.00	4811.0
169.00	722.0
171.00	394.0
172.00	752.0
172.90	865.0
174.00	1352.0
175.10	3000.0
175.90	838.0
177.00	1790.0
178.00	702.0
179.00	5971.0
179.90	3832.0
180.90	1972.0
182.00	381.0
183.90	510.0
185.00	3092.0
186.00	22576.0
187.00	6101.0
188.00	917.0
189.00	1988.0
189.80	387.0
191.00	778.0
192.00	1872.0
193.00	2554.0
194.10	368.0
194.90	408.0
196.00	3992.0
197.90	140160.0
198.90	9406.0
200.00	727.0
201.30	892.0
202.10	333.0
203.00	1589.0
204.00	6007.0
205.00	9233.0
206.00	40216.0
207.00	5986.0
207.90	1927.0
208.90	513.0
210.00	778.0
211.00	1602.0
211.80	568.0
214.80	447.0
216.00	850.0
217.00	10937.0
217.90	1188.0
221.00	5574.0
222.00	1627.0
223.00	2735.0
224.00	21840.0
225.00	5994.0
226.00	862.0
227.00	10487.0
228.00	1483.0
229.00	2013.0
231.10	813.0
234.00	774.0
234.90	907.0
236.00	402.0
237.00	962.0
239.10	372.0
239.90	354.0
240.80	525.0
242.00	1487.0
243.00	1276.0
244.00	17424.0
245.00	2107.0
246.00	4279.0
247.00	824.0
248.00	304.0
249.00	715.0
253.00	505.0
254.20	991.0
255.00	99824.0
256.00	13775.0
257.00	1512.0
258.00	6475.0
259.00	964.0
265.00	3019.0
265.90	435.0
270.90	324.0
272.90	3217.0
274.00	7837.0
275.00	46704.0
276.00	6044.0
277.00	3731.0
278.00	822.0
281.00	381.0
283.00	615.0
284.00	308.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

285.10	551.0
288.90	300.0
293.00	939.0
295.10	413.0
296.00	15540.0
296.90	2077.0
303.00	1733.0
304.00	429.0
313.90	763.0
314.90	2078.0
315.90	1018.0
316.90	344.0
321.00	705.0
323.00	5273.0
323.90	897.0
327.00	1101.0
328.00	524.0
332.00	391.0
333.00	503.0
334.00	3240.0
334.90	967.0
341.00	810.0
345.90	1459.0
352.00	1856.0
353.00	1095.0
354.00	1901.0
354.90	440.0
365.00	8694.0
366.00	1369.0
371.00	404.0
372.00	3526.0
373.00	561.0
383.00	1143.0
390.00	442.0
402.10	1095.0
403.00	1622.0
404.10	699.0
421.00	1498.0
422.00	1706.0
423.00	10420.0
424.10	3502.0
425.10	351.0
441.00	25048.0
442.00	170176.0
443.00	30944.0
444.10	3087.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60