

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM040220\
 Data File : BM025892.D
 Acq On : 01 Apr 2020 17:58
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV83

Quant Time: Apr 01 18:30:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM040220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 17:54:54 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
3	Naphthalene	1.176	1.067	9.3	97	0.00
4 SURR	2-Methylnaphthalene-d10	0.711	0.693	2.5	105	0.00
5	2-Methylnaphthalene	0.823	0.788	4.3	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
7	Acenaphthylene	1.710	1.766	-3.3	109	0.00
8 C	Acenaphthene	1.419	1.417	0.1	103	0.00
9	Fluorene	1.761	1.758	0.2	104	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
11	Pentachlorophenol	0.098	0.089	9.2	108	0.00
12	Phenanthrene	1.260	1.178	6.5	104	0.00
13	Anthracene	1.113	1.021	8.3	103	0.00
14 SURR	Fluoranthene-d10	1.253	1.247	0.5	112	0.00
15 C	Fluoranthene	1.557	1.425	8.5	103	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
17	Pyrene	1.433	1.379	3.8	103	0.00
18	Benzo(a)anthracene	1.430	1.326	7.3	100	0.00
19	Chrysene	1.561	1.485	4.9	102	0.00
20 I	Perylene-d12	1.000	1.000	0.0	82	0.00
21	Benzo(b)fluoranthene	1.472	1.454	1.2	100	0.00
22	Benzo(k)fluoranthene	1.578	1.497	5.1	97	0.00
23 C	Benzo(a)pyrene	1.316	1.337	-1.6	104	0.00
24	Indeno(1,2,3-cd)pyrene	1.762	1.653	6.2	95	0.00
25	Dibenzo(a,h)anthracene	1.471	1.385	5.8	95	0.00
26	Benzo(a,h,i)perylene	1.490	1.398	6.2	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0