

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM040420\
 Data File : BM025937.D
 Acq On : 04 Apr 2020 06:12
 Operator : CG/JU
 Sample : L2062-06RE
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F3E05RE

Manual Integrations
 APPROVED

mohammad

Quant Time: Apr 04 07:01:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM040220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 04 05:10:01 2020
 Response via : Initial Calibration

4/6/2020 1:41:53 PM

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	5537	0.40	ng/ul	0.00
2) Naphthalene-d8	10.32	136	24553	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.20	164	16677	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.94	188	38015m	0.40	ng/ul	0.00
16) Chrysene-d12	21.14	240	43201	0.40	ng/ul	0.00
20) Perylene-d12	23.29	264	36942	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.92	152	11065	0.25	ng/ul	0.00
14) Fluoranthene-d10	18.98	212	36069	0.30	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.37	128	4099	0.057	ng/ul#	95
5) 2-Methylnaphthalene	12.00	142	1702	0.034	ng/ul	98
7) Acenaphthylene	13.92	152	132431	1.857	ng/ul	98
8) Acenaphthene	14.26	153	12691	0.214	ng/ul	99
9) Fluorene	15.26	166	25616	0.349	ng/ul	100
11) Pentachlorophenol	16.61	266	464	0.050	ng/ul	94
12) Phenanthrene	16.99	178	85865	0.717	ng/ul	99
13) Anthracene	17.08	178	2446916	23.126	ng/ul	99
15) Fluoranthene	19.01	202	4643176	31.376	ng/ul	99
17) Pyrene	19.38	202	6846417	44.251	ng/ul	97
18) Benzo(a)anthracene	21.13	228	3828613	24.793	ng/ul	97
19) Chrysene	21.19	228	6416491	38.069	ng/ul	98
21) Benzo(b)fluoranthene	22.67	252	5793669m	42.615	ng/ul	
22) Benzo(k)fluoranthene	22.70	252	2224955m	15.263	ng/ul	
23) Benzo(a)pyrene	23.21	252	2554858	21.028	ng/ul	97
24) Indeno(1,2,3-cd)pyrene	25.41	276	986871	6.065	ng/ul#	99
25) Dibenzo(a,h)anthracene	25.41	278	295346	2.174	ng/ul	98
26) Benzo(a,h,i)perylene	26.05	276	570556	4.145	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM040420\
Data File : BM025937.D
Acq On : 04 Apr 2020 06:12
Operator : CG/JU
Sample : L2062-06RE
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
F3E05RE

Manual Integrations
APPROVED

mohammad

Quant Time: Apr 04 07:01:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM040220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 04 05:10:01 2020
Response via : Initial Calibration

4/6/2020 1:41:53 PM

