


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM040420\
 Data File : BM025905.D
 Acq On : 03 Apr 2020 08:17
 Operator : CG/JU
 Sample : DFTPP95
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP95

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM040220.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Apr 02 10:49:07 2020



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.8	155701	PASS
68	69	0.00	2	2.1	3458	PASS
69	198	0.00	100	43.0	168213	PASS
70	69	0.00	2	0.6	947	PASS
127	198	10	80	50.0	195648	PASS
197	198	0.00	2	0.7	2855	PASS
198	198	100	100	100.0	390954	PASS
199	198	5	9	6.8	26538	PASS
275	198	10	60	30.8	120480	PASS
365	198	1	100	3.8	14903	PASS
441	198	0.01	100	12.1	47197	PASS
442	198	50	100	81.2	317589	PASS
443	442	15	24	19.7	62450	PASS

Scan 2464 (17.380 min): BM025905.D
DFTPP95

m/z	Abundance
40.10	1003.0
44.00	1763.0
48.90	495.0
73.10	409.0
83.90	304.0
133.10	310.0
192.90	303.0
207.10	1667.0
207.90	488.0
281.10	409.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

m/z	Abundance
37.00	506.0
38.10	1856.0
39.10	9549.0
40.00	1211.0
44.00	1642.0
45.00	495.0
49.00	1139.0
50.10	34976.0
51.10	123936.0
52.10	6731.0
55.00	979.0
56.00	4450.0
57.00	8781.0
58.00	514.0
61.00	1573.0
62.00	1822.0
63.10	4734.0
64.00	988.0
65.00	1958.0
68.10	2685.0
69.00	132928.0
70.00	731.0
73.00	2673.0
74.10	12483.0
75.00	19832.0
76.00	6947.0
77.10	136640.0
78.10	10056.0
79.00	9112.0
80.00	7109.0
81.00	10532.0
82.00	2384.0
83.00	2419.0
84.00	450.0
85.10	1696.0
86.10	3025.0
87.00	1301.0
87.90	420.0
91.00	2209.0
92.00	2980.0
93.00	15572.0
94.00	1016.0
95.10	336.0
96.00	715.0
98.00	11416.0
99.00	11762.0
100.10	810.0
101.00	5323.0
103.00	1886.0
104.00	3439.0
105.00	3378.0
106.10	1154.0
107.00	42504.0
108.00	5992.0
109.00	1154.0
110.00	73432.0
111.00	11345.0
112.10	1762.0
112.90	705.0
116.00	2688.0
117.00	34632.0
118.00	2662.0
119.00	354.0
119.90	724.0
121.00	319.0
122.00	2455.0
123.00	3920.0
123.90	1813.0
125.00	1785.0
126.20	749.0
127.00	140160.0
128.00	12212.0
129.00	55264.0
130.00	5212.0
131.00	1102.0
132.00	524.0
133.00	768.0
134.00	1816.0
135.00	4705.0
136.10	1421.0
137.00	3184.0
137.90	364.0
140.00	562.0
141.00	7295.0
142.00	2502.0
143.00	1491.0
144.00	412.0
144.90	311.0
145.90	1064.0
147.00	4329.0
148.00	8292.0
149.00	2455.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

150.10	655.0
151.00	1038.0
151.80	495.0
153.00	2019.0
153.90	1618.0
155.00	4017.0
156.00	5491.0
157.00	1037.0
158.00	1244.0
158.90	1057.0
160.00	2349.0
161.00	3235.0
162.00	1041.0
163.00	318.0
164.10	329.0
165.00	3089.0
166.00	1708.0
167.00	14259.0
168.00	9177.0
169.00	1698.0
169.90	613.0
171.00	648.0
172.00	1264.0
173.00	1348.0
174.00	2283.0
175.00	5593.0
176.00	1735.0
176.90	2496.0
178.00	848.0
179.00	10010.0
180.00	8974.0
181.00	2817.0
182.00	402.0
183.10	358.0
184.00	906.0
185.00	4919.0
186.00	34792.0
187.10	9483.0
188.00	1363.0
189.00	2293.0
189.90	668.0
191.10	1486.0
192.00	3087.0
193.00	3703.0
193.90	846.0
195.10	534.0
196.00	7220.0
197.20	1861.0
198.00	250624.0
199.00	17960.0
200.00	1353.0
201.30	934.0
203.00	1977.0
204.00	10169.0
205.00	15158.0
206.00	63624.0
207.00	10625.0
208.00	2652.0
209.10	1171.0
210.10	1130.0
211.00	2537.0
212.10	565.0
213.00	313.0
214.90	576.0
216.00	1178.0
217.00	16308.0
218.00	2113.0
221.10	12080.0
222.00	2347.0
223.10	3367.0
224.00	35240.0
225.00	8467.0
226.00	1261.0
227.00	15363.0
228.00	2697.0
229.00	3199.0
230.10	439.0
231.00	1312.0
234.00	1079.0
235.00	1310.0
235.90	907.0
237.00	2049.0
239.00	589.0
240.10	522.0
241.00	1012.0
242.00	2128.0
243.00	2163.0
244.00	29792.0
245.10	4099.0
246.00	5615.0
246.90	911.0
249.00	1143.0
250.00	323.0
251.00	396.0
253.00	641.0
254.10	1302.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

255.00	153984.0
256.00	23640.0
257.00	2315.0
258.00	8341.0
259.00	1280.0
265.00	3268.0
265.90	664.0
271.00	459.0
272.00	503.0
273.00	4170.0
274.00	12975.0
275.00	71008.0
276.00	10340.0
277.00	5891.0
278.00	1544.0
281.10	935.0
283.00	861.0
284.00	428.0
284.90	937.0
292.00	335.0
293.00	1490.0
294.00	540.0
295.00	471.0
296.00	19312.0
297.00	2918.0
302.00	406.0
303.00	1868.0
304.00	499.0
307.90	336.0
310.00	425.0
314.00	1192.0
315.00	2161.0
316.00	1034.0
321.20	491.0
323.00	5700.0
324.00	953.0
326.90	1444.0
328.10	372.0
332.00	414.0
333.00	610.0
334.00	3585.0
335.10	863.0
341.10	753.0
342.10	325.0
345.90	1135.0
346.90	341.0
352.00	1768.0
353.00	1294.0
354.10	1627.0
355.10	946.0
365.10	8222.0
365.90	1003.0
371.10	317.0
372.10	2829.0
373.00	924.0
383.10	793.0
389.90	361.0
390.90	330.0
400.90	331.0
402.00	1170.0
403.10	1129.0
404.10	527.0
405.00	366.0
421.10	838.0
422.00	1520.0
423.10	8087.0
424.00	3315.0
425.10	484.0
441.10	21632.0
442.10	144256.0
443.10	28696.0
444.00	3205.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

m/z	Abundance
37.10	1070.0
38.10	2825.0
39.10	14198.0
40.00	1237.0
41.00	382.0
44.00	1836.0
44.90	418.0
49.10	1446.0
50.10	51624.0
51.10	188224.0
52.10	8608.0
53.00	473.0
55.00	821.0
56.00	6081.0
57.00	12927.0
58.00	438.0
61.00	1889.0
62.00	2415.0
63.00	7472.0
64.00	1081.0
65.00	3697.0
68.10	4531.0
69.00	199232.0
70.00	1223.0
73.10	3289.0
74.10	19216.0
75.00	32216.0
76.10	9614.0
77.10	209728.0
78.10	14177.0
79.00	14914.0
80.00	11515.0
81.00	15104.0
82.00	4275.0
83.00	4186.0
84.00	525.0
85.00	2509.0
86.00	4640.0
87.00	2591.0
87.90	662.0
89.20	471.0
91.00	3550.0
92.00	4146.0
93.00	26336.0
94.00	1731.0
96.00	1286.0
98.00	18240.0
99.00	16103.0
100.00	1229.0
101.00	7791.0
102.00	491.0
103.00	2812.0
104.00	4950.0
105.00	5272.0
106.10	1947.0
107.00	67160.0
108.10	9939.0
109.10	1707.0
110.00	122176.0
111.00	18864.0
112.00	2192.0
113.20	764.0
116.10	3498.0
117.00	54312.0
118.00	4763.0
119.00	710.0
120.00	805.0
120.90	478.0
122.00	4124.0
123.00	6904.0
123.90	3101.0
125.00	2483.0
126.20	1619.0
127.00	238144.0
128.00	18104.0
129.00	89264.0
130.00	8196.0
131.00	1540.0
132.10	707.0
132.90	559.0
134.00	3061.0
135.00	6870.0
136.00	2797.0
137.00	3845.0
137.90	713.0
139.10	300.0
140.00	1240.0
141.00	12195.0
142.00	4380.0
143.10	2295.0
143.90	990.0
145.00	613.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

146.00	2050.0
147.00	7677.0
148.00	14704.0
149.00	3288.0
149.90	592.0
150.10	594.0
151.00	1474.0
152.10	861.0
153.00	3406.0
154.00	2643.0
155.00	6408.0
156.00	9431.0
157.10	2333.0
158.00	2686.0
159.10	1747.0
160.00	3820.0
161.00	5364.0
161.90	1635.0
162.90	454.0
164.00	405.0
165.00	4818.0
166.00	2782.0
167.00	23032.0
168.00	13061.0
169.00	2382.0
170.00	1251.0
170.80	1109.0
172.00	1709.0
173.00	2940.0
174.00	5358.0
175.00	9100.0
176.10	2408.0
177.00	4045.0
178.00	1441.0
179.00	17584.0
180.00	12809.0
181.00	5078.0
182.00	771.0
183.00	374.0
184.00	1287.0
185.00	8072.0
186.00	62728.0
187.00	19080.0
188.10	2288.0
189.00	4375.0
189.90	902.0
191.00	2183.0
192.00	6195.0
193.00	6541.0
194.00	1387.0
194.90	750.0
196.00	12167.0
197.10	3518.0
198.00	470848.0
199.00	32912.0
200.00	2507.0
201.60	2423.0
203.00	3119.0
204.00	17000.0
205.00	28264.0
206.10	119968.0
207.00	17656.0
208.00	4063.0
209.00	1737.0
210.20	1864.0
211.00	4442.0
212.90	773.0
214.00	355.0
215.00	1287.0
216.10	2573.0
217.00	30088.0
218.00	4560.0
219.00	481.0
221.00	21776.0
221.70	4850.0
223.00	8796.0
224.00	66728.0
225.10	16432.0
226.10	2347.0
227.00	27576.0
228.00	4087.0
229.00	5156.0
230.00	561.0
231.00	2517.0
232.90	600.0
234.00	1897.0
235.00	2418.0
236.00	1469.0
237.00	3111.0
238.00	552.0
239.00	1107.0
240.00	936.0
241.00	1901.0
242.00	3405.0
243.10	4186.0
244.10	53328.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

245.00	7152.0
246.00	11012.0
247.00	2199.0
247.90	479.0
248.90	2402.0
250.00	409.0
251.00	394.0
251.90	478.0
252.90	1181.0
254.20	3344.0
255.00	303360.0
256.00	43616.0
257.00	3764.0
258.00	16432.0
259.00	2495.0
260.10	627.0
261.00	493.0
264.00	483.0
265.00	6681.0
265.90	1155.0
270.90	778.0
272.10	800.0
273.00	8705.0
274.00	22784.0
275.00	143168.0
276.00	18312.0
277.00	11392.0
278.00	2168.0
279.00	542.0
281.10	1066.0
282.00	530.0
283.10	1453.0
284.00	978.0
285.10	2096.0
285.90	461.0
289.00	439.0
292.10	500.0
293.00	2168.0
293.90	658.0
295.10	1156.0
296.00	37016.0
297.00	5369.0
298.10	463.0
301.00	703.0
302.00	710.0
303.00	4694.0
304.10	1099.0
308.00	578.0
310.00	659.0
314.10	2111.0
315.00	4015.0
316.10	2359.0
316.90	396.0
321.10	946.0
322.20	452.0
323.10	12436.0
324.10	2350.0
326.90	2301.0
328.00	1247.0
332.00	708.0
333.00	1429.0
334.10	8091.0
335.10	2339.0
341.00	1666.0
342.10	569.0
345.90	2704.0
346.90	452.0
352.10	3360.0
353.10	2305.0
354.00	4046.0
355.10	981.0
359.20	310.0
365.00	17656.0
366.00	2082.0
370.00	428.0
371.10	734.0
372.10	6337.0
373.00	1614.0
383.00	1286.0
384.00	388.0
390.10	818.0
391.00	446.0
392.00	336.0
401.10	350.0
402.00	2673.0
403.00	3602.0
404.00	1263.0
405.00	314.0
421.00	2712.0
422.10	2950.0
423.10	18952.0
424.10	5992.0
425.10	894.0
440.10	418.0
441.10	51872.0
442.10	353728.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

443.10	71848.0
444.10	6827.0
445.00	389.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

m/z	Abundance
35.90	321.0
37.00	838.0
38.00	1957.0
39.10	12867.0
40.00	1325.0
41.00	413.0
44.00	1848.0
45.10	374.0
49.10	1704.0
50.10	40976.0
51.10	154944.0
52.00	8377.0
52.90	361.0
54.90	849.0
56.00	4580.0
57.00	10541.0
58.00	486.0
61.00	2028.0
62.00	2370.0
63.00	6152.0
64.00	911.0
65.00	2926.0
66.10	303.0
68.10	3159.0
69.00	172480.0
70.00	888.0
73.00	2833.0
74.10	16576.0
75.00	26488.0
76.10	8987.0
77.10	183872.0
78.10	12378.0
79.00	12635.0
80.00	10093.0
81.00	13627.0
82.00	3453.0
83.00	3054.0
84.00	648.0
85.00	1912.0
86.00	3624.0
87.00	1884.0
88.00	565.0
89.00	486.0
91.00	3225.0
92.00	3738.0
93.00	21872.0
93.90	1485.0
96.00	924.0
97.10	447.0
98.00	16584.0
99.00	14357.0
100.00	886.0
101.00	6769.0
101.90	466.0
103.00	2126.0
104.00	4977.0
105.00	4294.0
106.10	1636.0
107.00	57488.0
108.00	9218.0
109.20	1944.0
110.00	107344.0
111.00	16185.0
112.00	1596.0
113.00	543.0
116.00	3062.0
117.00	48744.0
118.00	3455.0
118.90	356.0
120.00	790.0
120.90	305.0
122.00	5128.0
123.00	6688.0
124.00	3173.0
125.00	2945.0
127.00	208640.0
128.00	15044.0
129.00	80232.0
130.00	7174.0
131.00	1246.0
132.00	698.0
133.00	628.0
134.00	2427.0
135.10	6050.0
135.90	2276.0
137.00	3938.0
138.10	754.0
139.00	387.0
140.00	752.0
141.00	12009.0
141.90	3189.0
143.00	2609.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

143.90	528.0
145.00	604.0
146.00	1944.0
147.00	6238.0
148.00	12335.0
149.00	2423.0
149.90	661.0
151.20	1249.0
152.00	854.0
153.00	3770.0
154.00	2633.0
155.00	5449.0
156.10	7872.0
157.10	1745.0
158.00	2383.0
159.00	1246.0
160.00	3702.0
161.00	5910.0
161.90	1564.0
163.00	523.0
164.10	622.0
165.00	4304.0
166.00	3629.0
167.00	21112.0
168.00	11799.0
169.00	1849.0
170.00	746.0
171.10	788.0
172.00	1747.0
173.00	2786.0
174.00	5015.0
175.00	8945.0
176.00	2143.0
177.00	4187.0
178.10	1361.0
179.00	17000.0
180.00	12514.0
181.00	5958.0
182.00	826.0
182.80	527.0
184.00	1451.0
185.00	8385.0
186.00	61016.0
187.00	16968.0
188.10	1944.0
189.00	3884.0
190.10	862.0
191.00	1757.0
192.00	4995.0
193.00	5997.0
193.90	1083.0
195.10	820.0
196.00	14150.0
197.20	3186.0
198.00	451392.0
199.00	28744.0
200.10	2061.0
201.40	2293.0
202.90	2816.0
204.10	16520.0
205.00	27480.0
206.10	118288.0
207.10	15621.0
208.00	4241.0
208.90	1470.0
211.00	4603.0
212.90	323.0
215.00	1319.0
216.00	2579.0
217.00	29896.0
218.00	4071.0
220.10	823.0
221.00	20248.0
221.80	4696.0
223.10	7157.0
224.00	65504.0
225.00	15377.0
226.00	2658.0
227.00	26712.0
228.00	3598.0
229.00	6330.0
230.00	754.0
231.10	2209.0
232.00	600.0
233.00	434.0
234.00	1700.0
235.00	2410.0
236.00	1355.0
237.00	2861.0
239.00	1152.0
240.00	1065.0
241.00	1578.0
242.00	3462.0
243.10	3811.0
244.10	52240.0
245.00	7863.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

246.00	11019.0
247.00	2223.0
247.90	703.0
249.00	1847.0
249.90	336.0
250.90	641.0
252.00	429.0
253.00	1001.0
254.20	2994.0
255.00	301504.0
256.00	45848.0
257.00	3809.0
258.00	17200.0
259.10	2761.0
260.10	564.0
261.00	611.0
264.00	441.0
265.00	7079.0
266.00	1240.0
270.00	394.0
271.00	899.0
272.00	789.0
273.00	10207.0
274.00	24552.0
275.00	147264.0
276.00	18248.0
277.00	12450.0
278.00	2210.0
279.00	488.0
281.00	914.0
281.90	566.0
283.00	1518.0
284.10	1115.0
285.00	1991.0
286.10	317.0
287.90	361.0
289.00	409.0
290.00	431.0
291.90	581.0
293.00	2850.0
294.00	679.0
295.00	1133.0
296.00	39184.0
297.00	5585.0
301.00	584.0
302.00	757.0
303.10	5283.0
304.10	1076.0
308.10	647.0
309.10	388.0
310.00	328.0
313.10	396.0
314.10	1885.0
315.00	4801.0
316.00	2735.0
317.10	425.0
321.00	1483.0
322.10	1124.0
323.00	14517.0
324.00	2589.0
325.00	447.0
325.90	477.0
327.00	3099.0
328.00	1391.0
332.10	826.0
333.00	1782.0
334.10	8344.0
335.10	2061.0
335.90	462.0
339.10	390.0
341.00	1790.0
342.10	495.0
346.00	2447.0
347.00	406.0
352.10	4787.0
353.10	3289.0
354.10	4050.0
355.10	1331.0
364.20	451.0
365.00	18832.0
366.00	2599.0
369.90	509.0
371.00	1127.0
372.00	6941.0
373.10	1873.0
383.10	1819.0
384.10	685.0
390.10	700.0
391.00	464.0
392.10	427.0
400.90	607.0
402.10	2679.0
403.10	3266.0
404.10	1363.0
405.10	312.0
421.00	3116.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95

422.10	3317.0
423.10	23704.0
424.10	6256.0
425.10	782.0
440.20	410.0
441.10	68088.0
442.10	454784.0
443.10	86808.0
444.10	8403.0
445.00	373.0

Instrument :
BNA_M
ClientSampleId :
DFTPP95