

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
 Data File : BM044910.D
 Acq On : 04 Apr 2024 12:23
 Operator : MA/JU
 Sample : P1948-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AH6

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/04/2024
 Supervised By :mohammad ahmed 04/06/2024

Quant Time: Apr 04 13:08:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 03 23:34:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.648	152	1765	0.400	ng/ul	0.00
4) Naphthalene-d8	10.418	136	5431	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.297	164	3120	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.051	188	7341m	0.400	ng/ul	-0.02
17) Chrysene-d12	21.275	240	7052m	0.400	ng/ul	0.00
23) Perylene-d12	23.496	264	6623m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	9852	4.249	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.046	152	2699	0.367	ng/ul	0.00
18) Fluoranthene-d10	19.094	212	7556	0.424	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.251	88	2566	1.155	ng/ul#	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
Data File : BM044910.D
Acq On : 04 Apr 2024 12:23
Operator : MA/JU
Sample : P1948-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AH6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/04/2024
Supervised By :mohammad ahmed 04/06/2024

Quant Time: Apr 04 13:08:01 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 03 23:34:03 2024
Response via : Initial Calibration

