

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
 Data File : BM044939.D
 Acq On : 05 Apr 2024 08:55
 Operator : MA/JU
 Sample : PB159840BS
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS840

Quant Time: Apr 05 10:56:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 03 23:34:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	1568	0.400	ng/ul	0.00
4) Naphthalene-d8	10.430	136	4847	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.297	164	2770	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.056	188	6276	0.400	ng/ul	-0.01
17) Chrysene-d12	21.263	240	5661	0.400	ng/ul	0.00
23) Perylene-d12	23.490	264	7209	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	1589	0.771	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.035	152	2586	0.394	ng/ul	-0.02
18) Fluoranthene-d10	19.094	212	6758	0.472	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.251	88	4366	2.211	ng/ul#	80
5) Naphthalene	10.479	128	5321	0.414	ng/ul	98
7) 2-Methylnaphthalene	12.112	142	3258	0.424	ng/ul	98
8) 1-Methylnaphthalene	12.327	142	3446	0.411	ng/ul	100
10) Acenaphthylene	14.015	152	5772	0.418	ng/ul	98
11) Acenaphthene	14.358	153	4192	0.424	ng/ul	99
12) Fluorene	15.361	166	4756	0.429	ng/ul	100
14) Pentachlorophenol	16.718	266	1544	1.049	ng/ul	99
15) Phenanthrene	17.098	178	7792	0.441	ng/ul	99
16) Anthracene	17.199	178	7803	0.448	ng/ul	100
19) Fluoranthene	19.127	202	9936	0.491	ng/ul	99
20) Pyrene	19.489	202	10498	0.491	ng/ul	98
21) Benzo(a)anthracene	21.249	228	9573	0.486	ng/ul	100
22) Chrysene	21.301	228	11305	0.460	ng/ul	99
24) Benzo(b)fluoranthene	22.824	252	11730	0.451	ng/ul	98
25) Benzo(k)fluoranthene	22.868	252	13170	0.452	ng/ul	100
26) Benzo(a)pyrene	23.394	252	11376	0.458	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	25.733	276	15434	0.453	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.753	278	12171	0.451	ng/ul	99
29) Benzo(g,h,i)perylene	26.417	276	12922	0.440	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
Data File : BM044939.D
Acq On : 05 Apr 2024 08:55
Operator : MA/JU
Sample : PB159840BS
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS840

Quant Time: Apr 05 10:56:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 03 23:34:03 2024
Response via : Initial Calibration

