

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
 Data File : BM044907.D
 Acq On : 04 Apr 2024 10:35
 Operator : MA/JU
 Sample : P1948-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AH3

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/04/2024
 Supervised By :mohammad ahmed 04/06/2024

Quant Time: Apr 04 11:22:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 03 23:34:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.648	152	1676	0.400	ng/u1	0.00
4) Naphthalene-d8	10.424	136	5084	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.298	164	2851	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.056	188	6651m	0.400	ng/u1	-0.01
17) Chrysene-d12	21.280	240	6453m	0.400	ng/u1	0.00
23) Perylene-d12	23.505	264	5967m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.218	96	8941	4.061	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.046	152	2191	0.318	ng/u1	0.00
18) Fluoranthene-d10	19.094	212	5459	0.335	ng/u1	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.251	88	1209	0.573	ng/u1	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
Data File : BM044907.D
Acq On : 04 Apr 2024 10:35
Operator : MA/JU
Sample : P1948-01
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
C0AH3

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/04/2024
Supervised By :mohammad ahmed 04/06/2024


Quant Time: Apr 04 11:22:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 03 23:34:03 2024
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.573 ng/ul
 RT: 3.251 min Scan# 40
 Delta R.T. -0.004 min
 Lab File: BM044907.D
 Acq: 04 Apr 2024 10:35

Instrument :
 BNA_M
 ClientSampleId :
 C0AH3



Tgt Ion: 88 Resp: 1209
 Ion Ratio Lower Upper
 88 100
 43 42.9 30.6 46.0
 58 47.4 32.6 48.8

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/04/2024
 Supervised By :mohammad ahmed 04/06/2024

