

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
 Data File : BM044913.D
 Acq On : 04 Apr 2024 14:11
 Operator : MA/JU
 Sample : P1849-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E0LY5

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/04/2024
 Supervised By :mohammad ahmed 04/06/2024

Quant Time: Apr 04 15:08:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 03 23:34:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.648	152	2578	0.400	ng/u1	0.00
4) Naphthalene-d8	10.419	136	8148	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.294	164	4683	0.400	ng/u1	-0.01
13) Phenanthrene-d10	17.052	188	11059m	0.400	ng/u1	-0.02
17) Chrysene-d12	21.262	240	9222m	0.400	ng/u1	0.00
23) Perylene-d12	23.492	264	10917m	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.218	96	7856	2.320	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.024	152	2315	0.210	ng/u1	-0.03
18) Fluoranthene-d10	19.090	212	6065m	0.260	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
Data File : BM044913.D
Acq On : 04 Apr 2024 14:11
Operator : MA/JU
Sample : P1849-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E0LY5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/04/2024
Supervised By :mohammad ahmed 04/06/2024

Quant Time: Apr 04 15:08:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 03 23:34:03 2024
Response via : Initial Calibration

