

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
 Data File : BM044938.D
 Acq On : 05 Apr 2024 08:19
 Operator : MA/JU
 Sample : PB160015BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK015

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/05/2024
 Supervised By :mohammad ahmed 04/06/2024

Quant Time: Apr 05 10:52:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 03 23:34:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.644	152	1930	0.400	ng/u1	-0.01
4) Naphthalene-d8	10.418	136	6062	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.293	164	3550	0.400	ng/u1	-0.01
13) Phenanthrene-d10	17.051	188	8478m	0.400	ng/u1	-0.02
17) Chrysene-d12	21.272	240	7737m	0.400	ng/u1	0.00
23) Perylene-d12	23.496	264	7555m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	12449	4.910	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.030	152	2336	0.284	ng/u1	-0.02
18) Fluoranthene-d10	19.094	212	5677	0.290	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040424\
Data File : BM044938.D
Acq On : 05 Apr 2024 08:19
Operator : MA/JU
Sample : PB160015BL
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK015

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/05/2024
Supervised By :mohammad ahmed 04/06/2024

Quant Time: Apr 05 10:52:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 03 23:34:03 2024
Response via : Initial Calibration

