

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040524\
 Data File : BM044956.D
 Acq On : 05 Apr 2024 19:44
 Operator : MA/JU
 Sample : SST01007
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST010028

Quant Time: Apr 06 00:15:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 06 00:13:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	63327	20.000	ng/ul	0.00
20) Naphthalene-d8	10.410	136	263764	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.286	164	165523	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.045	188	355144	20.000	ng/ul	0.00
79) Chrysene-d12	21.251	240	376416	20.000	ng/ul	0.00
88) Perylene-d12	23.480	264	471953	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	6536	3.501	ng/uL	0.00
4) Pyridine-d5	3.616	84	39317	7.154	ng/ul	0.00
7) Phenol-d5	6.816	99	47817	7.344	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.987	67	30242	7.520	ng/ul	0.00
11) 2-Chlorophenol-d4	7.169	132	39483	8.252	ng/ul	0.00
15) 4-Methylphenol-d8	8.345	113	36742	7.186	ng/ul	0.00
21) Nitrobenzene-d5	8.798	128	18015	8.111	ng/ul	0.00
24) 2-Nitrophenol-d4	9.510	143	19366	8.088	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.045	165	35858	8.799	ng/ul	0.00
31) 4-Chloroaniline-d4	10.569	131	53530	7.909	ng/ul	0.00
46) Dimethylphthalate-d6	13.710	166	109441	8.429	ng/ul	0.00
49) Acenaphthylene-d8	13.975	160	128912	8.681	ng/ul	0.00
54) 4-Nitrophenol-d4	14.516	143	16409	6.094	ng/ul	0.00
60) Fluorene-d10	15.286	176	98086	8.989	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.427	200	15024	6.517	ng/ul	0.00
73) Anthracene-d10	17.145	188	150599	8.871	ng/ul	0.00
81) Pyrene-d10	19.451	212	177049	7.669	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.333	264	214848	8.661	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	7286	3.769	ng/uL	96
5) Pyridine	3.634	79	43163	7.573	ng/ul	96
6) Benzaldehyde	6.804	77	27869	9.827	ng/ul	89
8) Phenol	6.846	94	49189	7.403	ng/ul	91
10) Bis(2-Chloroethyl)ether	7.081	93	40978	7.510	ng/ul	96
12) 2-Chlorophenol	7.204	128	42387	8.589	ng/ul	93
13) 2-Methylphenol	8.081	108	36692	7.277	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.157	45	32926	4.570	ng/ul	95
16) Acetophenone	8.463	105	61891	7.804	ng/ul	95
17) N-Nitroso-di-n-propyla...	8.445	70	30519	7.449	ng/ul#	91
18) 4-Methylphenol	8.410	108	39149	7.132	ng/ul	95
19) Hexachloroethane	8.687	117	18676	9.239	ng/ul	99
22) Nitrobenzene	8.840	77	47862	9.087	ng/ul	94
23) Isophorone	9.363	82	84475	7.861	ng/ul#	97
25) 2-Nitrophenol	9.540	139	21369	8.186	ng/ul	93
26) 2,4-Dimethylphenol	9.604	107	43697	9.200	ng/ul	95
27) Bis(2-Chloroethoxy)met...	9.845	93	52266	7.733	ng/ul	98
29) 2,4-Dichlorophenol	10.069	162	36910	9.073	ng/ul	100
30) Naphthalene	10.457	128	133681	8.949	ng/ul	99
32) 4-Chloroaniline	10.592	127	53331	8.174	ng/ul	96
33) Hexachlorobutadiene	10.728	225	24036	10.541	ng/ul	97
34) Caprolactam	11.386	113	9930	6.268	ng/ul	96
35) 4-Chloro-3-methylphenol	11.722	107	35458	7.516	ng/ul	96
36) 2-Methylnaphthalene	12.081	142	87520	8.810	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040524\
 Data File : BM044956.D
 Acq On : 05 Apr 2024 19:44
 Operator : MA/JU
 Sample : SSTD01007
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD010028


Quant Time: Apr 06 00:15:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 06 00:13:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.304	142	88277	9.008	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.457	216	44359	9.420	ng/ul	97
40) Hexachlorocyclopentadiene	12.416	237	22174	7.094	ng/ul	91
41) 2,4,6-Trichlorophenol	12.710	196	27901	8.452	ng/ul	98
42) 2,4,5-Trichlorophenol	12.786	196	32048	9.059	ng/ul	95
43) 1,1'-Biphenyl	13.110	154	114058	8.678	ng/ul	97
44) 2-Chloronaphthalene	13.151	162	94262	9.139	ng/ul	100
45) 2-Nitroaniline	13.380	65	22965	7.092	ng/ul	96
47) Dimethylphthalate	13.757	163	114283	8.650	ng/ul#	96
48) 2,6-Dinitrotoluene	13.886	165	20967	7.945	ng/ul#	90
50) Acenaphthylene	14.004	152	151284	8.741	ng/ul	97
51) 3-Nitroaniline	14.222	138	21022	7.375	ng/ul	92
52) Acenaphthene	14.351	153	102005	8.965	ng/ul	98
53) 2,4-Dinitrophenol	14.439	184	8301	4.917	ng/ul	89
55) 4-Nitrophenol	14.533	109	13625	7.222	ng/ul	97
56) Dibenzofuran	14.692	168	139111	9.192	ng/ul	100
57) 2,4-Dinitrotoluene	14.680	165	30981	8.133	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.922	232	26118	8.921	ng/ul	91
59) Diethylphthalate	15.127	149	114196	8.328	ng/ul	99
61) Fluorene	15.339	166	114362	9.239	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.345	204	54317	9.455	ng/ul	98
63) 4-Nitroaniline	15.386	138	17227	6.323	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.439	198	17013	6.906	ng/ul#	95
67) N-Nitrosodiphenylamine	15.563	169	92156	8.616	ng/ul	99
68) 4-Bromophenyl-phenylether	16.239	248	33708	9.489	ng/ul	95
69) Hexachlorobenzene	16.339	284	43677	10.795	ng/ul	97
70) Atrazine	16.527	200	30767	8.802	ng/ul	98
71) Pentachlorophenol	16.698	266	20177	7.552	ng/ul	100
72) Phenanthrene	17.086	178	183314	9.099	ng/ul	99
74) Anthracene	17.180	178	183774	9.020	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.069	216	44186	9.315	ng/uL	97
76) Pentachlorobenzene	14.598	250	47625	10.385	ng/uL	98
77) Carbazole	17.463	167	167983	8.774	ng/ul	99
78) Di-n-butylphthalate	18.033	149	184028	7.146	ng/ul	100
80) Fluoranthene	19.115	202	207287	7.539	ng/ul	98
82) Pyrene	19.480	202	227598	7.951	ng/ul	99
83) Butylbenzylphthalate	20.404	149	84426	5.998	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.186	252	72747	8.934	ng/ul	99
85) Benzo(a)anthracene	21.239	228	240683	8.515	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.186	149	125990	6.139	ng/ul	99
87) Chrysene	21.292	228	228522	8.732	ng/ul	98
89) Di-n-octyl phthalate	22.068	149	218750	5.179	ng/ul	100
90) Benzo(b)fluoranthene	22.809	252	257927	8.167	ng/ul	99
91) Benzo(k)fluoranthene	22.856	252	254543	8.230	ng/ul	98
93) Benzo(a)pyrene	23.380	252	247563	8.380	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.715	276	326049	9.840	ng/ul	99
95) Dibenzo(a,h)anthracene	25.733	278	268420	9.716	ng/ul	99
96) Benzo(g,h,i)perylene	26.397	276	259752	9.824	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 06 00:15:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 06 00:13:49 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.769 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion	88	Resp	7286
Ion	Ratio	Lower	Upper
88	100		
43	27.9	24.0	36.0
58	59.9	45.4	68.2



#5
Pyridine
Concen: 7.573 ng/uL
RT: 3.634 min Scan# 93
Delta R.T. 0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion	79	Resp	43163
Ion	Ratio	Lower	Upper
79	100		
52	71.7	56.0	84.0
51	38.6	27.6	41.4









#6
Benzaldehyde
Concen: 9.827 ng/ul
RT: 6.804 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion: 77 Resp: 27869
Ion Ratio Lower Upper
77 100
105 108.3 75.8 113.6
106 101.2 74.6 112.0



#8
Phenol
Concen: 7.403 ng/ul
RT: 6.846 min Scan# 639
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion: 94 Resp: 49189
Ion Ratio Lower Upper
94 100
65 31.1 21.7 32.5
66 33.9 31.9 47.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 7.510 ng/ul
 RT: 7.081 min Scan# 679
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44




Instrument :
 BNA_M
 ClientSampleId :
 SST010028


Tgt Ion:	93	Resp:	40978
Ion	Ratio	Lower	Upper
93	100		
63	74.2	61.3	91.9
95	31.2	27.6	41.4



#12
 2-Chlorophenol
 Concen: 8.589 ng/ul
 RT: 7.204 min Scan# 700
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44

Tgt Ion:	128	Resp:	42387
Ion	Ratio	Lower	Upper
128	100		
64	37.5	33.3	49.9
130	29.1	26.2	39.4









#13
2-Methylphenol
Concen: 7.277 ng/ul
RT: 8.081 min Scan# 849
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:108 Resp: 36692
Ion Ratio Lower Upper
108 100
107 90.2 69.1 103.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.570 ng/ul
RT: 8.157 min Scan# 862
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion: 45 Resp: 32926
Ion Ratio Lower Upper
45 100
77 18.8 13.4 20.0
79 10.7 10.1 15.1









#16
Acetophenone
Concen: 7.804 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:105 Resp: 61891
Ion Ratio Lower Upper
105 100
77 76.8 65.8 98.6
51 26.6 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 7.449 ng/ul
RT: 8.445 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion: 70 Resp: 30519
Ion Ratio Lower Upper
70 100
42 63.3 43.2 64.8
101 10.1 6.4 9.6#
130 20.6 16.2 24.2









#18
4-Methylphenol
Concen: 7.132 ng/ul
RT: 8.410 min Scan# 905
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:108 Resp: 39149
Ion Ratio Lower Upper
108 100
107 117.7 90.0 135.0



#19
Hexachloroethane
Concen: 9.239 ng/ul
RT: 8.687 min Scan# 952
Delta R.T. -0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:117 Resp: 18676
Ion Ratio Lower Upper
117 100
201 89.4 73.0 109.4
199 55.4 44.4 66.6









#22
Nitrobenzene
Concen: 9.087 ng/ul
RT: 8.840 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion: 77 Resp: 47862
Ion Ratio Lower Upper
77 100
123 42.5 37.4 56.2
65 12.0 10.7 16.1



#23
Isophorone
Concen: 7.861 ng/ul
RT: 9.363 min Scan# 1067
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion: 82 Resp: 84475
Ion Ratio Lower Upper
82 100
95 7.9 5.1 7.7#
138 18.8 14.2 21.4









#25
2-Nitrophenol
Concen: 8.186 ng/ul
RT: 9.540 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:139 Resp: 21369
Ion Ratio Lower Upper
139 100
65 50.9 34.9 52.3
109 30.5 24.4 36.6



#26
2,4-Dimethylphenol
Concen: 9.200 ng/ul
RT: 9.604 min Scan# 1108
Delta R.T. 0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:107 Resp: 43697
Ion Ratio Lower Upper
107 100
121 47.0 41.1 61.7
122 83.1 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 7.733 ng/ul
 RT: 9.845 min Scan# 1149
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44




Instrument :
 BNA_M
 ClientSampleId :
 SST010028


Tgt Ion:	93	Resp:	52266
Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.9	40.3
123	12.7	9.1	13.7



#29
 2,4-Dichlorophenol
 Concen: 9.073 ng/ul
 RT: 10.069 min Scan# 1187
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44

Tgt Ion:	162	Resp:	36910
Ion	Ratio	Lower	Upper
162	100		
164	63.1	50.4	75.6
98	33.5	27.0	40.4









#30
Naphthalene
Concen: 8.949 ng/ul
RT: 10.457 min Scan# 1253
Delta R.T. -0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:128 Resp: 133681
Ion Ratio Lower Upper
128 100
129 11.4 8.5 12.7
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 8.174 ng/ul
RT: 10.592 min Scan# 1276
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:127 Resp: 53331
Ion Ratio Lower Upper
127 100
129 30.4 26.2 39.2









#33
Hexachlorobutadiene
Concen: 10.541 ng/ul
RT: 10.728 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:225 Resp: 24036
Ion Ratio Lower Upper
225 100
223 58.1 45.3 67.9
227 62.8 47.6 71.4



#34
Caprolactam
Concen: 6.268 ng/ul
RT: 11.386 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:113 Resp: 9930
Ion Ratio Lower Upper
113 100
55 147.6 125.4 188.0
56 114.2 91.8 137.8









#35
 4-Chloro-3-methylphenol
 Concen: 7.516 ng/ul
 RT: 11.722 min Scan# 1468
 Delta R.T. 0.006 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010028


Tgt Ion:107 Resp: 35458
 Ion Ratio Lower Upper
 107 100
 144 27.4 22.8 34.2
 142 90.9 69.8 104.6



#36
 2-Methylnaphthalene
 Concen: 8.810 ng/ul
 RT: 12.081 min Scan# 1529
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44

Tgt Ion:142 Resp: 87520
 Ion Ratio Lower Upper
 142 100
 141 90.8 72.3 108.5









#37
1-Methylnaphthalene
Concen: 9.008 ng/ul
RT: 12.304 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:142 Resp: 88277
Ion Ratio Lower Upper
142 100
141 91.7 77.2 115.8
116 3.9 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.420 ng/ul
RT: 12.457 min Scan# 1593
Delta R.T. 0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:216 Resp: 44359
Ion Ratio Lower Upper
216 100
214 76.7 63.5 95.3
179 21.5 16.6 25.0
108 17.5 15.0 22.4









#40
Hexachlorocyclopentadiene
Concen: 7.094 ng/ul
RT: 12.416 min Scan# 1586
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:237 Resp: 22174
Ion Ratio Lower Upper
237 100
235 72.6 51.3 76.9
272 13.0 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 8.452 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:196 Resp: 27901
Ion Ratio Lower Upper
196 100
198 97.1 79.4 119.2
200 33.1 25.0 37.4





#42

2,4,5-Trichlorophenol

Concen: 9.059 ng/ul

RT: 12.786 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM044956.D


Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028






Tgt Ion:196 Resp: 32048

Ion Ratio Lower Upper

196 100

198 95.2 79.9 119.9

200 29.4 25.6 38.4



#43

1,1'-Biphenyl

Concen: 8.678 ng/ul

RT: 13.110 min Scan# 1704

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:154 Resp: 114058


Ion Ratio Lower Upper

154 100

153 42.3 31.6 47.4

76 12.9 10.4 15.6









#44
2-Chloronaphthalene
Concen: 9.139 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:162 Resp: 94262
Ion Ratio Lower Upper
162 100
164 32.4 26.1 39.1
127 37.6 30.1 45.1



#45
2-Nitroaniline
Concen: 7.092 ng/ul
RT: 13.380 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion: 65 Resp: 22965
Ion Ratio Lower Upper
65 100
92 66.8 55.2 82.8
138 111.2 93.0 139.4









#47
Dimethylphthalate
Concen: 8.650 ng/ul
RT: 13.757 min Scan# 1814
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:163 Resp: 114283
Ion Ratio Lower Upper
163 100
194 4.3 2.8 4.2#
164 11.3 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 7.945 ng/ul
RT: 13.886 min Scan# 1836
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:165 Resp: 20967
Ion Ratio Lower Upper
165 100
89 63.0 45.0 67.4
121 27.7 18.0 27.0#





#50

Acenaphthylene

Concen: 8.741 ng/ul

RT: 14.004 min Scan# 1856

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:152 Resp: 151284

Ion Ratio Lower Upper

152 100

151 21.6 16.0 24.0

153 13.2 10.2 15.2



#51

3-Nitroaniline

Concen: 7.375 ng/ul

RT: 14.222 min Scan# 1893

Delta R.T. 0.006 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:138 Resp: 21022


Ion Ratio Lower Upper

138 100

108 8.2 6.5 9.7

92 110.3 81.4 122.0





#52

Acenaphthene

Concen: 8.965 ng/ul

RT: 14.351 min Scan# 1915

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:153 Resp: 102005

Ion Ratio Lower Upper

153 100

152 50.7 38.6 58.0

154 85.5 69.4 104.2



#53

2,4-Dinitrophenol

Concen: 4.917 ng/ul

RT: 14.439 min Scan# 1930

Delta R.T. 0.006 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:184 Resp: 8301


Ion Ratio Lower Upper

184 100

63 60.6 56.9 85.3

154 55.3 50.3 75.5









#55
4-Nitrophenol
Concen: 7.222 ng/ul
RT: 14.533 min Scan# 1946
Delta R.T. 0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:109 Resp: 13625
Ion Ratio Lower Upper
109 100
139 130.2 107.7 161.5
65 119.0 96.3 144.5



#56
Dibenzofuran
Concen: 9.192 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. 0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:168 Resp: 139111
Ion Ratio Lower Upper
168 100
139 38.1 30.6 46.0









#57
 2,4-Dinitrotoluene
 Concen: 8.133 ng/ul
 RT: 14.680 min Scan# 1971
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44




Instrument :
 BNA_M
 ClientSampleId :
 SST010028


Tgt Ion:165 Resp: 30981
 Ion Ratio Lower Upper
 165 100
 63 65.5 52.0 78.0
 182 2.7 2.3 3.5



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 8.921 ng/ul
 RT: 14.922 min Scan# 2012
 Delta R.T. 0.006 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44

Tgt Ion:232 Resp: 26118
 Ion Ratio Lower Upper
 232 100
 131 44.4 43.4 65.0
 130 3.3 2.6 4.0
 166 30.7 25.6 38.4









#59
Diethylphthalate
Concen: 8.328 ng/ul
RT: 15.127 min Scan# 2047
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:149 Resp: 114196
Ion Ratio Lower Upper
149 100
177 21.4 16.4 24.6
150 12.7 10.3 15.5



#61
Fluorene
Concen: 9.239 ng/ul
RT: 15.339 min Scan# 2083
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:166 Resp: 114362
Ion Ratio Lower Upper
166 100
165 98.8 80.5 120.7
167 13.1 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 9.455 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. 0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:204 Resp: 54317
Ion Ratio Lower Upper
204 100
206 32.5 24.9 37.3
141 56.0 46.2 69.2



#63
4-Nitroaniline
Concen: 6.323 ng/ul
RT: 15.386 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:138 Resp: 17227
Ion Ratio Lower Upper
138 100
92 52.6 39.8 59.6
108 86.1 66.4 99.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 6.906 ng/ul
 RT: 15.439 min Scan# 2100
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010028


Tgt Ion:198 Resp: 17013
 Ion Ratio Lower Upper
 198 100
 182 8.3 3.8 5.6#
 77 32.7 24.6 37.0



#67
 N-Nitrosodiphenylamine
 Concen: 8.616 ng/ul
 RT: 15.563 min Scan# 2121
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44

Tgt Ion:169 Resp: 92156
 Ion Ratio Lower Upper
 169 100
 168 69.8 56.7 85.1
 167 38.3 31.2 46.8









#68
4-Bromophenyl-phenylether
Concen: 9.489 ng/ul
RT: 16.239 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:248 Resp: 33708
Ion Ratio Lower Upper
248 100
250 92.9 81.0 121.6
141 68.5 55.1 82.7



#69
Hexachlorobenzene
Concen: 10.795 ng/ul
RT: 16.339 min Scan# 2253
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:284 Resp: 43677
Ion Ratio Lower Upper
284 100
142 34.0 27.3 40.9
249 26.7 23.6 35.4





#70

Atrazine

Concen: 8.802 ng/ul

RT: 16.527 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:200 Resp: 30767

Ion Ratio Lower Upper

200 100

173 26.6 20.3 30.5

215 46.4 37.9 56.9



#71

Pentachlorophenol

Concen: 7.552 ng/ul

RT: 16.698 min Scan# 2314

Delta R.T. 0.006 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:266 Resp: 20177


Ion Ratio Lower Upper

266 100

264 61.6 49.3 73.9

268 67.0 53.3 79.9





#72

Phenanthrene

Concen: 9.099 ng/ul

RT: 17.086 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:178 Resp: 183314

Ion Ratio Lower Upper

178 100

179 15.4 11.9 17.9

176 19.3 15.4 23.2



#74

Anthracene

Concen: 9.020 ng/ul

RT: 17.180 min Scan# 2396

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:178 Resp: 183774


Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 20.8 15.8 23.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.315 ng/uL

RT: 13.069 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:216 Resp: 44186

Ion Ratio Lower Upper

216 100

214 78.2 64.1 96.1

218 50.7 38.6 58.0



#76

Pentachlorobenzene

Concen: 10.385 ng/uL

RT: 14.598 min Scan# 1957

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:250 Resp: 47625


Ion Ratio Lower Upper

250 100

248 62.2 49.8 74.6

252 65.1 49.4 74.2









#77
 Carbazole
 Concen: 8.774 ng/ul
 RT: 17.463 min Scan# 2444
 Delta R.T. 0.000 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44




Instrument :
 BNA_M
 ClientSampleId :
 SST010028


Tgt Ion:167 Resp: 167983
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.3 25.9
 139 12.9 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 7.146 ng/ul
 RT: 18.033 min Scan# 2541
 Delta R.T. 0.006 min
 Lab File: BM044956.D
 Acq: 05 Apr 2024 19:44

Tgt Ion:149 Resp: 184028
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.2 10.8
 104 5.5 4.3 6.5





#80

Fluoranthene

Concen: 7.539 ng/ul

RT: 19.115 min Scan# 2725

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:202 Resp: 207287

Ion Ratio Lower Upper

202 100

101 11.7 8.7 13.1

100 8.3 6.3 9.5



#82

Pyrene

Concen: 7.951 ng/ul

RT: 19.480 min Scan# 2787

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:202 Resp: 227598


Ion Ratio Lower Upper

202 100

101 13.2 10.2 15.4

100 10.1 8.5 12.7





#83

Butylbenzylphthalate

Concen: 5.998 ng/ul

RT: 20.404 min Scan# 2944

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:149 Resp: 84426

Ion Ratio Lower Upper

149 100

91 73.3 56.2 84.4

206 17.6 14.3 21.5



#84

3,3'-Dichlorobenzidine

Concen: 8.934 ng/ul

RT: 21.186 min Scan# 3077

Delta R.T. 0.006 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:252 Resp: 72747


Ion Ratio Lower Upper

252 100

254 68.0 53.9 80.9

126 11.9 9.0 13.4









#85
Benzo(a)anthracene
Concen: 8.515 ng/ul
RT: 21.239 min Scan# 3086
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:228 Resp: 240683
Ion Ratio Lower Upper
228 100
229 18.5 15.5 23.3
226 26.1 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 6.139 ng/ul
RT: 21.186 min Scan# 3077
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:149 Resp: 125990
Ion Ratio Lower Upper
149 100
167 24.3 19.7 29.5
279 4.3 3.3 4.9





#87

Chrysene

Concen: 8.732 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:228 Resp: 228522

Ion Ratio Lower Upper

228 100

226 29.6 24.3 36.5

229 18.5 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 5.179 ng/ul




RT: 22.068 min Scan# 3227


Delta R.T. 0.000 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Tgt Ion:149 Resp: 218750









#90
Benzo(b)fluoranthene
Concen: 8.167 ng/ul
RT: 22.809 min Scan# 3353
Delta R.T. -0.006 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44




Instrument :
BNA_M
ClientSampleId :
SSTD010028


Tgt Ion:252 Resp: 257927
Ion Ratio Lower Upper
252 100
253 22.1 18.0 27.0
125 10.1 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 8.230 ng/ul
RT: 22.856 min Scan# 3361
Delta R.T. 0.000 min
Lab File: BM044956.D
Acq: 05 Apr 2024 19:44

Tgt Ion:252 Resp: 254543
Ion Ratio Lower Upper
252 100
253 22.6 17.5 26.3
125 9.9 8.5 12.7





#93

Benzo(a)pyrene

Concen: 8.380 ng/ul

RT: 23.380 min Scan# 34

Delta R.T. -0.006 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:252 Resp: 247563

Ion Ratio Lower Upper

252 100

253 21.1 17.6 26.4

125 11.3 8.8 13.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.840 ng/ul

RT: 25.715 min Scan# 3847

Delta R.T. -0.006 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44




Tgt Ion:276 Resp: 326049


Ion Ratio Lower Upper

276 100

138 22.9 19.0 28.6

277 24.7 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 9.716 ng/ul

RT: 25.733 min Scan# 3851

Delta R.T. -0.006 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Instrument :

BNA_M

ClientSampleId :

SSTD010028





Tgt Ion:278 Resp: 268420

Ion Ratio Lower Upper

278 100

139 15.0 11.9 17.9

279 25.0 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 9.824 ng/ul

RT: 26.397 min Scan# 3963

Delta R.T. -0.006 min

Lab File: BM044956.D

Acq: 05 Apr 2024 19:44

Tgt Ion:276 Resp: 259752

Ion Ratio Lower Upper

276 100

138 21.8 17.8 26.6

277 23.2 18.7 28.1

