

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040524\
 Data File : BM044959.D
 Acq On : 05 Apr 2024 21:33
 Operator : MA/JU
 Sample : SSTD08010
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080031

Quant Time: Apr 06 00:16:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 06 00:13:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.633	152	74778	20.000	ng/u1	0.00
20) Naphthalene-d8	10.410	136	326236	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.286	164	209236	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.045	188	442443	20.000	ng/u1	0.00
79) Chrysene-d12	21.256	240	463353	20.000	ng/u1	0.00
88) Perylene-d12	23.480	264	544078	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.204	96	58458	26.515	ng/uL	0.00
4) Pyridine-d5	3.604	84	455851	70.247	ng/u1	0.00
7) Phenol-d5	6.816	99	527867	68.661	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.992	67	302998	63.805	ng/u1	0.00
11) 2-Chlorophenol-d4	7.175	132	424451	75.128	ng/u1	0.00
15) 4-Methylphenol-d8	8.351	113	423046	70.066	ng/u1	0.00
21) Nitrobenzene-d5	8.798	128	216804	78.921	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	245641	82.946	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.039	165	429198	85.156	ng/u1	0.00
31) 4-Chloroaniline-d4	10.569	131	643143	76.831	ng/u1	0.00
46) Dimethylphthalate-d6	13.715	166	1202230	73.248	ng/u1	0.00
49) Acenaphthylene-d8	13.980	160	1472576	78.449	ng/u1	0.00
54) 4-Nitrophenol-d4	14.521	143	257794	75.737	ng/u1	0.01
60) Fluorene-d10	15.292	176	1081368	78.400	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.433	200	236229	82.252	ng/u1	0.00
73) Anthracene-d10	17.151	188	1696451	80.216	ng/u1	0.00
81) Pyrene-d10	19.456	212	1988442	69.969	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.344	264	2279965	79.726	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	59960	26.270	ng/uL#	85
5) Pyridine	3.622	79	474608	70.516	ng/u1	97
6) Benzaldehyde	6.804	77	177633	53.046	ng/u1	98
8) Phenol	6.845	94	547477	69.775	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.081	93	413623	64.199	ng/u1	97
12) 2-Chlorophenol	7.204	128	435672	74.763	ng/u1	98
13) 2-Methylphenol	8.081	108	405129	68.041	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.163	45	492851	57.925	ng/u1	99
16) Acetophenone	8.469	105	692685	73.972	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.463	70	348112	71.953	ng/u1#	93
18) 4-Methylphenol	8.416	108	443014	68.351	ng/u1	99
19) Hexachloroethane	8.692	117	186785	78.251	ng/u1	97
22) Nitrobenzene	8.845	77	519638	79.763	ng/u1	99
23) Isophorone	9.369	82	981186	73.820	ng/u1	99
25) 2-Nitrophenol	9.545	139	260396	80.653	ng/u1	93
26) 2,4-Dimethylphenol	9.604	107	470868	80.152	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.851	93	566391	67.757	ng/u1	98
29) 2,4-Dichlorophenol	10.069	162	422062	83.878	ng/u1	98
30) Naphthalene	10.463	128	1402764	75.919	ng/u1	98
32) 4-Chloroaniline	10.592	127	629205	77.968	ng/u1	100
33) Hexachlorobutadiene	10.727	225	240365	85.227	ng/u1	95
34) Caprolactam	11.410	113	74957	38.252	ng/u1	99
35) 4-Chloro-3-methylphenol	11.721	107	438904	75.216	ng/u1	99
36) 2-Methylnaphthalene	12.086	142	946795	77.056	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040524\
 Data File : BM044959.D
 Acq On : 05 Apr 2024 21:33
 Operator : MA/JU
 Sample : SSTD08010
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080031


Quant Time: Apr 06 00:16:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 06 00:13:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.304	142	950541	78.421	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.457	216	464822	78.086	ng/ul	99
40) Hexachlorocyclopentadiene	12.421	237	295334	74.740	ng/ul	97
41) 2,4,6-Trichlorophenol	12.710	196	330764	79.262	ng/ul	99
42) 2,4,5-Trichlorophenol	12.780	196	352305	78.778	ng/ul	96
43) 1,1'-Biphenyl	13.116	154	1209807	72.814	ng/ul	99
44) 2-Chloronaphthalene	13.157	162	986057	75.626	ng/ul	100
45) 2-Nitroaniline	13.380	65	306054	74.770	ng/ul	98
47) Dimethylphthalate	13.763	163	1249760	74.828	ng/ul	99
48) 2,6-Dinitrotoluene	13.892	165	266677	79.938	ng/ul	93
50) Acenaphthylene	14.010	152	1701731	77.783	ng/ul	99
51) 3-Nitroaniline	14.221	138	254499	70.627	ng/ul	94
52) Acenaphthene	14.351	153	1118299	77.754	ng/ul	99
53) 2,4-Dinitrophenol	14.433	184	162209	76.014	ng/ul	93
55) 4-Nitrophenol	14.539	109	229951	96.421	ng/ul	92
56) Dibenzofuran	14.692	168	1533270	80.144	ng/ul	97
57) 2,4-Dinitrotoluene	14.686	165	409593	85.056	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.921	232	308154	83.265	ng/ul	98
59) Diethylphthalate	15.139	149	1315575	75.899	ng/ul	99
61) Fluorene	15.345	166	1340088	85.646	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.345	204	622425	85.708	ng/ul	94
63) 4-Nitroaniline	15.398	138	221545	64.328	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.451	198	245724	80.065	ng/ul#	93
67) N-Nitrosodiphenylamine	15.568	169	1034667	77.645	ng/ul	99
68) 4-Bromophenyl-phenylether	16.245	248	357591	80.804	ng/ul	97
69) Hexachlorobenzene	16.339	284	459449	91.149	ng/ul#	92
70) Atrazine	16.533	200	359633	82.588	ng/ul	99
71) Pentachlorophenol	16.698	266	295779	88.865	ng/ul	99
72) Phenanthrene	17.092	178	1993654	79.433	ng/ul	98
74) Anthracene	17.186	178	2077215	81.834	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.068	216	465517	78.773	ng/ul	98
76) Pentachlorobenzene	14.604	250	489438	85.666	ng/ul	96
77) Carbazole	17.462	167	1879949	78.816	ng/ul	100
78) Di-n-butylphthalate	18.033	149	2391110	74.528	ng/ul	100
80) Fluoranthene	19.115	202	2355751	69.606	ng/ul	98
82) Pyrene	19.486	202	2516698	71.426	ng/ul	99
83) Butylbenzylphthalate	20.403	149	1139372	65.763	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.186	252	816637	81.476	ng/ul	97
85) Benzo(a)anthracene	21.238	228	2636521	75.771	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.186	149	1826323	72.298	ng/ul	98
87) Chrysene	21.297	228	2417092	75.031	ng/ul	100
89) Di-n-octyl phthalate	22.068	149	2985201	61.309	ng/ul	100
90) Benzo(b)fluoranthene	22.821	252	2725672	74.869	ng/ul	99
91) Benzo(k)fluoranthene	22.868	252	2781993	78.022	ng/ul	99
93) Benzo(a)pyrene	23.397	252	2675908	78.574	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	25.738	276	3396139	88.910	ng/ul	99
95) Dibenzo(a,h)anthracene	25.750	278	2864873	89.950	ng/ul	99
96) Benzo(g,h,i)perylene	26.421	276	2637138	86.515	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 06 00:16:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 06 00:13:49 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 26.270 ng/uL
RT: 3.240 min Scan# 20
Delta R.T. -0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion: 88 Resp: 59960
Ion Ratio Lower Upper
88 100
43 33.1 24.0 36.0
58 71.0 45.4 68.2#



#5
Pyridine
Concen: 70.516 ng/uL
RT: 3.622 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion: 79 Resp: 474608
Ion Ratio Lower Upper
79 100
52 66.4 56.0 84.0
51 34.6 27.6 41.4









#6
Benzaldehyde
Concen: 53.046 ng/ul
RT: 6.804 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion: 77 Resp: 177633
Ion Ratio Lower Upper
77 100
105 96.3 75.8 113.6
106 91.6 74.6 112.0



#8
Phenol
Concen: 69.775 ng/ul
RT: 6.845 min Scan# 639
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion: 94 Resp: 547477
Ion Ratio Lower Upper
94 100
65 29.2 21.7 32.5
66 37.6 31.9 47.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 64.199 ng/ul
 RT: 7.081 min Scan# 679
 Delta R.T. -0.000 min
 Lab File: BM044959.D
 Acq: 05 Apr 2024 21:33

Instrument :
 BNA_M
 ClientSampleId :
 SST080031

Tgt Ion: 93 Resp: 413623




Ion	Ratio	Lower	Upper
93	100		
63	74.8	61.3	91.9
95	32.2	27.6	41.4




#12
 2-Chlorophenol
 Concen: 74.763 ng/ul
 RT: 7.204 min Scan# 700
 Delta R.T. -0.000 min
 Lab File: BM044959.D
 Acq: 05 Apr 2024 21:33

Tgt Ion: 128 Resp: 435672

Ion	Ratio	Lower	Upper
128	100		
64	43.3	33.3	49.9
130	33.1	26.2	39.4





#13

2-Methylphenol

Concen: 68.041 ng/ul

RT: 8.081 min Scan# 849

Delta R.T. -0.000 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :





SSTD080031

Tgt Ion:108 Resp: 405129

Ion Ratio Lower Upper

108 100

107 89.3 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 57.925 ng/ul

RT: 8.163 min Scan# 863

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion: 45 Resp: 492851


Ion Ratio Lower Upper

45 100

77 17.1 13.4 20.0

79 12.8 10.1 15.1









#16
Acetophenone
Concen: 73.972 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:105 Resp: 692685
Ion Ratio Lower Upper
105 100
77 79.0 65.8 98.6
51 26.6 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 71.953 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. 0.018 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion: 70 Resp: 348112
Ion Ratio Lower Upper
70 100
42 59.7 43.2 64.8
101 10.1 6.4 9.6#
130 22.9 16.2 24.2









#18
4-Methylphenol
Concen: 68.351 ng/ul
RT: 8.416 min Scan# 906
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:108 Resp: 443014
Ion Ratio Lower Upper
108 100
107 113.8 90.0 135.0



#19
Hexachloroethane
Concen: 78.251 ng/ul
RT: 8.692 min Scan# 953
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:117 Resp: 186785
Ion Ratio Lower Upper
117 100
201 93.8 73.0 109.4
199 58.2 44.4 66.6





#22

Nitrobenzene

Concen: 79.763 ng/ul

RT: 8.845 min Scan# 91

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion: 77 Resp: 519638

Ion Ratio Lower Upper

77 100

123 46.1 37.4 56.2

65 12.3 10.7 16.1



#23

Isophorone

Concen: 73.820 ng/ul

RT: 9.369 min Scan# 1068

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion: 82 Resp: 981186


Ion Ratio Lower Upper

82 100

95 6.6 5.1 7.7

138 17.5 14.2 21.4









#25
2-Nitrophenol
Concen: 80.653 ng/ul
RT: 9.545 min Scan# 1108
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:139 Resp: 260396
Ion Ratio Lower Upper
139 100
65 49.5 34.9 52.3
109 32.8 24.4 36.6



#26
2,4-Dimethylphenol
Concen: 80.152 ng/ul
RT: 9.604 min Scan# 1108
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:107 Resp: 470868
Ion Ratio Lower Upper
107 100
121 52.2 41.1 61.7
122 85.8 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 67.757 ng/ul
 RT: 9.851 min Scan# 1149
 Delta R.T. 0.006 min
 Lab File: BM044959.D
 Acq: 05 Apr 2024 21:33




Instrument :
 BNA_M
 ClientSampleId :
 SST080031


Tgt Ion: 93 Resp: 566391
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.9 40.3
 123 11.3 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 83.878 ng/ul
 RT: 10.069 min Scan# 1187
 Delta R.T. -0.000 min
 Lab File: BM044959.D
 Acq: 05 Apr 2024 21:33

Tgt Ion: 162 Resp: 422062
 Ion Ratio Lower Upper
 162 100
 164 64.3 50.4 75.6
 98 32.7 27.0 40.4









#30
Naphthalene
Concen: 75.919 ng/ul
RT: 10.463 min Scan# 1254
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:128 Resp: 1402764
Ion Ratio Lower Upper
128 100
129 11.1 8.5 12.7
127 13.4 10.2 15.4



#32
4-Chloroaniline
Concen: 77.968 ng/ul
RT: 10.592 min Scan# 1276
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:127 Resp: 629205
Ion Ratio Lower Upper
127 100
129 32.6 26.2 39.2









#33
Hexachlorobutadiene
Concen: 85.227 ng/ul
RT: 10.727 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:225 Resp: 240365
Ion Ratio Lower Upper
225 100
223 61.0 45.3 67.9
227 62.0 47.6 71.4



#34
Caprolactam
Concen: 38.252 ng/ul
RT: 11.410 min Scan# 1415
Delta R.T. 0.023 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:113 Resp: 74957
Ion Ratio Lower Upper
113 100
55 156.6 125.4 188.0
56 111.1 91.8 137.8









#35
4-Chloro-3-methylphenol
Concen: 75.216 ng/ul
RT: 11.721 min Scan# 1468
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:107 Resp: 438904
Ion Ratio Lower Upper
107 100
144 27.7 22.8 34.2
142 86.0 69.8 104.6



#36
2-Methylnaphthalene
Concen: 77.056 ng/ul
RT: 12.086 min Scan# 1530
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:142 Resp: 946795
Ion Ratio Lower Upper
142 100
141 91.0 72.3 108.5









#37
1-Methylnaphthalene
Concen: 78.421 ng/ul
RT: 12.304 min Scan# 1567
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:142 Resp: 950541
Ion Ratio Lower Upper
142 100
141 96.2 77.2 115.8
116 4.2 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 78.086 ng/ul
RT: 12.457 min Scan# 1593
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:216 Resp: 464822
Ion Ratio Lower Upper
216 100
214 78.0 63.5 95.3
179 21.1 16.6 25.0
108 19.0 15.0 22.4









#40
Hexachlorocyclopentadiene
Concen: 74.740 ng/ul
RT: 12.421 min Scan# 1586
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:237 Resp: 295334
Ion Ratio Lower Upper
237 100
235 61.8 51.3 76.9
272 13.4 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 79.262 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:196 Resp: 330764
Ion Ratio Lower Upper
196 100
198 98.5 79.4 119.2
200 31.9 25.0 37.4









#42
2,4,5-Trichlorophenol
Concen: 78.778 ng/ul
RT: 12.780 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:196 Resp: 352305
Ion Ratio Lower Upper
196 100
198 94.9 79.9 119.9
200 30.8 25.6 38.4



#43
1,1'-Biphenyl
Concen: 72.814 ng/ul
RT: 13.116 min Scan# 1705
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:154 Resp: 1209807
Ion Ratio Lower Upper
154 100
153 40.2 31.6 47.4
76 12.6 10.4 15.6









#44
2-Chloronaphthalene
Concen: 75.626 ng/ul
RT: 13.157 min Scan# 1711
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:162 Resp: 986057
Ion Ratio Lower Upper
162 100
164 32.6 26.1 39.1
127 37.4 30.1 45.1



#45
2-Nitroaniline
Concen: 74.770 ng/ul
RT: 13.380 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion: 65 Resp: 306054
Ion Ratio Lower Upper
65 100
92 69.0 55.2 82.8
138 112.0 93.0 139.4





#47

Dimethylphthalate

Concen: 74.828 ng/ul

RT: 13.763 min Scan# 1815

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:163 Resp: 1249760

Ion Ratio Lower Upper

163 100

194 3.8 2.8 4.2

164 10.0 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 79.938 ng/ul

RT: 13.892 min Scan# 1837

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:165 Resp: 266677


Ion Ratio Lower Upper

165 100

89 63.0 45.0 67.4

121 21.7 18.0 27.0





#50

Acenaphthylene

Concen: 77.783 ng/ul

RT: 14.010 min Scan# 1856

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:152 Resp: 1701731

Ion Ratio Lower Upper

152 100

151 20.4 16.0 24.0

153 13.1 10.2 15.2



#51

3-Nitroaniline

Concen: 70.627 ng/ul

RT: 14.221 min Scan# 1893

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:138 Resp: 254499


Ion Ratio Lower Upper

138 100

108 8.2 6.5 9.7

92 108.3 81.4 122.0





#52

Acenaphthene

Concen: 77.754 ng/ul

RT: 14.351 min Scan# 1915

Delta R.T. -0.000 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:153 Resp: 1118299

Ion Ratio Lower Upper

153 100

152 48.8 38.6 58.0

154 87.3 69.4 104.2



#53

2,4-Dinitrophenol

Concen: 76.014 ng/ul

RT: 14.433 min Scan# 1929

Delta R.T. -0.000 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:184 Resp: 162209


Ion Ratio Lower Upper

184 100

63 76.2 56.9 85.3

154 68.5 50.3 75.5









#55
4-Nitrophenol
Concen: 96.421 ng/ul
RT: 14.539 min Scan# 1947
Delta R.T. 0.012 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:109 Resp: 229951
Ion Ratio Lower Upper
109 100
139 119.9 107.7 161.5
65 116.4 96.3 144.5



#56
Dibenzofuran
Concen: 80.144 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:168 Resp: 1533270
Ion Ratio Lower Upper
168 100
139 40.0 30.6 46.0









#57
2,4-Dinitrotoluene
Concen: 85.056 ng/ul
RT: 14.686 min Scan# 1971
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:165 Resp: 409593
Ion Ratio Lower Upper
165 100
63 65.6 52.0 78.0
182 2.7 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 83.265 ng/ul
RT: 14.921 min Scan# 2012
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:232 Resp: 308154
Ion Ratio Lower Upper
232 100
131 52.3 43.4 65.0
130 3.1 2.6 4.0
166 31.6 25.6 38.4









#59
Diethylphthalate
Concen: 75.899 ng/ul
RT: 15.139 min Scan# 2049
Delta R.T. 0.012 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:149 Resp: 1315575
Ion Ratio Lower Upper
149 100
177 20.9 16.4 24.6
150 12.3 10.3 15.5



#61
Fluorene
Concen: 85.646 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:166 Resp: 1340088
Ion Ratio Lower Upper
166 100
165 99.9 80.5 120.7
167 13.5 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 85.708 ng/ul
RT: 15.345 min Scan# 2083
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:204 Resp: 622425
Ion Ratio Lower Upper
204 100
206 32.1 24.9 37.3
141 63.5 46.2 69.2



#63
4-Nitroaniline
Concen: 64.328 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. 0.012 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:138 Resp: 221545
Ion Ratio Lower Upper
138 100
92 49.2 39.8 59.6
108 87.8 66.4 99.6









#66
4,6-Dinitro-2-methylphenol
Concen: 80.065 ng/ul
RT: 15.451 min Scan# 2102
Delta R.T. 0.012 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:198 Resp: 245724
Ion Ratio Lower Upper
198 100
182 3.3 3.8 5.6#
77 27.0 24.6 37.0



#67
N-Nitrosodiphenylamine
Concen: 77.645 ng/ul
RT: 15.568 min Scan# 2122
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:169 Resp: 1034667
Ion Ratio Lower Upper
169 100
168 70.3 56.7 85.1
167 38.3 31.2 46.8









#68
4-Bromophenyl-phenylether
Concen: 80.804 ng/ul
RT: 16.245 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:248 Resp: 357591
Ion Ratio Lower Upper
248 100
250 99.0 81.0 121.6
141 71.7 55.1 82.7



#69
Hexachlorobenzene
Concen: 91.149 ng/ul
RT: 16.339 min Scan# 2253
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:284 Resp: 459449
Ion Ratio Lower Upper
284 100
142 41.1 27.3 40.9#
249 28.0 23.6 35.4









#70
Atrazine
Concen: 82.588 ng/ul
RT: 16.533 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:200 Resp: 359633
Ion Ratio Lower Upper
200 100
173 26.1 20.3 30.5
215 46.6 37.9 56.9



#71
Pentachlorophenol
Concen: 88.865 ng/ul
RT: 16.698 min Scan# 2314
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:266 Resp: 295779
Ion Ratio Lower Upper
266 100
264 62.6 49.3 73.9
268 65.8 53.3 79.9





#72

Phenanthrene

Concen: 79.433 ng/ul

RT: 17.092 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:178 Resp: 1993654

Ion Ratio Lower Upper

178 100

179 15.3 11.9 17.9

176 20.2 15.4 23.2



#74

Anthracene

Concen: 81.834 ng/ul

RT: 17.186 min Scan# 2397

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:178 Resp: 2077215


Ion Ratio Lower Upper

178 100

179 15.1 12.3 18.5

176 19.1 15.8 23.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 78.773 ng/uL

RT: 13.068 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:216 Resp: 465517

Ion Ratio Lower Upper

216 100

214 78.3 64.1 96.1

218 47.8 38.6 58.0



#76

Pentachlorobenzene

Concen: 85.666 ng/uL

RT: 14.604 min Scan# 1958

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:250 Resp: 489438


Ion Ratio Lower Upper

250 100

248 64.5 49.8 74.6

252 64.9 49.4 74.2









#77
 Carbazole
 Concen: 78.816 ng/ul
 RT: 17.462 min Scan# 2444
 Delta R.T. -0.000 min
 Lab File: BM044959.D
 Acq: 05 Apr 2024 21:33




Instrument :
 BNA_M
 ClientSampleId :
 SSTD080031


Tgt Ion:167 Resp: 1879949
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.3 25.9
 139 13.6 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 74.528 ng/ul
 RT: 18.033 min Scan# 2541
 Delta R.T. 0.006 min
 Lab File: BM044959.D
 Acq: 05 Apr 2024 21:33

Tgt Ion:149 Resp: 2391110
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.2 10.8
 104 5.5 4.3 6.5









#80
Fluoranthene
Concen: 69.606 ng/uI
RT: 19.115 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33




Instrument :
BNA_M
ClientSampleId :
SSTD080031


Tgt Ion:202 Resp: 2355751
Ion Ratio Lower Upper
202 100
101 11.2 8.7 13.1
100 8.8 6.3 9.5



#82
Pyrene
Concen: 71.426 ng/uI
RT: 19.486 min Scan# 2788
Delta R.T. 0.006 min
Lab File: BM044959.D
Acq: 05 Apr 2024 21:33

Tgt Ion:202 Resp: 2516698
Ion Ratio Lower Upper
202 100
101 12.2 10.2 15.4
100 10.2 8.5 12.7





#83

Butylbenzylphthalate

Concen: 65.763 ng/ul

RT: 20.403 min Scan# 2944

Delta R.T. -0.000 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:149 Resp: 1139372

Ion Ratio Lower Upper

149 100

91 74.1 56.2 84.4

206 17.1 14.3 21.5



#84

3,3'-Dichlorobenzidine

Concen: 81.476 ng/ul

RT: 21.186 min Scan# 3077

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:252 Resp: 816637


Ion Ratio Lower Upper

252 100

254 65.1 53.9 80.9

126 10.8 9.0 13.4





#85

Benzo(a)anthracene

Concen: 75.771 ng/ul

RT: 21.238 min Scan# 3086

Delta R.T. -0.000 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:228 Resp: 2636521

Ion Ratio Lower Upper

228 100

229 20.2 15.5 23.3

226 27.8 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 72.298 ng/ul

RT: 21.186 min Scan# 3077

Delta R.T. -0.000 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:149 Resp: 1826323


Ion Ratio Lower Upper

149 100

167 25.4 19.7 29.5

279 3.8 3.3 4.9





#87

Chrysene

Concen: 75.031 ng/ul

RT: 21.297 min Scan# 3095

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:228 Resp: 2417092

Ion Ratio Lower Upper

228 100

226 30.7 24.3 36.5

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 61.309 ng/ul




RT: 22.068 min Scan# 3227


Delta R.T. -0.000 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Tgt Ion:149 Resp: 2985201





#90

Benzo(b)fluoranthene

Concen: 74.869 ng/ul

RT: 22.821 min Scan# 3355

Delta R.T. 0.006 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:252 Resp: 2725672

Ion Ratio Lower Upper

252 100

253 22.1 18.0 27.0

125 10.1 8.2 12.4



#91

Benzo(k)fluoranthene

Concen: 78.022 ng/ul

RT: 22.868 min Scan# 3363

Delta R.T. 0.012 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:252 Resp: 2781993


Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 9.8 8.5 12.7





#93

Benzo(a)pyrene

Concen: 78.574 ng/ul

RT: 23.397 min Scan# 34

Delta R.T. 0.012 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33

Instrument :

BNA_M

ClientSampleId :

SSTD080031





Tgt Ion:252 Resp: 2675908

Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4

125 11.0 8.8 13.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 88.910 ng/ul

RT: 25.738 min Scan# 3851

Delta R.T. 0.018 min

Lab File: BM044959.D

Acq: 05 Apr 2024 21:33




Tgt Ion:276 Resp: 3396139


Ion Ratio Lower Upper

276 100

138 24.6 19.0 28.6

277 25.1 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 89.950 ng/ul
 RT: 25.750 min Scan# 3853
 Delta R.T. 0.012 min
 Lab File: BM044959.D
 Acq: 05 Apr 2024 21:33

Instrument :
 BNA_M
 ClientSampleId :
 SST080031

Tgt Ion:278 Resp: 2864873
 Ion Ratio Lower Upper
 278 100
 139 15.6 11.9 17.9
 279 24.2 19.4 29.2



#96
 Benzo(g,h,i)perylene
 Concen: 86.515 ng/ul
 RT: 26.421 min Scan# 3967
 Delta R.T. 0.018 min
 Lab File: BM044959.D
 Acq: 05 Apr 2024 21:33

Tgt Ion:276 Resp: 2637138
 Ion Ratio Lower Upper
 276 100
 138 21.2 17.8 26.6
 277 23.4 18.7 28.1

