

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040524\
 Data File : BM044961.D
 Acq On : 05 Apr 2024 22:45
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV033

Quant Time: Apr 06 00:56:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 06 00:56:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	69991	20.000	ng/u1	0.00
20) Naphthalene-d8	10.410	136	302853	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.286	164	194922	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.045	188	405948	20.000	ng/u1	0.00
79) Chrysene-d12	21.250	240	435503	20.000	ng/u1	0.00
88) Perylene-d12	23.480	264	544978	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	15458	8.247	ng/uL	0.00
4) Pyridine-d5	3.610	84	113274	22.224	ng/u1	0.00
7) Phenol-d5	6.816	99	126906	21.388	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.987	67	84741	23.971	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	102754	22.104	ng/u1	0.00
15) 4-Methylphenol-d8	8.345	113	100006	21.533	ng/u1	0.00
21) Nitrobenzene-d5	8.798	128	51483	22.131	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	55017	22.407	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.039	165	98461	21.871	ng/u1	0.00
31) 4-Chloroaniline-d4	10.569	131	145545	20.306	ng/u1	0.00
46) Dimethylphthalate-d6	13.710	166	287887	21.212	ng/u1	0.00
49) Acenaphthylene-d8	13.974	160	334900	20.888	ng/u1	0.00
54) 4-Nitrophenol-d4	14.515	143	55498	19.594	ng/u1	0.00
60) Fluorene-d10	15.286	176	242675	20.056	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.427	200	47358	19.675	ng/u1	0.00
73) Anthracene-d10	17.145	188	384391	21.049	ng/u1	0.00
81) Pyrene-d10	19.450	212	457753	21.346	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.338	264	560128	21.069	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	15722	8.032	ng/uL#	89
5) Pyridine	3.628	79	101955	18.878	ng/u1	96
6) Benzaldehyde	6.798	77	67869	24.674	ng/u1	98
8) Phenol	6.839	94	126640	20.611	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.081	93	98514	20.544	ng/u1	98
12) 2-Chlorophenol	7.198	128	103525	21.224	ng/u1	97
13) 2-Methylphenol	8.081	108	90535	19.962	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.151	45	119557	20.903	ng/u1	97
16) Acetophenone	8.463	105	157978	20.922	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.445	70	79115	22.191	ng/u1	95
18) 4-Methylphenol	8.410	108	102324	20.932	ng/u1	100
19) Hexachloroethane	8.686	117	43221	20.146	ng/u1	98
22) Nitrobenzene	8.839	77	117876	20.653	ng/u1	97
23) Isophorone	9.363	82	210030	20.029	ng/u1	98
25) 2-Nitrophenol	9.539	139	56986	20.910	ng/u1	93
26) 2,4-Dimethylphenol	9.598	107	86035	16.375	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.845	93	127531	20.515	ng/u1	100
29) 2,4-Dichlorophenol	10.063	162	94916	20.868	ng/u1	99
30) Naphthalene	10.457	128	317145	19.782	ng/u1	99
32) 4-Chloroaniline	10.592	127	133834	19.190	ng/u1	99
33) Hexachlorobutadiene	10.727	225	54736	19.355	ng/u1	93
34) Caprolactam	11.386	113	32119	20.917	ng/u1	91
35) 4-Chloro-3-methylphenol	11.716	107	98836	21.860	ng/u1	98
36) 2-Methylnaphthalene	12.080	142	211994	20.252	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040524\
 Data File : BM044961.D
 Acq On : 05 Apr 2024 22:45
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SICV033


Quant Time: Apr 06 00:56:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 06 00:56:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.304	142	199607	18.843	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.451	216	106470	19.973	ng/ul	99
40) Hexachlorocyclopentadiene	12.416	237	43625	13.519	ng/ul	98
41) 2,4,6-Trichlorophenol	12.704	196	71919	20.682	ng/ul	94
42) 2,4,5-Trichlorophenol	12.780	196	77259	20.164	ng/ul	97
43) 1,1'-Biphenyl	13.110	154	278643	20.244	ng/ul	99
44) 2-Chloronaphthalene	13.151	162	217293	19.431	ng/ul	100
45) 2-Nitroaniline	13.380	65	67091	21.734	ng/ul	93
47) Dimethylphthalate	13.757	163	273362	19.270	ng/ul	98
48) 2,6-Dinitrotoluene	13.886	165	59626	21.445	ng/ul	93
50) Acenaphthylene	14.004	152	367386	19.619	ng/ul	99
51) 3-Nitroaniline	14.215	138	67087	22.120	ng/ul	91
52) Acenaphthene	14.351	153	234826	18.793	ng/ul	99
53) 2,4-Dinitrophenol	14.433	184	29373	17.769	ng/ul	95
55) 4-Nitrophenol	14.527	109	46340	18.603	ng/ul	94
56) Dibenzofuran	14.686	168	335643	19.630	ng/ul	98
57) 2,4-Dinitrotoluene	14.680	165	83533	20.071	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	14.915	232	66980	20.426	ng/ul	95
59) Diethylphthalate	15.127	149	276215	19.086	ng/ul	99
61) Fluorene	15.339	166	271482	18.838	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.345	204	124337	18.407	ng/ul	98
63) 4-Nitroaniline	15.386	138	61938	22.463	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.439	198	49579	19.234	ng/ul	100
67) N-Nitrosodiphenylamine	15.562	169	227113	20.739	ng/ul	99
68) 4-Bromophenyl-phenylether	16.239	248	77727	19.660	ng/ul	97
69) Hexachlorobenzene	16.339	284	100730	19.586	ng/ul	97
70) Atrazine	16.527	200	75476	18.737	ng/ul	99
71) Pentachlorophenol	16.692	266	59636	19.089	ng/ul	98
72) Phenanthrene	17.086	178	424014	19.410	ng/ul	98
74) Anthracene	17.180	178	434457	19.442	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.068	216	105983	20.714	ng/uL	98
76) Pentachlorobenzene	14.598	250	107692	19.782	ng/uL	97
77) Carbazole	17.462	167	418043	19.401	ng/ul	99
78) Di-n-butylphthalate	18.027	149	469219	19.310	ng/ul	100
80) Fluoranthene	19.115	202	501862	19.842	ng/ul	99
82) Pyrene	19.480	202	543524	19.914	ng/ul	99
83) Butylbenzylphthalate	20.403	149	219886	19.229	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.180	252	216447	22.352	ng/ul	95
85) Benzo(a)anthracene	21.239	228	579227	19.602	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.180	149	338730	18.950	ng/ul	99
87) Chrysene	21.291	228	542787	19.704	ng/ul	100
89) Di-n-octyl phthalate	22.068	149	582565	16.990	ng/ul	100
90) Benzo(b)fluoranthene	22.809	252	633478	19.944	ng/ul	99
91) Benzo(k)fluoranthene	22.856	252	610203	19.086	ng/ul	99
93) Benzo(a)pyrene	23.380	252	599523	19.394	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.715	276	804506	20.091	ng/ul	99
95) Dibenzo(a,h)anthracene	25.732	278	655405	19.576	ng/ul	98
96) Benzo(g,h,i)perylene	26.397	276	612068	19.375	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 06 00:56:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 06 00:56:29 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.032 ng/uL
 RT: 3.246 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BM044961.D
 Acq: 05 Apr 2024 22:45




Instrument :
 BNA_M
 ClientSampleId :
 SICV033


Tgt Ion: 88 Resp: 15722
 Ion Ratio Lower Upper
 88 100
 43 29.8 24.0 36.0
 58 69.2 45.4 68.2#



#5
 Pyridine
 Concen: 18.878 ng/uL
 RT: 3.628 min Scan# 92
 Delta R.T. -0.000 min
 Lab File: BM044961.D
 Acq: 05 Apr 2024 22:45

Tgt Ion: 79 Resp: 101955
 Ion Ratio Lower Upper
 79 100
 52 72.8 56.0 84.0
 51 37.1 27.6 41.4









#6
Benzaldehyde
Concen: 24.674 ng/ul
RT: 6.798 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion: 77 Resp: 67869
Ion Ratio Lower Upper
77 100
105 96.4 75.8 113.6
106 90.2 74.6 112.0



#8
Phenol
Concen: 20.611 ng/ul
RT: 6.839 min Scan# 638
Delta R.T. -0.006 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion: 94 Resp: 126640
Ion Ratio Lower Upper
94 100
65 29.7 21.7 32.5
66 39.5 31.9 47.9









#10
Bis(2-Chloroethyl)ether
Concen: 20.544 ng/ul
RT: 7.081 min Scan# 610
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion: 93 Resp: 98514
Ion Ratio Lower Upper
93 100
63 75.9 61.3 91.9
95 32.6 27.6 41.4



#12
2-Chlorophenol
Concen: 21.224 ng/ul
RT: 7.198 min Scan# 699
Delta R.T. -0.006 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:128 Resp: 103525
Ion Ratio Lower Upper
128 100
64 44.4 33.3 49.9
130 31.8 26.2 39.4









#13
2-Methylphenol
Concen: 19.962 ng/ul
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:108 Resp: 90535
Ion Ratio Lower Upper
108 100
107 88.5 69.1 103.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.903 ng/ul
RT: 8.151 min Scan# 861
Delta R.T. -0.006 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion: 45 Resp: 119557
Ion Ratio Lower Upper
45 100
77 18.3 13.4 20.0
79 13.3 10.1 15.1









#16
Acetophenone
Concen: 20.922 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:105 Resp: 157978
Ion Ratio Lower Upper
105 100
77 78.6 65.8 98.6
51 25.5 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 22.191 ng/ul
RT: 8.445 min Scan# 911
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion: 70 Resp: 79115
Ion Ratio Lower Upper
70 100
42 58.2 43.2 64.8
101 8.9 6.4 9.6
130 22.2 16.2 24.2









#18
4-Methylphenol
Concen: 20.932 ng/ul
RT: 8.410 min Scan# 905
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:108 Resp: 102324
Ion Ratio Lower Upper
108 100
107 112.4 90.0 135.0



#19
Hexachloroethane
Concen: 20.146 ng/ul
RT: 8.686 min Scan# 952
Delta R.T. -0.006 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:117 Resp: 43221
Ion Ratio Lower Upper
117 100
201 90.6 73.0 109.4
199 58.0 44.4 66.6









#22
Nitrobenzene
Concen: 20.653 ng/ul
RT: 8.839 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion: 77 Resp: 117876
Ion Ratio Lower Upper
77 100
123 44.4 37.4 56.2
65 12.8 10.7 16.1



#23
Isophorone
Concen: 20.029 ng/ul
RT: 9.363 min Scan# 1067
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion: 82 Resp: 210030
Ion Ratio Lower Upper
82 100
95 7.4 5.1 7.7
138 18.4 14.2 21.4









#25
2-Nitrophenol
Concen: 20.910 ng/ul
RT: 9.539 min Scan# 1097
Delta R.T. -0.006 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:139 Resp: 56986
Ion Ratio Lower Upper
139 100
65 51.0 34.9 52.3
109 30.2 24.4 36.6



#26
2,4-Dimethylphenol
Concen: 16.375 ng/ul
RT: 9.598 min Scan# 1107
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:107 Resp: 86035
Ion Ratio Lower Upper
107 100
121 48.6 41.1 61.7
122 87.2 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.515 ng/ul
 RT: 9.845 min Scan# 1149
 Delta R.T. -0.000 min
 Lab File: BM044961.D
 Acq: 05 Apr 2024 22:45




Instrument :
 BNA_M
 ClientSampleId :
 SICV033


Tgt Ion: 93 Resp: 127531
 Ion Ratio Lower Upper
 93 100
 95 33.5 26.9 40.3
 123 10.9 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 20.868 ng/ul
 RT: 10.063 min Scan# 1186
 Delta R.T. -0.006 min
 Lab File: BM044961.D
 Acq: 05 Apr 2024 22:45

Tgt Ion: 162 Resp: 94916
 Ion Ratio Lower Upper
 162 100
 164 63.3 50.4 75.6
 98 32.5 27.0 40.4





#30

Naphthalene

Concen: 19.782 ng/ul

RT: 10.457 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:128 Resp: 317145

Ion Ratio Lower Upper

128 100

129 10.7 8.5 12.7

127 13.5 10.2 15.4



#32

4-Chloroaniline

Concen: 19.190 ng/ul

RT: 10.592 min Scan# 1276

Delta R.T. -0.000 min

Lab File: BM044961.D




Acq: 05 Apr 2024 22:45


Tgt Ion:127 Resp: 133834

Ion Ratio Lower Upper

127 100

129 32.3 26.2 39.2









#33
Hexachlorobutadiene
Concen: 19.355 ng/ul
RT: 10.727 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:225 Resp: 54736
Ion Ratio Lower Upper
225 100
223 61.7 45.3 67.9
227 64.4 47.6 71.4



#34
Caprolactam
Concen: 20.917 ng/ul
RT: 11.386 min Scan# 1411
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:113 Resp: 32119
Ion Ratio Lower Upper
113 100
55 143.3 125.4 188.0
56 107.1 91.8 137.8









#35
4-Chloro-3-methylphenol
Concen: 21.860 ng/ul
RT: 11.716 min Scan# 1467
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:107 Resp: 98836
Ion Ratio Lower Upper
107 100
144 29.2 22.8 34.2
142 85.1 69.8 104.6



#36
2-Methylnaphthalene
Concen: 20.252 ng/ul
RT: 12.080 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:142 Resp: 211994
Ion Ratio Lower Upper
142 100
141 91.2 72.3 108.5





#37

1-Methylnaphthalene

Concen: 18.843 ng/ul

RT: 12.304 min Scan# 1567

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:142 Resp: 199607

Ion Ratio Lower Upper

142 100

141 93.0 77.2 115.8

116 4.2 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.973 ng/ul

RT: 12.451 min Scan# 1592

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Tgt Ion:216 Resp: 106470




Ion Ratio Lower Upper


216 100

214 79.1 63.5 95.3

179 20.0 16.6 25.0

108 20.9 15.0 22.4









#40
Hexachlorocyclopentadiene
Concen: 13.519 ng/ul
RT: 12.416 min Scan# 1586
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:237 Resp: 43625
Ion Ratio Lower Upper
237 100
235 62.3 51.3 76.9
272 13.3 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 20.682 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.006 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:196 Resp: 71919
Ion Ratio Lower Upper
196 100
198 91.6 79.4 119.2
200 29.8 25.0 37.4





#42

2,4,5-Trichlorophenol

Concen: 20.164 ng/ul

RT: 12.780 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:196 Resp: 77259

Ion Ratio Lower Upper

196 100

198 96.6 79.9 119.9

200 29.9 25.6 38.4



#43

1,1'-Biphenyl

Concen: 20.244 ng/ul

RT: 13.110 min Scan# 1704

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45




Tgt Ion:154 Resp: 278643


Ion Ratio Lower Upper

154 100

153 38.7 31.6 47.4

76 12.7 10.4 15.6





#44

2-Chloronaphthalene

Concen: 19.431 ng/ul

RT: 13.151 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:162 Resp: 217293

Ion Ratio Lower Upper

162 100

164 32.3 26.1 39.1

127 37.6 30.1 45.1



#45

2-Nitroaniline

Concen: 21.734 ng/ul

RT: 13.380 min Scan# 1750

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45




Tgt Ion: 65 Resp: 67091


Ion Ratio Lower Upper

65 100

92 66.8 55.2 82.8

138 106.1 93.0 139.4









#47
Dimethylphthalate
Concen: 19.270 ng/ul
RT: 13.757 min Scan# 1814
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:163 Resp: 273362
Ion Ratio Lower Upper
163 100
194 3.8 2.8 4.2
164 10.7 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 21.445 ng/ul
RT: 13.886 min Scan# 1836
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:165 Resp: 59626
Ion Ratio Lower Upper
165 100
89 62.5 45.0 67.4
121 24.6 18.0 27.0





#50

Acenaphthylene

Concen: 19.619 ng/ul

RT: 14.004 min Scan# 1856

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:152 Resp: 367386

Ion Ratio Lower Upper

152 100

151 20.3 16.0 24.0

153 13.0 10.2 15.2



#51

3-Nitroaniline

Concen: 22.120 ng/ul

RT: 14.215 min Scan# 1892

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45




Tgt Ion:138 Resp: 67087


Ion Ratio Lower Upper

138 100

108 9.1 6.5 9.7

92 110.9 81.4 122.0









#52
Acenaphthene
Concen: 18.793 ng/ul
RT: 14.351 min Scan# 1915
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:153 Resp: 234826
Ion Ratio Lower Upper
153 100
152 47.8 38.6 58.0
154 87.5 69.4 104.2



#53
2,4-Dinitrophenol
Concen: 17.769 ng/ul
RT: 14.433 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:184 Resp: 29373
Ion Ratio Lower Upper
184 100
63 75.1 56.9 85.3
154 67.5 50.3 75.5









#55
4-Nitrophenol
Concen: 18.603 ng/ul
RT: 14.527 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:109 Resp: 46340
Ion Ratio Lower Upper
109 100
139 124.9 107.7 161.5
65 123.7 96.3 144.5



#56
Dibenzofuran
Concen: 19.630 ng/ul
RT: 14.686 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:168 Resp: 335643
Ion Ratio Lower Upper
168 100
139 39.6 30.6 46.0









#57
2,4-Dinitrotoluene
Concen: 20.071 ng/ul
RT: 14.680 min Scan# 1971
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:165 Resp: 83533
Ion Ratio Lower Upper
165 100
63 62.3 52.0 78.0
182 2.3 2.3 3.5#



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.426 ng/ul
RT: 14.915 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:232 Resp: 66980
Ion Ratio Lower Upper
232 100
131 49.6 43.4 65.0
130 3.2 2.6 4.0
166 33.8 25.6 38.4









#59
Diethylphthalate
Concen: 19.086 ng/ul
RT: 15.127 min Scan# 2047
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:149 Resp: 276215
Ion Ratio Lower Upper
149 100
177 20.9 16.4 24.6
150 12.0 10.3 15.5



#61
Fluorene
Concen: 18.838 ng/ul
RT: 15.339 min Scan# 2083
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:166 Resp: 271482
Ion Ratio Lower Upper
166 100
165 98.8 80.5 120.7
167 13.1 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 18.407 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. 0.006 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:204 Resp: 124337
Ion Ratio Lower Upper
204 100
206 33.1 24.9 37.3
141 58.9 46.2 69.2



#63
4-Nitroaniline
Concen: 22.463 ng/ul
RT: 15.386 min Scan# 2091
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:138 Resp: 61938
Ion Ratio Lower Upper
138 100
92 52.6 39.8 59.6
108 92.0 66.4 99.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.234 ng/ul
RT: 15.439 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:198 Resp: 49579
Ion Ratio Lower Upper
198 100
182 4.7 3.8 5.6
77 30.7 24.6 37.0



#67
N-Nitrosodiphenylamine
Concen: 20.739 ng/ul
RT: 15.562 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:169 Resp: 227113
Ion Ratio Lower Upper
169 100
168 71.3 56.7 85.1
167 40.3 31.2 46.8









#68
4-Bromophenyl-phenylether
Concen: 19.660 ng/ul
RT: 16.239 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:248 Resp: 77727
Ion Ratio Lower Upper
248 100
250 98.4 81.0 121.6
141 72.2 55.1 82.7



#69
Hexachlorobenzene
Concen: 19.586 ng/ul
RT: 16.339 min Scan# 2253
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:284 Resp: 100730
Ion Ratio Lower Upper
284 100
142 36.6 27.3 40.9
249 28.9 23.6 35.4





#70

Atrazine

Concen: 18.737 ng/ul

RT: 16.527 min Scan# 2285

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:200 Resp: 75476

Ion Ratio Lower Upper

200 100

173 26.3 20.3 30.5

215 47.8 37.9 56.9



#71

Pentachlorophenol

Concen: 19.089 ng/ul

RT: 16.692 min Scan# 2313

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45




Tgt Ion:266 Resp: 59636


Ion Ratio Lower Upper

266 100

264 62.5 49.3 73.9

268 64.1 53.3 79.9





#72

Phenanthrene

Concen: 19.410 ng/ul

RT: 17.086 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:178 Resp: 424014

Ion Ratio Lower Upper

178 100

179 15.6 11.9 17.9

176 19.9 15.4 23.2



#74

Anthracene

Concen: 19.442 ng/ul

RT: 17.180 min Scan# 2396

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45




Tgt Ion:178 Resp: 434457


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.6 15.8 23.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.714 ng/uL
RT: 13.068 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:216 Resp: 105983
Ion Ratio Lower Upper
216 100
214 77.1 64.1 96.1
218 48.3 38.6 58.0



#76
Pentachlorobenzene
Concen: 19.782 ng/uL
RT: 14.598 min Scan# 1957
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:250 Resp: 107692
Ion Ratio Lower Upper
250 100
248 63.6 49.8 74.6
252 64.8 49.4 74.2





#77

Carbazole

Concen: 19.401 ng/ul

RT: 17.462 min Scan# 2444

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:167 Resp: 418043

Ion Ratio Lower Upper

167 100

166 22.2 17.3 25.9

139 13.7 10.7 16.1



#78

Di-n-butylphthalate

Concen: 19.310 ng/ul

RT: 18.027 min Scan# 2540

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45




Tgt Ion:149 Resp: 469219


Ion Ratio Lower Upper

149 100

150 9.1 7.2 10.8

104 5.6 4.3 6.5









#80
 Fluoranthene
 Concen: 19.842 ng/ul
 RT: 19.115 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BM044961.D
 Acq: 05 Apr 2024 22:45




Instrument :
 BNA_M
 ClientSampleId :
 SICV033


Tgt Ion:202 Resp: 501862
 Ion Ratio Lower Upper
 202 100
 101 10.4 8.7 13.1
 100 8.5 6.3 9.5



#82
 Pyrene
 Concen: 19.914 ng/ul
 RT: 19.480 min Scan# 2787
 Delta R.T. -0.000 min
 Lab File: BM044961.D
 Acq: 05 Apr 2024 22:45

Tgt Ion:202 Resp: 543524
 Ion Ratio Lower Upper
 202 100
 101 12.5 10.2 15.4
 100 10.4 8.5 12.7









#83
 Butylbenzylphthalate
 Concen: 19.229 ng/ul
 RT: 20.403 min Scan# 2944
 Delta R.T. -0.000 min
 Lab File: BM044961.D
 Acq: 05 Apr 2024 22:45




Instrument :
 BNA_M
 ClientSampleId :
 SICV033


Tgt Ion:	149	Resp:	219886
Ion Ratio	Lower	Upper	
149	100		
91	70.3	56.2	84.4
206	17.8	14.3	21.5



#84
 3,3'-Dichlorobenzidine
 Concen: 22.352 ng/ul
 RT: 21.180 min Scan# 3076
 Delta R.T. -0.000 min
 Lab File: BM044961.D
 Acq: 05 Apr 2024 22:45

Tgt Ion:	252	Resp:	216447
Ion Ratio	Lower	Upper	
252	100		
254	62.7	53.9	80.9
126	11.4	9.0	13.4





#85

Benzo(a)anthracene

Concen: 19.602 ng/ul

RT: 21.239 min Scan# 3086

Delta R.T. -0.000 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:228 Resp: 579227

Ion Ratio Lower Upper

228 100

229 19.8 15.5 23.3

226 27.7 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.950 ng/ul

RT: 21.180 min Scan# 3076

Delta R.T. -0.006 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45




Tgt Ion:149 Resp: 338730


Ion Ratio Lower Upper

149 100

167 24.9 19.7 29.5

279 3.5 3.3 4.9









#87
Chrysene
Concen: 19.704 ng/ul
RT: 21.291 min Scan# 3095
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:228 Resp: 542787
Ion Ratio Lower Upper
228 100
226 30.5 24.3 36.5
229 19.3 15.7 23.5



#89
Di-n-octyl phthalate
Concen: 16.990 ng/ul
RT: 22.068 min Scan# 3227
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:149 Resp: 582565









#90
Benzo(b)fluoranthene
Concen: 19.944 ng/ul
RT: 22.809 min Scan# 3353
Delta R.T. -0.006 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45




Instrument :
BNA_M
ClientSampleId :
SICV033


Tgt Ion:252 Resp: 633478
Ion Ratio Lower Upper
252 100
253 22.3 18.0 27.0
125 10.8 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 19.086 ng/ul
RT: 22.856 min Scan# 3361
Delta R.T. -0.000 min
Lab File: BM044961.D
Acq: 05 Apr 2024 22:45

Tgt Ion:252 Resp: 610203
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 10.6 8.5 12.7





#93

Benzo(a)pyrene

Concen: 19.394 ng/ul

RT: 23.380 min Scan# 34

Delta R.T. -0.006 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:252 Resp: 599523

Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 11.9 8.8 13.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.091 ng/ul

RT: 25.715 min Scan# 3847

Delta R.T. -0.006 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45




Tgt Ion:276 Resp: 804506


Ion Ratio Lower Upper

276 100

138 23.2 19.0 28.6

277 24.3 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 19.576 ng/ul

RT: 25.732 min Scan# 3851

Delta R.T. -0.006 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Instrument :

BNA_M

ClientSampleId :

SICV033





Tgt Ion:278 Resp: 655405

Ion Ratio Lower Upper

278 100

139 15.8 11.9 17.9

279 23.1 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 19.375 ng/ul

RT: 26.397 min Scan# 3963

Delta R.T. -0.006 min

Lab File: BM044961.D

Acq: 05 Apr 2024 22:45

Tgt Ion:276 Resp: 612068

Ion Ratio Lower Upper

276 100

138 20.6 17.8 26.6

277 23.6 18.7 28.1

