


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM040620\
 Data File : BM025941.D
 Acq On : 06 Apr 2020 08:40
 Operator : CG/JU
 Sample : DFTPP96
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP96

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM040220.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Apr 02 10:49:07 2020



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.9	205824	PASS
68	69	0.00	2	1.7	3840	PASS
69	198	0.00	100	38.1	224938	PASS
70	69	0.00	2	0.6	1244	PASS
127	198	10	80	47.3	278890	PASS
197	198	0.00	2	0.8	4779	PASS
198	198	100	100	100.0	590101	PASS
199	198	5	9	6.6	39205	PASS
275	198	10	60	31.9	188096	PASS
365	198	1	100	3.8	22320	PASS
441	198	0.01	100	13.7	80944	PASS
442	198	50	100	94.1	555114	PASS
443	442	15	24	19.4	107845	PASS

Scan 2463 (17.374 min): BM025941.D
DFTPP96

m/z	Abundance
40.00	960.0
44.00	1000.0
49.00	762.0
73.10	497.0
83.90	883.0
85.80	411.0
207.00	965.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

m/z	Abundance
37.00	972.0
38.10	3180.0
39.10	18072.0
40.00	1542.0
41.00	718.0
44.00	1355.0
44.90	392.0
49.00	2729.0
50.10	58392.0
51.10	233536.0
52.10	11693.0
53.00	532.0
55.00	946.0
56.00	6669.0
57.00	14922.0
58.00	837.0
61.00	2675.0
62.00	3249.0
63.00	9215.0
64.00	994.0
65.10	4272.0
66.00	454.0
67.10	583.0
68.00	4264.0
69.00	248576.0
70.00	1486.0
73.10	2273.0
74.10	24240.0
75.00	39192.0
76.00	12241.0
77.10	262592.0
78.10	18568.0
79.00	17448.0
80.00	14847.0
81.00	18480.0
82.00	4284.0
83.00	4376.0
84.00	1618.0
85.00	2939.0
86.00	5503.0
87.10	2368.0
88.00	838.0
89.00	420.0
91.00	4220.0
92.00	5444.0
93.00	30360.0
94.00	1999.0
95.10	537.0
96.00	1492.0
97.00	467.0
98.00	23304.0
99.00	17928.0
99.90	1454.0
101.00	11001.0
101.90	772.0
103.00	3564.0
104.00	6500.0
105.00	6353.0
106.10	2200.0
107.00	79344.0
108.00	12346.0
109.10	2153.0
110.00	147200.0
111.00	20352.0
112.00	3032.0
113.10	943.0
115.10	420.0
116.00	4317.0
117.00	61232.0
118.00	4897.0
119.00	536.0
120.00	1060.0
120.90	487.0
122.00	5751.0
123.00	8903.0
124.00	4312.0
125.00	3458.0
126.20	2167.0
127.00	293184.0
128.00	22736.0
129.00	110008.0
130.00	7834.0
131.00	2254.0
132.10	1276.0
132.80	721.0
134.00	3238.0
135.00	9826.0
136.00	3471.0
137.10	4380.0
138.10	892.0
139.00	618.0
140.10	1365.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

141.00	13887.0
142.00	4688.0
143.00	3222.0
144.10	854.0
145.10	704.0
146.00	2327.0
147.00	7391.0
148.00	16325.0
149.00	3302.0
150.00	883.0
151.00	1822.0
151.90	1069.0
153.00	4401.0
154.00	2902.0
155.00	7978.0
156.00	11984.0
157.10	2375.0
158.00	2834.0
159.00	2091.0
160.00	4158.0
161.10	6631.0
162.10	2041.0
163.00	510.0
163.90	770.0
165.00	5217.0
166.10	4456.0
167.00	27880.0
168.00	14018.0
169.00	2134.0
169.90	908.0
170.90	1146.0
172.00	2485.0
173.00	3578.0
174.00	5433.0
175.00	10616.0
176.10	3459.0
177.00	5281.0
178.00	1502.0
179.00	21040.0
180.00	15453.0
181.00	6097.0
182.10	981.0
183.10	558.0
184.00	1537.0
185.00	10230.0
186.00	74928.0
187.00	22632.0
188.00	2574.0
189.00	4933.0
190.00	976.0
191.00	2788.0
192.00	5884.0
193.00	6525.0
194.00	1667.0
195.00	1186.0
196.00	17776.0
197.10	4822.0
198.00	578432.0
199.00	40376.0
200.00	3303.0
201.40	2171.0
203.00	3846.0
204.10	18624.0
205.00	33464.0
206.00	142272.0
207.00	20808.0
208.00	4955.0
209.00	1514.0
210.40	2940.0
211.00	5442.0
212.10	1088.0
213.00	383.0
214.10	342.0
215.10	1383.0
216.00	2962.0
217.00	36096.0
218.00	5343.0
219.10	552.0
221.00	25592.0
223.00	8295.0
224.00	78696.0
225.00	20352.0
226.00	2240.0
227.00	31296.0
228.00	4843.0
229.00	7014.0
230.00	816.0
231.00	3117.0
232.00	619.0
233.00	563.0
234.00	2199.0
235.00	2307.0
236.00	1611.0
237.00	3297.0
237.90	522.0
239.00	1414.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

240.00	879.0
240.90	2151.0
242.10	3681.0
243.10	4256.0
244.00	67232.0
245.10	9134.0
246.00	12404.0
247.00	2668.0
248.10	704.0
249.00	2455.0
249.90	552.0
251.10	533.0
252.00	704.0
253.10	1281.0
254.20	3392.0
255.00	357312.0
256.00	51816.0
257.00	4119.0
258.00	19256.0
259.00	3484.0
260.00	585.0
261.00	636.0
264.00	658.0
265.00	7792.0
266.00	1373.0
270.00	490.0
271.00	1125.0
272.00	969.0
273.00	10463.0
274.00	28440.0
275.00	167680.0
276.00	22632.0
277.00	13216.0
278.00	2385.0
279.10	867.0
281.00	493.0
282.10	448.0
283.00	1448.0
284.00	843.0
285.00	2594.0
286.00	359.0
289.00	587.0
290.00	430.0
290.90	563.0
292.10	549.0
293.00	2783.0
294.00	934.0
295.10	824.0
296.00	39208.0
297.10	6530.0
298.00	500.0
301.00	812.0
302.00	687.0
303.10	5734.0
304.00	1446.0
308.10	735.0
309.00	423.0
309.90	341.0
314.00	2310.0
314.90	4544.0
316.10	2855.0
317.10	591.0
321.10	1165.0
322.10	530.0
323.00	14107.0
324.00	2884.0
326.10	312.0
327.00	2690.0
328.00	1516.0
332.10	1196.0
333.00	1157.0
334.00	8485.0
335.00	2059.0
335.90	318.0
341.00	1554.0
342.10	416.0
346.00	2925.0
347.00	400.0
351.00	385.0
352.00	3710.0
353.00	2474.0
354.10	4425.0
355.10	966.0
365.00	19584.0
366.00	2655.0
370.00	423.0
371.00	1039.0
372.10	6811.0
373.00	2054.0
383.00	1963.0
384.00	548.0
389.90	778.0
391.00	706.0
392.00	498.0
401.00	505.0
402.00	2985.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

403.10	3779.0
404.00	1491.0
421.00	3041.0
422.10	2833.0
423.10	21800.0
424.10	6018.0
425.00	640.0
441.10	60688.0
442.10	413568.0
443.10	83456.0
444.10	8296.0
445.10	613.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

m/z	Abundance
37.10	999.0
38.00	3191.0
39.10	19152.0
40.00	1277.0
41.10	737.0
43.90	1020.0
44.90	621.0
49.00	1880.0
50.10	64344.0
51.10	240128.0
52.10	12372.0
53.00	408.0
55.10	1331.0
56.00	7738.0
57.00	16896.0
57.90	789.0
59.00	361.0
61.00	3514.0
62.00	3260.0
63.10	10065.0
64.00	1177.0
65.10	4937.0
66.00	401.0
68.10	4383.0
69.00	268992.0
70.00	1328.0
70.90	342.0
73.10	2718.0
74.00	25288.0
75.00	41224.0
76.10	12997.0
77.10	285824.0
78.10	18624.0
79.00	18432.0
80.00	14291.0
81.00	20944.0
82.00	5274.0
83.00	5126.0
83.90	1186.0
85.00	3661.0
86.00	6034.0
87.00	2661.0
88.10	1140.0
88.80	471.0
91.00	4528.0
92.00	5053.0
93.00	32296.0
94.00	2382.0
95.10	589.0
96.00	1578.0
97.00	668.0
98.00	23864.0
99.00	19208.0
100.00	1865.0
101.00	10902.0
102.10	716.0
103.00	3884.0
104.00	7115.0
105.00	7330.0
106.00	1755.0
107.00	89728.0
108.00	13098.0
109.20	2940.0
110.00	167168.0
111.00	23400.0
112.00	2742.0
112.90	935.0
115.00	617.0
116.00	4975.0
117.00	68040.0
118.00	5279.0
119.10	583.0
119.90	971.0
121.10	454.0
122.00	6183.0
123.00	9058.0
124.00	4394.0
125.00	3884.0
127.00	330240.0
128.00	25456.0
129.00	128088.0
130.00	10361.0
130.90	1793.0
132.00	1459.0
133.00	545.0
134.00	3562.0
135.00	10105.0
136.10	4192.0
137.10	5361.0
137.80	1106.0
138.80	523.0
140.00	1611.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

141.00	17736.0
142.00	4737.0
143.00	4207.0
144.10	1178.0
145.00	714.0
146.00	3315.0
147.00	9050.0
148.00	18008.0
149.00	3713.0
150.10	997.0
151.20	1781.0
151.90	1131.0
153.00	5579.0
154.00	4012.0
155.00	9811.0
156.00	13903.0
157.00	2958.0
157.90	2746.0
158.90	2647.0
160.00	4864.0
161.00	7650.0
162.00	2269.0
162.90	401.0
164.00	968.0
165.00	5829.0
166.00	5358.0
167.00	30416.0
168.00	14554.0
169.00	2670.0
170.00	1179.0
171.00	1429.0
171.90	3547.0
173.00	3950.0
174.00	7186.0
175.10	12915.0
176.00	4197.0
177.00	5974.0
178.00	2334.0
179.00	24288.0
180.00	18400.0
181.10	7482.0
182.00	1334.0
183.00	715.0
184.00	1967.0
185.00	13010.0
186.00	89312.0
187.00	25984.0
188.00	2716.0
189.00	5995.0
190.00	1068.0
191.00	3356.0
192.00	7569.0
193.00	8954.0
194.10	1753.0
195.10	1167.0
196.00	22888.0
197.20	5383.0
198.00	704128.0
199.00	46192.0
200.00	4329.0
201.60	3707.0
203.00	4433.0
204.00	24712.0
205.00	39752.0
206.00	176064.0
207.00	25000.0
208.00	6490.0
209.00	2010.0
210.40	2966.0
211.10	6876.0
213.00	501.0
214.10	317.0
214.90	2147.0
216.10	3899.0
217.00	45232.0
218.00	5230.0
219.10	644.0
221.00	34928.0
223.00	10483.0
224.00	97152.0
225.10	25544.0
226.00	2855.0
227.00	37992.0
228.00	4900.0
229.00	9453.0
229.90	1096.0
231.00	3873.0
232.00	657.0
233.00	963.0
233.90	2558.0
235.00	3343.0
236.00	1756.0
237.10	3717.0
238.10	391.0
239.00	1821.0
240.00	1008.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

241.00	2234.0
242.10	5054.0
243.10	6373.0
244.10	80424.0
245.10	10430.0
246.00	16584.0
247.00	3865.0
248.00	629.0
249.00	3423.0
250.10	536.0
251.00	755.0
252.00	907.0
253.00	1797.0
254.20	3945.0
255.00	451712.0
256.00	66224.0
257.10	4990.0
258.00	22760.0
259.00	4186.0
259.90	490.0
261.00	750.0
263.90	807.0
265.00	10248.0
266.00	1524.0
270.00	459.0
271.10	951.0
272.00	958.0
273.00	14438.0
274.00	36624.0
275.00	222208.0
276.10	27976.0
277.00	17456.0
278.00	3602.0
279.00	831.0
280.90	512.0
282.10	447.0
283.00	1749.0
284.00	1107.0
285.00	3401.0
286.10	575.0
288.90	691.0
289.90	591.0
292.10	864.0
293.00	3730.0
294.10	1104.0
295.10	1239.0
296.00	56864.0
297.00	7892.0
298.00	602.0
301.10	793.0
301.90	1009.0
303.10	6785.0
304.00	1968.0
308.00	1076.0
308.90	368.0
310.00	856.0
313.10	522.0
314.00	3087.0
315.00	6955.0
316.10	3590.0
317.00	671.0
321.10	1617.0
322.00	1218.0
323.10	19088.0
324.10	3287.0
326.10	430.0
327.00	3417.0
328.00	1988.0
328.90	331.0
332.00	1205.0
333.10	1736.0
334.00	13054.0
335.00	3154.0
336.10	415.0
341.10	2614.0
342.10	722.0
346.00	4225.0
347.10	589.0
351.20	332.0
352.10	5748.0
353.10	4143.0
354.00	6079.0
355.00	1156.0
359.00	505.0
365.00	25600.0
366.10	3944.0
370.10	637.0
371.00	1376.0
372.10	10210.0
373.10	2838.0
383.00	2563.0
384.00	726.0
390.00	1341.0
391.00	910.0
391.90	609.0
401.10	626.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

402.10	4413.0
403.00	5263.0
404.10	1789.0
421.10	4232.0
422.10	4645.0
423.10	33464.0
424.10	8254.0
425.00	1133.0
441.10	94024.0
442.10	644480.0
443.10	124800.0
444.10	12858.0
445.00	938.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

m/z	Abundance
36.90	501.0
38.10	1838.0
39.10	11711.0
40.00	1146.0
41.00	405.0
44.00	1434.0
45.00	318.0
49.10	1779.0
50.00	36056.0
51.10	143808.0
52.10	6836.0
55.10	690.0
56.00	3982.0
57.00	10324.0
58.00	433.0
61.00	2116.0
62.00	2296.0
63.00	6499.0
64.00	889.0
65.00	2986.0
67.00	333.0
68.10	2873.0
69.00	157248.0
70.00	919.0
73.10	1428.0
74.00	15359.0
75.00	23824.0
76.10	8609.0
77.10	172416.0
78.10	11808.0
79.00	11254.0
80.00	9560.0
81.00	12782.0
82.00	3417.0
83.00	3352.0
84.10	733.0
85.00	2030.0
86.00	3807.0
87.00	1699.0
88.00	633.0
91.00	2635.0
92.00	3247.0
93.00	19648.0
94.00	1205.0
95.10	376.0
96.00	904.0
98.00	15075.0
99.00	12063.0
100.00	942.0
101.00	7297.0
103.00	2413.0
104.00	4625.0
105.00	4459.0
106.00	1754.0
107.00	56360.0
108.00	8905.0
109.00	1689.0
110.00	101816.0
111.00	15524.0
112.00	1753.0
113.00	617.0
116.00	3028.0
117.00	43448.0
118.00	3299.0
119.10	413.0
120.00	687.0
120.90	301.0
122.00	3845.0
123.00	5554.0
124.00	2564.0
125.00	2945.0
127.00	213248.0
128.00	15560.0
129.00	82288.0
130.00	6485.0
131.00	1315.0
132.10	679.0
132.90	453.0
133.90	2481.0
135.00	6916.0
136.00	2721.0
137.00	3120.0
137.80	915.0
139.10	368.0
139.90	871.0
141.00	11095.0
142.00	3521.0
143.00	2336.0
144.10	700.0
145.00	537.0
145.90	1925.0
147.00	5826.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

148.00	11862.0
149.00	3022.0
150.00	766.0
151.10	1326.0
152.00	848.0
153.00	3730.0
154.00	2687.0
155.00	5502.0
156.00	9582.0
157.00	1746.0
157.90	2035.0
158.90	1286.0
160.00	3403.0
161.00	4546.0
162.00	1560.0
163.00	420.0
164.20	660.0
165.00	3983.0
166.00	3667.0
167.00	21736.0
168.00	10826.0
169.00	1913.0
169.90	835.0
170.90	1183.0
172.00	1982.0
173.00	2911.0
174.00	5100.0
175.10	9081.0
176.00	2969.0
177.00	4137.0
178.10	1243.0
179.00	17240.0
180.00	11610.0
181.00	6240.0
182.00	1176.0
182.90	548.0
183.90	1412.0
185.00	7735.0
186.00	65888.0
187.00	17520.0
188.10	1839.0
189.00	4181.0
190.00	783.0
191.00	1587.0
192.00	5413.0
193.00	6762.0
194.10	1292.0
195.00	1058.0
196.00	15680.0
197.20	4133.0
198.00	487744.0
199.00	31048.0
199.90	2137.0
201.50	2407.0
203.00	3267.0
204.00	17440.0
205.00	30728.0
206.00	129792.0
207.00	17208.0
208.00	3918.0
209.00	1035.0
210.10	2250.0
211.00	4710.0
213.00	582.0
215.00	1029.0
216.00	2591.0
217.00	31776.0
218.00	3324.0
219.00	397.0
221.00	23592.0
223.00	6948.0
224.10	69072.0
225.00	18448.0
226.10	1864.0
227.00	28672.0
228.00	4050.0
229.00	5843.0
229.90	838.0
231.00	3181.0
232.00	435.0
232.90	587.0
234.00	1929.0
235.10	2195.0
236.00	1615.0
237.00	2472.0
238.00	439.0
239.00	1503.0
240.00	817.0
241.00	1837.0
242.00	3541.0
243.10	4055.0
244.10	60616.0
245.10	8290.0
246.00	11522.0
247.00	2217.0
248.10	481.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

249.00	2366.0
249.90	506.0
250.90	676.0
252.10	650.0
253.00	1319.0
254.20	2975.0
255.00	336064.0
256.00	47488.0
257.00	3514.0
258.00	18112.0
259.00	2705.0
260.00	514.0
261.00	655.0
263.10	301.0
264.00	654.0
265.00	6779.0
266.00	954.0
270.00	310.0
270.90	796.0
272.00	1179.0
273.00	10164.0
274.00	27104.0
275.00	174400.0
276.00	21888.0
277.00	12752.0
278.00	2486.0
279.00	553.0
281.00	549.0
282.00	581.0
283.00	1798.0
284.00	1423.0
285.00	2095.0
290.20	326.0
291.00	304.0
292.00	643.0
293.00	3338.0
294.00	1011.0
295.10	1201.0
296.00	44680.0
297.00	6477.0
297.90	365.0
300.90	730.0
302.10	1077.0
303.10	4750.0
304.10	1384.0
308.00	859.0
309.00	640.0
309.90	491.0
313.00	401.0
314.10	2320.0
315.00	5466.0
316.10	3008.0
317.00	628.0
321.00	1562.0
322.00	977.0
323.10	16159.0
324.10	2690.0
325.00	334.0
327.00	2921.0
328.00	1534.0
332.00	897.0
333.00	1308.0
334.00	10272.0
335.00	2830.0
336.00	393.0
341.00	1858.0
342.10	518.0
346.00	3667.0
347.10	690.0
351.10	460.0
352.10	4230.0
353.10	3535.0
354.00	5436.0
355.00	960.0
359.00	350.0
365.00	21776.0
366.00	2897.0
371.00	1329.0
372.00	8216.0
373.00	2223.0
374.10	389.0
383.00	2194.0
384.00	543.0
385.10	368.0
390.00	1225.0
391.00	798.0
392.10	678.0
400.90	568.0
402.00	3145.0
403.10	5497.0
404.00	1332.0
421.10	4049.0
422.00	3822.0
423.10	30344.0
424.10	7127.0
425.10	579.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

440.20	384.0
441.10	88120.0
442.10	607296.0
443.10	115280.0
444.10	9803.0
445.00	703.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96