

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040624\
 Data File : BM045000.D
 Acq On : 07 Apr 2024 10:00
 Operator : MA/JU
 Sample : PB160034BL
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :


Quant Time: Apr 08 05:54:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 08 05:51:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	57328	20.000	ng/ul	0.00
20) Naphthalene-d8	10.404	136	256165	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.280	164	174758	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.039	188	384166	20.000	ng/ul	0.00
79) Chrysene-d12	21.245	240	366480	20.000	ng/ul	0.00
88) Perylene-d12	23.468	264	452160	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	11037	7.189	ng/uL	0.00
4) Pyridine-d5	3.605	84	147478	35.326	ng/ul	0.00
7) Phenol-d5	6.816	99	176336	36.283	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.987	67	112210	38.753	ng/ul	0.00
11) 2-Chlorophenol-d4	7.163	132	148183	38.918	ng/ul	0.00
15) 4-Methylphenol-d8	8.339	113	143568	37.740	ng/ul	0.00
21) Nitrobenzene-d5	8.792	128	72355	36.772	ng/ul	0.00
24) 2-Nitrophenol-d4	9.504	143	80417	38.720	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.039	165	131328	34.489	ng/ul	0.00
31) 4-Chloroaniline-d4	10.563	131	246392	40.642	ng/ul	0.00
46) Dimethylphthalate-d6	13.704	166	468638	38.514	ng/ul	0.00
49) Acenaphthylene-d8	13.969	160	524333	36.476	ng/ul	0.00
54) 4-Nitrophenol-d4	14.521	143	74136	29.194	ng/ul	0.00
60) Fluorene-d10	15.280	176	413378	38.107	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.421	200	61835	27.147	ng/ul	0.00
73) Anthracene-d10	17.139	188	664682	38.461	ng/ul	0.00
81) Pyrene-d10	19.445	212	786191	43.567	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.327	264	874694	39.655	ng/ul	0.00


Target Compounds					Qvalue
17) N-Nitroso-di-n-propyla...	8.339	70	3731	1.278 ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed



N-Nitroso-di-n-propylamine
 Concen: 1.278 ng/ul
 RT: 8.339 min Scan# 893
 Delta R.T. -0.100 min
 Lab File: BM045000.D
 Acq: 07 Apr 2024 10:00

Instrument :
 BNA_M
 ClientSampleId :



Tgt Ion: 70	Resp: 3731
Ion Ratio	Lower Upper
70	100
42	330.4 43.2 64.8#
101	0.0 6.4 9.6#
130	0.0 16.2 24.2#

