

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040724\
 Data File : BM045036.D
 Acq On : 08 Apr 2024 10:18
 Operator : MA/JU
 Sample : PB160051BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK051

Quant Time: Apr 08 11:13:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 08 05:51:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	89295	20.000	ng/u1	0.00
20) Naphthalene-d8	10.398	136	388052	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.274	164	248778	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.033	188	529619	20.000	ng/u1	0.00
79) Chrysene-d12	21.244	240	475903	20.000	ng/u1	0.00
88) Perylene-d12	23.462	264	557528	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.204	96	12595	5.267	ng/uL	0.00
4) Pyridine-d5	3.604	84	186780	28.723	ng/u1	0.00
7) Phenol-d5	6.804	99	233790	30.884	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	6.981	67	138996	30.819	ng/u1	0.00
11) 2-Chlorophenol-d4	7.157	132	191364	32.267	ng/u1	0.00
15) 4-Methylphenol-d8	8.334	113	192488	32.485	ng/u1	0.00
21) Nitrobenzene-d5	8.786	128	95282	31.966	ng/u1	0.00
24) 2-Nitrophenol-d4	9.498	143	105462	33.521	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.028	165	182046	31.560	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	265917	28.955	ng/u1	0.00
46) Dimethylphthalate-d6	13.704	166	555021	32.042	ng/u1	0.00
49) Acenaphthylene-d8	13.968	160	671049	32.793	ng/u1	0.00
54) 4-Nitrophenol-d4	14.515	143	86820	24.016	ng/u1	0.00
60) Fluorene-d10	15.274	176	499817	32.366	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.421	200	65654	20.907	ng/u1	0.00
73) Anthracene-d10	17.133	188	796923	33.449	ng/u1	0.00
81) Pyrene-d10	19.439	212	917391	39.149	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.321	264	934139	34.346	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040724\
Data File : BM045036.D
Acq On : 08 Apr 2024 10:18
Operator : MA/JU
Sample : PB160051BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK051

Quant Time: Apr 08 11:13:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Apr 08 05:51:38 2024
Response via : Initial Calibration

