



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040724\
Data File : BM045004.D
Acq On : 07 Apr 2024 13:07
Operator : MA/JU
Sample : DFTPP018
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP018

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
Title : SVOA CALIBRATION
Last Update : Mon Apr 08 05:51:38 2024



AutoFind: Scans 2447, 2448, 2449; Background Corrected with Scan 2441

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.1	45651	PASS
68	69	0.00	2	1.6	809	PASS
69	198	0.00	100	48.9	49475	PASS
70	69	0.00	2	0.4	188	PASS
127	198	10	80	57.1	57813	PASS
197	198	0.00	2	0.6	591	PASS
198	198	100	100	100.0	101219	PASS
199	198	5	9	6.6	6721	PASS
275	198	10	60	28.4	28739	PASS
365	198	1	100	4.9	4951	PASS
441	198	0.01	100	14.4	14566	PASS
442	198	50	100	85.3	86320	PASS
443	442	15	24	19.4	16769	PASS

m/z	Abundance
40.00	765.0
40.90	190.0
44.00	1223.0
44.90	190.0
48.80	286.0
73.00	759.0
76.90	217.0
83.90	165.0
132.90	279.0
147.00	191.0
191.00	256.0
192.80	168.0
207.00	1718.0
208.00	328.0
209.00	200.0
281.00	372.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

m/z	Abundance
37.00	301.0
38.00	820.0
39.00	4120.0
39.90	836.0
41.00	278.0
44.00	1027.0
45.00	161.0
49.00	582.0
50.00	15593.0
51.00	54520.0
52.00	2982.0
52.90	207.0
54.90	545.0
55.90	1705.0
57.00	3800.0
57.80	158.0
61.10	703.0
62.00	1150.0
63.00	2178.0
64.00	384.0
65.00	1053.0
67.90	1071.0
69.00	58888.0
70.00	249.0
73.00	1296.0
74.00	6710.0
75.00	10405.0
76.00	3871.0
77.00	76464.0
78.00	5841.0
79.00	3907.0
80.00	3308.0
81.00	5207.0
81.90	1186.0
83.00	835.0
83.90	331.0
85.00	478.0
86.00	1623.0
87.00	659.0
87.90	281.0
88.90	230.0
90.90	1023.0
92.00	1387.0
93.00	8901.0
94.00	610.0
95.00	218.0
96.00	529.0
98.00	6861.0
99.00	5185.0
99.90	456.0
101.00	2855.0
102.90	536.0
103.90	1690.0
105.00	1669.0
106.00	525.0
107.00	20904.0
108.00	2921.0
109.00	476.0
110.00	30768.0
111.00	4943.0
112.00	668.0
112.90	197.0
116.00	844.0
117.00	22664.0
118.00	1697.0
118.90	202.0
120.00	248.0
122.00	1245.0
122.90	2016.0
123.90	999.0
124.90	843.0
126.10	238.0
127.00	65312.0
128.00	4834.0
128.90	26800.0
129.90	2112.0
130.90	463.0
132.10	241.0
133.10	328.0
133.90	856.0
135.00	2011.0
135.80	782.0
136.90	1538.0
137.80	160.0
139.80	269.0
141.00	3329.0
141.90	1095.0
142.90	824.0
143.80	185.0
144.90	224.0
145.90	467.0
146.90	2312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

147.90	5777.0
148.90	1386.0
150.00	237.0
150.80	420.0
151.80	200.0
152.90	1446.0
153.80	752.0
155.00	1957.0
156.00	2849.0
157.00	553.0
157.80	458.0
158.80	493.0
159.90	974.0
161.00	1547.0
162.00	353.0
163.00	277.0
163.90	297.0
165.00	1169.0
165.90	1013.0
166.90	8124.0
168.00	5265.0
169.00	864.0
169.90	252.0
170.80	380.0
171.90	597.0
172.90	963.0
174.00	1463.0
175.00	2683.0
175.90	616.0
176.90	1425.0
177.80	462.0
178.90	5436.0
180.00	3492.0
181.00	1790.0
181.90	331.0
183.90	476.0
185.00	2641.0
186.00	16536.0
186.90	5493.0
187.80	330.0
188.90	939.0
190.90	740.0
191.90	2124.0
193.00	1974.0
194.00	436.0
195.90	3334.0
197.00	530.0
197.90	104432.0
198.90	7182.0
199.90	685.0
201.40	996.0
202.90	1070.0
204.00	4692.0
205.00	7815.0
206.00	31200.0
207.00	6180.0
207.90	1168.0
208.90	471.0
209.90	553.0
211.00	1374.0
211.80	370.0
214.90	318.0
215.90	540.0
216.90	9541.0
217.90	1254.0
219.90	246.0
221.00	4393.0
222.00	1127.0
222.90	2332.0
224.00	17728.0
225.00	4809.0
225.90	513.0
226.90	7796.0
227.90	1290.0
228.90	1606.0
230.00	199.0
230.90	582.0
232.80	176.0
233.90	596.0
234.90	596.0
236.00	606.0
237.00	970.0
239.00	334.0
239.80	215.0
240.90	597.0
241.90	1113.0
242.90	1338.0
244.00	14670.0
245.00	2185.0
245.90	2786.0
246.90	440.0
248.90	481.0
251.00	192.0
251.80	174.0
252.90	281.0
254.00	451.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

255.00	70472.0
255.90	10023.0
256.90	1112.0
258.00	4906.0
259.00	763.0
264.90	1905.0
265.80	227.0
273.00	2054.0
273.90	5246.0
275.00	27640.0
276.00	3870.0
276.90	2649.0
277.80	571.0
280.90	415.0
281.80	224.0
282.90	420.0
283.90	287.0
284.90	400.0
292.90	576.0
293.90	215.0
296.00	9427.0
296.90	1371.0
301.00	160.0
303.00	1188.0
304.00	250.0
310.10	172.0
313.90	425.0
314.90	1048.0
316.00	584.0
320.90	534.0
322.00	219.0
323.00	2780.0
323.90	466.0
327.00	926.0
327.70	179.0
331.90	194.0
333.00	434.0
333.90	1698.0
334.90	414.0
340.90	396.0
345.90	859.0
346.80	193.0
352.00	874.0
352.90	629.0
354.00	992.0
355.10	211.0
365.00	4527.0
365.90	622.0
370.90	202.0
372.00	1672.0
373.00	562.0
383.00	335.0
389.90	260.0
401.90	660.0
403.00	686.0
404.00	248.0
405.00	166.0
421.00	682.0
421.80	450.0
423.00	4077.0
424.00	1308.0
424.90	325.0
441.00	11186.0
442.00	72144.0
443.00	14020.0
444.00	1264.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

m/z	Abundance
36.90	301.0
38.00	772.0
39.00	3954.0
39.90	874.0
41.00	205.0
43.90	1183.0
45.00	231.0
49.10	550.0
50.00	16496.0
51.00	54504.0
52.00	3225.0
52.90	201.0
55.00	418.0
56.00	1662.0
57.00	3395.0
57.80	196.0
60.90	715.0
62.00	995.0
63.00	2632.0
64.00	331.0
65.00	1264.0
68.00	922.0
69.00	59632.0
69.90	315.0
73.00	1241.0
74.00	7032.0
75.00	10269.0
76.00	4016.0
77.00	79384.0
78.00	5540.0
79.00	4407.0
80.00	3752.0
81.00	4689.0
81.90	1499.0
83.00	978.0
83.80	248.0
84.90	550.0
86.00	1626.0
86.90	882.0
87.90	209.0
88.90	205.0
91.00	900.0
92.00	861.0
93.00	8233.0
94.00	526.0
95.90	609.0
96.80	154.0
98.00	7196.0
99.00	5217.0
100.00	371.0
101.00	3185.0
101.90	200.0
103.00	655.0
104.00	1764.0
104.90	1444.0
106.00	504.0
107.00	22480.0
108.00	3108.0
108.90	451.0
110.00	32336.0
110.90	5755.0
112.00	523.0
112.80	356.0
114.80	202.0
115.90	1064.0
117.00	24232.0
117.90	1801.0
118.90	187.0
119.90	262.0
120.90	150.0
121.90	1131.0
122.90	2322.0
123.90	1033.0
124.80	778.0
126.10	154.0
127.00	69008.0
128.00	5160.0
129.00	26864.0
129.90	2440.0
131.10	474.0
131.90	264.0
132.80	373.0
133.80	692.0
134.90	2230.0
135.90	858.0
137.00	1175.0
137.80	390.0
140.00	334.0
140.90	3429.0
142.00	1332.0
143.00	827.0
144.00	232.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

144.90	241.0
145.80	544.0
146.90	2364.0
148.00	6134.0
148.90	1287.0
150.00	483.0
151.00	538.0
151.70	367.0
152.90	1297.0
154.00	791.0
155.00	2200.0
155.90	3099.0
157.00	580.0
157.90	816.0
158.90	673.0
159.90	999.0
160.90	1464.0
161.80	352.0
162.80	297.0
163.80	194.0
164.90	1330.0
166.00	1066.0
167.00	9841.0
167.90	5840.0
169.00	768.0
169.80	300.0
170.90	395.0
172.00	839.0
172.90	973.0
173.90	1591.0
175.00	3236.0
175.90	650.0
176.90	1332.0
177.90	405.0
178.90	6227.0
180.00	4129.0
181.00	1864.0
182.10	190.0
183.90	431.0
184.90	3066.0
186.00	19728.0
187.00	5114.0
187.90	563.0
188.90	1231.0
190.90	870.0
191.90	2166.0
192.90	2881.0
193.90	495.0
195.00	393.0
195.90	4064.0
197.00	780.0
197.90	124272.0
198.90	8306.0
199.90	841.0
201.40	455.0
203.00	927.0
204.00	6488.0
205.00	9317.0
206.00	37128.0
207.00	6899.0
207.90	1544.0
208.90	720.0
209.90	629.0
211.00	1629.0
212.00	225.0
214.90	421.0
216.00	802.0
216.90	10132.0
217.90	1455.0
219.00	194.0
221.00	4706.0
222.00	1308.0
222.90	1824.0
224.00	21760.0
225.00	5156.0
225.90	620.0
227.00	9715.0
227.90	1429.0
228.90	2120.0
229.90	210.0
230.90	782.0
231.90	184.0
232.90	183.0
233.90	739.0
234.90	757.0
235.90	534.0
236.90	1139.0
238.80	328.0
240.00	303.0
240.80	543.0
241.90	1278.0
242.90	1154.0
244.00	17080.0
245.00	2561.0
245.90	3229.0
247.00	638.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

248.90	806.0
249.90	165.0
250.90	265.0
251.90	198.0
252.90	490.0
253.90	332.0
255.00	88192.0
256.00	13146.0
257.00	1027.0
257.90	5185.0
258.90	729.0
263.80	172.0
264.90	2356.0
265.80	331.0
271.00	222.0
271.90	210.0
273.00	3163.0
273.90	6771.0
275.00	34824.0
275.90	4406.0
276.90	3351.0
277.90	799.0
280.90	504.0
283.10	466.0
283.80	386.0
284.90	762.0
285.90	153.0
289.00	272.0
292.90	728.0
294.00	207.0
295.00	239.0
295.90	12635.0
296.90	2207.0
300.90	192.0
301.90	196.0
303.00	1287.0
304.00	401.0
314.00	547.0
314.90	1434.0
315.90	658.0
320.90	463.0
321.90	194.0
323.00	3600.0
324.10	525.0
327.00	1200.0
328.00	399.0
331.90	210.0
332.90	386.0
334.00	2574.0
335.00	651.0
340.90	497.0
346.00	610.0
352.00	1236.0
352.90	854.0
353.90	815.0
354.90	195.0
364.90	6068.0
365.90	889.0
371.00	323.0
372.00	1808.0
373.00	497.0
373.90	158.0
383.00	483.0
384.00	161.0
390.00	232.0
390.90	267.0
401.90	757.0
403.00	960.0
404.00	354.0
421.00	828.0
422.00	761.0
423.00	5246.0
424.00	1475.0
441.00	17128.0
442.00	100744.0
443.00	19824.0
444.00	2267.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

m/z	Abundance
36.90	159.0
38.00	276.0
39.00	2068.0
39.90	708.0
42.90	156.0
44.00	1052.0
45.00	158.0
48.90	326.0
50.00	8538.0
51.00	27928.0
52.00	1387.0
55.00	171.0
56.00	744.0
57.00	1626.0
57.90	173.0
61.00	413.0
61.90	520.0
62.90	1204.0
64.00	255.0
64.90	476.0
68.00	434.0
69.00	29904.0
73.00	1148.0
74.00	4097.0
75.00	5615.0
76.00	2342.0
77.00	42600.0
78.00	2855.0
79.00	2253.0
79.90	1899.0
81.00	2199.0
81.90	615.0
83.00	591.0
83.90	237.0
85.00	390.0
86.00	693.0
86.80	203.0
87.90	179.0
89.00	163.0
90.90	753.0
91.90	621.0
93.00	4539.0
93.90	350.0
96.00	382.0
97.20	151.0
98.00	3491.0
99.00	3271.0
100.00	353.0
100.90	1305.0
103.00	537.0
103.90	849.0
105.00	1018.0
105.90	294.0
107.00	12218.0
107.90	1787.0
108.90	322.0
110.00	18600.0
110.90	2830.0
111.90	392.0
112.90	196.0
116.00	495.0
117.00	13653.0
117.90	840.0
118.90	231.0
119.90	205.0
121.90	821.0
122.90	1098.0
124.00	429.0
124.90	576.0
127.00	39120.0
128.00	3278.0
129.00	15576.0
129.90	1334.0
131.00	255.0
132.90	238.0
134.00	421.0
134.90	1366.0
135.90	579.0
136.90	712.0
137.80	226.0
140.00	174.0
140.90	2256.0
142.00	408.0
142.90	428.0
143.90	230.0
144.90	195.0
145.90	465.0
146.90	1330.0
147.90	2953.0
148.90	663.0
150.10	161.0
150.80	222.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

151.30	207.0
152.00	239.0
152.80	557.0
153.90	543.0
154.90	1324.0
155.90	1408.0
157.20	312.0
157.90	272.0
158.90	419.0
159.90	702.0
161.00	1097.0
161.80	330.0
163.90	168.0
164.80	841.0
166.00	622.0
167.00	5350.0
168.00	4117.0
168.90	447.0
172.00	387.0
173.00	449.0
173.80	860.0
175.00	1899.0
176.00	364.0
177.00	605.0
177.80	325.0
178.90	3656.0
179.90	2214.0
180.90	795.0
182.00	198.0
183.90	429.0
184.90	1694.0
186.00	12108.0
186.90	3127.0
187.90	327.0
189.00	678.0
190.90	626.0
191.90	1324.0
192.90	1859.0
194.00	342.0
195.90	2740.0
196.90	463.0
197.90	74952.0
198.90	4674.0
199.90	358.0
200.90	227.0
201.50	312.0
203.00	757.0
204.00	3390.0
205.00	5952.0
206.00	22808.0
207.00	4942.0
207.90	968.0
208.80	319.0
209.80	285.0
211.00	929.0
211.60	198.0
214.90	240.0
215.90	635.0
216.90	7322.0
217.90	883.0
220.90	3846.0
221.70	1070.0
222.90	1451.0
224.00	13292.0
225.00	3486.0
226.00	370.0
226.90	5515.0
227.90	631.0
228.90	1385.0
229.90	222.0
230.80	523.0
233.90	291.0
235.00	370.0
235.80	301.0
237.00	633.0
238.90	241.0
240.00	264.0
240.80	365.0
241.90	1005.0
242.90	788.0
244.00	11286.0
245.00	1415.0
246.00	2024.0
246.90	806.0
248.90	335.0
251.90	177.0
252.90	237.0
253.90	363.0
255.00	55984.0
256.00	8465.0
256.90	501.0
257.90	3290.0
258.90	543.0
264.90	1826.0
266.00	160.0
266.90	172.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018

272.90	2174.0
273.90	4669.0
275.00	23752.0
276.00	3021.0
276.90	2172.0
278.00	390.0
280.90	437.0
281.80	164.0
283.00	389.0
283.90	247.0
285.00	525.0
293.00	687.0
294.90	168.0
295.90	8014.0
296.90	1130.0
301.10	204.0
301.80	170.0
302.90	886.0
304.00	282.0
313.90	390.0
314.90	892.0
315.90	442.0
320.80	286.0
323.00	2458.0
323.90	499.0
326.90	689.0
327.80	228.0
331.90	231.0
333.00	201.0
333.90	1540.0
334.90	617.0
341.00	325.0
345.90	599.0
352.00	845.0
352.90	507.0
353.90	914.0
354.90	274.0
365.00	4259.0
365.90	832.0
371.10	221.0
372.00	2015.0
373.00	223.0
382.90	515.0
390.90	189.0
401.90	714.0
402.90	618.0
403.90	235.0
421.00	745.0
421.90	548.0
423.00	4032.0
424.00	1062.0
441.00	15383.0
442.00	86072.0
443.00	16464.0
443.90	1472.0

Instrument :
BNA_M
ClientSampleId :
DFTPP018