


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040819\
 Data File : BM019786.D
 Acq On : 08 Apr 2019 09:15
 Operator : JU/SJ
 Sample : DFTPP99
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP99

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Apr 10 03:34:56 2019



AutoFind: Scans 2451, 2452, 2453; Background Corrected with Scan 2445

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.6	67709	PASS
68	69	0.00	2	1.8	1900	PASS
69	198	0.00	100	47.1	107585	PASS
70	69	0.00	2	0.8	815	PASS
127	198	10	80	49.5	113058	PASS
197	198	0.00	2	0.7	1506	PASS
198	198	100	100	100.0	228394	PASS
199	198	5	9	6.7	15256	PASS
275	198	10	60	31.5	71952	PASS
365	198	1	100	5.2	11941	PASS
441	198	0.01	100	14.4	32805	PASS
442	198	50	100	94.7	216320	PASS
443	442	15	24	19.3	41752	PASS

m/z	Abundance
39.90	484.0
41.00	519.0
43.10	394.0
55.00	621.0
57.00	396.0
67.00	346.0
69.00	460.0
73.00	1267.0
76.90	427.0
79.10	348.0
80.90	395.0
83.00	355.0
91.00	381.0
105.00	358.0
115.10	316.0
147.10	498.0
165.00	337.0
193.10	321.0
207.00	1831.0
208.00	307.0
281.10	564.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

m/z	Abundance
37.00	458.0
38.00	1079.0
39.00	7194.0
40.00	709.0
41.00	695.0
43.00	643.0
49.00	877.0
50.00	21232.0
51.00	81840.0
52.00	4629.0
52.90	315.0
55.00	1126.0
56.00	2858.0
57.00	7553.0
57.90	366.0
61.00	1344.0
62.00	1691.0
63.00	5027.0
64.10	658.0
65.00	2283.0
67.00	477.0
68.10	2432.0
69.00	129608.0
70.00	905.0
71.00	539.0
73.10	2058.0
74.00	12822.0
75.00	20104.0
76.10	6441.0
77.00	141120.0
78.10	9541.0
79.00	10049.0
80.00	7744.0
81.00	10535.0
82.00	2455.0
83.00	2341.0
83.90	405.0
85.00	2027.0
86.00	2774.0
87.00	1357.0
87.90	579.0
90.90	3053.0
92.00	2733.0
93.00	16760.0
94.00	1097.0
95.00	545.0
96.10	511.0
97.10	536.0
98.00	11743.0
99.00	9596.0
100.00	836.0
101.00	5135.0
102.90	1998.0
104.00	3444.0
105.00	3197.0
105.90	1131.0
107.00	39040.0
108.00	6577.0
109.10	1236.0
110.00	67776.0
111.00	11265.0
112.00	1502.0
113.00	473.0
115.00	357.0
116.00	2356.0
117.00	33296.0
118.00	2439.0
119.00	513.0
119.90	510.0
120.90	414.0
121.90	2825.0
123.00	4281.0
123.90	1642.0
125.00	1699.0
126.10	921.0
127.00	126040.0
128.00	10492.0
129.00	50768.0
130.00	5133.0
130.90	917.0
132.00	434.0
133.00	460.0
133.90	1543.0
135.00	4433.0
136.00	1973.0
137.00	2539.0
138.00	470.0
140.10	505.0
141.00	7371.0
141.90	2138.0
143.00	1671.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

144.00	468.0
145.00	489.0
145.90	1065.0
147.00	4164.0
148.00	8684.0
149.00	2138.0
150.00	382.0
151.00	909.0
151.90	1056.0
153.00	2203.0
153.90	1520.0
155.10	3854.0
156.00	5524.0
157.10	809.0
157.90	1311.0
158.90	1033.0
159.90	1996.0
161.00	3690.0
162.00	1000.0
163.00	317.0
163.90	428.0
165.00	2724.0
165.90	2351.0
167.00	14653.0
167.90	7630.0
169.00	1096.0
170.10	575.0
171.10	727.0
171.90	1180.0
173.10	1411.0
174.00	2695.0
175.00	5691.0
176.00	1473.0
176.90	2283.0
178.00	1108.0
178.90	10694.0
180.00	7274.0
181.00	3009.0
181.80	522.0
182.20	408.0
183.90	995.0
185.00	4893.0
186.00	37232.0
187.00	11771.0
188.00	1165.0
189.00	2399.0
190.00	368.0
191.00	1631.0
191.90	3437.0
193.00	3563.0
194.00	685.0
195.00	505.0
196.00	7348.0
197.10	1670.0
197.90	248384.0
198.90	17384.0
200.00	1494.0
201.40	835.0
201.70	837.0
203.00	2105.0
204.00	9750.0
205.00	17040.0
206.00	68712.0
207.00	10688.0
208.00	2848.0
208.90	1231.0
210.20	1302.0
211.00	2580.0
215.00	622.0
216.00	1565.0
217.00	17928.0
217.90	2434.0
221.00	11043.0
221.80	2817.0
223.00	3984.0
224.00	38040.0
225.00	9405.0
226.00	1190.0
227.00	16760.0
227.90	2489.0
229.00	3852.0
229.90	542.0
231.00	1340.0
233.90	1502.0
235.00	1167.0
236.00	740.0
237.00	1866.0
239.00	661.0
239.90	581.0
240.90	1118.0
241.90	1901.0
243.00	2030.0
244.00	30568.0
245.00	4671.0
246.00	6137.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

247.00	1278.0
248.90	1066.0
253.00	658.0
254.10	1212.0
255.00	160576.0
256.00	22792.0
257.00	2019.0
258.00	9039.0
258.90	1757.0
260.80	368.0
264.10	410.0
264.90	3600.0
266.00	634.0
271.90	451.0
273.00	4737.0
274.00	12290.0
275.00	71424.0
276.00	10184.0
277.00	6962.0
278.00	1193.0
278.90	304.0
281.00	546.0
283.00	1070.0
283.90	543.0
285.00	973.0
285.90	366.0
291.90	327.0
293.00	1301.0
293.90	347.0
295.10	608.0
296.00	23088.0
297.00	2998.0
300.90	366.0
302.00	371.0
303.00	2613.0
303.90	789.0
309.80	317.0
314.00	1034.0
314.90	2202.0
316.00	1396.0
317.10	312.0
321.00	690.0
322.10	331.0
323.00	6285.0
324.00	1583.0
326.90	1574.0
328.00	564.0
332.00	530.0
332.90	768.0
334.00	4367.0
335.00	1142.0
340.90	688.0
346.00	1478.0
352.00	2529.0
353.00	1473.0
354.00	2049.0
355.00	675.0
365.00	11447.0
365.90	1404.0
371.00	594.0
372.00	3827.0
373.00	1124.0
383.00	1048.0
383.90	300.0
390.00	393.0
401.00	410.0
402.00	1376.0
403.00	2080.0
404.00	580.0
404.90	399.0
421.00	1792.0
421.90	1857.0
423.00	10824.0
424.10	3516.0
425.00	442.0
441.10	28640.0
442.10	184192.0
443.10	36408.0
444.00	3521.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

m/z	Abundance
36.90	487.0
38.00	893.0
39.00	6291.0
39.90	758.0
41.00	842.0
42.90	701.0
44.00	369.0
49.00	876.0
50.00	20520.0
51.00	79232.0
52.10	3929.0
55.00	1052.0
56.00	3060.0
57.00	6930.0
57.90	438.0
61.00	1423.0
62.00	1348.0
63.00	4724.0
64.00	675.0
65.00	2358.0
67.00	524.0
68.00	2116.0
69.00	125008.0
70.00	991.0
71.00	509.0
73.10	2547.0
74.00	13722.0
75.00	20672.0
76.10	6753.0
77.00	135424.0
78.00	9061.0
79.00	10156.0
80.00	7121.0
81.00	10646.0
82.00	2359.0
83.00	2382.0
84.00	499.0
85.00	1961.0
86.00	2630.0
87.00	1379.0
88.10	380.0
91.00	2864.0
92.00	2818.0
93.00	16060.0
94.00	1211.0
95.10	473.0
96.00	855.0
97.10	427.0
98.00	11763.0
99.00	9291.0
100.00	713.0
101.00	4548.0
101.90	408.0
103.00	2014.0
103.90	3161.0
105.00	3426.0
106.00	1206.0
107.00	39160.0
108.00	6320.0
109.10	1377.0
110.00	68704.0
111.00	11396.0
111.90	1413.0
112.90	548.0
115.00	408.0
116.00	2586.0
117.00	35904.0
118.00	2743.0
119.00	595.0
120.00	656.0
120.80	331.0
122.00	2597.0
123.00	4171.0
124.00	1757.0
125.00	1570.0
127.00	136512.0
128.00	10289.0
129.00	54856.0
129.90	4592.0
130.90	970.0
132.10	488.0
132.90	468.0
133.90	1611.0
135.00	4228.0
136.00	1769.0
137.00	2599.0
138.00	568.0
139.10	365.0
140.00	834.0
141.00	7191.0
142.00	2613.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

143.00	1715.0
143.90	492.0
144.90	434.0
146.00	1403.0
147.00	4095.0
148.00	9120.0
148.90	1858.0
149.90	428.0
151.10	926.0
151.80	663.0
152.90	2397.0
154.00	1681.0
155.00	3904.0
156.00	6014.0
157.00	1193.0
158.00	1134.0
159.00	1402.0
159.90	2080.0
161.00	3480.0
161.90	997.0
163.10	354.0
164.00	355.0
164.90	2948.0
165.90	2122.0
167.00	15740.0
168.00	7793.0
169.00	1478.0
170.00	692.0
170.90	719.0
171.90	1300.0
173.00	1702.0
174.00	2936.0
175.00	5793.0
176.00	1787.0
177.00	3269.0
177.90	1119.0
179.00	11279.0
180.00	8398.0
181.00	3655.0
182.00	723.0
182.90	382.0
184.00	1035.0
185.00	5835.0
186.00	39696.0
187.00	12021.0
188.10	1225.0
189.00	2946.0
190.00	528.0
191.00	1474.0
192.00	3534.0
193.00	4249.0
193.90	1019.0
194.90	564.0
196.00	8988.0
197.10	1768.0
197.90	266816.0
198.90	17344.0
200.00	1452.0
201.60	1248.0
203.00	2470.0
204.00	11404.0
205.00	18424.0
206.00	77392.0
207.00	11875.0
208.00	2931.0
208.90	1233.0
210.00	1214.0
211.00	3051.0
211.70	683.0
215.00	818.0
216.00	1618.0
217.00	22072.0
218.00	2794.0
221.00	13129.0
221.80	3229.0
223.00	5021.0
224.00	43152.0
225.00	10617.0
226.00	1249.0
227.00	18600.0
228.00	2803.0
229.00	3915.0
229.90	731.0
231.00	1511.0
232.90	356.0
233.90	1270.0
234.90	1333.0
235.90	1084.0
237.00	1997.0
238.90	746.0
240.00	666.0
240.90	1114.0
242.00	2362.0
243.00	2403.0
244.00	33080.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

245.10	3878.0
246.00	7411.0
247.00	1331.0
249.00	1673.0
250.90	346.0
252.90	764.0
254.10	1674.0
255.00	188544.0
256.00	26520.0
256.90	2066.0
258.00	12161.0
259.00	1637.0
259.90	309.0
265.00	4769.0
265.90	750.0
270.00	306.0
271.00	467.0
271.90	367.0
273.00	5572.0
274.00	14624.0
275.00	86384.0
276.00	11254.0
277.00	7879.0
277.90	1371.0
281.00	735.0
282.10	360.0
283.00	996.0
283.90	649.0
285.00	1356.0
292.00	424.0
292.90	1884.0
294.00	344.0
294.80	430.0
296.00	28240.0
297.00	3786.0
301.00	330.0
302.00	419.0
303.00	3352.0
304.10	733.0
307.90	402.0
310.10	377.0
312.90	383.0
314.00	1225.0
314.90	3079.0
316.00	2043.0
321.00	952.0
322.10	589.0
323.00	8223.0
324.00	1586.0
326.90	1868.0
327.90	926.0
332.00	656.0
332.90	668.0
334.00	6506.0
335.00	1692.0
341.00	1088.0
342.10	340.0
346.00	1866.0
346.90	415.0
352.00	2492.0
353.00	1829.0
354.00	2945.0
355.00	503.0
358.90	349.0
365.00	13868.0
365.90	1845.0
371.00	667.0
372.00	4856.0
373.00	1059.0
383.00	1398.0
384.00	408.0
390.00	573.0
391.00	471.0
402.00	1901.0
403.00	2915.0
404.10	948.0
421.00	2182.0
422.00	2244.0
423.00	15674.0
424.00	3965.0
425.00	552.0
441.10	38240.0
442.10	258880.0
443.00	49504.0
444.10	4532.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

m/z	Abundance
38.00	555.0
39.00	3578.0
40.00	503.0
41.00	559.0
43.10	488.0
44.00	302.0
49.00	539.0
50.00	11378.0
51.00	42056.0
52.10	1949.0
55.00	920.0
56.00	1673.0
57.00	4074.0
60.90	779.0
62.00	829.0
63.00	2509.0
64.00	387.0
65.00	1099.0
66.90	323.0
68.00	1154.0
69.00	69520.0
70.00	550.0
71.10	506.0
73.10	3597.0
74.00	7589.0
75.00	11310.0
76.10	3650.0
77.00	74280.0
78.00	5431.0
79.00	5151.0
80.00	4227.0
81.00	6040.0
82.00	1591.0
83.00	1856.0
84.10	339.0
85.00	1300.0
86.00	1833.0
87.00	924.0
91.00	1743.0
92.00	1481.0
93.00	9081.0
93.90	814.0
95.00	359.0
96.00	608.0
97.10	498.0
98.00	6835.0
99.00	5005.0
100.00	480.0
100.90	2558.0
103.00	1131.0
104.00	2239.0
105.00	2314.0
105.90	699.0
107.00	22096.0
108.00	3434.0
109.10	993.0
110.00	38368.0
111.00	5558.0
112.10	697.0
115.00	354.0
115.90	1344.0
117.00	21216.0
118.00	1495.0
118.90	464.0
119.90	385.0
120.90	437.0
121.90	1410.0
123.00	2385.0
124.00	1082.0
124.90	912.0
127.00	76624.0
128.00	6593.0
129.00	31168.0
129.90	2594.0
130.80	425.0
131.10	441.0
132.00	346.0
132.90	533.0
134.00	1141.0
135.00	2786.0
136.10	850.0
137.00	1484.0
138.10	345.0
139.90	358.0
141.00	4011.0
142.00	1548.0
142.90	1237.0
144.90	470.0
146.00	864.0
147.00	3041.0
148.00	5783.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

149.00	1402.0
151.10	721.0
151.90	372.0
153.00	1556.0
154.00	990.0
155.00	2476.0
156.00	3883.0
156.90	786.0
157.90	809.0
158.90	810.0
160.00	1250.0
161.00	2122.0
162.00	547.0
163.90	323.0
164.90	2176.0
166.00	1778.0
167.00	9939.0
168.00	4803.0
168.90	649.0
169.90	318.0
170.20	318.0
171.00	420.0
171.90	930.0
173.10	1188.0
174.00	1751.0
175.00	3807.0
176.00	1104.0
177.00	1970.0
178.00	756.0
179.00	6812.0
180.00	5206.0
181.00	2158.0
183.90	647.0
184.90	3303.0
186.00	25288.0
187.00	7796.0
188.00	875.0
189.00	1629.0
190.00	513.0
190.90	1053.0
192.00	2871.0
193.00	2763.0
194.00	558.0
194.90	471.0
195.90	4820.0
197.10	1081.0
197.90	169984.0
199.00	11042.0
199.90	929.0
201.40	968.0
202.90	1205.0
204.00	7067.0
205.00	11858.0
206.00	49464.0
207.00	7578.0
208.00	1998.0
209.10	1007.0
209.90	808.0
210.30	506.0
210.90	1881.0
211.80	511.0
214.90	433.0
216.00	1146.0
216.90	13519.0
218.00	1500.0
221.00	7653.0
221.80	1932.0
223.00	2891.0
224.00	26352.0
225.00	6761.0
226.10	680.0
227.00	11277.0
227.90	1615.0
228.90	2069.0
229.90	312.0
231.00	1056.0
233.90	648.0
234.90	733.0
236.00	545.0
237.00	965.0
238.90	704.0
240.00	463.0
240.90	853.0
242.00	1517.0
243.10	1549.0
244.00	22048.0
245.00	2713.0
246.00	5096.0
247.00	806.0
248.80	802.0
252.90	563.0
254.00	966.0
255.00	122096.0
256.00	16752.0
257.00	1170.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99

258.00	7242.0
259.00	1003.0
264.90	3200.0
266.00	435.0
272.90	3797.0
274.00	9279.0
275.00	58048.0
276.00	7642.0
277.00	5513.0
278.00	885.0
281.00	795.0
281.90	373.0
283.00	641.0
284.00	515.0
285.00	873.0
292.90	1285.0
294.00	390.0
295.10	480.0
296.00	18320.0
297.00	2449.0
303.00	2281.0
304.00	492.0
310.00	408.0
314.00	1046.0
314.90	2172.0
316.00	1179.0
321.00	703.0
321.90	394.0
323.00	5778.0
324.10	828.0
326.90	1277.0
328.00	601.0
332.00	515.0
333.00	739.0
334.00	4539.0
335.00	1358.0
341.00	855.0
346.00	1423.0
352.00	2006.0
353.00	1272.0
354.00	2443.0
355.10	490.0
365.00	10509.0
366.00	1422.0
371.00	707.0
372.00	3515.0
373.10	970.0
382.90	1131.0
383.90	362.0
390.00	628.0
401.00	368.0
402.00	1621.0
403.00	2014.0
404.00	992.0
421.00	1710.0
422.00	1883.0
423.10	12155.0
424.00	3360.0
441.00	31536.0
442.00	205888.0
443.10	39344.0
444.00	3145.0

Instrument :
BNA_M
ClientSampleId :
DFTPP99