

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034557.D
 Acq On : 08 Apr 2022 10:49
 Operator : CG/JU
 Sample : SSTD01005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010012

Quant Time: Apr 08 13:20:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 13:18:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.878	152	99658	20.000	ng/u1	0.00
20) Naphthalene-d8	10.689	136	397761	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.524	164	231540	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.265	188	454559	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.442	240	457058	20.000	ng/u1	# 0.00
88) Perylene-d12	23.824	264	482856	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.254	96	11365	4.978	ng/uL	0.00
4) Pyridine-d5	3.684	84	57680	10.612	ng/u1	0.00
7) Phenol-d5	7.037	99	72792	10.915	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.201	67	53575	13.693	ng/u1	0.00
11) 2-Chlorophenol-d4	7.407	132	62609	10.863	ng/u1	0.00
15) 4-Methylphenol-d8	8.589	113	56831	9.964	ng/u1	0.00
21) Nitrobenzene-d5	9.042	128	27365	9.591	ng/u1	0.00
24) 2-Nitrophenol-d4	9.766	143	27629	8.733	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.307	165	54193	8.544	ng/u1	0.00
31) 4-Chloroaniline-d4	10.830	131	63004	7.988	ng/u1	0.00
46) Dimethylphthalate-d6	13.936	166	164709	9.777	ng/u1	0.00
49) Acenaphthylene-d8	14.218	160	209366	9.784	ng/u1	0.00
54) 4-Nitrophenol-d4	14.724	143	13465	5.131	ng/u1	0.00
60) Fluorene-d10	15.512	176	143698	9.744	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.636	200	20267	7.047	ng/u1	0.00
73) Anthracene-d10	17.365	188	213950	9.965	ng/u1	0.00
81) Pyrene-d10	19.653	212	245409	9.863	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.671	264	231127	9.702	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.290	88	11827	5.228	ng/uL	89
5) Pyridine	3.701	79	69265	12.296	ng/u1#	72
6) Benzaldehyde	7.013	77	52359	15.085	ng/u1	91
8) Phenol	7.066	94	75144	11.419	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.295	93	64903	11.831	ng/u1	87
12) 2-Chlorophenol	7.437	128	63585	10.934	ng/u1#	84
13) 2-Methylphenol	8.325	108	55999	10.634	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.407	45	103125	20.386	ng/u1#	86
16) Acetophenone	8.701	105	93471	10.436	ng/u1#	87
17) N-Nitroso-di-n-propyla...	8.684	70	51597	12.244	ng/u1#	84
18) 4-Methylphenol	8.654	108	58864	10.369	ng/u1	98
19) Hexachloroethane	8.966	117	29411	11.214	ng/u1	83
22) Nitrobenzene	9.084	77	72567	11.718	ng/u1	89
23) Isophorone	9.613	82	129868	11.096	ng/u1	94
25) 2-Nitrophenol	9.801	139	30590	9.442	ng/u1#	89
26) 2,4-Dimethylphenol	9.860	107	69818	10.430	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.095	93	79305	11.015	ng/u1	97
29) 2,4-Dichlorophenol	10.336	162	54312	8.873	ng/u1#	95
30) Naphthalene	10.736	128	208209	10.481	ng/u1	99
32) 4-Chloroaniline	10.854	127	62835	8.010	ng/u1	94
33) Hexachlorobutadiene	11.030	225	39063	7.724	ng/u1	95
34) Caprolactam	11.613	113	13783	8.049	ng/u1#	78
35) 4-Chloro-3-methylphenol	11.977	107	55299	9.739	ng/u1	91
36) 2-Methylnaphthalene	12.354	142	129147	9.467	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034557.D
 Acq On : 08 Apr 2022 10:49
 Operator : CG/JU
 Sample : SSTD01005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010012

Quant Time: Apr 08 13:20:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 13:18:10 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.572	142	134865	9.679	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.724	216	69041	8.643	ng/ul#	96
40) Hexachlorocyclopentadiene	12.707	237	39875	6.997	ng/ul	96
41) 2,4,6-Trichlorophenol	12.966	196	39486	8.436	ng/ul	92
42) 2,4,5-Trichlorophenol	13.036	196	41203	8.122	ng/ul	97
43) 1,1'-Biphenyl	13.360	154	173074	10.063	ng/ul#	96
44) 2-Chloronaphthalene	13.401	162	134577	9.899	ng/ul	95
45) 2-Nitroaniline	13.607	65	31002	11.498	ng/ul#	76
47) Dimethylphthalate	13.983	163	164852	10.023	ng/ul#	98
48) 2,6-Dinitrotoluene	14.101	165	29086	9.070	ng/ul#	84
50) Acenaphthylene	14.248	152	215294	10.271	ng/ul	99
51) 3-Nitroaniline	14.436	138	19604	7.021	ng/ul	86
52) Acenaphthene	14.589	153	145915	10.613	ng/ul	99
53) 2,4-Dinitrophenol	14.642	184	10264	4.923	ng/ul#	86
55) 4-Nitrophenol	14.736	109	11867	5.434	ng/ul#	82
56) Dibenzofuran	14.924	168	195119	9.799	ng/ul	93
57) 2,4-Dinitrotoluene	14.889	165	38557	8.737	ng/ul#	71
58) 2,3,4,6-Tetrachlorophenol	15.148	232	36510	8.418	ng/ul#	88
59) Diethylphthalate	15.342	149	164542	10.471	ng/ul	96
61) Fluorene	15.571	166	161503	9.973	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.565	204	80011	9.113	ng/ul	91
63) 4-Nitroaniline	15.601	138	11573	4.896	ng/ul#	90
66) 4,6-Dinitro-2-methylph...	15.648	198	21673	7.710	ng/ul#	80
67) N-Nitrosodiphenylamine	15.777	169	132607	10.087	ng/ul	98
68) 4-Bromophenyl-phenylether	16.459	248	45905	8.964	ng/ul#	91
69) Hexachlorobenzene	16.577	284	56816	8.894	ng/ul#	93
70) Atrazine	16.730	200	45606	8.804	ng/ul	97
71) Pentachlorophenol	16.918	266	27670	7.201	ng/ul	96
72) Phenanthrene	17.307	178	252655	10.440	ng/ul	98
74) Anthracene	17.401	178	249144	10.373	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.324	216	71411	8.885	ng/uL	99
76) Pentachlorobenzene	14.848	250	67277	8.898	ng/uL	99
77) Carbazole	17.671	167	168701	8.643	ng/ul	99
78) Di-n-butylphthalate	18.236	149	246373	10.557	ng/ul#	98
80) Fluoranthene	19.324	202	265574	9.380	ng/ul#	89
82) Pyrene	19.683	202	301503	10.262	ng/ul#	89
83) Butylbenzylphthalate	20.583	149	107452	10.925	ng/ul#	85
84) 3,3'-Dichlorobenzidine	21.371	252	62480	6.895	ng/ul#	94
85) Benzo(a)anthracene	21.430	228	256914	9.618	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.359	149	160692	10.985	ng/ul	98
87) Chrysene	21.483	228	278832	9.696	ng/ul	99
89) Di-n-octyl phthalate	22.277	149	280401	10.838	ng/ul	100
90) Benzo(b)fluoranthene	23.100	252	270593	9.559	ng/ul#	98
91) Benzo(k)fluoranthene	23.147	252	287176	9.953	ng/ul#	96
93) Benzo(a)pyrene	23.718	252	269917	9.560	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.300	276	268300	8.691	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.324	278	218176	8.435	ng/ul#	93
96) Benzo(g,h,i)perylene	27.171	276	1238	0.043	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
Data File : BM034557.D
Acq On : 08 Apr 2022 10:49
Operator : CG/JU
Sample : SST01005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD010012

Quant Time: Apr 08 13:20:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 08 13:18:10 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 5.228 ng/uL
RT: 3.290 min Scan# 41
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion: 88 Resp: 11827
Ion Ratio Lower Upper
88 100
43 31.3 21.0 31.6
58 68.3 47.8 71.8



#5
Pyridine
Concen: 12.296 ng/uL
RT: 3.701 min Scan# 113
Delta R.T. 0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion: 79 Resp: 69265
Ion Ratio Lower Upper
79 100
52 81.6 47.7 71.5#
51 46.2 25.6 38.4#









#6
Benzaldehyde
Concen: 15.085 ng/ul
RT: 7.013 min Scan# 676
Delta R.T. 0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion: 77 Resp: 52359
Ion Ratio Lower Upper
77 100
105 91.5 80.8 121.2
106 86.9 76.6 115.0



#8
Phenol
Concen: 11.419 ng/ul
RT: 7.066 min Scan# 685
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion: 94 Resp: 75144
Ion Ratio Lower Upper
94 100
65 31.2 22.9 34.3
66 44.7 33.7 50.5









#10
Bis(2-Chloroethyl)ether
Concen: 11.831 ng/ul
RT: 7.295 min Scan# 724
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion: 93 Resp: 64903
Ion Ratio Lower Upper
93 100
63 81.5 54.7 82.1
95 30.2 26.6 39.8



#12
2-Chlorophenol
Concen: 10.934 ng/ul
RT: 7.437 min Scan# 748
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:128 Resp: 63585
Ion Ratio Lower Upper
128 100
64 55.6 30.6 46.0#
130 32.3 25.8 38.6









#13
2-Methylphenol
Concen: 10.634 ng/ul
RT: 8.325 min Scan# 899
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:108 Resp: 55999
Ion Ratio Lower Upper
108 100
107 86.3 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.386 ng/ul
RT: 8.407 min Scan# 913
Delta R.T. -0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion: 45 Resp: 103125
Ion Ratio Lower Upper
45 100
77 14.5 17.1 25.7#
79 10.8 12.8 19.2#









#16
Acetophenone
Concen: 10.436 ng/ul
RT: 8.701 min Scan# 963
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:105 Resp: 93471
Ion Ratio Lower Upper
105 100
77 83.6 59.4 89.2
51 34.4 19.3 28.9#



#17
N-Nitroso-di-n-propylamine
Concen: 12.244 ng/ul
RT: 8.684 min Scan# 960
Delta R.T. -0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion: 70 Resp: 51597
Ion Ratio Lower Upper
70 100
42 64.2 40.6 60.8#
101 8.8 8.2 12.2
130 18.5 19.9 29.9#









#18
4-Methylphenol
Concen: 10.369 ng/ul
RT: 8.654 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:108 Resp: 58864
Ion Ratio Lower Upper
108 100
107 112.9 89.0 133.4



#19
Hexachloroethane
Concen: 11.214 ng/ul
RT: 8.966 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:117 Resp: 29411
Ion Ratio Lower Upper
117 100
201 91.4 89.9 134.9
199 60.3 57.2 85.8









#22
Nitrobenzene
Concen: 11.718 ng/ul
RT: 9.084 min Scan# 1028
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion: 77	Resp: 72567
Ion Ratio	Lower Upper
77 100	
123 38.5	38.2 57.2
65 12.6	10.6 15.8



#23
Isophorone
Concen: 11.096 ng/ul
RT: 9.613 min Scan# 1118
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion: 82	Resp: 129868
Ion Ratio	Lower Upper
82 100	
95 7.7	6.8 10.2
138 17.1	16.5 24.7









#25
2-Nitrophenol
Concen: 9.442 ng/ul
RT: 9.801 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:139 Resp: 30590
Ion Ratio Lower Upper
139 100
65 58.6 37.8 56.6#
109 33.8 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 10.430 ng/ul
RT: 9.860 min Scan# 1160
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:107 Resp: 69818
Ion Ratio Lower Upper
107 100
121 49.8 40.8 61.2
122 79.4 69.0 103.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 11.015 ng/ul
 RT: 10.095 min Scan# 1200
 Delta R.T. -0.000 min
 Lab File: BM034557.D
 Acq: 08 Apr 2022 10:49




Instrument :
 BNA_M
 ClientSampleId :
 SST010012


Tgt Ion: 93 Resp: 79305
 Ion Ratio Lower Upper
 93 100
 95 30.6 25.4 38.2
 123 9.7 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 8.873 ng/ul
 RT: 10.336 min Scan# 1241
 Delta R.T. -0.000 min
 Lab File: BM034557.D
 Acq: 08 Apr 2022 10:49

Tgt Ion:162 Resp: 54312
 Ion Ratio Lower Upper
 162 100
 164 66.7 52.8 79.2
 98 39.8 26.0 39.0#









#30
Naphthalene
Concen: 10.481 ng/ul
RT: 10.736 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:128 Resp: 208209
Ion Ratio Lower Upper
128 100
129 10.5 8.7 13.1
127 12.7 9.9 14.9



#32
4-Chloroaniline
Concen: 8.010 ng/ul
RT: 10.854 min Scan# 1329
Delta R.T. 0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:127 Resp: 62835
Ion Ratio Lower Upper
127 100
129 34.6 25.1 37.7









#33
Hexachlorobutadiene
Concen: 7.724 ng/ul
RT: 11.030 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:225 Resp: 39063
Ion Ratio Lower Upper
225 100
223 62.5 50.6 75.8
227 69.8 50.3 75.5



#34
Caprolactam
Concen: 8.049 ng/ul
RT: 11.613 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:113 Resp: 13783
Ion Ratio Lower Upper
113 100
55 160.5 100.9 151.3#
56 114.3 81.0 121.4









#35
4-Chloro-3-methylphenol
Concen: 9.739 ng/ul
RT: 11.977 min Scan# 1520
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:107 Resp: 55299
Ion Ratio Lower Upper
107 100
144 24.9 23.0 34.4
142 78.1 69.8 104.8



#36
2-Methylnaphthalene
Concen: 9.467 ng/ul
RT: 12.354 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:142 Resp: 129147
Ion Ratio Lower Upper
142 100
141 90.5 71.4 107.0









#37
1-Methylnaphthalene
Concen: 9.679 ng/ul
RT: 12.572 min Scan# 1621
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:142 Resp: 134865
Ion Ratio Lower Upper
142 100
141 93.8 74.5 111.7
116 3.8 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 8.643 ng/ul
RT: 12.724 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49


Tgt Ion:216 Resp: 69041
Ion Ratio Lower Upper
216 100
214 78.7 62.2 93.2
179 22.4 15.4 23.2
108 19.5 11.8 17.6#








#40
Hexachlorocyclopentadiene
Concen: 6.997 ng/ul
RT: 12.707 min Scan# 1644
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49


Instrument :
BNA_M
ClientSampleId :
SSTD010012





Tgt Ion:237 Resp: 39875
Ion Ratio Lower Upper
237 100
235 63.0 52.9 79.3
272 13.2 11.8 17.8




#41
2,4,6-Trichlorophenol
Concen: 8.436 ng/ul
RT: 12.966 min Scan# 1688
Delta R.T. 0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49



Tgt Ion:196 Resp: 39486
Ion Ratio Lower Upper
196 100
198 92.6 80.8 121.2
200 29.7 26.5 39.7









#42
2,4,5-Trichlorophenol
Concen: 8.122 ng/ul
RT: 13.036 min Scan# 1700
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:196 Resp: 41203
Ion Ratio Lower Upper
196 100
198 100.1 78.0 117.0
200 31.8 24.5 36.7



#43
1,1'-Biphenyl
Concen: 10.063 ng/ul
RT: 13.360 min Scan# 1755
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:154 Resp: 173074
Ion Ratio Lower Upper
154 100
153 39.8 32.6 49.0
76 13.9 7.7 11.5#





#44

2-Chloronaphthalene

Concen: 9.899 ng/ul

RT: 13.401 min Scan# 1762

Delta R.T. -0.000 min

Lab File: BM034557.D

Acq: 08 Apr 2022 10:49

Instrument :

BNA_M

ClientSampleId :

SSTD010012





Tgt Ion:162 Resp: 134577

Ion Ratio Lower Upper

162 100

164 32.1 26.6 40.0

127 41.4 29.6 44.4



#45

2-Nitroaniline

Concen: 11.498 ng/ul

RT: 13.607 min Scan# 1797

Delta R.T. -0.000 min

Lab File: BM034557.D

Acq: 08 Apr 2022 10:49




Tgt Ion: 65 Resp: 31002


Ion Ratio Lower Upper

65 100

92 64.4 57.4 86.2

138 81.4 94.8 142.2#









#47
Dimethylphthalate
Concen: 10.023 ng/ul
RT: 13.983 min Scan# 1861
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:163 Resp: 164852
Ion Ratio Lower Upper
163 100
194 3.4 3.4 5.2#
164 9.1 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 9.070 ng/ul
RT: 14.101 min Scan# 1881
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:165 Resp: 29086
Ion Ratio Lower Upper
165 100
89 69.3 44.2 66.4#
121 24.1 16.7 25.1









#50
Acenaphthylene
Concen: 10.271 ng/ul
RT: 14.248 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:152 Resp: 215294
Ion Ratio Lower Upper
152 100
151 19.4 15.8 23.8
153 12.5 10.2 15.2



#51
3-Nitroaniline
Concen: 7.021 ng/ul
RT: 14.436 min Scan# 1938
Delta R.T. 0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:138 Resp: 19604
Ion Ratio Lower Upper
138 100
108 12.2 8.3 12.5
92 124.6 87.0 130.4





#52

Acenaphthene

Concen: 10.613 ng/ul

RT: 14.589 min Scan# 1964

Delta R.T. -0.000 min

Lab File: BM034557.D

Acq: 08 Apr 2022 10:49

Instrument :

BNA_M

ClientSampleId :

SSTD010012





Tgt Ion:153 Resp: 145915

Ion Ratio Lower Upper

153 100

152 46.4 37.4 56.0

154 89.3 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 4.923 ng/ul

RT: 14.642 min Scan# 1973

Delta R.T. 0.006 min

Lab File: BM034557.D

Acq: 08 Apr 2022 10:49




Tgt Ion:184 Resp: 10264


Ion Ratio Lower Upper

184 100

63 73.4 43.8 65.6#

154 65.6 49.9 74.9









#55
4-Nitrophenol
Concen: 5.434 ng/ul
RT: 14.736 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:109 Resp: 11867
Ion Ratio Lower Upper
109 100
139 107.3 94.6 142.0
65 131.2 83.5 125.3#



#56
Dibenzofuran
Concen: 9.799 ng/ul
RT: 14.924 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:168 Resp: 195119
Ion Ratio Lower Upper
168 100
139 41.2 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 8.737 ng/ul
RT: 14.889 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:165 Resp: 38557
Ion Ratio Lower Upper
165 100
63 56.3 30.2 45.4#
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.418 ng/ul
RT: 15.148 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:232 Resp: 36510
Ion Ratio Lower Upper
232 100
131 48.8 32.2 48.4#
130 3.4 1.9 2.9#
166 34.6 23.0 34.4#





#59

Diethylphthalate

Concen: 10.471 ng/ul

RT: 15.342 min Scan# 2092

Delta R.T. -0.000 min

Lab File: BM034557.D

Acq: 08 Apr 2022 10:49

Instrument :

BNA_M

ClientSampleId :

SSTD010012





Tgt Ion:149 Resp: 164542

Ion Ratio Lower Upper

149 100

177 20.2 17.9 26.9

150 11.8 10.2 15.2



#61

Fluorene

Concen: 9.973 ng/ul

RT: 15.571 min Scan# 2131

Delta R.T. -0.000 min

Lab File: BM034557.D

Acq: 08 Apr 2022 10:49




Tgt Ion:166 Resp: 161503


Ion Ratio Lower Upper

166 100

165 97.6 76.9 115.3

167 14.0 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 9.113 ng/ul
RT: 15.565 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:204 Resp: 80011
Ion Ratio Lower Upper
204 100
206 32.2 25.5 38.3
141 65.0 44.2 66.2



#63
4-Nitroaniline
Concen: 4.896 ng/ul
RT: 15.601 min Scan# 2136
Delta R.T. 0.012 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:138 Resp: 11573
Ion Ratio Lower Upper
138 100
92 60.2 36.3 54.5#
108 82.4 68.9 103.3









#66
4,6-Dinitro-2-methylphenol
Concen: 7.710 ng/ul
RT: 15.648 min Scan# 2144
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:198 Resp: 21673
Ion Ratio Lower Upper
198 100
182 7.5 3.8 5.6#
77 34.0 18.5 27.7#



#67
N-Nitrosodiphenylamine
Concen: 10.087 ng/ul
RT: 15.777 min Scan# 2166
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:169 Resp: 132607
Ion Ratio Lower Upper
169 100
168 67.6 52.8 79.2
167 35.8 28.1 42.1









#68
4-Bromophenyl-phenylether
Concen: 8.964 ng/ul
RT: 16.459 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:248 Resp: 45905
Ion Ratio Lower Upper
248 100
250 94.9 78.4 117.6
141 78.9 51.4 77.2#



#69
Hexachlorobenzene
Concen: 8.894 ng/ul
RT: 16.577 min Scan# 2302
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:284 Resp: 56816
Ion Ratio Lower Upper
284 100
142 37.7 24.6 36.8#
249 29.7 24.1 36.1









#70
 Atrazine
 Concen: 8.804 ng/ul
 RT: 16.730 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BM034557.D
 Acq: 08 Apr 2022 10:49




Instrument :
 BNA_M
 ClientSampleId :
 SST010012


Tgt Ion:200 Resp: 45606
 Ion Ratio Lower Upper
 200 100
 173 27.8 20.2 30.4
 215 47.0 38.4 57.6



#71
 Pentachlorophenol
 Concen: 7.201 ng/ul
 RT: 16.918 min Scan# 2360
 Delta R.T. -0.000 min
 Lab File: BM034557.D
 Acq: 08 Apr 2022 10:49

Tgt Ion:266 Resp: 27670
 Ion Ratio Lower Upper
 266 100
 264 65.7 49.3 73.9
 268 61.0 50.3 75.5









#72
Phenanthrene
Concen: 10.440 ng/ul
RT: 17.307 min Scan# 2426
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:178 Resp: 252655
Ion Ratio Lower Upper
178 100
179 15.9 11.7 17.5
176 18.3 15.0 22.4



#74
Anthracene
Concen: 10.373 ng/ul
RT: 17.401 min Scan# 2442
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:178 Resp: 249144
Ion Ratio Lower Upper
178 100
179 15.8 11.9 17.9
176 18.2 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 8.885 ng/uL

RT: 13.324 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM034557.D

Acq: 08 Apr 2022 10:49

Instrument :

BNA_M

ClientSampleId :

SSTD010012





Tgt Ion:216 Resp: 71411

Ion Ratio Lower Upper

216 100

214 76.8 62.0 93.0

218 46.3 37.8 56.8



#76

Pentachlorobenzene

Concen: 8.898 ng/uL

RT: 14.848 min Scan# 2008

Delta R.T. -0.000 min

Lab File: BM034557.D

Acq: 08 Apr 2022 10:49




Tgt Ion:250 Resp: 67277


Ion Ratio Lower Upper

250 100

248 63.2 50.6 75.8

252 64.2 50.2 75.4









#77
 Carbazole
 Concen: 8.643 ng/ul
 RT: 17.671 min Scan# 2488
 Delta R.T. -0.000 min
 Lab File: BM034557.D
 Acq: 08 Apr 2022 10:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010012


Tgt Ion:167 Resp: 168701
 Ion Ratio Lower Upper
 167 100
 166 20.7 16.6 25.0
 139 12.9 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 10.557 ng/ul
 RT: 18.236 min Scan# 2584
 Delta R.T. -0.000 min
 Lab File: BM034557.D
 Acq: 08 Apr 2022 10:49

Tgt Ion:149 Resp: 246373
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 6.5 4.2 6.2#









#80
Fluoranthene
Concen: 9.380 ng/ul
RT: 19.324 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:202 Resp: 265574
Ion Ratio Lower Upper
202 100
101 11.9 6.2 9.2#
100 8.9 4.7 7.1#



#82
Pyrene
Concen: 10.262 ng/ul
RT: 19.683 min Scan# 2830
Delta R.T. -0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:202 Resp: 301503
Ion Ratio Lower Upper
202 100
101 13.1 7.0 10.4#
100 10.7 5.8 8.6#









#83
Butylbenzylphthalate
Concen: 10.925 ng/ul
RT: 20.583 min Scan# 2983
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:149 Resp: 107452
Ion Ratio Lower Upper
149 100
91 76.6 50.9 76.3#
206 19.5 19.3 28.9



#84
3,3'-Dichlorobenzidine
Concen: 6.895 ng/ul
RT: 21.371 min Scan# 3117
Delta R.T. 0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:252 Resp: 62480
Ion Ratio Lower Upper
252 100
254 66.9 50.7 76.1
126 13.2 6.6 9.8#









#85
Benzo(a)anthracene
Concen: 9.618 ng/ul
RT: 21.430 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:228 Resp: 256914
Ion Ratio Lower Upper
228 100
229 19.7 15.4 23.2
226 25.8 21.0 31.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 10.985 ng/ul
RT: 21.359 min Scan# 3115
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:149 Resp: 160692
Ion Ratio Lower Upper
149 100
167 26.4 22.1 33.1
279 4.6 4.2 6.2









#87
Chrysene
Concen: 9.696 ng/ul
RT: 21.483 min Scan# 3136
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:228 Resp: 278832
Ion Ratio Lower Upper
228 100
226 30.1 23.4 35.2
229 19.5 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 10.838 ng/ul
RT: 22.277 min Scan# 3271
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:149 Resp: 280401









#90
Benzo(b)fluoranthene
Concen: 9.559 ng/ul
RT: 23.100 min Scan# 3411
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:252 Resp: 270593
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 9.7 5.8 8.6#



#91
Benzo(k)fluoranthene
Concen: 9.953 ng/ul
RT: 23.147 min Scan# 3419
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:252 Resp: 287176
Ion Ratio Lower Upper
252 100
253 22.6 17.5 26.3
125 10.5 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 9.560 ng/ul
RT: 23.718 min Scan# 3516
Delta R.T. -0.000 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49




Instrument :
BNA_M
ClientSampleId :
SSTD010012


Tgt Ion:252 Resp: 269917
Ion Ratio Lower Upper
252 100
253 20.8 17.5 26.3
125 11.1 6.2 9.4#



#94
Indeno(1,2,3-cd)pyrene
Concen: 8.691 ng/ul
RT: 26.300 min Scan# 3955
Delta R.T. -0.006 min
Lab File: BM034557.D
Acq: 08 Apr 2022 10:49

Tgt Ion:276 Resp: 268300
Ion Ratio Lower Upper
276 100
138 21.9 13.6 20.4#
277 24.3 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 8.435 ng/ul
 RT: 26.324 min Scan# 3958
 Delta R.T. 0.006 min
 Lab File: BM034557.D
 Acq: 08 Apr 2022 10:49

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010012

Tgt Ion:278 Resp: 218176
 Ion Ratio Lower Upper
 278 100
 139 16.1 9.6 14.4#
 279 21.8 19.3 28.9



#96
 Benzo(g,h,i)perylene
 Concen: 0.043 ng/ul
 RT: 27.171 min Scan# 4103
 Delta R.T. 0.106 min
 Lab File: BM034557.D
 Acq: 08 Apr 2022 10:49

Tgt Ion:276 Resp: 1238
 Ion Ratio Lower Upper
 276 100
 138 15.2 11.8 17.6
 277 22.6 18.8 28.2

