

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034559.D
 Acq On : 08 Apr 2022 12:02
 Operator : CG/JU
 Sample : SST04007
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040014

Quant Time: Apr 08 13:22:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 13:18:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.878	152	113852	20.000	ng/u1	0.00
20) Naphthalene-d8	10.689	136	459885	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.524	164	269950	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.265	188	537080	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.447	240	541787	20.000	ng/u1	# 0.00
88) Perylene-d12	23.830	264	540598	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.254	96	49554	18.999	ng/uL	0.00
4) Pyridine-d5	3.678	84	318375	51.270	ng/u1	0.00
7) Phenol-d5	7.036	99	381556	50.080	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.201	67	257126	57.526	ng/u1	0.00
11) 2-Chlorophenol-d4	7.407	132	308912	46.915	ng/u1	0.00
15) 4-Methylphenol-d8	8.595	113	300893	46.176	ng/u1	0.00
21) Nitrobenzene-d5	9.183	128	155	0.047	ng/u1	0.14
24) 2-Nitrophenol-d4	9.766	143	158651	43.372	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.307	165	285301	38.906	ng/u1	0.00
31) 4-Chloroaniline-d4	10.824	131	367874	40.339	ng/u1	0.00
46) Dimethylphthalate-d6	13.936	166	791484	40.296	ng/u1	0.00
49) Acenaphthylene-d8	14.218	160	1040640	41.713	ng/u1	0.00
54) 4-Nitrophenol-d4	14.718	143	128640	42.046	ng/u1	0.00
60) Fluorene-d10	15.518	176	684673	39.820	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.636	200	133102	39.170	ng/u1	0.00
73) Anthracene-d10	17.365	188	1040758	41.027	ng/u1	0.00
81) Pyrene-d10	19.659	212	1205517	40.871	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.671	264	1178008	44.169	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.290	88	48841	18.900	ng/uL#	82
5) Pyridine	3.696	79	341312	53.036	ng/u1#	79
6) Benzaldehyde	7.007	77	248263	62.609	ng/u1	90
8) Phenol	7.066	94	387227	51.509	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.295	93	310206	49.497	ng/u1#	88
12) 2-Chlorophenol	7.436	128	314534	47.346	ng/u1#	86
13) 2-Methylphenol	8.325	108	288212	47.908	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.419	45	457022	79.080	ng/u1#	86
16) Acetophenone	8.701	105	463765	45.326	ng/u1#	87
17) N-Nitroso-di-n-propyla...	8.695	70	254546	52.872	ng/u1#	87
18) 4-Methylphenol	8.654	108	304865	47.006	ng/u1	98
19) Hexachloroethane	8.966	117	137293	45.823	ng/u1	84
22) Nitrobenzene	9.083	77	373154	52.118	ng/u1	92
23) Isophorone	9.613	82	656780	48.535	ng/u1#	93
25) 2-Nitrophenol	9.801	139	166710	44.507	ng/u1	91
26) 2,4-Dimethylphenol	9.860	107	345449	44.636	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.095	93	395637	47.527	ng/u1	99
29) 2,4-Dichlorophenol	10.336	162	275693	38.958	ng/u1	96
30) Naphthalene	10.736	128	959055	41.758	ng/u1	98
32) 4-Chloroaniline	10.848	127	369788	40.769	ng/u1	98
33) Hexachlorobutadiene	11.030	225	183577	31.396	ng/u1	98
34) Caprolactam	11.819	113	404	0.204	ng/u1	91
35) 4-Chloro-3-methylphenol	11.977	107	294842	44.913	ng/u1	92
36) 2-Methylnaphthalene	12.354	142	632405	40.095	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034559.D
 Acq On : 08 Apr 2022 12:02
 Operator : CG/JU
 Sample : SST04007
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040014

Quant Time: Apr 08 13:22:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 13:18:10 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.571	142	638626	39.642	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.724	216	328187	35.237	ng/ul#	97
40) Hexachlorocyclopentadiene	12.707	237	207601	31.244	ng/ul	98
41) 2,4,6-Trichlorophenol	12.960	196	210330	38.540	ng/ul	96
42) 2,4,5-Trichlorophenol	13.036	196	220980	37.362	ng/ul	100
43) 1,1'-Biphenyl	13.360	154	827212	41.253	ng/ul#	98
44) 2-Chloronaphthalene	13.401	162	637374	40.214	ng/ul	96
45) 2-Nitroaniline	13.607	65	200675	63.836	ng/ul	86
47) Dimethylphthalate	13.983	163	770458	40.179	ng/ul	99
48) 2,6-Dinitrotoluene	14.101	165	163350	43.689	ng/ul	91
50) Acenaphthylene	14.248	152	1044611	42.743	ng/ul	99
51) 3-Nitroaniline	14.430	138	149489	45.917	ng/ul	86
52) Acenaphthene	14.589	153	681005	42.484	ng/ul	100
53) 2,4-Dinitrophenol	14.636	184	85068	34.993	ng/ul#	80
55) 4-Nitrophenol	14.736	109	112496	44.182	ng/ul#	82
56) Dibenzofuran	14.924	168	936463	40.338	ng/ul	94
57) 2,4-Dinitrotoluene	14.889	165	230270	44.757	ng/ul#	81
58) 2,3,4,6-Tetrachlorophenol	15.154	232	193037	38.174	ng/ul#	90
59) Diethylphthalate	15.348	149	790484	43.145	ng/ul	99
61) Fluorene	15.571	166	791519	41.924	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.565	204	375659	36.699	ng/ul	92
63) 4-Nitroaniline	15.589	138	135188	49.053	ng/ul#	93
66) 4,6-Dinitro-2-methylph...	15.654	198	133461	40.181	ng/ul#	91
67) N-Nitrosodiphenylamine	15.777	169	663536	42.716	ng/ul	99
68) 4-Bromophenyl-phenylether	16.459	248	224100	37.035	ng/ul#	92
69) Hexachlorobenzene	16.577	284	266887	35.359	ng/ul#	91
70) Atrazine	16.730	200	242090	39.555	ng/ul	98
71) Pentachlorophenol	16.924	266	166594	36.694	ng/ul	99
72) Phenanthrene	17.312	178	1207190	42.216	ng/ul	99
74) Anthracene	17.401	178	1227226	43.243	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.330	216	336640	35.449	ng/uL	99
76) Pentachlorobenzene	14.848	250	315927	35.365	ng/uL	99
77) Carbazole	17.671	167	1020006	44.227	ng/ul	98
78) Di-n-butylphthalate	18.242	149	1306411	47.376	ng/ul	99
80) Fluoranthene	19.324	202	1370789	40.845	ng/ul#	89
82) Pyrene	19.689	202	1445549	41.506	ng/ul#	88
83) Butylbenzylphthalate	20.583	149	597775	51.275	ng/ul#	84
84) 3,3'-Dichlorobenzidine	21.365	252	425790	39.643	ng/ul#	97
85) Benzo(a)anthracene	21.430	228	1323223	41.790	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.365	149	887448	51.181	ng/ul	97
87) Chrysene	21.483	228	1387074	40.690	ng/ul	99
89) Di-n-octyl phthalate	22.277	149	1503447	51.902	ng/ul	100
90) Benzo(b)fluoranthene	23.100	252	1390780	43.882	ng/ul#	98
91) Benzo(k)fluoranthene	23.147	252	1398747	43.299	ng/ul#	97
93) Benzo(a)pyrene	23.724	252	1378061	43.594	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.312	276	1547705	44.778	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.323	278	1285041	44.376	ng/ul#	96
96) Benzo(g,h,i)perylene	27.070	276	1411846	43.925	ng/ul#	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034559.D
 Acq On : 08 Apr 2022 12:02
 Operator : CG/JU
 Sample : SST04007
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040014

Quant Time: Apr 08 13:22:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 13:18:10 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 18.900 ng/uL
RT: 3.290 min Scan# 41
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion: 88 Resp: 48841
Ion Ratio Lower Upper
88 100
43 34.0 21.0 31.6#
58 74.8 47.8 71.8#



#5
Pyridine
Concen: 53.036 ng/uL
RT: 3.696 min Scan# 112
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion: 79 Resp: 341312
Ion Ratio Lower Upper
79 100
52 78.6 47.7 71.5#
51 39.6 25.6 38.4#









#6
Benzaldehyde
Concen: 62.609 ng/ul
RT: 7.007 min Scan# 675
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion: 77 Resp: 248263
Ion Ratio Lower Upper
77 100
105 90.5 80.8 121.2
106 85.6 76.6 115.0



#8
Phenol
Concen: 51.509 ng/ul
RT: 7.066 min Scan# 685
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion: 94 Resp: 387227
Ion Ratio Lower Upper
94 100
65 31.4 22.9 34.3
66 41.5 33.7 50.5









#10
Bis(2-Chloroethyl)ether
Concen: 49.497 ng/ul
RT: 7.295 min Scan# 724
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion: 93 Resp: 310206
Ion Ratio Lower Upper
93 100
63 82.5 54.7 82.1#
95 33.0 26.6 39.8



#12
2-Chlorophenol
Concen: 47.346 ng/ul
RT: 7.436 min Scan# 748
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:128 Resp: 314534
Ion Ratio Lower Upper
128 100
64 53.3 30.6 46.0#
130 31.6 25.8 38.6





#13

2-Methylphenol

Concen: 47.908 ng/ul

RT: 8.325 min Scan# 899

Delta R.T. -0.000 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Instrument :

BNA_M

ClientSampleId :





SSTD040014

Tgt Ion:108 Resp: 288212

Ion Ratio Lower Upper

108 100

107 89.4 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 79.080 ng/ul

RT: 8.419 min Scan# 915

Delta R.T. 0.006 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02




Tgt Ion: 45 Resp: 457022


Ion Ratio Lower Upper

45 100

77 13.8 17.1 25.7#

79 10.9 12.8 19.2#









#16
 Acetophenone
 Concen: 45.326 ng/ul
 RT: 8.701 min Scan# 963
 Delta R.T. -0.000 min
 Lab File: BM034559.D
 Acq: 08 Apr 2022 12:02




Instrument :
 BNA_M
 ClientSampleId :
 SST040014


Tgt Ion:105 Resp: 463765
 Ion Ratio Lower Upper
 105 100
 77 84.1 59.4 89.2
 51 32.7 19.3 28.9#



#17
 N-Nitroso-di-n-propylamine
 Concen: 52.872 ng/ul
 RT: 8.695 min Scan# 962
 Delta R.T. 0.006 min
 Lab File: BM034559.D
 Acq: 08 Apr 2022 12:02

Tgt Ion: 70 Resp: 254546
 Ion Ratio Lower Upper
 70 100
 42 62.4 40.6 60.8#
 101 9.8 8.2 12.2
 130 19.1 19.9 29.9#









#18
4-Methylphenol
Concen: 47.006 ng/ul
RT: 8.654 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:108 Resp: 304865
Ion Ratio Lower Upper
108 100
107 112.9 89.0 133.4



#19
Hexachloroethane
Concen: 45.823 ng/ul
RT: 8.966 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:117 Resp: 137293
Ion Ratio Lower Upper
117 100
201 94.9 89.9 134.9
199 58.3 57.2 85.8









#22
Nitrobenzene
Concen: 52.118 ng/ul
RT: 9.083 min Scan# 1028
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion: 77 Resp: 373154
Ion Ratio Lower Upper
77 100
123 41.2 38.2 57.2
65 14.1 10.6 15.8



#23
Isophorone
Concen: 48.535 ng/ul
RT: 9.613 min Scan# 1118
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion: 82 Resp: 656780
Ion Ratio Lower Upper
82 100
95 7.4 6.8 10.2
138 16.4 16.5 24.7#









#25
2-Nitrophenol
Concen: 44.507 ng/ul
RT: 9.801 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:139 Resp: 166710
Ion Ratio Lower Upper
139 100
65 56.4 37.8 56.6
109 32.4 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 44.636 ng/ul
RT: 9.860 min Scan# 1160
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:107 Resp: 345449
Ion Ratio Lower Upper
107 100
121 48.9 40.8 61.2
122 81.2 69.0 103.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 47.527 ng/ul
 RT: 10.095 min Scan# 1200
 Delta R.T. -0.000 min
 Lab File: BM034559.D
 Acq: 08 Apr 2022 12:02




Instrument :
 BNA_M
 ClientSampleId :
 SST040014


Tgt Ion: 93 Resp: 395637
 Ion Ratio Lower Upper
 93 100
 95 32.0 25.4 38.2
 123 10.4 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 38.958 ng/ul
 RT: 10.336 min Scan# 1241
 Delta R.T. -0.000 min
 Lab File: BM034559.D
 Acq: 08 Apr 2022 12:02

Tgt Ion:162 Resp: 275693
 Ion Ratio Lower Upper
 162 100
 164 65.3 52.8 79.2
 98 38.0 26.0 39.0









#30
Naphthalene
Concen: 41.758 ng/ul
RT: 10.736 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:128 Resp: 959055
Ion Ratio Lower Upper
128 100
129 11.4 8.7 13.1
127 13.1 9.9 14.9



#32
4-Chloroaniline
Concen: 40.769 ng/ul
RT: 10.848 min Scan# 1328
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:127 Resp: 369788
Ion Ratio Lower Upper
127 100
129 32.8 25.1 37.7









#33
Hexachlorobutadiene
Concen: 31.396 ng/ul
RT: 11.030 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:225 Resp: 183577
Ion Ratio Lower Upper
225 100
223 61.3 50.6 75.8
227 63.9 50.3 75.5



#34
Caprolactam
Concen: 0.204 ng/ul
RT: 11.819 min Scan# 1493
Delta R.T. 0.206 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:113 Resp: 404
Ion Ratio Lower Upper
113 100
55 118.1 100.9 151.3
56 113.2 81.0 121.4









#35
4-Chloro-3-methylphenol
Concen: 44.913 ng/ul
RT: 11.977 min Scan# 1520
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:107 Resp: 294842
Ion Ratio Lower Upper
107 100
144 25.6 23.0 34.4
142 79.1 69.8 104.8



#36
2-Methylnaphthalene
Concen: 40.095 ng/ul
RT: 12.354 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:142 Resp: 632405
Ion Ratio Lower Upper
142 100
141 87.7 71.4 107.0









#37
1-Methylnaphthalene
Concen: 39.642 ng/ul
RT: 12.571 min Scan# 1621
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:142 Resp: 638626
Ion Ratio Lower Upper
142 100
141 92.2 74.5 111.7
116 4.2 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 35.237 ng/ul
RT: 12.724 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:216 Resp: 328187
Ion Ratio Lower Upper
216 100
214 78.9 62.2 93.2
179 21.7 15.4 23.2
108 19.0 11.8 17.6#









#40
Hexachlorocyclopentadiene
Concen: 31.244 ng/ul
RT: 12.707 min Scan# 1644
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:237 Resp: 207601
Ion Ratio Lower Upper
237 100
235 65.1 52.9 79.3
272 12.9 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 38.540 ng/ul
RT: 12.960 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:196 Resp: 210330
Ion Ratio Lower Upper
196 100
198 97.3 80.8 121.2
200 29.8 26.5 39.7





#42

2,4,5-Trichlorophenol

Concen: 37.362 ng/ul

RT: 13.036 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM034559.D


Acq: 08 Apr 2022 12:02

Instrument :

BNA_M

ClientSampleId :

SSTD040014






Tgt Ion:196 Resp: 220980

Ion Ratio Lower Upper

196 100

198 97.8 78.0 117.0

200 31.2 24.5 36.7



#43

1,1'-Biphenyl

Concen: 41.253 ng/ul

RT: 13.360 min Scan# 1755

Delta R.T. -0.000 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02




Tgt Ion:154 Resp: 827212


Ion Ratio Lower Upper

154 100

153 40.9 32.6 49.0

76 13.8 7.7 11.5#





#44

2-Chloronaphthalene

Concen: 40.214 ng/ul

RT: 13.401 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Instrument :

BNA_M

ClientSampleId :

SSTD040014





Tgt Ion:162 Resp: 637374

Ion Ratio Lower Upper

162 100

164 32.5 26.6 40.0

127 40.5 29.6 44.4



#45

2-Nitroaniline

Concen: 63.836 ng/ul

RT: 13.607 min Scan# 1797

Delta R.T. -0.000 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02




Tgt Ion: 65 Resp: 200675


Ion Ratio Lower Upper

65 100

92 68.9 57.4 86.2

138 95.3 94.8 142.2









#47
Dimethylphthalate
Concen: 40.179 ng/ul
RT: 13.983 min Scan# 1861
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:163 Resp: 770458
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.2
164 10.1 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 43.689 ng/ul
RT: 14.101 min Scan# 1881
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:165 Resp: 163350
Ion Ratio Lower Upper
165 100
89 62.6 44.2 66.4
121 23.5 16.7 25.1









#50
Acenaphthylene
Concen: 42.743 ng/ul
RT: 14.248 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:152 Resp: 1044611
Ion Ratio Lower Upper
152 100
151 20.3 15.8 23.8
153 13.5 10.2 15.2



#51
3-Nitroaniline
Concen: 45.917 ng/ul
RT: 14.430 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:138 Resp: 149489
Ion Ratio Lower Upper
138 100
108 10.2 8.3 12.5
92 125.2 87.0 130.4





#52

Acenaphthene

Concen: 42.484 ng/ul

RT: 14.589 min Scan# 1964

Delta R.T. -0.000 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Instrument :

BNA_M

ClientSampleId :

SSTD040014





Tgt Ion:153 Resp: 681005

Ion Ratio Lower Upper

153 100

152 47.0 37.4 56.0

154 88.9 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 34.993 ng/ul

RT: 14.636 min Scan# 1972

Delta R.T. -0.000 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02




Tgt Ion:184 Resp: 85068


Ion Ratio Lower Upper

184 100

63 80.3 43.8 65.6#

154 67.8 49.9 74.9









#55
4-Nitrophenol
Concen: 44.182 ng/ul
RT: 14.736 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:109 Resp: 112496
Ion Ratio Lower Upper
109 100
139 121.4 94.6 142.0
65 140.8 83.5 125.3#



#56
Dibenzofuran
Concen: 40.338 ng/ul
RT: 14.924 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:168 Resp: 936463
Ion Ratio Lower Upper
168 100
139 40.5 29.8 44.6





#57

2,4-Dinitrotoluene

Concen: 44.757 ng/ul

RT: 14.889 min Scan# 2015

Delta R.T. -0.000 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Instrument :

BNA_M

ClientSampleId :

SSTD040014





Tgt Ion:165 Resp: 230270

Ion Ratio Lower Upper

165 100

63 49.7 30.2 45.4#

182 3.2 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 38.174 ng/ul

RT: 15.154 min Scan# 2060

Delta R.T. 0.006 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Tgt Ion:232 Resp: 193037




Ion Ratio Lower Upper


232 100

131 49.0 32.2 48.4#

130 3.0 1.9 2.9#

166 30.9 23.0 34.4









#59
Diethylphthalate
Concen: 43.145 ng/ul
RT: 15.348 min Scan# 2092
Delta R.T. 0.006 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:149 Resp: 790484
Ion Ratio Lower Upper
149 100
177 21.5 17.9 26.9
150 12.6 10.2 15.2



#61
Fluorene
Concen: 41.924 ng/ul
RT: 15.571 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:166 Resp: 791519
Ion Ratio Lower Upper
166 100
165 97.8 76.9 115.3
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 36.699 ng/ul
RT: 15.565 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:204 Resp: 375659
Ion Ratio Lower Upper
204 100
206 31.5 25.5 38.3
141 64.4 44.2 66.2



#63
4-Nitroaniline
Concen: 49.053 ng/ul
RT: 15.589 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:138 Resp: 135188
Ion Ratio Lower Upper
138 100
92 56.9 36.3 54.5#
108 85.3 68.9 103.3









#66
4,6-Dinitro-2-methylphenol
Concen: 40.181 ng/ul
RT: 15.654 min Scan# 2145
Delta R.T. 0.006 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:198 Resp: 133461
Ion Ratio Lower Upper
198 100
182 4.0 3.8 5.6
77 28.3 18.5 27.7#



#67
N-Nitrosodiphenylamine
Concen: 42.716 ng/ul
RT: 15.777 min Scan# 2166
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:169 Resp: 663536
Ion Ratio Lower Upper
169 100
168 66.5 52.8 79.2
167 35.8 28.1 42.1









#68
4-Bromophenyl-phenylether
Concen: 37.035 ng/ul
RT: 16.459 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:248 Resp: 224100
Ion Ratio Lower Upper
248 100
250 97.7 78.4 117.6
141 80.0 51.4 77.2#



#69
Hexachlorobenzene
Concen: 35.359 ng/ul
RT: 16.577 min Scan# 2302
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:284 Resp: 266887
Ion Ratio Lower Upper
284 100
142 38.7 24.6 36.8#
249 31.6 24.1 36.1





#70

Atrazine

Concen: 39.555 ng/ul

RT: 16.730 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Instrument :

BNA_M

ClientSampleId :

SSTD040014





Tgt Ion:200 Resp: 242090

Ion Ratio Lower Upper

200 100

173 26.1 20.2 30.4

215 46.5 38.4 57.6



#71

Pentachlorophenol

Concen: 36.694 ng/ul

RT: 16.924 min Scan# 2361

Delta R.T. 0.006 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02




Tgt Ion:266 Resp: 166594


Ion Ratio Lower Upper

266 100

264 63.5 49.3 73.9

268 62.9 50.3 75.5









#72
Phenanthrene
Concen: 42.216 ng/u1
RT: 17.312 min Scan# 2427
Delta R.T. 0.006 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:178 Resp: 1207190
Ion Ratio Lower Upper
178 100
179 15.3 11.7 17.5
176 19.1 15.0 22.4



#74
Anthracene
Concen: 43.243 ng/u1
RT: 17.401 min Scan# 2442
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:178 Resp: 1227226
Ion Ratio Lower Upper
178 100
179 15.4 11.9 17.9
176 18.4 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.449 ng/uL
RT: 13.330 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:216 Resp: 336640
Ion Ratio Lower Upper
216 100
214 78.4 62.0 93.0
218 48.2 37.8 56.8



#76
Pentachlorobenzene
Concen: 35.365 ng/uL
RT: 14.848 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:250 Resp: 315927
Ion Ratio Lower Upper
250 100
248 62.1 50.6 75.8
252 63.7 50.2 75.4









#77
 Carbazole
 Concen: 44.227 ng/ul
 RT: 17.671 min Scan# 2488
 Delta R.T. -0.000 min
 Lab File: BM034559.D
 Acq: 08 Apr 2022 12:02




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040014


Tgt Ion:167 Resp: 1020006
 Ion Ratio Lower Upper
 167 100
 166 21.5 16.6 25.0
 139 13.3 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 47.376 ng/ul
 RT: 18.242 min Scan# 2585
 Delta R.T. 0.006 min
 Lab File: BM034559.D
 Acq: 08 Apr 2022 12:02

Tgt Ion:149 Resp: 1306411
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.0
 104 5.6 4.2 6.2









#80
Fluoranthene
Concen: 40.845 ng/ul
RT: 19.324 min Scan# 2769
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:202 Resp: 1370789
Ion Ratio Lower Upper
202 100
101 11.9 6.2 9.2#
100 9.3 4.7 7.1#



#82
Pyrene
Concen: 41.506 ng/ul
RT: 19.689 min Scan# 2831
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:202 Resp: 1445549
Ion Ratio Lower Upper
202 100
101 13.2 7.0 10.4#
100 11.4 5.8 8.6#









#83
Butylbenzylphthalate
Concen: 51.275 ng/ul
RT: 20.583 min Scan# 2983
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:149 Resp: 597775
Ion Ratio Lower Upper
149 100
91 77.9 50.9 76.3#
206 19.3 19.3 28.9



#84
3,3'-Dichlorobenzidine
Concen: 39.643 ng/ul
RT: 21.365 min Scan# 3116
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:252 Resp: 425790
Ion Ratio Lower Upper
252 100
254 65.2 50.7 76.1
126 10.9 6.6 9.8#









#85
Benzo(a)anthracene
Concen: 41.790 ng/ul
RT: 21.430 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:228 Resp: 1323223
Ion Ratio Lower Upper
228 100
229 19.9 15.4 23.2
226 26.8 21.0 31.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 51.181 ng/ul
RT: 21.365 min Scan# 3116
Delta R.T. 0.006 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:149 Resp: 887448
Ion Ratio Lower Upper
149 100
167 26.0 22.1 33.1
279 4.5 4.2 6.2









#87
Chrysene
Concen: 40.690 ng/ul
RT: 21.483 min Scan# 3136
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:228 Resp: 1387074
Ion Ratio Lower Upper
228 100
226 29.8 23.4 35.2
229 19.4 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 51.902 ng/ul
RT: 22.277 min Scan# 3271
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:149 Resp: 1503447









#90
Benzo(b)fluoranthene
Concen: 43.882 ng/ul
RT: 23.100 min Scan# 3411
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02




Instrument :
BNA_M
ClientSampleId :
SSTD040014


Tgt Ion:252 Resp: 1390780
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 10.2 5.8 8.6#



#91
Benzo(k)fluoranthene
Concen: 43.299 ng/ul
RT: 23.147 min Scan# 3419
Delta R.T. -0.000 min
Lab File: BM034559.D
Acq: 08 Apr 2022 12:02

Tgt Ion:252 Resp: 1398747
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 10.4 5.1 7.7#





#93

Benzo(a)pyrene

Concen: 43.594 ng/ul

RT: 23.724 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Instrument :

BNA_M

ClientSampleId :

SSTD040014





Tgt Ion:252 Resp: 1378061

Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 10.9 6.2 9.4#



#94

Indeno(1,2,3-cd)pyrene

Concen: 44.778 ng/ul

RT: 26.312 min Scan# 3957

Delta R.T. 0.006 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02




Tgt Ion:276 Resp: 1547705


Ion Ratio Lower Upper

276 100

138 23.9 13.6 20.4#

277 24.9 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 44.376 ng/ul

RT: 26.323 min Scan# 3959

Delta R.T. 0.006 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Instrument :

BNA_M

ClientSampleId :

SSTD040014





Tgt Ion:278 Resp: 1285041

Ion Ratio Lower Upper

278 100

139 15.5 9.6 14.4#

279 23.6 19.3 28.9



#96

Benzo(g,h,i)perylene

Concen: 43.925 ng/ul

RT: 27.070 min Scan# 4086

Delta R.T. 0.006 min

Lab File: BM034559.D

Acq: 08 Apr 2022 12:02

Tgt Ion:276 Resp: 1411846

Ion Ratio Lower Upper

276 100

138 20.2 11.8 17.6#

277 23.6 18.8 28.2

