

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034560.D
 Acq On : 08 Apr 2022 12:39
 Operator : CG/JU
 Sample : SSTD08008
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080015

Quant Time: Apr 08 14:09:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 13:18:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.878	152	114898	20.000	ng/u1	0.00
20) Naphthalene-d8	10.689	136	468246	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.524	164	278089	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.271	188	555759	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.447	240	563886	20.000	ng/u1	# 0.00
88) Perylene-d12	23.830	264	578141	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.249	96	100612	38.223	ng/uL	0.00
4) Pyridine-d5	3.672	84	673563	107.481	ng/u1	0.00
7) Phenol-d5	7.042	99	826124	107.442	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.207	67	515991	114.390	ng/u1	0.00
11) 2-Chlorophenol-d4	7.407	132	649500	97.742	ng/u1	0.00
15) 4-Methylphenol-d8	8.595	113	646021	98.238	ng/u1	0.00
21) Nitrobenzene-d5	9.042	128	315255	93.862	ng/u1	0.00
24) 2-Nitrophenol-d4	9.772	143	343228	92.157	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.313	165	614036	82.240	ng/u1	0.00
31) 4-Chloroaniline-d4	10.825	131	828871	89.267	ng/u1	0.00
46) Dimethylphthalate-d6	13.942	166	1692786	83.660	ng/u1	0.00
49) Acenaphthylene-d8	14.218	160	2236787	87.036	ng/u1	0.00
54) 4-Nitrophenol-d4	14.730	143	305859	97.044	ng/u1	0.01
60) Fluorene-d10	15.518	176	1452680	82.014	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.642	200	309874	88.126	ng/u1	0.00
73) Anthracene-d10	17.371	188	2237631	85.244	ng/u1	0.00
81) Pyrene-d10	19.659	212	2565240	83.563	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.677	264	2529730	88.691	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.290	88	102704	39.381	ng/uL#	82
5) Pyridine	3.696	79	703928	108.387	ng/u1#	81
6) Benzaldehyde	7.007	77	296036	73.977	ng/u1	92
8) Phenol	7.072	94	819354	107.998	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.301	93	643636	101.764	ng/u1	89
12) 2-Chlorophenol	7.442	128	655977	97.843	ng/u1#	89
13) 2-Methylphenol	8.325	108	608652	100.253	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.419	45	939111	161.018	ng/u1#	86
16) Acetophenone	8.707	105	984158	95.310	ng/u1#	90
17) N-Nitroso-di-n-propyla...	8.701	70	542021	111.559	ng/u1#	86
18) 4-Methylphenol	8.660	108	649651	99.256	ng/u1	97
19) Hexachloroethane	8.966	117	278508	92.109	ng/u1	84
22) Nitrobenzene	9.089	77	778572	106.800	ng/u1	93
23) Isophorone	9.619	82	1410151	102.347	ng/u1#	92
25) 2-Nitrophenol	9.801	139	356404	93.451	ng/u1#	89
26) 2,4-Dimethylphenol	9.866	107	732075	92.904	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.101	93	832297	98.196	ng/u1	99
29) 2,4-Dichlorophenol	10.336	162	586192	81.355	ng/u1	96
30) Naphthalene	10.742	128	2004939	85.738	ng/u1	99
32) 4-Chloroaniline	10.848	127	824771	89.307	ng/u1	98
33) Hexachlorobutadiene	11.030	225	374972	62.983	ng/u1	99
34) Caprolactam	11.630	113	102818	51.006	ng/u1#	54
35) 4-Chloro-3-methylphenol	11.983	107	650038	97.251	ng/u1	92
36) 2-Methylnaphthalene	12.354	142	1339768	83.426	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034560.D
 Acq On : 08 Apr 2022 12:39
 Operator : CG/JU
 Sample : SST08008
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST080015

Quant Time: Apr 08 14:09:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 13:18:10 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.571	142	1353799	82.535	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.724	216	693293	72.259	ng/ul#	97
40) Hexachlorocyclopentadiene	12.707	237	465167	67.958	ng/ul	96
41) 2,4,6-Trichlorophenol	12.966	196	456259	81.157	ng/ul	95
42) 2,4,5-Trichlorophenol	13.036	196	489191	80.289	ng/ul	100
43) 1,1'-Biphenyl	13.366	154	1764997	85.443	ng/ul#	97
44) 2-Chloronaphthalene	13.407	162	1354066	82.932	ng/ul	97
45) 2-Nitroaniline	13.607	65	456677	141.019	ng/ul#	83
47) Dimethylphthalate	13.989	163	1646431	83.347	ng/ul	99
48) 2,6-Dinitrotoluene	14.107	165	361992	93.983	ng/ul	93
50) Acenaphthylene	14.248	152	2234138	88.739	ng/ul	98
51) 3-Nitroaniline	14.430	138	301855	90.005	ng/ul	87
52) Acenaphthene	14.589	153	1453072	87.996	ng/ul	100
53) 2,4-Dinitrophenol	14.636	184	212192	84.731	ng/ul#	81
55) 4-Nitrophenol	14.742	109	263245	100.362	ng/ul#	83
56) Dibenzofuran	14.924	168	1983921	82.956	ng/ul	95
57) 2,4-Dinitrotoluene	14.889	165	510665	96.351	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.154	232	411497	78.994	ng/ul	90
59) Diethylphthalate	15.348	149	1706170	90.399	ng/ul	98
61) Fluorene	15.577	166	1698822	87.348	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.565	204	805424	76.381	ng/ul	91
63) 4-Nitroaniline	15.730	138	524	0.185	ng/ul	89
66) 4,6-Dinitro-2-methylph...	15.654	198	301859	87.827	ng/ul#	86
67) N-Nitrosodiphenylamine	15.783	169	1418755	88.265	ng/ul	98
68) 4-Bromophenyl-phenylether	16.459	248	486503	77.698	ng/ul#	91
69) Hexachlorobenzene	16.583	284	566964	72.591	ng/ul	93
70) Atrazine	16.736	200	521241	82.303	ng/ul	99
71) Pentachlorophenol	16.924	266	367508	78.228	ng/ul	98
72) Phenanthrene	17.312	178	2578471	87.140	ng/ul	99
74) Anthracene	17.406	178	2628319	89.501	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.330	216	715053	72.766	ng/uL	99
76) Pentachlorobenzene	14.854	250	665949	72.041	ng/uL	99
77) Carbazole	17.671	167	2227119	93.321	ng/ul	98
78) Di-n-butylphthalate	18.242	149	2864555	100.389	ng/ul	100
80) Fluoranthene	19.324	202	2980816	85.338	ng/ul#	88
82) Pyrene	19.689	202	3085951	85.135	ng/ul#	86
83) Butylbenzylphthalate	20.583	149	1344847	110.835	ng/ul#	84
84) 3,3'-Dichlorobenzidine	21.365	252	884532	79.126	ng/ul#	98
85) Benzo(a)anthracene	21.430	228	2832620	85.954	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.359	149	1977029	109.552	ng/ul#	97
87) Chrysene	21.483	228	2903923	81.849	ng/ul	99
89) Di-n-octyl phthalate	22.277	149	3295904	106.392	ng/ul	100
90) Benzo(b)fluoranthene	23.106	252	3069765	90.567	ng/ul#	97
91) Benzo(k)fluoranthene	23.153	252	2958538	85.636	ng/ul#	97
93) Benzo(a)pyrene	23.730	252	2985538	88.312	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.318	276	3336373	90.260	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.335	278	2824161	91.193	ng/ul#	96
96) Benzo(g,h,i)perylene	27.082	276	2951313	85.859	ng/ul#	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034560.D
 Acq On : 08 Apr 2022 12:39
 Operator : CG/JU
 Sample : SST080008
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST080015

Quant Time: Apr 08 14:09:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 13:18:10 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 39.381 ng/uL
RT: 3.290 min Scan# 43
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion: 88 Resp: 102704
Ion Ratio Lower Upper
88 100
43 32.0 21.0 31.6#
58 75.7 47.8 71.8#



#5
Pyridine
Concen: 108.387 ng/uL
RT: 3.696 min Scan# 112
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion: 79 Resp: 703928
Ion Ratio Lower Upper
79 100
52 76.7 47.7 71.5#
51 38.1 25.6 38.4









#6
Benzaldehyde
Concen: 73.977 ng/ul
RT: 7.007 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion: 77 Resp: 296036
Ion Ratio Lower Upper
77 100
105 92.3 80.8 121.2
106 88.4 76.6 115.0



#8
Phenol
Concen: 107.998 ng/ul
RT: 7.072 min Scan# 686
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion: 94 Resp: 819354
Ion Ratio Lower Upper
94 100
65 31.1 22.9 34.3
66 41.3 33.7 50.5









#10
Bis(2-Chloroethyl)ether
Concen: 101.764 ng/ul
RT: 7.301 min Scan# 725
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion: 93 Resp: 643636
Ion Ratio Lower Upper
93 100
63 81.6 54.7 82.1
95 32.3 26.6 39.8



#12
2-Chlorophenol
Concen: 97.843 ng/ul
RT: 7.442 min Scan# 749
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:128 Resp: 655977
Ion Ratio Lower Upper
128 100
64 50.7 30.6 46.0#
130 32.5 25.8 38.6









#13
2-Methylphenol
Concen: 100.253 ng/ul
RT: 8.325 min Scan# 899
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:108 Resp: 608652
Ion Ratio Lower Upper
108 100
107 90.9 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 161.018 ng/ul
RT: 8.419 min Scan# 915
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion: 45 Resp: 939111
Ion Ratio Lower Upper
45 100
77 13.9 17.1 25.7#
79 11.1 12.8 19.2#









#16
Acetophenone
Concen: 95.310 ng/ul
RT: 8.707 min Scan# 963
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:105 Resp: 984158
Ion Ratio Lower Upper
105 100
77 82.5 59.4 89.2
51 30.3 19.3 28.9#



#17
N-Nitroso-di-n-propylamine
Concen: 111.559 ng/ul
RT: 8.701 min Scan# 963
Delta R.T. 0.012 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion: 70 Resp: 542021
Ion Ratio Lower Upper
70 100
42 63.4 40.6 60.8#
101 9.7 8.2 12.2
130 19.6 19.9 29.9#









#18
4-Methylphenol
Concen: 99.256 ng/ul
RT: 8.660 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:108 Resp: 649651
Ion Ratio Lower Upper
108 100
107 114.1 89.0 133.4



#19
Hexachloroethane
Concen: 92.109 ng/ul
RT: 8.966 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:117 Resp: 278508
Ion Ratio Lower Upper
117 100
201 95.1 89.9 134.9
199 58.6 57.2 85.8









#22
Nitrobenzene
Concen: 106.800 ng/ul
RT: 9.089 min Scan# 1102
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion: 77 Resp: 778572
Ion Ratio Lower Upper
77 100
123 42.0 38.2 57.2
65 14.0 10.6 15.8



#23
Isophorone
Concen: 102.347 ng/ul
RT: 9.619 min Scan# 1119
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion: 82 Resp: 1410151
Ion Ratio Lower Upper
82 100
95 7.3 6.8 10.2
138 16.2 16.5 24.7#









#25
2-Nitrophenol
Concen: 93.451 ng/ul
RT: 9.801 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:139 Resp: 356404
Ion Ratio Lower Upper
139 100
65 59.8 37.8 56.6#
109 32.6 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 92.904 ng/ul
RT: 9.866 min Scan# 1161
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:107 Resp: 732075
Ion Ratio Lower Upper
107 100
121 49.3 40.8 61.2
122 82.8 69.0 103.6





#27

Bis(2-Chloroethoxy)methane

Concen: 98.196 ng/ul

RT: 10.101 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Instrument :

BNA_M

ClientSampleId :

SSTD080015





Tgt Ion: 93 Resp: 832297

Ion Ratio Lower Upper

93 100

95 32.2 25.4 38.2

123 10.7 9.5 14.3



#29

2,4-Dichlorophenol

Concen: 81.355 ng/ul

RT: 10.336 min Scan# 1241

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39




Tgt Ion:162 Resp: 586192


Ion Ratio Lower Upper

162 100

164 64.9 52.8 79.2

98 38.5 26.0 39.0









#30
Naphthalene
Concen: 85.738 ng/ul
RT: 10.742 min Scan# 1310
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:128 Resp: 2004939
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.2 9.9 14.9



#32
4-Chloroaniline
Concen: 89.307 ng/ul
RT: 10.848 min Scan# 1328
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:127 Resp: 824771
Ion Ratio Lower Upper
127 100
129 32.4 25.1 37.7









#33
Hexachlorobutadiene
Concen: 62.983 ng/ul
RT: 11.030 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:225 Resp: 374972
Ion Ratio Lower Upper
225 100
223 61.6 50.6 75.8
227 63.6 50.3 75.5



#34
Caprolactam
Concen: 51.006 ng/ul
RT: 11.630 min Scan# 1461
Delta R.T. 0.018 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:113 Resp: 102818
Ion Ratio Lower Upper
113 100
55 186.4 100.9 151.3#
56 139.0 81.0 121.4#









#35
4-Chloro-3-methylphenol
Concen: 97.251 ng/ul
RT: 11.983 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:107 Resp: 650038
Ion Ratio Lower Upper
107 100
144 25.7 23.0 34.4
142 79.5 69.8 104.8



#36
2-Methylnaphthalene
Concen: 83.426 ng/ul
RT: 12.354 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:142 Resp: 1339768
Ion Ratio Lower Upper
142 100
141 88.2 71.4 107.0





#37

1-Methylnaphthalene

Concen: 82.535 ng/ul

RT: 12.571 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Instrument :

BNA_M

ClientSampleId :

SSTD080015





Tgt Ion:142 Resp: 1353799

Ion Ratio Lower Upper

142 100

141 92.7 74.5 111.7

116 4.2 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 72.259 ng/ul

RT: 12.724 min Scan# 1647

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Tgt Ion:216 Resp: 693293




Ion Ratio Lower Upper


216 100

214 78.0 62.2 93.2

179 21.9 15.4 23.2

108 19.3 11.8 17.6#









#40
Hexachlorocyclopentadiene
Concen: 67.958 ng/ul
RT: 12.707 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:237 Resp: 465167
Ion Ratio Lower Upper
237 100
235 62.7 52.9 79.3
272 12.7 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 81.157 ng/ul
RT: 12.966 min Scan# 1688
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:196 Resp: 456259
Ion Ratio Lower Upper
196 100
198 95.2 80.8 121.2
200 31.1 26.5 39.7









#42
2,4,5-Trichlorophenol
Concen: 80.289 ng/ul
RT: 13.036 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:196 Resp: 489191
Ion Ratio Lower Upper
196 100
198 97.2 78.0 117.0
200 30.9 24.5 36.7



#43
1,1'-Biphenyl
Concen: 85.443 ng/ul
RT: 13.366 min Scan# 1756
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:154 Resp: 1764997
Ion Ratio Lower Upper
154 100
153 40.0 32.6 49.0
76 13.9 7.7 11.5#









#44
2-Chloronaphthalene
Concen: 82.932 ng/ul
RT: 13.407 min Scan# 1762
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:162 Resp: 1354066
Ion Ratio Lower Upper
162 100
164 32.6 26.6 40.0
127 39.9 29.6 44.4



#45
2-Nitroaniline
Concen: 141.019 ng/ul
RT: 13.607 min Scan# 1797
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion: 65 Resp: 456677
Ion Ratio Lower Upper
65 100
92 65.8 57.4 86.2
138 93.8 94.8 142.2#









#47
Dimethylphthalate
Concen: 83.347 ng/ul
RT: 13.989 min Scan# 1862
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:163 Resp: 1646431
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.0 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 93.983 ng/ul
RT: 14.107 min Scan# 1882
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:165 Resp: 361992
Ion Ratio Lower Upper
165 100
89 61.9 44.2 66.4
121 21.3 16.7 25.1





#50

Acenaphthylene

Concen: 88.739 ng/ul

RT: 14.248 min Scan# 1906

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Instrument :

BNA_M

ClientSampleId :

SSTD080015





Tgt Ion:152 Resp: 2234138

Ion Ratio Lower Upper

152 100

151 20.5 15.8 23.8

153 13.4 10.2 15.2



#51

3-Nitroaniline

Concen: 90.005 ng/ul

RT: 14.430 min Scan# 1937

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39




Tgt Ion:138 Resp: 301855


Ion Ratio Lower Upper

138 100

108 10.6 8.3 12.5

92 124.0 87.0 130.4









#52
Acenaphthene
Concen: 87.996 ng/ul
RT: 14.589 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:153 Resp: 1453072
Ion Ratio Lower Upper
153 100
152 46.4 37.4 56.0
154 88.7 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 84.731 ng/ul
RT: 14.636 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:184 Resp: 212192
Ion Ratio Lower Upper
184 100
63 81.9 43.8 65.6#
154 64.7 49.9 74.9









#55
4-Nitrophenol
Concen: 100.362 ng/ul
RT: 14.742 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:109 Resp: 263245
Ion Ratio Lower Upper
109 100
139 121.4 94.6 142.0
65 137.5 83.5 125.3#



#56
Dibenzofuran
Concen: 82.956 ng/ul
RT: 14.924 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:168 Resp: 1983921
Ion Ratio Lower Upper
168 100
139 40.2 29.8 44.6





#57

2,4-Dinitrotoluene

Concen: 96.351 ng/ul

RT: 14.889 min Scan# 2015

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Instrument :

BNA_M

ClientSampleId :

SSTD080015





Tgt Ion:165 Resp: 510665

Ion Ratio Lower Upper

165 100

63 51.2 30.2 45.4#

182 3.0 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 78.994 ng/ul

RT: 15.154 min Scan# 2060

Delta R.T. 0.006 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Tgt Ion:232 Resp: 411497




Ion Ratio Lower Upper


232 100

131 48.0 32.2 48.4

130 2.7 1.9 2.9

166 32.7 23.0 34.4









#59
Diethylphthalate
Concen: 90.399 ng/ul
RT: 15.348 min Scan# 2092
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:149 Resp: 1706170
Ion Ratio Lower Upper
149 100
177 20.8 17.9 26.9
150 12.4 10.2 15.2



#61
Fluorene
Concen: 87.348 ng/ul
RT: 15.577 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:166 Resp: 1698822
Ion Ratio Lower Upper
166 100
165 96.9 76.9 115.3
167 13.2 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 76.381 ng/ul
RT: 15.565 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:204 Resp: 805424
Ion Ratio Lower Upper
204 100
206 32.6 25.5 38.3
141 64.8 44.2 66.2



#63
4-Nitroaniline
Concen: 0.185 ng/ul
RT: 15.730 min Scan# 2158
Delta R.T. 0.141 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:138 Resp: 524
Ion Ratio Lower Upper
138 100
92 54.1 36.3 54.5
108 95.6 68.9 103.3









#66
4,6-Dinitro-2-methylphenol
Concen: 87.827 ng/ul
RT: 15.654 min Scan# 2145
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:198 Resp: 301859
Ion Ratio Lower Upper
198 100
182 3.8 3.8 5.6
77 31.1 18.5 27.7#



#67
N-Nitrosodiphenylamine
Concen: 88.265 ng/ul
RT: 15.783 min Scan# 2167
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:169 Resp: 1418755
Ion Ratio Lower Upper
169 100
168 67.7 52.8 79.2
167 36.3 28.1 42.1









#68
4-Bromophenyl-phenylether
Concen: 77.698 ng/ul
RT: 16.459 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:248 Resp: 486503
Ion Ratio Lower Upper
248 100
250 98.8 78.4 117.6
141 80.4 51.4 77.2#



#69
Hexachlorobenzene
Concen: 72.591 ng/ul
RT: 16.583 min Scan# 2303
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:284 Resp: 566964
Ion Ratio Lower Upper
284 100
142 36.4 24.6 36.8
249 31.6 24.1 36.1









#70
Atrazine
Concen: 82.303 ng/ul
RT: 16.736 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:200 Resp: 521241
Ion Ratio Lower Upper
200 100
173 26.2 20.2 30.4
215 47.5 38.4 57.6



#71
Pentachlorophenol
Concen: 78.228 ng/ul
RT: 16.924 min Scan# 2361
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:266 Resp: 367508
Ion Ratio Lower Upper
266 100
264 62.6 49.3 73.9
268 64.9 50.3 75.5









#72
 Phenanthrene
 Concen: 87.140 ng/u1
 RT: 17.312 min Scan# 2426
 Delta R.T. 0.006 min
 Lab File: BM034560.D
 Acq: 08 Apr 2022 12:39




Instrument :
 BNA_M
 ClientSampleId :
 SSTD080015


Tgt Ion:178 Resp: 2578471
 Ion Ratio Lower Upper
 178 100
 179 15.3 11.7 17.5
 176 18.8 15.0 22.4



#74
 Anthracene
 Concen: 89.501 ng/u1
 RT: 17.406 min Scan# 2443
 Delta R.T. 0.006 min
 Lab File: BM034560.D
 Acq: 08 Apr 2022 12:39

Tgt Ion:178 Resp: 2628319
 Ion Ratio Lower Upper
 178 100
 179 15.4 11.9 17.9
 176 18.4 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 72.766 ng/uL

RT: 13.330 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Instrument :

BNA_M

ClientSampleId :

SSTD080015





Tgt Ion:216 Resp: 715053

Ion Ratio Lower Upper

216 100

214 78.5 62.0 93.0

218 47.6 37.8 56.8



#76

Pentachlorobenzene

Concen: 72.041 ng/uL

RT: 14.854 min Scan# 2009

Delta R.T. 0.006 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39




Tgt Ion:250 Resp: 665949


Ion Ratio Lower Upper

250 100

248 63.5 50.6 75.8

252 63.5 50.2 75.4









#77
 Carbazole
 Concen: 93.321 ng/ul
 RT: 17.671 min Scan# 2488
 Delta R.T. -0.000 min
 Lab File: BM034560.D
 Acq: 08 Apr 2022 12:39




Instrument :
 BNA_M
 ClientSampleId :
 SST080015


Tgt Ion:167 Resp: 2227119
 Ion Ratio Lower Upper
 167 100
 166 21.6 16.6 25.0
 139 13.5 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 100.389 ng/ul
 RT: 18.242 min Scan# 2585
 Delta R.T. 0.006 min
 Lab File: BM034560.D
 Acq: 08 Apr 2022 12:39

Tgt Ion:149 Resp: 2864555
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.0
 104 5.6 4.2 6.2









#80
Fluoranthene
Concen: 85.338 ng/ul
RT: 19.324 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:202 Resp: 2980816
Ion Ratio Lower Upper
202 100
101 12.1 6.2 9.2#
100 9.7 4.7 7.1#



#82
Pyrene
Concen: 85.135 ng/ul
RT: 19.689 min Scan# 2831
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:202 Resp: 3085951
Ion Ratio Lower Upper
202 100
101 13.8 7.0 10.4#
100 11.9 5.8 8.6#





#83

Butylbenzylphthalate

Concen: 110.835 ng/ul

RT: 20.583 min Scan# 2983

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Instrument :

BNA_M

ClientSampleId :

SSTD080015





Tgt Ion:149 Resp: 1344847

Ion Ratio Lower Upper

149 100

91 78.2 50.9 76.3#

206 19.3 19.3 28.9#



#84

3,3'-Dichlorobenzidine

Concen: 79.126 ng/ul

RT: 21.365 min Scan# 3116

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39




Tgt Ion:252 Resp: 884532


Ion Ratio Lower Upper

252 100

254 64.7 50.7 76.1

126 11.1 6.6 9.8#









#85
Benzo(a)anthracene
Concen: 85.954 ng/ul
RT: 21.430 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:228 Resp: 2832620
Ion Ratio Lower Upper
228 100
229 20.0 15.4 23.2
226 27.0 21.0 31.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 109.552 ng/ul
RT: 21.359 min Scan# 3115
Delta R.T. -0.000 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:149 Resp: 1977029
Ion Ratio Lower Upper
149 100
167 26.4 22.1 33.1
279 3.9 4.2 6.2#





#87

Chrysene

Concen: 81.849 ng/ul

RT: 21.483 min Scan# 3136

Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Instrument :

BNA_M

ClientSampleId :

SSTD080015





Tgt Ion:228 Resp: 2903923

Ion Ratio Lower Upper

228 100

226 29.5 23.4 35.2

229 19.8 15.4 23.0



#89

Di-n-octyl phthalate

Concen: 106.392 ng/ul




RT: 22.277 min Scan# 3271


Delta R.T. -0.000 min

Lab File: BM034560.D

Acq: 08 Apr 2022 12:39

Tgt Ion:149 Resp: 3295904









#90
Benzo(b)fluoranthene
Concen: 90.567 ng/ul
RT: 23.106 min Scan# 3411
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:252 Resp: 3069765
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 10.4 5.8 8.6#



#91
Benzo(k)fluoranthene
Concen: 85.636 ng/ul
RT: 23.153 min Scan# 3420
Delta R.T. 0.006 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:252 Resp: 2958538
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 10.3 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 88.312 ng/ul
RT: 23.730 min Scan# 3516
Delta R.T. 0.012 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39




Instrument :
BNA_M
ClientSampleId :
SSTD080015


Tgt Ion:252 Resp: 2985538
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 11.3 6.2 9.4#



#94
Indeno(1,2,3-cd)pyrene
Concen: 90.260 ng/ul
RT: 26.318 min Scan# 3958
Delta R.T. 0.012 min
Lab File: BM034560.D
Acq: 08 Apr 2022 12:39

Tgt Ion:276 Resp: 3336373
Ion Ratio Lower Upper
276 100
138 23.8 13.6 20.4#
277 25.2 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 91.193 ng/ul
 RT: 26.335 min Scan# 3958
 Delta R.T. 0.018 min
 Lab File: BM034560.D
 Acq: 08 Apr 2022 12:39

Instrument :
 BNA_M
 ClientSampleId :
 SSTD080015

Tgt Ion:278 Resp: 2824161
 Ion Ratio Lower Upper
 278 100
 139 15.6 9.6 14.4#
 279 23.3 19.3 28.9



#96
 Benzo(g,h,i)perylene
 Concen: 85.859 ng/ul
 RT: 27.082 min Scan# 4088
 Delta R.T. 0.018 min
 Lab File: BM034560.D
 Acq: 08 Apr 2022 12:39

Tgt Ion:276 Resp: 2951313
 Ion Ratio Lower Upper
 276 100
 138 20.8 11.8 17.6#
 277 23.5 18.8 28.2

