

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034562.D
 Acq On : 08 Apr 2022 15:27
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV017

Quant Time: Apr 08 18:20:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.878	152	109705	20.000	ng/u1	0.00
20) Naphthalene-d8	10.684	136	439432	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.524	164	255352	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.265	188	501454	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.442	240	488682	20.000	ng/u1	# 0.00
88) Perylene-d12	23.824	264	535704	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.255	96	23937	7.781	ng/uL	0.00
4) Pyridine-d5	3.678	84	142363	19.175	ng/u1	0.00
7) Phenol-d5	7.037	99	173070	18.954	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.201	67	120157	19.748	ng/u1	0.00
11) 2-Chlorophenol-d4	7.407	132	141953	20.111	ng/u1	0.00
15) 4-Methylphenol-d8	8.590	113	136268	19.105	ng/u1	0.00
21) Nitrobenzene-d5	9.037	128	68406	21.181	ng/u1	0.00
24) 2-Nitrophenol-d4	9.766	143	69786	20.755	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.307	165	130435	20.601	ng/u1	0.00
31) 4-Chloroaniline-d4	10.825	131	165176	18.910	ng/u1	0.00
46) Dimethylphthalate-d6	13.936	166	372396	20.183	ng/u1	0.00
49) Acenaphthylene-d8	14.219	160	482596	20.350	ng/u1	0.00
54) 4-Nitrophenol-d4	14.719	143	48997	17.416	ng/u1	0.00
60) Fluorene-d10	15.513	176	321781	20.181	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.636	200	56233	18.369	ng/u1	0.00
73) Anthracene-d10	17.365	188	480993	20.119	ng/u1	0.00
81) Pyrene-d10	19.659	212	545663	20.458	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.671	264	530306	19.679	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.290	88	24773	7.997	ng/uL#	85
5) Pyridine	3.696	79	154569	19.053	ng/u1#	74
6) Benzaldehyde	7.007	77	117242	20.944	ng/u1	88
8) Phenol	7.066	94	174963	18.968	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.295	93	147343	19.793	ng/u1#	86
12) 2-Chlorophenol	7.437	128	147801	20.542	ng/u1#	88
13) 2-Methylphenol	8.325	108	131527	19.288	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.413	45	224137	20.138	ng/u1#	85
16) Acetophenone	8.701	105	218043	19.585	ng/u1#	89
17) N-Nitroso-di-n-propyla...	8.690	70	120344	20.308	ng/u1#	88
18) 4-Methylphenol	8.654	108	140574	19.378	ng/u1	99
19) Hexachloroethane	8.960	117	64783	20.018	ng/u1	81
22) Nitrobenzene	9.084	77	170538	20.143	ng/u1	92
23) Isophorone	9.613	82	303192	20.166	ng/u1#	93
25) 2-Nitrophenol	9.795	139	74548	20.700	ng/u1#	88
26) 2,4-Dimethylphenol	9.860	107	162805	20.366	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.095	93	185135	20.466	ng/u1	98
29) 2,4-Dichlorophenol	10.336	162	126166	20.190	ng/u1	97
30) Naphthalene	10.736	128	458071	20.023	ng/u1	99
32) 4-Chloroaniline	10.848	127	161374	18.441	ng/u1	100
33) Hexachlorobutadiene	11.031	225	88011	20.033	ng/u1	99
34) Caprolactam	11.760	113	319	0.157	ng/u1	86
35) 4-Chloro-3-methylphenol	11.978	107	133985	20.655	ng/u1	92
36) 2-Methylnaphthalene	12.354	142	293731	20.033	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034562.D
 Acq On : 08 Apr 2022 15:27
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV017

Quant Time: Apr 08 18:20:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.572	142	302754	20.127	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.725	216	154722	20.243	ng/ul#	97
40) Hexachlorocyclopentadiene	12.707	237	94661	18.873	ng/ul	96
41) 2,4,6-Trichlorophenol	12.960	196	93970	20.239	ng/ul	95
42) 2,4,5-Trichlorophenol	13.036	196	99022	20.581	ng/ul	99
43) 1,1'-Biphenyl	13.360	154	386350	19.973	ng/ul#	98
44) 2-Chloronaphthalene	13.401	162	300184	20.214	ng/ul	95
45) 2-Nitroaniline	13.607	65	83029	20.504	ng/ul	84
47) Dimethylphthalate	13.983	163	365569	20.133	ng/ul	99
48) 2,6-Dinitrotoluene	14.101	165	73377	21.035	ng/ul	90
50) Acenaphthylene	14.248	152	493985	20.518	ng/ul	99
51) 3-Nitroaniline	14.430	138	57767	18.941	ng/ul	83
52) Acenaphthene	14.589	153	324653	20.254	ng/ul	100
53) 2,4-Dinitrophenol	14.636	184	29952	15.236	ng/ul#	80
55) 4-Nitrophenol	14.736	109	40592	16.896	ng/ul#	75
56) Dibenzofuran	14.924	168	442136	20.295	ng/ul	97
57) 2,4-Dinitrotoluene	14.889	165	98821	20.354	ng/ul#	82
58) 2,3,4,6-Tetrachlorophenol	15.148	232	86228	20.412	ng/ul#	87
59) Diethylphthalate	15.342	149	375314	20.349	ng/ul	99
61) Fluorene	15.571	166	362541	19.724	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.566	204	176285	19.870	ng/ul	92
63) 4-Nitroaniline	15.748	138	261	0.097	ng/ul	87
66) 4,6-Dinitro-2-methylph...	15.648	198	58122	19.042	ng/ul#	86
67) N-Nitrosodiphenylamine	15.777	169	308053	20.619	ng/ul	99
68) 4-Bromophenyl-phenylether	16.530	248	472	0.091	ng/ul	93
69) Hexachlorobenzene	16.577	284	125533	20.165	ng/ul#	91
70) Atrazine	16.730	200	109783	19.636	ng/ul	99
71) Pentachlorophenol	16.918	266	70873	18.593	ng/ul	98
72) Phenanthrene	17.307	178	562666	20.109	ng/ul	99
74) Anthracene	17.401	178	561098	20.237	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.325	216	160423	20.369	ng/ul	98
76) Pentachlorobenzene	14.848	250	148875	20.143	ng/ul	99
77) Carbazole	17.671	167	424704	18.516	ng/ul	98
78) Di-n-butylphthalate	18.236	149	596759	20.498	ng/ul	99
80) Fluoranthene	19.324	202	619821	20.999	ng/ul#	89
82) Pyrene	19.689	202	675948	21.069	ng/ul#	89
83) Butylbenzylphthalate	20.583	149	267805	21.322	ng/ul	86
84) 3,3'-Dichlorobenzidine	21.365	252	177554	20.731	ng/ul#	98
85) Benzo(a)anthracene	21.430	228	581306	20.458	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.359	149	401910	21.381	ng/ul#	95
87) Chrysene	21.483	228	631325	20.605	ng/ul	100
89) Di-n-octyl phthalate	22.277	149	687180	19.268	ng/ul	100
90) Benzo(b)fluoranthene	23.100	252	630966	19.822	ng/ul#	97
91) Benzo(k)fluoranthene	23.147	252	661240	19.781	ng/ul#	97
93) Benzo(a)pyrene	23.718	252	623086	19.805	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.300	276	678738	20.160	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.318	278	525761	19.048	ng/ul#	97
96) Benzo(g,h,i)perylene	27.065	276	625096	19.343	ng/ul#	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034562.D
 Acq On : 08 Apr 2022 15:27
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV017

Quant Time: Apr 08 18:20:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.997 ng/uL
RT: 3.290 min Scan# 41
Delta R.T. -0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion: 88 Resp: 24773
Ion Ratio Lower Upper
88 100
43 31.8 21.0 31.6#
58 72.9 47.8 71.8#



#5
Pyridine
Concen: 19.053 ng/uL
RT: 3.696 min Scan# 112
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion: 79 Resp: 154569
Ion Ratio Lower Upper
79 100
52 82.8 47.7 71.5#
51 42.4 25.6 38.4#









#6
Benzaldehyde
Concen: 20.944 ng/ul
RT: 7.007 min Scan# 675
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion: 77 Resp: 117242
Ion Ratio Lower Upper
77 100
105 89.5 80.8 121.2
106 84.2 76.6 115.0



#8
Phenol
Concen: 18.968 ng/ul
RT: 7.066 min Scan# 685
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion: 94 Resp: 174963
Ion Ratio Lower Upper
94 100
65 29.9 22.9 34.3
66 40.6 33.7 50.5









#10
Bis(2-Chloroethyl)ether
Concen: 19.793 ng/ul
RT: 7.295 min Scan# 724
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion: 93 Resp: 147343
Ion Ratio Lower Upper
93 100
63 84.0 54.7 82.1#
95 32.2 26.6 39.8



#12
2-Chlorophenol
Concen: 20.542 ng/ul
RT: 7.437 min Scan# 748
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:128 Resp: 147801
Ion Ratio Lower Upper
128 100
64 51.8 30.6 46.0#
130 32.0 25.8 38.6





#13

2-Methylphenol

Concen: 19.288 ng/ul

RT: 8.325 min Scan# 899

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Instrument :

BNA_M

ClientSampleId :





SICV017

Tgt Ion:108 Resp: 131527

Ion Ratio Lower Upper

108 100

107 90.9 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.138 ng/ul

RT: 8.413 min Scan# 914

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27




Tgt Ion: 45 Resp: 224137


Ion Ratio Lower Upper

45 100

77 13.8 17.1 25.7#

79 10.7 12.8 19.2#









#16
Acetophenone
Concen: 19.585 ng/ul
RT: 8.701 min Scan# 963
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:105 Resp: 218043
Ion Ratio Lower Upper
105 100
77 82.1 59.4 89.2
51 32.0 19.3 28.9#



#17
N-Nitroso-di-n-propylamine
Concen: 20.308 ng/ul
RT: 8.690 min Scan# 961
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion: 70 Resp: 120344
Ion Ratio Lower Upper
70 100
42 60.1 40.6 60.8
101 8.6 8.2 12.2
130 19.3 19.9 29.9#









#18
4-Methylphenol
Concen: 19.378 ng/ul
RT: 8.654 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:108 Resp: 140574
Ion Ratio Lower Upper
108 100
107 111.8 89.0 133.4



#19
Hexachloroethane
Concen: 20.018 ng/ul
RT: 8.960 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:117 Resp: 64783
Ion Ratio Lower Upper
117 100
201 90.5 89.9 134.9
199 58.3 57.2 85.8









#22
Nitrobenzene
Concen: 20.143 ng/ul
RT: 9.084 min Scan# 1028
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion: 77 Resp: 170538
Ion Ratio Lower Upper
77 100
123 41.9 38.2 57.2
65 14.8 10.6 15.8



#23
Isophorone
Concen: 20.166 ng/ul
RT: 9.613 min Scan# 1118
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion: 82 Resp: 303192
Ion Ratio Lower Upper
82 100
95 7.4 6.8 10.2
138 16.5 16.5 24.7#









#25
2-Nitrophenol
Concen: 20.700 ng/ul
RT: 9.795 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:139 Resp: 74548
Ion Ratio Lower Upper
139 100
65 61.1 37.8 56.6#
109 33.2 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 20.366 ng/ul
RT: 9.860 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:107 Resp: 162805
Ion Ratio Lower Upper
107 100
121 48.4 40.8 61.2
122 80.9 69.0 103.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.466 ng/ul
 RT: 10.095 min Scan# 1200
 Delta R.T. 0.000 min
 Lab File: BM034562.D
 Acq: 08 Apr 2022 15:27




Instrument :
 BNA_M
 ClientSampleId :
 SICV017


Tgt Ion: 93 Resp: 185135
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.4 38.2
 123 10.8 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 20.190 ng/ul
 RT: 10.336 min Scan# 1241
 Delta R.T. 0.000 min
 Lab File: BM034562.D
 Acq: 08 Apr 2022 15:27

Tgt Ion:162 Resp: 126166
 Ion Ratio Lower Upper
 162 100
 164 65.5 52.8 79.2
 98 37.1 26.0 39.0









#30
Naphthalene
Concen: 20.023 ng/ul
RT: 10.736 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:128 Resp: 458071
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.4 9.9 14.9



#32
4-Chloroaniline
Concen: 18.441 ng/ul
RT: 10.848 min Scan# 1328
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:127 Resp: 161374
Ion Ratio Lower Upper
127 100
129 31.3 25.1 37.7





#33

Hexachlorobutadiene

Concen: 20.033 ng/u1

RT: 11.031 min Scan# 11

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV017





Tgt Ion:225 Resp: 88011

Ion Ratio Lower Upper

225 100

223 63.5 50.6 75.8

227 64.8 50.3 75.5



#34

Caprolactam

Concen: 0.157 ng/u1

RT: 11.760 min Scan# 1483

Delta R.T. 0.147 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27




Tgt Ion:113 Resp: 319


Ion Ratio Lower Upper

113 100

55 108.5 100.9 151.3

56 88.0 81.0 121.4









#35
4-Chloro-3-methylphenol
Concen: 20.655 ng/ul
RT: 11.978 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:107 Resp: 133985
Ion Ratio Lower Upper
107 100
144 25.3 23.0 34.4
142 79.4 69.8 104.8



#36
2-Methylnaphthalene
Concen: 20.033 ng/ul
RT: 12.354 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:142 Resp: 293731
Ion Ratio Lower Upper
142 100
141 89.0 71.4 107.0





#37

1-Methylnaphthalene

Concen: 20.127 ng/ul

RT: 12.572 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV017





Tgt Ion:142 Resp: 302754

Ion Ratio Lower Upper

142 100

141 92.9 74.5 111.7

116 4.2 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.243 ng/ul

RT: 12.725 min Scan# 1647

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Tgt Ion:216 Resp: 154722




Ion Ratio Lower Upper


216 100

214 79.4 62.2 93.2

179 21.7 15.4 23.2

108 18.2 11.8 17.6#









#40
Hexachlorocyclopentadiene
Concen: 18.873 ng/ul
RT: 12.707 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:237 Resp: 94661
Ion Ratio Lower Upper
237 100
235 63.5 52.9 79.3
272 12.9 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 20.239 ng/ul
RT: 12.960 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:196 Resp: 93970
Ion Ratio Lower Upper
196 100
198 96.4 80.8 121.2
200 29.6 26.5 39.7









#42
2,4,5-Trichlorophenol
Concen: 20.581 ng/ul
RT: 13.036 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:196 Resp: 99022
Ion Ratio Lower Upper
196 100
198 96.9 78.0 117.0
200 31.1 24.5 36.7



#43
1,1'-Biphenyl
Concen: 19.973 ng/ul
RT: 13.360 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:154 Resp: 386350
Ion Ratio Lower Upper
154 100
153 40.6 32.6 49.0
76 13.9 7.7 11.5#









#44
2-Chloronaphthalene
Concen: 20.214 ng/ul
RT: 13.401 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:162 Resp: 300184
Ion Ratio Lower Upper
162 100
164 31.2 26.6 40.0
127 40.4 29.6 44.4



#45
2-Nitroaniline
Concen: 20.504 ng/ul
RT: 13.607 min Scan# 1797
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion: 65 Resp: 83029
Ion Ratio Lower Upper
65 100
92 65.9 57.4 86.2
138 94.9 94.8 142.2









#47
Dimethylphthalate
Concen: 20.133 ng/ul
RT: 13.983 min Scan# 1861
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:163 Resp: 365569
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.2
164 10.1 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 21.035 ng/ul
RT: 14.101 min Scan# 1881
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:165 Resp: 73377
Ion Ratio Lower Upper
165 100
89 64.2 44.2 66.4
121 22.5 16.7 25.1









#50
Acenaphthylene
Concen: 20.518 ng/ul
RT: 14.248 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:152 Resp: 493985
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.0 10.2 15.2



#51
3-Nitroaniline
Concen: 18.941 ng/ul
RT: 14.430 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:138 Resp: 57767
Ion Ratio Lower Upper
138 100
108 11.3 8.3 12.5
92 128.4 87.0 130.4





#52

Acenaphthene

Concen: 20.254 ng/ul

RT: 14.589 min Scan# 1964

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV017





Tgt Ion:153 Resp: 324653

Ion Ratio Lower Upper

153 100

152 47.0 37.4 56.0

154 88.4 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 15.236 ng/ul

RT: 14.636 min Scan# 1972

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27




Tgt Ion:184 Resp: 29952


Ion Ratio Lower Upper

184 100

63 84.5 43.8 65.6#

154 63.8 49.9 74.9









#55
4-Nitrophenol
Concen: 16.896 ng/ul
RT: 14.736 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:109 Resp: 40592
Ion Ratio Lower Upper
109 100
139 126.6 94.6 142.0
65 150.0 83.5 125.3#



#56
Dibenzofuran
Concen: 20.295 ng/ul
RT: 14.924 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:168 Resp: 442136
Ion Ratio Lower Upper
168 100
139 39.2 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 20.354 ng/ul
RT: 14.889 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:165 Resp: 98821
Ion Ratio Lower Upper
165 100
63 49.4 30.2 45.4#
182 2.8 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.412 ng/ul
RT: 15.148 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:232 Resp: 86228
Ion Ratio Lower Upper
232 100
131 50.2 32.2 48.4#
130 3.1 1.9 2.9#
166 33.3 23.0 34.4





#59

Diethylphthalate

Concen: 20.349 ng/ul

RT: 15.342 min Scan# 2092

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV017





Tgt Ion:149 Resp: 375314

Ion Ratio Lower Upper

149 100

177 21.4 17.9 26.9

150 12.9 10.2 15.2



#61

Fluorene

Concen: 19.724 ng/ul

RT: 15.571 min Scan# 2131

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27




Tgt Ion:166 Resp: 362541


Ion Ratio Lower Upper

166 100

165 98.4 76.9 115.3

167 13.6 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 19.870 ng/ul
RT: 15.566 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:204 Resp: 176285
Ion Ratio Lower Upper
204 100
206 33.1 25.5 38.3
141 62.9 44.2 66.2



#63
4-Nitroaniline
Concen: 0.097 ng/ul
RT: 15.748 min Scan# 2161
Delta R.T. 0.159 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:138 Resp: 261
Ion Ratio Lower Upper
138 100
92 51.9 36.3 54.5
108 72.4 68.9 103.3









#66
4,6-Dinitro-2-methylphenol
Concen: 19.042 ng/ul
RT: 15.648 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:198 Resp: 58122
Ion Ratio Lower Upper
198 100
182 6.3 3.8 5.6#
77 30.8 18.5 27.7#



#67
N-Nitrosodiphenylamine
Concen: 20.619 ng/ul
RT: 15.777 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:169 Resp: 308053
Ion Ratio Lower Upper
169 100
168 64.6 52.8 79.2
167 34.9 28.1 42.1









#68
4-Bromophenyl-phenylether
Concen: 0.091 ng/ul
RT: 16.530 min Scan# 21
Delta R.T. 0.071 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:248 Resp: 472
Ion Ratio Lower Upper
248 100
250 88.0 78.4 117.6
141 62.3 51.4 77.2



#69
Hexachlorobenzene
Concen: 20.165 ng/ul
RT: 16.577 min Scan# 2302
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:284 Resp: 125533
Ion Ratio Lower Upper
284 100
142 37.4 24.6 36.8#
249 32.8 24.1 36.1





#70

Atrazine

Concen: 19.636 ng/ul

RT: 16.730 min Scan# 2328

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV017





Tgt Ion:200 Resp: 109783

Ion Ratio Lower Upper

200 100

173 25.2 20.2 30.4

215 47.2 38.4 57.6



#71

Pentachlorophenol

Concen: 18.593 ng/ul

RT: 16.918 min Scan# 2360

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27




Tgt Ion:266 Resp: 70873


Ion Ratio Lower Upper

266 100

264 62.6 49.3 73.9

268 64.3 50.3 75.5









#72
Phenanthrene
Concen: 20.109 ng/u1
RT: 17.307 min Scan# 2426
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:178 Resp: 562666
Ion Ratio Lower Upper
178 100
179 15.1 11.7 17.5
176 18.8 15.0 22.4



#74
Anthracene
Concen: 20.237 ng/u1
RT: 17.401 min Scan# 2442
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:178 Resp: 561098
Ion Ratio Lower Upper
178 100
179 14.8 11.9 17.9
176 18.1 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.369 ng/uL
RT: 13.325 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:216 Resp: 160423
Ion Ratio Lower Upper
216 100
214 79.7 62.0 93.0
218 48.1 37.8 56.8



#76
Pentachlorobenzene
Concen: 20.143 ng/uL
RT: 14.848 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:250 Resp: 148875
Ion Ratio Lower Upper
250 100
248 62.0 50.6 75.8
252 63.3 50.2 75.4









#77
 Carbazole
 Concen: 18.516 ng/ul
 RT: 17.671 min Scan# 2488
 Delta R.T. 0.000 min
 Lab File: BM034562.D
 Acq: 08 Apr 2022 15:27




Instrument :
 BNA_M
 ClientSampleId :
 SICV017


Tgt Ion:167 Resp: 424704
 Ion Ratio Lower Upper
 167 100
 166 21.7 16.6 25.0
 139 13.3 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 20.498 ng/ul
 RT: 18.236 min Scan# 2584
 Delta R.T. 0.000 min
 Lab File: BM034562.D
 Acq: 08 Apr 2022 15:27

Tgt Ion:149 Resp: 596759
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.0
 104 5.9 4.2 6.2









#80
Fluoranthene
Concen: 20.999 ng/u1
RT: 19.324 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:202 Resp: 619821
Ion Ratio Lower Upper
202 100
101 11.9 6.2 9.2#
100 9.2 4.7 7.1#



#82
Pyrene
Concen: 21.069 ng/u1
RT: 19.689 min Scan# 2831
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:202 Resp: 675948
Ion Ratio Lower Upper
202 100
101 12.9 7.0 10.4#
100 11.0 5.8 8.6#









#83
Butylbenzylphthalate
Concen: 21.322 ng/ul
RT: 20.583 min Scan# 2983
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:149 Resp: 267805
Ion Ratio Lower Upper
149 100
91 75.7 50.9 76.3
206 19.8 19.3 28.9



#84
3,3'-Dichlorobenzidine
Concen: 20.731 ng/ul
RT: 21.365 min Scan# 3116
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:252 Resp: 177554
Ion Ratio Lower Upper
252 100
254 64.2 50.7 76.1
126 11.7 6.6 9.8









#85
Benzo(a)anthracene
Concen: 20.458 ng/ul
RT: 21.430 min Scan# 3127
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:228 Resp: 581306
Ion Ratio Lower Upper
228 100
229 19.3 15.4 23.2
226 26.5 21.0 31.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.381 ng/ul
RT: 21.359 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:149 Resp: 401910
Ion Ratio Lower Upper
149 100
167 25.2 22.1 33.1
279 3.7 4.2 6.2#





#87

Chrysene

Concen: 20.605 ng/u1

RT: 21.483 min Scan# 3136

Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Instrument :

BNA_M

ClientSampleId :

SICV017





Tgt Ion:228 Resp: 631325

Ion Ratio Lower Upper

228 100

226 29.3 23.4 35.2

229 19.7 15.4 23.0



#89

Di-n-octyl phthalate

Concen: 19.268 ng/u1




RT: 22.277 min Scan# 3271


Delta R.T. 0.000 min

Lab File: BM034562.D

Acq: 08 Apr 2022 15:27

Tgt Ion:149 Resp: 687180









#90
Benzo(b)fluoranthene
Concen: 19.822 ng/ul
RT: 23.100 min Scan# 3411
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:252 Resp: 630966
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 10.3 5.8 8.6#



#91
Benzo(k)fluoranthene
Concen: 19.781 ng/ul
RT: 23.147 min Scan# 3419
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:252 Resp: 661240
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 10.2 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 19.805 ng/ul
RT: 23.718 min Scan# 3516
Delta R.T. 0.000 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27




Instrument :
BNA_M
ClientSampleId :
SICV017


Tgt Ion:252 Resp: 623086
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 10.9 6.2 9.4#



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.160 ng/ul
RT: 26.300 min Scan# 3955
Delta R.T. -0.006 min
Lab File: BM034562.D
Acq: 08 Apr 2022 15:27

Tgt Ion:276 Resp: 678738
Ion Ratio Lower Upper
276 100
138 23.0 13.6 20.4#
277 24.8 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 19.048 ng/ul
 RT: 26.318 min Scan# 3958
 Delta R.T. 0.000 min
 Lab File: BM034562.D
 Acq: 08 Apr 2022 15:27

Instrument :
 BNA_M
 ClientSampleId :
 SICV017

Tgt Ion:278 Resp: 525761
 Ion Ratio Lower Upper
 278 100
 139 15.3 9.6 14.4#
 279 23.7 19.3 28.9



#96
 Benzo(g,h,i)perylene
 Concen: 19.343 ng/ul
 RT: 27.065 min Scan# 4085
 Delta R.T. 0.000 min
 Lab File: BM034562.D
 Acq: 08 Apr 2022 15:27

Tgt Ion:276 Resp: 625096
 Ion Ratio Lower Upper
 276 100
 138 20.8 11.8 17.6#
 277 24.3 18.8 28.2

