

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034584.D
 Acq On : 09 Apr 2022 05:27
 Operator : CG/JU
 Sample : PB143966BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS966

Quant Time: Apr 09 06:27:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.878	152	99822	20.000	ng/u1	0.00
20) Naphthalene-d8	10.683	136	407354	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.524	164	239440	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.265	188	480069	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.442	240	469015	20.000	ng/u1	# 0.00
88) Perylene-d12	23.824	264	479992	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.255	96	22965	8.204	ng/uL	0.00
4) Pyridine-d5	3.678	84	283632	41.984	ng/u1	0.00
7) Phenol-d5	7.037	99	355432	42.780	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.201	67	241466	43.615	ng/u1	0.00
11) 2-Chlorophenol-d4	7.407	132	291195	45.339	ng/u1	0.00
15) 4-Methylphenol-d8	8.589	113	286748	44.183	ng/u1	0.00
21) Nitrobenzene-d5	9.037	128	139344	46.544	ng/u1	0.00
24) 2-Nitrophenol-d4	9.766	143	147972	47.473	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.307	165	267212	45.528	ng/u1	0.00
31) 4-Chloroaniline-d4	10.819	131	323954	40.008	ng/u1	0.00
46) Dimethylphthalate-d6	13.936	166	794679	45.932	ng/u1	0.00
49) Acenaphthylene-d8	14.219	160	970758	43.654	ng/u1	0.00
54) 4-Nitrophenol-d4	14.719	143	126694	48.025	ng/u1	0.00
60) Fluorene-d10	15.513	176	674402	45.107	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.636	200	133711	45.624	ng/u1	0.00
73) Anthracene-d10	17.365	188	1022179	44.660	ng/u1	0.00
81) Pyrene-d10	19.653	212	1204772	47.064	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.671	264	1158249	47.971	ng/u1	0.00
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.290	88	42211	14.976	ng/uL#	81
5) Pyridine	3.696	79	281202	38.094	ng/u1#	78
6) Benzaldehyde	7.007	77	225784	44.327	ng/u1	91
8) Phenol	7.066	94	340012	40.510	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.295	93	275284	40.641	ng/u1#	86
12) 2-Chlorophenol	7.437	128	275665m	42.106	ng/u1	
13) 2-Methylphenol	8.325	108	254283	40.982	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.413	45	410368	40.520	ng/u1#	84
16) Acetophenone	8.701	105	404036	39.884	ng/u1#	86
17) N-Nitroso-di-n-propyla...	8.689	70	226647	42.034	ng/u1#	86
18) 4-Methylphenol	8.654	108	270936	41.047	ng/u1	100
19) Hexachloroethane	8.966	117	118337	40.186	ng/u1	86
22) Nitrobenzene	9.084	77	325718	41.501	ng/u1	91
23) Isophorone	9.613	82	586903	42.110	ng/u1#	92
25) 2-Nitrophenol	9.795	139	147390	44.148	ng/u1#	89
26) 2,4-Dimethylphenol	9.860	107	303192	40.915	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.095	93	351971	41.974	ng/u1	99
29) 2,4-Dichlorophenol	10.336	162	242298	41.828	ng/u1	96
30) Naphthalene	10.736	128	850039	40.083	ng/u1	99
32) 4-Chloroaniline	10.848	127	295833	36.468	ng/u1	98
33) Hexachlorobutadiene	11.031	225	161839	39.739	ng/u1	98
34) Caprolactam	11.613	113	82973m	44.092	ng/u1	
35) 4-Chloro-3-methylphenol	11.977	107	261268	43.448	ng/u1	92
36) 2-Methylnaphthalene	12.354	142	554821	40.821	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034584.D
 Acq On : 09 Apr 2022 05:27
 Operator : CG/JU
 Sample : PB143966BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS966

Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022

Quant Time: Apr 09 06:27:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.572	142	574076	41.169	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.725	216	288406	40.241	ng/ul#	96
40) Hexachlorocyclopentadiene	12.707	237	178304	37.911	ng/ul	95
41) 2,4,6-Trichlorophenol	12.960	196	188617	43.323	ng/ul	97
42) 2,4,5-Trichlorophenol	13.030	196	204796	45.394	ng/ul	99
43) 1,1'-Biphenyl	13.360	154	745035	41.076	ng/ul#	98
44) 2-Chloronaphthalene	13.401	162	571311	41.028	ng/ul	97
45) 2-Nitroaniline	13.607	65	178697	47.062	ng/ul	84
47) Dimethylphthalate	13.983	163	715113	42.001	ng/ul	99
48) 2,6-Dinitrotoluene	14.101	165	150890	46.131	ng/ul	91
50) Acenaphthylene	14.248	152	933248	41.339	ng/ul	100
51) 3-Nitroaniline	14.430	138	124640	43.585	ng/ul	84
52) Acenaphthene	14.589	153	617311	41.071	ng/ul	100
53) 2,4-Dinitrophenol	14.636	184	78799	42.747	ng/ul#	83
55) 4-Nitrophenol	14.736	109	101709	45.147	ng/ul#	83
56) Dibenzofuran	14.924	168	847047	41.466	ng/ul	97
57) 2,4-Dinitrotoluene	14.889	165	212868	46.758	ng/ul#	82
58) 2,3,4,6-Tetrachlorophenol	15.148	232	175678	44.351	ng/ul#	87
59) Diethylphthalate	15.348	149	742904	42.957	ng/ul	99
61) Fluorene	15.571	166	712599	41.346	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.566	204	343405	41.278	ng/ul	94
63) 4-Nitroaniline	15.589	138	127306	50.556	ng/ul#	94
66) 4,6-Dinitro-2-methylph...	15.648	198	122938	42.070	ng/ul#	85
67) N-Nitrosodiphenylamine	15.777	169	606671	42.416	ng/ul	98
68) 4-Bromophenyl-phenylether	16.460	248	206473	41.606	ng/ul#	93
69) Hexachlorobenzene	16.577	284	249303	41.830	ng/ul#	92
70) Atrazine	16.730	200	221703	41.421	ng/ul	98
71) Pentachlorophenol	16.918	266	151869	41.615	ng/ul	99
72) Phenanthrene	17.307	178	1117231	41.706	ng/ul	99
74) Anthracene	17.401	178	1121207	42.239	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.324	216	302776	40.157	ng/uL	99
76) Pentachlorobenzene	14.848	250	283165	40.019	ng/uL	98
77) Carbazole	17.671	167	935434	42.599	ng/ul	99
78) Di-n-butylphthalate	18.236	149	1247772	44.770	ng/ul	100
80) Fluoranthene	19.324	202	1278409	45.127	ng/ul#	90
82) Pyrene	19.683	202	1346828	43.741	ng/ul#	86
83) Butylbenzylphthalate	20.583	149	558855	46.361	ng/ul	87
84) 3,3'-Dichlorobenzidine	21.359	252	349314	42.496	ng/ul#	97
85) Benzo(a)anthracene	21.430	228	1189784	43.627	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.359	149	824107	45.679	ng/ul#	98
87) Chrysene	21.477	228	1241444	42.216	ng/ul	100
89) Di-n-octyl phthalate	22.277	149	1417207	44.350	ng/ul	100
90) Benzo(b)fluoranthene	23.100	252	1324312	46.432	ng/ul#	97
91) Benzo(k)fluoranthene	23.147	252	1286952	42.967	ng/ul#	97
93) Benzo(a)pyrene	23.718	252	1289205	45.735	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.300	276	1397036	46.311	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.318	278	1164191	47.075	ng/ul#	95
96) Benzo(g,h,i)perylene	27.065	276	1265144	43.693	ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Abundance Scan 43 (3.290 min): BM034570.D\data.ms (-38)



#2
1,4-Dioxane
Concen: 14.976 ng/uL
RT: 3.290 min Scan# 41
Delta R.T. -0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27



Instrument :
BNA_M
ClientSampleId :
SLCS966

Tgt Ion: 88 Resp: 4221
Ion Ratio Lower Upper
88 100
43 33.3 21.0 31.6
58 75.8 47.8 71.8


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022


Abundance Scan 43 (3.290 min): BM034584.D\data.ms



Abundance Scan 43 (3.290 min): BM034584.D\data.ms (-9)





Abundance Scan 113 (3.701 min): BM034570.D\data.ms (-10)




#5
Pyridine
Concen: 38.094 ng/uL
RT: 3.696 min Scan# 112
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27


Tgt Ion: 79 Resp: 281202
Ion Ratio Lower Upper
79 100
52 80.0 47.7 71.5
51 38.8 25.6 38.4

Abundance Scan 112 (3.696 min): BM034584.D\data.ms




Abundance Scan 112 (3.696 min): BM034584.D\data.ms (-95)





#6
Benzaldehyde
Concen: 44.327 ng/ul
RT: 7.007 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion: 77 Resp: 225784
Ion Ratio Lower Upper
77 100
105 90.7 80.8 121.2
106 87.5 76.6 115.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#8
Phenol
Concen: 40.510 ng/ul
RT: 7.066 min Scan# 685
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion: 94 Resp: 340012
Ion Ratio Lower Upper
94 100
65 31.0 22.9 34.3
66 40.5 33.7 50.5





#10
Bis(2-Chloroethyl)ether
Concen: 40.641 ng/ul
RT: 7.295 min Scan# 724
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966



Tgt Ion: 93 Resp: 275284
Ion Ratio Lower Upper
93 100
63 84.6 54.7 82.1
95 31.4 26.6 39.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#12
2-Chlorophenol
Concen: 42.106 ng/ul m
RT: 7.437 min Scan# 748
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27


Tgt Ion:128 Resp: 275665
Ion Ratio Lower Upper
128 100
64 52.3 30.6 46.0
130 32.9 25.8 38.6





#13
2-Methylphenol
Concen: 40.982 ng/ul
RT: 8.325 min Scan# 899
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:108 Resp: 25428
Ion Ratio Lower Upper
108 100
107 91.9 71.9 107.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 40.520 ng/ul
RT: 8.413 min Scan# 914
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion: 45 Resp: 410368
Ion Ratio Lower Upper
45 100
77 13.5 17.1 25.7#
79 10.0 12.8 19.2#





#16
Acetophenone
Concen: 39.884 ng/ul
RT: 8.701 min Scan# 963
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966



Tgt Ion: 105 Resp: 404036
Ion Ratio Lower Upper
105 100
77 86.3 59.4 89.2
51 31.6 19.3 28.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#17
N-Nitroso-di-n-propylamine
Concen: 42.034 ng/ul
RT: 8.689 min Scan# 961
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27


Tgt Ion: 70 Resp: 226647
Ion Ratio Lower Upper
70 100
42 63.5 40.6 60.8#
101 9.5 8.2 12.2
130 19.8 19.9 29.9#





#18
4-Methylphenol
Concen: 41.047 ng/ul
RT: 8.654 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:108 Resp: 270930
Ion Ratio Lower Upper
108 100
107 111.7 89.0 133.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#19
Hexachloroethane
Concen: 40.186 ng/ul
RT: 8.966 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:117 Resp: 118337
Ion Ratio Lower Upper
117 100
201 98.3 89.9 134.9
199 58.9 57.2 85.8





#22
Nitrobenzene
Concen: 41.501 ng/u1
RT: 9.084 min Scan# 1028
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966



Tgt Ion: 77 Resp: 325718
Ion Ratio Lower Upper
77 100
123 40.4 38.2 57.2
65 14.4 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#23
Isophorone
Concen: 42.110 ng/u1
RT: 9.613 min Scan# 1118
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27


Tgt Ion: 82 Resp: 586903
Ion Ratio Lower Upper
82 100
95 7.1 6.8 10.2
138 16.0 16.5 24.7#





#25
2-Nitrophenol
Concen: 44.148 ng/ul
RT: 9.795 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:139 Resp: 147390
Ion Ratio Lower Upper
139 100
65 58.8 37.8 56.6
109 31.8 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#26
2,4-Dimethylphenol
Concen: 40.915 ng/ul
RT: 9.860 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:107 Resp: 303192
Ion Ratio Lower Upper
107 100
121 50.2 40.8 61.2
122 83.8 69.0 103.6





#27
Bis(2-Chloroethoxy)methane
Concen: 41.974 ng/ul
RT: 10.095 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion: 93 Resp: 35197
Ion Ratio Lower Upper
93 100
95 31.5 25.4 38.2
123 10.7 9.5 14.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#29
2,4-Dichlorophenol
Concen: 41.828 ng/ul
RT: 10.336 min Scan# 1241
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:162 Resp: 242298
Ion Ratio Lower Upper
162 100
164 66.5 52.8 79.2
98 38.2 26.0 39.0





#30
Naphthalene
Concen: 40.083 ng/u1
RT: 10.736 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:128 Resp: 850039
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.1 9.9 14.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#32
4-Chloroaniline
Concen: 36.468 ng/u1
RT: 10.848 min Scan# 1328
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:127 Resp: 295833
Ion Ratio Lower Upper
127 100
129 32.3 25.1 37.7





#33
Hexachlorobutadiene
Concen: 39.739 ng/ul
RT: 11.031 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:225 Resp: 161839
Ion Ratio Lower Upper
225 100
223 64.4 50.6 75.8
227 64.3 50.3 75.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#34
Caprolactam
Concen: 44.092 ng/ul m
RT: 11.613 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:113 Resp: 82973
Ion Ratio Lower Upper
113 100
55 186.5 100.9 151.3#
56 135.1 81.0 121.4#





#35
4-Chloro-3-methylphenol
Concen: 43.448 ng/ul
RT: 11.977 min Scan# 1520
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966



Tgt Ion:107 Resp: 261268
Ion Ratio Lower Upper
107 100
144 26.3 23.0 34.4
142 78.7 69.8 104.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#36
2-Methylnaphthalene
Concen: 40.821 ng/ul
RT: 12.354 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27


Tgt Ion:142 Resp: 554821
Ion Ratio Lower Upper
142 100
141 90.6 71.4 107.0





#37
1-Methylnaphthalene
Concen: 41.169 ng/ul
RT: 12.572 min Scan# 1621
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:142 Resp: 574070
Ion Ratio Lower Upper
142 100
141 91.9 74.5 111.7
116 3.9 3.0 4.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.241 ng/ul
RT: 12.725 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:216 Resp: 288406
Ion Ratio Lower Upper
216 100
214 80.1 62.2 93.2
179 22.1 15.4 23.2
108 19.2 11.8 17.6#





#40
Hexachlorocyclopentadiene
Concen: 37.911 ng/ul
RT: 12.707 min Scan# 1644
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:237 Resp: 178304
Ion Ratio Lower Upper
237 100
235 62.0 52.9 79.3
272 13.3 11.8 17.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#41
2,4,6-Trichlorophenol
Concen: 43.323 ng/ul
RT: 12.960 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:196 Resp: 188617
Ion Ratio Lower Upper
196 100
198 97.8 80.8 121.2
200 32.6 26.5 39.7





#42
2,4,5-Trichlorophenol
Concen: 45.394 ng/ul
RT: 13.030 min Scan# 1700
Delta R.T. -0.006 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:196 Resp: 204790
Ion Ratio Lower Upper
196 100
198 96.0 78.0 117.0
200 30.2 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#43
1,1'-Biphenyl
Concen: 41.076 ng/ul
RT: 13.360 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:154 Resp: 745035
Ion Ratio Lower Upper
154 100
153 40.8 32.6 49.0
76 14.2 7.7 11.5





#44
2-Chloronaphthalene
Concen: 41.028 ng/ul
RT: 13.401 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:162 Resp: 57131
Ion Ratio Lower Upper
162 100
164 32.6 26.6 40.0
127 40.0 29.6 44.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#45
2-Nitroaniline
Concen: 47.062 ng/ul
RT: 13.607 min Scan# 1797
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion: 65 Resp: 178697
Ion Ratio Lower Upper
65 100
92 64.2 57.4 86.2
138 96.7 94.8 142.2





#47
Dimethylphthalate
Concen: 42.001 ng/ul
RT: 13.983 min Scan# 1861
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:163 Resp: 71511
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 9.9 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#48
2,6-Dinitrotoluene
Concen: 46.131 ng/ul
RT: 14.101 min Scan# 1881
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:165 Resp: 150890
Ion Ratio Lower Upper
165 100
89 63.4 44.2 66.4
121 22.2 16.7 25.1





#50
Acenaphthylene
Concen: 41.339 ng/ul
RT: 14.248 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:152 Resp: 933248
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.8
153 12.8 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#51
3-Nitroaniline
Concen: 43.585 ng/ul
RT: 14.430 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:138 Resp: 124640
Ion Ratio Lower Upper
138 100
108 11.0 8.3 12.5
92 127.2 87.0 130.4





#52
Acenaphthene
Concen: 41.071 ng/ul
RT: 14.589 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:153 Resp: 61731
Ion Ratio Lower Upper
153 100
152 46.3 37.4 56.0
154 88.5 71.0 106.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#53
2,4-Dinitrophenol
Concen: 42.747 ng/ul
RT: 14.636 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:184 Resp: 78799
Ion Ratio Lower Upper
184 100
63 79.8 43.8 65.6#
154 63.8 49.9 74.9





#55
4-Nitrophenol
Concen: 45.147 ng/ul
RT: 14.736 min Scan# 1988
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966



Tgt Ion:109 Resp: 101709
Ion Ratio Lower Upper
109 100
139 120.6 94.6 142.0
65 139.8 83.5 125.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#56
Dibenzofuran
Concen: 41.466 ng/ul
RT: 14.924 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27


Tgt Ion:168 Resp: 847047
Ion Ratio Lower Upper
168 100
139 39.1 29.8 44.6





#57
2,4-Dinitrotoluene
Concen: 46.758 ng/ul
RT: 14.889 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:165 Resp: 212868
Ion Ratio Lower Upper
165 100
63 49.4 30.2 45.4
182 2.5 2.2 3.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 44.351 ng/ul
RT: 15.148 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:232 Resp: 175678
Ion Ratio Lower Upper
232 100
131 51.5 32.2 48.4
130 2.8 1.9 2.9
166 33.0 23.0 34.4





#59
Diethylphthalate
Concen: 42.957 ng/ul
RT: 15.348 min Scan# 2092
Delta R.T. 0.006 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:149 Resp: 742904
Ion Ratio Lower Upper
149 100
177 21.6 17.9 26.9
150 12.6 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#61
Fluorene
Concen: 41.346 ng/ul
RT: 15.571 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:166 Resp: 712599
Ion Ratio Lower Upper
166 100
165 97.7 76.9 115.3
167 13.6 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 41.278 ng/ul
RT: 15.566 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:204 Resp: 34340
Ion Ratio Lower Upper
204 100
206 32.0 25.5 38.3
141 61.9 44.2 66.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#63
4-Nitroaniline
Concen: 50.556 ng/ul
RT: 15.589 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:138 Resp: 127306
Ion Ratio Lower Upper
138 100
92 54.7 36.3 54.5#
108 87.7 68.9 103.3





#66
4,6-Dinitro-2-methylphenol
Concen: 42.070 ng/ul
RT: 15.648 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:198 Resp: 122938
Ion Ratio Lower Upper
198 100
182 4.2 3.8 5.6
77 31.5 18.5 27.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#67
N-Nitrosodiphenylamine
Concen: 42.416 ng/ul
RT: 15.777 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:169 Resp: 606671
Ion Ratio Lower Upper
169 100
168 67.0 52.8 79.2
167 36.4 28.1 42.1





#68
4-Bromophenyl-phenylether
Concen: 41.606 ng/ul
RT: 16.460 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:248 Resp: 206473
Ion Ratio Lower Upper
248 100
250 97.9 78.4 117.6
141 77.6 51.4 77.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#69
Hexachlorobenzene
Concen: 41.830 ng/ul
RT: 16.577 min Scan# 2302
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:284 Resp: 249303
Ion Ratio Lower Upper
284 100
142 38.9 24.6 36.8#
249 30.7 24.1 36.1





#70
Atrazine
Concen: 41.421 ng/ul
RT: 16.730 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:200 Resp: 22170
Ion Ratio Lower Upper
200 100
173 26.2 20.2 30.4
215 46.3 38.4 57.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#71
Pentachlorophenol
Concen: 41.615 ng/ul
RT: 16.918 min Scan# 2360
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:266 Resp: 151869
Ion Ratio Lower Upper
266 100
264 62.6 49.3 73.9
268 63.0 50.3 75.5





#72
Phenanthrene
Concen: 41.706 ng/u1
RT: 17.307 min Scan# 2426
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:178 Resp: 111723
Ion Ratio Lower Upper
178 100
179 15.2 11.7 17.5
176 18.9 15.0 22.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#74
Anthracene
Concen: 42.239 ng/u1
RT: 17.401 min Scan# 2442
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:178 Resp: 1121207
Ion Ratio Lower Upper
178 100
179 15.2 11.9 17.9
176 18.3 14.4 21.6





#75
1,2,3,4-Tetrachlorobenzene
Concen: 40.157 ng/uL
RT: 13.324 min Scan# 1750
Delta R.T. -0.006 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:216 Resp: 302770
Ion Ratio Lower Upper
216 100
214 78.6 62.0 93.0
218 47.2 37.8 56.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#76
Pentachlorobenzene
Concen: 40.019 ng/uL
RT: 14.848 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:250 Resp: 283165
Ion Ratio Lower Upper
250 100
248 64.4 50.6 75.8
252 63.9 50.2 75.4





#77
 Carbazole
 Concen: 42.599 ng/ul
 RT: 17.671 min Scan# 2488
 Delta R.T. 0.000 min
 Lab File: BM034584.D
 Acq: 09 Apr 2022 05:27




Instrument :
 BNA_M
 ClientSampleId :
 SLCS966




Tgt Ion:167 Resp: 935434
 Ion Ratio Lower Upper
 167 100
 166 21.0 16.6 25.0
 139 13.1 9.9 14.9

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022




#78
 Di-n-butylphthalate
 Concen: 44.770 ng/ul
 RT: 18.236 min Scan# 2584
 Delta R.T. 0.000 min
 Lab File: BM034584.D
 Acq: 09 Apr 2022 05:27




Tgt Ion:149 Resp: 1247772
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.0
 104 5.6 4.2 6.2





#80
Fluoranthene
Concen: 45.127 ng/ul
RT: 19.324 min Scan# 2769
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:202 Resp: 1278409
Ion Ratio Lower Upper
202 100
101 11.2 6.2 9.2#
100 8.8 4.7 7.1#

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#82
Pyrene
Concen: 43.741 ng/ul
RT: 19.683 min Scan# 2830
Delta R.T. -0.006 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:202 Resp: 1346828
Ion Ratio Lower Upper
202 100
101 13.6 7.0 10.4#
100 11.9 5.8 8.6#





#83
Butylbenzylphthalate
Concen: 46.361 ng/ul
RT: 20.583 min Scan# 2983
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:149 Resp: 55885
Ion Ratio Lower Upper
149 100
91 74.9 50.9 76.3
206 20.3 19.3 28.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#84
3,3'-Dichlorobenzidine
Concen: 42.496 ng/ul
RT: 21.359 min Scan# 3115
Delta R.T. -0.006 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:252 Resp: 349314
Ion Ratio Lower Upper
252 100
254 64.9 50.7 76.1
126 11.5 6.6 9.8#





#85
Benzo(a)anthracene
Concen: 43.627 ng/u1
RT: 21.430 min Scan# 3127
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:228 Resp: 1189784
Ion Ratio Lower Upper
228 100
229 19.4 15.4 23.2
226 27.2 21.0 31.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 45.679 ng/u1
RT: 21.359 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:149 Resp: 824107
Ion Ratio Lower Upper
149 100
167 26.9 22.1 33.1
279 4.1 4.2 6.2#





#87
Chrysene
Concen: 42.216 ng/ul
RT: 21.477 min Scan# 3136
Delta R.T. -0.006 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966



Tgt Ion:228 Resp: 1241444
Ion Ratio Lower Upper
228 100
226 29.3 23.4 35.2
229 19.7 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#89
Di-n-octyl phthalate
Concen: 44.350 ng/ul
RT: 22.277 min Scan# 3271
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27


Tgt Ion:149 Resp: 1417207





#90
Benzo(b)fluoranthene
Concen: 46.432 ng/ul
RT: 23.100 min Scan# 3411
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:252 Resp: 1324312
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.0
125 9.7 5.8 8.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#91
Benzo(k)fluoranthene
Concen: 42.967 ng/ul
RT: 23.147 min Scan# 3419
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:252 Resp: 1286952
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 9.9 5.1 7.7





#93
Benzo(a)pyrene
Concen: 45.735 ng/ul
RT: 23.718 min Scan# 3516
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:252 Resp: 128920
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 10.9 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 46.311 ng/ul
RT: 26.300 min Scan# 3955
Delta R.T. -0.006 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Tgt Ion:276 Resp: 1397036
Ion Ratio Lower Upper
276 100
138 23.9 13.6 20.4
277 24.5 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 47.075 ng/ul
RT: 26.318 min Scan# 3958
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27




Instrument :
BNA_M
ClientSampleId :
SLCS966




Tgt Ion:278 Resp: 116419
Ion Ratio Lower Upper
278 100
139 15.9 9.6 14.4
279 23.2 19.3 28.9

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#96
Benzo(g,h,i)perylene
Concen: 43.693 ng/ul
RT: 27.065 min Scan# 4085
Delta R.T. 0.000 min
Lab File: BM034584.D
Acq: 09 Apr 2022 05:27



Tgt Ion:276 Resp: 1265144
Ion Ratio Lower Upper
276 100
138 20.4 11.8 17.6
277 24.0 18.8 28.2

