

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034601.D
 Acq On : 09 Apr 2022 16:21
 Operator : CG/JU
 Sample : PB143975BS
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS975

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022

Quant Time: Apr 11 01:01:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.878	152	107494	20.000	ng/u1	0.00
20) Naphthalene-d8	10.683	136	442109	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.524	164	268592	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.265	188	543094	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.442	240	549952	20.000	ng/u1	# 0.00
88) Perylene-d12	23.824	264	577705	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.255	96	22247	7.380	ng/uL	0.00
4) Pyridine-d5	3.678	84	281415	38.683	ng/u1	0.00
7) Phenol-d5	7.037	99	351757	39.316	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.201	67	244003	40.927	ng/u1	0.00
11) 2-Chlorophenol-d4	7.407	132	289962	41.925	ng/u1	0.00
15) 4-Methylphenol-d8	8.589	113	289824	41.470	ng/u1	0.00
21) Nitrobenzene-d5	9.042	128	140085m	43.113	ng/u1	0.00
24) 2-Nitrophenol-d4	9.766	143	147760	43.679	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.307	165	270371	42.445	ng/u1	0.00
31) 4-Chloroaniline-d4	10.825	131	328359	37.364	ng/u1	0.00
46) Dimethylphthalate-d6	13.936	166	840907	43.328	ng/u1	0.00
49) Acenaphthylene-d8	14.219	160	1018301	40.822	ng/u1	0.00
54) 4-Nitrophenol-d4	14.719	143	133673m	45.171	ng/u1	0.00
60) Fluorene-d10	15.518	176	716009	42.692	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.636	200	140461	42.365	ng/u1	0.00
73) Anthracene-d10	17.365	188	1100696	42.509	ng/u1	0.00
81) Pyrene-d10	19.659	212	1289147	42.948	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.671	264	1272585	43.792	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.290	88	38371	12.642	ng/uL#	80
5) Pyridine	3.702	79	265785	33.435	ng/u1#	81
6) Benzaldehyde	7.007	77	215437m	39.277	ng/u1	
8) Phenol	7.066	94	321770	35.600	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.295	93	258713	35.468	ng/u1#	86
12) 2-Chlorophenol	7.437	128	260440	36.941	ng/u1#	85
13) 2-Methylphenol	8.325	108	239271	35.810	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.413	45	393703	36.100	ng/u1#	85
16) Acetophenone	8.701	105	388678	35.630	ng/u1#	89
17) N-Nitroso-di-n-propyla...	8.689	70	222265	38.279	ng/u1#	84
18) 4-Methylphenol	8.654	108	258246	36.332	ng/u1	99
19) Hexachloroethane	8.966	117	113438	35.773	ng/u1	85
22) Nitrobenzene	9.084	77	318951	37.444	ng/u1	92
23) Isophorone	9.613	82	576625	38.120	ng/u1#	92
25) 2-Nitrophenol	9.795	139	139438	38.483	ng/u1#	87
26) 2,4-Dimethylphenol	9.860	107	292688	36.393	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.095	93	343557	37.750	ng/u1	98
29) 2,4-Dichlorophenol	10.336	162	238717	37.971	ng/u1	96
30) Naphthalene	10.736	128	820075	35.630	ng/u1	99
32) 4-Chloroaniline	10.848	127	286294	32.518	ng/u1	98
33) Hexachlorobutadiene	11.031	225	156061	35.308	ng/u1	98
34) Caprolactam	11.613	113	81142m	39.729	ng/u1	
35) 4-Chloro-3-methylphenol	11.977	107	261088	40.005	ng/u1	91
36) 2-Methylnaphthalene	12.354	142	547539	37.118	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034601.D
 Acq On : 09 Apr 2022 16:21
 Operator : CG/JU
 Sample : PB143975BS
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS975

Quant Time: Apr 11 01:01:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.572	142	566814	37.453	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.725	216	287724	35.788	ng/ul#	97
40) Hexachlorocyclopentadiene	12.707	237	174884	33.148	ng/ul	94
41) 2,4,6-Trichlorophenol	12.960	196	190878	39.084	ng/ul	94
42) 2,4,5-Trichlorophenol	13.030	196	207684	41.038	ng/ul	98
43) 1,1'-Biphenyl	13.360	154	738502	36.296	ng/ul#	97
44) 2-Chloronaphthalene	13.401	162	565371	36.194	ng/ul	97
45) 2-Nitroaniline	13.601	65	180184	42.303	ng/ul#	81
47) Dimethylphthalate	13.983	163	723475	37.880	ng/ul	99
48) 2,6-Dinitrotoluene	14.101	165	151019	41.159	ng/ul	92
50) Acenaphthylene	14.248	152	942076	37.201	ng/ul	99
51) 3-Nitroaniline	14.430	138	125427	39.099	ng/ul	88
52) Acenaphthene	14.589	153	619032	36.716	ng/ul	99
53) 2,4-Dinitrophenol	14.636	184	78271	37.852	ng/ul#	78
55) 4-Nitrophenol	14.736	109	102187	40.436	ng/ul#	84
56) Dibenzofuran	14.924	168	859053	37.489	ng/ul	95
57) 2,4-Dinitrotoluene	14.889	165	214880	42.077	ng/ul#	82
58) 2,3,4,6-Tetrachlorophenol	15.148	232	176896	39.811	ng/ul#	86
59) Diethylphthalate	15.348	149	761455	39.251	ng/ul	99
61) Fluorene	15.571	166	719113	37.195	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.566	204	350913	37.603	ng/ul	95
63) 4-Nitroaniline	15.589	138	128077m	45.342	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.648	198	122802	37.147	ng/ul#	81
67) N-Nitrosodiphenylamine	15.777	169	615141	38.017	ng/ul	99
68) 4-Bromophenyl-phenylether	16.460	248	212117	37.783	ng/ul	95
69) Hexachlorobenzene	16.577	284	252346	37.427	ng/ul#	92
70) Atrazine	16.730	200	225317	37.211	ng/ul	99
71) Pentachlorophenol	16.918	266	155495	37.664	ng/ul	97
72) Phenanthrene	17.307	178	1139153	37.590	ng/ul	100
74) Anthracene	17.401	178	1138172	37.902	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.330	216	303274	35.555	ng/uL	98
76) Pentachlorobenzene	14.848	250	284813	35.581	ng/uL	98
77) Carbazole	17.671	167	959717	38.633	ng/ul	99
78) Di-n-butylphthalate	18.236	149	1295859	41.099	ng/ul	99
80) Fluoranthene	19.324	202	1324825	39.883	ng/ul#	90
82) Pyrene	19.689	202	1373749	38.049	ng/ul#	88
83) Butylbenzylphthalate	20.583	149	592008	41.884	ng/ul#	85
84) 3,3'-Dichlorobenzidine	21.359	252	357344	37.075	ng/ul#	97
85) Benzo(a)anthracene	21.430	228	1226195	38.345	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.359	149	892628	42.196	ng/ul#	98
87) Chrysene	21.483	228	1286847	37.320	ng/ul	99
89) Di-n-octyl phthalate	22.277	149	1517653	39.461	ng/ul	100
90) Benzo(b)fluoranthene	23.100	252	1337424	38.961	ng/ul#	98
91) Benzo(k)fluoranthene	23.147	252	1352088	37.507	ng/ul#	97
93) Benzo(a)pyrene	23.724	252	1316508	38.804	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.306	276	1445675	39.817	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.318	278	1196514	40.198	ng/ul#	96
96) Benzo(g,h,i)perylene	27.065	276	1317538	37.806	ng/ul#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040922\
 Data File : BM034601.D
 Acq On : 09 Apr 2022 16:21
 Operator : CG/JU
 Sample : PB143975BS
 Misc :
 ALS Vial : 48 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SLCS975

Quant Time: Apr 11 01:01:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022





1,4-Dioxane

Concen: 12.642 ng/uL

RT: 3.290 min Scan# 41

Delta R.T. -0.000 min

Lab File: BM034601.D

Acq: 09 Apr 2022 16:21

Instrument :

BNA_M

ClientSampleId :

SLCS975





Tgt Ion: 88 Resp: 3837

Ion	Ratio	Lower	Upper
88	100		
43	33.1	21.0	31.6
58	77.2	47.8	71.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#5

Pyridine

Concen: 33.435 ng/uL

RT: 3.702 min Scan# 113




Delta R.T. 0.006 min


Lab File: BM034601.D

Acq: 09 Apr 2022 16:21

Tgt Ion: 79 Resp: 265785

Ion	Ratio	Lower	Upper
79	100		
52	75.8	47.7	71.5
51	40.4	25.6	38.4









#6
Benzaldehyde
Concen: 39.277 ng/ul m
RT: 7.007 min Scan# 675
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Instrument :
BNA_M
ClientSampleId :
SLCS975

Tgt Ion: 77 Resp: 21543
Ion Ratio Lower Upper
77 100
105 93.8 80.8 121.2
106 88.5 76.6 115.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#8
Phenol
Concen: 35.600 ng/ul
RT: 7.066 min Scan# 685
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21


Tgt Ion: 94 Resp: 321770
Ion Ratio Lower Upper
94 100
65 31.9 22.9 34.3
66 41.1 33.7 50.5





#10
Bis(2-Chloroethyl)ether
Concen: 35.468 ng/ul
RT: 7.295 min Scan# 724
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion: 93 Resp: 258711
Ion Ratio Lower Upper
93 100
63 84.2 54.7 82.1
95 32.4 26.6 39.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#12
2-Chlorophenol
Concen: 36.941 ng/ul
RT: 7.437 min Scan# 748
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:128 Resp: 260440
Ion Ratio Lower Upper
128 100
64 53.3 30.6 46.0
130 30.8 25.8 38.6





#13
2-Methylphenol
Concen: 35.810 ng/ul
RT: 8.325 min Scan# 899
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21



Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion: 108 Resp: 23927
Ion Ratio Lower Upper
108 100
107 87.2 71.9 107.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 36.100 ng/ul
RT: 8.413 min Scan# 914
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion: 45 Resp: 393703
Ion Ratio Lower Upper
45 100
77 13.7 17.1 25.7#
79 10.2 12.8 19.2#





#16
Acetophenone
Concen: 35.630 ng/ul
RT: 8.701 min Scan# 963
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975



Tgt Ion: 105 Resp: 388678
Ion Ratio Lower Upper
105 100
77 82.6 59.4 89.2
51 31.7 19.3 28.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#17
N-Nitroso-di-n-propylamine
Concen: 38.279 ng/ul
RT: 8.689 min Scan# 961
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Tgt Ion: 70 Resp: 222265
Ion Ratio Lower Upper
70 100
42 64.6 40.6 60.8#
101 9.8 8.2 12.2
130 18.7 19.9 29.9#









#18
4-Methylphenol
Concen: 36.332 ng/ul
RT: 8.654 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Instrument :
BNA_M
ClientSampleId :
SLCS975

Tgt Ion:108 Resp: 258240
Ion Ratio Lower Upper
108 100
107 110.4 89.0 133.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#19
Hexachloroethane
Concen: 35.773 ng/ul
RT: 8.966 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Tgt Ion:117 Resp: 113438
Ion Ratio Lower Upper
117 100
201 95.4 89.9 134.9
199 61.0 57.2 85.8









#22
Nitrobenzene
Concen: 37.444 ng/ul
RT: 9.084 min Scan# 1027
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Instrument :
BNA_M
ClientSampleId :
SLCS975

Tgt Ion: 77 Resp: 31895
Ion Ratio Lower Upper
77 100
123 41.0 38.2 57.2
65 14.1 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#23
Isophorone
Concen: 38.120 ng/ul
RT: 9.613 min Scan# 1118
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21


Tgt Ion: 82 Resp: 576625
Ion Ratio Lower Upper
82 100
95 7.3 6.8 10.2
138 16.1 16.5 24.7#





#25
2-Nitrophenol
Concen: 38.483 ng/ul
RT: 9.795 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:139 Resp: 139438
Ion Ratio Lower Upper
139 100
65 61.6 37.8 56.6
109 32.5 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#26
2,4-Dimethylphenol
Concen: 36.393 ng/ul
RT: 9.860 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21



Tgt Ion:107 Resp: 292688
Ion Ratio Lower Upper
107 100
121 48.8 40.8 61.2
122 82.4 69.0 103.6









#27
Bis(2-Chloroethoxy)methane
Concen: 37.750 ng/ul
RT: 10.095 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Instrument :
BNA_M
ClientSampleId :
SLCS975

Tgt Ion: 93 Resp: 34355
Ion Ratio Lower Upper
93 100
95 32.7 25.4 38.2
123 10.6 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#29
2,4-Dichlorophenol
Concen: 37.971 ng/ul
RT: 10.336 min Scan# 1241
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21


Tgt Ion:162 Resp: 238717
Ion Ratio Lower Upper
162 100
164 65.1 52.8 79.2
98 37.3 26.0 39.0





#30
Naphthalene
Concen: 35.630 ng/u1
RT: 10.736 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975



Tgt Ion:128 Resp: 820075
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 13.2 9.9 14.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#32
4-Chloroaniline
Concen: 32.518 ng/u1
RT: 10.848 min Scan# 1328
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21


Tgt Ion:127 Resp: 286294
Ion Ratio Lower Upper
127 100
129 32.5 25.1 37.7





#33
Hexachlorobutadiene
Concen: 35.308 ng/ul
RT: 11.031 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:225 Resp: 15606
Ion Ratio Lower Upper
225 100
223 61.8 50.6 75.8
227 64.0 50.3 75.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#34
Caprolactam
Concen: 39.729 ng/ul m
RT: 11.613 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:113 Resp: 81142
Ion Ratio Lower Upper
113 100
55 185.2 100.9 151.3#
56 139.2 81.0 121.4#





#35
4-Chloro-3-methylphenol
Concen: 40.005 ng/ul
RT: 11.977 min Scan# 1520
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:107 Resp: 261088
Ion Ratio Lower Upper
107 100
144 25.0 23.0 34.4
142 78.8 69.8 104.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#36
2-Methylnaphthalene
Concen: 37.118 ng/ul
RT: 12.354 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21



Tgt Ion:142 Resp: 547539
Ion Ratio Lower Upper
142 100
141 87.7 71.4 107.0





#37

1-Methylnaphthalene

Concen: 37.453 ng/ul

RT: 12.572 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034601.D

Acq: 09 Apr 2022 16:21

Instrument :

BNA_M

ClientSampleId :





SLCS975

Tgt Ion:142 Resp: 566814
 Ion Ratio Lower Upper
 142 100
 141 92.7 74.5 111.7
 116 4.0 3.0 4.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022



#39

1,2,4,5-Tetrachlorobenzene

Concen: 35.788 ng/ul




RT: 12.725 min Scan# 1647


Delta R.T. 0.000 min

Lab File: BM034601.D

Acq: 09 Apr 2022 16:21


Tgt Ion:216 Resp: 287724
 Ion Ratio Lower Upper
 216 100
 214 75.9 62.2 93.2
 179 21.2 15.4 23.2
 108 17.7 11.8 17.6#





#40
Hexachlorocyclopentadiene
Concen: 33.148 ng/ul
RT: 12.707 min Scan# 1644
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:237 Resp: 174884
Ion Ratio Lower Upper
237 100
235 60.9 52.9 79.3
272 13.0 11.8 17.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#41
2,4,6-Trichlorophenol
Concen: 39.084 ng/ul
RT: 12.960 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21



Tgt Ion:196 Resp: 190878
Ion Ratio Lower Upper
196 100
198 93.8 80.8 121.2
200 31.7 26.5 39.7






#42
2,4,5-Trichlorophenol
Concen: 41.038 ng/ul
RT: 13.030 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Instrument :

BNA_M

ClientSampleId :

SLCS975






Tgt Ion:196 Resp: 207684
Ion Ratio Lower Upper
196 100
198 95.9 78.0 117.0
200 31.2 24.5 36.7


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#43
1,1'-Biphenyl
Concen: 36.296 ng/ul
RT: 13.360 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:154 Resp: 738502
Ion Ratio Lower Upper
154 100
153 40.5 32.6 49.0
76 14.2 7.7 11.5#





#44
2-Chloronaphthalene
Concen: 36.194 ng/ul
RT: 13.401 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:162 Resp: 56537
Ion Ratio Lower Upper
162 100
164 32.9 26.6 40.0
127 40.4 29.6 44.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#45
2-Nitroaniline
Concen: 42.303 ng/ul
RT: 13.601 min Scan# 1796
Delta R.T. -0.006 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion: 65 Resp: 180184
Ion Ratio Lower Upper
65 100
92 64.9 57.4 86.2
138 90.0 94.8 142.2#





#47
Dimethylphthalate
Concen: 37.880 ng/ul
RT: 13.983 min Scan# 1861
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:163 Resp: 723475
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.2 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#48
2,6-Dinitrotoluene
Concen: 41.159 ng/ul
RT: 14.101 min Scan# 1881
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:165 Resp: 151019
Ion Ratio Lower Upper
165 100
89 62.4 44.2 66.4
121 22.5 16.7 25.1





#50
Acenaphthylene
Concen: 37.201 ng/ul
RT: 14.248 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:152 Resp: 942070
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.0 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#51
3-Nitroaniline
Concen: 39.099 ng/ul
RT: 14.430 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:138 Resp: 125427
Ion Ratio Lower Upper
138 100
108 11.8 8.3 12.5
92 122.0 87.0 130.4





#52
Acenaphthene
Concen: 36.716 ng/ul
RT: 14.589 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:153 Resp: 61903
Ion Ratio Lower Upper
153 100
152 46.2 37.4 56.0
154 87.7 71.0 106.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#53
2,4-Dinitrophenol
Concen: 37.852 ng/ul
RT: 14.636 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:184 Resp: 78271
Ion Ratio Lower Upper
184 100
63 82.9 43.8 65.6#
154 67.7 49.9 74.9





#55
4-Nitrophenol
Concen: 40.436 ng/ul
RT: 14.736 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:109 Resp: 10218
Ion Ratio Lower Upper
109 100
139 116.6 94.6 142.0
65 136.8 83.5 125.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#56
Dibenzofuran
Concen: 37.489 ng/ul
RT: 14.924 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:168 Resp: 859053
Ion Ratio Lower Upper
168 100
139 39.9 29.8 44.6





#57
2,4-Dinitrotoluene
Concen: 42.077 ng/ul
RT: 14.889 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:165 Resp: 214880
Ion Ratio Lower Upper
165 100
63 49.2 30.2 45.4
182 2.9 2.2 3.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 39.811 ng/ul
RT: 15.148 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:232 Resp: 176896
Ion Ratio Lower Upper
232 100
131 52.1 32.2 48.4
130 2.9 1.9 2.9
166 32.4 23.0 34.4





#59
Diethylphthalate
Concen: 39.251 ng/ul
RT: 15.348 min Scan# 2093
Delta R.T. 0.006 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:149 Resp: 761458
Ion Ratio Lower Upper
149 100
177 21.6 17.9 26.9
150 12.3 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#61
Fluorene
Concen: 37.195 ng/ul
RT: 15.571 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:166 Resp: 719113
Ion Ratio Lower Upper
166 100
165 96.8 76.9 115.3
167 13.6 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 37.603 ng/ul
RT: 15.566 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:204 Resp: 35091
Ion Ratio Lower Upper
204 100
206 31.7 25.5 38.3
141 60.4 44.2 66.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#63
4-Nitroaniline
Concen: 45.342 ng/ul m
RT: 15.589 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21



Tgt Ion:138 Resp: 128077
Ion Ratio Lower Upper
138 100
92 58.2 36.3 54.5#
108 86.0 68.9 103.3









#66
4,6-Dinitro-2-methylphenol
Concen: 37.147 ng/ul
RT: 15.648 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Instrument :
BNA_M
ClientSampleId :
SLCS975

Tgt Ion:198 Resp: 12280
Ion Ratio Lower Upper
198 100
182 4.8 3.8 5.6
77 34.2 18.5 27.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#67
N-Nitrosodiphenylamine
Concen: 38.017 ng/ul
RT: 15.777 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Tgt Ion:169 Resp: 615141
Ion Ratio Lower Upper
169 100
168 66.4 52.8 79.2
167 35.5 28.1 42.1






#68
4-Bromophenyl-phenylether
Concen: 37.783 ng/ul
RT: 16.460 min Scan# 211
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Instrument :

BNA_M

ClientSampleId :

SLCS975






Tgt Ion:248 Resp: 21211
Ion Ratio Lower Upper
248 100
250 97.9 78.4 117.6
141 74.5 51.4 77.2


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#69
Hexachlorobenzene
Concen: 37.427 ng/ul
RT: 16.577 min Scan# 2302
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:284 Resp: 252346
Ion Ratio Lower Upper
284 100
142 38.1 24.6 36.8#
249 31.4 24.1 36.1





#70
Atrazine
Concen: 37.211 ng/ul
RT: 16.730 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:200 Resp: 22531
Ion Ratio Lower Upper
200 100
173 26.2 20.2 30.4
215 48.3 38.4 57.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#71
Pentachlorophenol
Concen: 37.664 ng/ul
RT: 16.918 min Scan# 2360
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:266 Resp: 155495
Ion Ratio Lower Upper
266 100
264 63.6 49.3 73.9
268 65.5 50.3 75.5





#72
Phenanthrene
Concen: 37.590 ng/u1
RT: 17.307 min Scan# 2426
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:178 Resp: 113915
Ion Ratio Lower Upper
178 100
179 14.8 11.7 17.5
176 18.9 15.0 22.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#74
Anthracene
Concen: 37.902 ng/u1
RT: 17.401 min Scan# 2442
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:178 Resp: 1138172
Ion Ratio Lower Upper
178 100
179 15.3 11.9 17.9
176 18.2 14.4 21.6





#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.555 ng/uL
RT: 13.330 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:216 Resp: 303274
Ion Ratio Lower Upper
216 100
214 78.3 62.0 93.0
218 49.2 37.8 56.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#76
Pentachlorobenzene
Concen: 35.581 ng/uL
RT: 14.848 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:250 Resp: 284813
Ion Ratio Lower Upper
250 100
248 63.9 50.6 75.8
252 64.9 50.2 75.4





#77
 Carbazole
 Concen: 38.633 ng/ul
 RT: 17.671 min Scan# 2488
 Delta R.T. 0.000 min
 Lab File: BM034601.D
 Acq: 09 Apr 2022 16:21




Instrument :
 BNA_M
 ClientSampleId :
 SLCS975




Tgt Ion:167 Resp: 95971
 Ion Ratio Lower Upper
 167 100
 166 21.3 16.6 25.0
 139 13.1 9.9 14.9

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022




#78
 Di-n-butylphthalate
 Concen: 41.099 ng/ul
 RT: 18.236 min Scan# 2584
 Delta R.T. 0.000 min
 Lab File: BM034601.D
 Acq: 09 Apr 2022 16:21




Tgt Ion:149 Resp: 1295859
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 5.7 4.2 6.2





#80
Fluoranthene
Concen: 39.883 ng/u1
RT: 19.324 min Scan# 2769
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975



Tgt Ion:202 Resp: 1324825
Ion Ratio Lower Upper
202 100
101 11.2 6.2 9.2#
100 9.2 4.7 7.1#




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#82
Pyrene
Concen: 38.049 ng/u1
RT: 19.689 min Scan# 2831
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21


Tgt Ion:202 Resp: 1373749
Ion Ratio Lower Upper
202 100
101 13.2 7.0 10.4#
100 11.4 5.8 8.6#





#83
Butylbenzylphthalate
Concen: 41.884 ng/ul
RT: 20.583 min Scan# 2983
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:149 Resp: 592000
Ion Ratio Lower Upper
149 100
91 77.5 50.9 76.3
206 20.0 19.3 28.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#84
3,3'-Dichlorobenzidine
Concen: 37.075 ng/ul
RT: 21.359 min Scan# 3115
Delta R.T. -0.006 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:252 Resp: 357344
Ion Ratio Lower Upper
252 100
254 64.8 50.7 76.1
126 12.1 6.6 9.8





#85
Benzo(a)anthracene
Concen: 38.345 ng/ul
RT: 21.430 min Scan# 3126
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975



Tgt Ion:228 Resp: 122619
Ion Ratio Lower Upper
228 100
229 19.4 15.4 23.2
226 27.2 21.0 31.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 42.196 ng/ul
RT: 21.359 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21


Tgt Ion:149 Resp: 892628
Ion Ratio Lower Upper
149 100
167 26.7 22.1 33.1
279 3.8 4.2 6.2#





#87
Chrysene
Concen: 37.320 ng/ul
RT: 21.483 min Scan# 3135
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975



Tgt Ion:228 Resp: 128684
Ion Ratio Lower Upper
228 100
226 29.6 23.4 35.2
229 19.4 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#89
Di-n-octyl phthalate
Concen: 39.461 ng/ul
RT: 22.277 min Scan# 3271
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21


Tgt Ion:149 Resp: 1517653





#90
Benzo(b)fluoranthene
Concen: 38.961 ng/ul
RT: 23.100 min Scan# 3410
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:252 Resp: 1337424
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 10.2 5.8 8.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#91
Benzo(k)fluoranthene
Concen: 37.507 ng/ul
RT: 23.147 min Scan# 3419
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Tgt Ion:252 Resp: 1352088
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 10.2 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 38.804 ng/ul
RT: 23.724 min Scan# 3516
Delta R.T. 0.006 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21




Instrument :
BNA_M
ClientSampleId :
SLCS975




Tgt Ion:252 Resp: 1316508
Ion Ratio Lower Upper
252 100
253 21.3 17.5 26.3
125 10.9 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 39.817 ng/ul
RT: 26.306 min Scan# 3956
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21



Tgt Ion:276 Resp: 1445675
Ion Ratio Lower Upper
276 100
138 23.4 13.6 20.4
277 25.5 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 40.198 ng/ul
RT: 26.318 min Scan# 3958
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Instrument :
BNA_M
ClientSampleId :
SLCS975

Tgt Ion:278 Resp: 1196514
Ion Ratio Lower Upper
278 100
139 16.2 9.6 14.4
279 23.9 19.3 28.9

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#96
Benzo(g,h,i)perylene
Concen: 37.806 ng/ul
RT: 27.065 min Scan# 4085
Delta R.T. 0.000 min
Lab File: BM034601.D
Acq: 09 Apr 2022 16:21

Tgt Ion:276 Resp: 1317538
Ion Ratio Lower Upper
276 100
138 21.2 11.8 17.6
277 23.9 18.8 28.2

