

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM041119\
 Data File : BM019901.D
 Acq On : 11 Apr 2019 19:29
 Operator : JU/SJ
 Sample : K2304-11DL 5X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BF895DL

Manual Integrations
 APPROVED

Sohil
 4/12/2019 5:21:06 PM

Quant Time: Apr 12 05:37:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 12 04:12:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.54	152	129191	20.00	ng/ul	0.00
18) Naphthalene-d8	10.32	136	526995	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.20	164	288964	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.97	188	621697	20.00	ng/ul	0.00
78) Chrysene-d12	21.19	240	665709	20.00	ng/ul	0.00
86) Perylene-d12	23.39	264	748903	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.11	96	983	0.30	ng/uL	0.01
5) Phenol-d5	6.76	99	22405	1.95	ng/ul	0.02
7) Bis-(2-Chloroethyl)ether-d	6.90	67	17170	2.23	ng/ul	0.01
9) 2-Chlorophenol-d4	7.09	132	19262	2.21	ng/ul	0.02
13) 4-Methylphenol-d8	8.29	113	11578	1.34	ng/ul	0.03
19) Nitrobenzene-d5	8.73	128	8149	2.17	ng/ul	0.02
22) 2-Nitrophenol-d4	9.45	143	10274	2.46	ng/ul	0.02
26) 2,4-Dichlorophenol-d3	10.02	165	17005m	2.15	ng/ul	0.06
29) 4-Chloroaniline-d4	10.54	131	14219	1.50	ng/ul	0.05
44) Dimethylphthalate-d6	13.63	166	63499	2.72	ng/ul	0.00
47) Acenaphthylene-d8	13.90	160	73660	2.53	ng/ul	0.00
52) 4-Nitrophenol-d4	14.63	143	3394m	0.93	ng/ul	0.15
58) Fluorene-d10	15.22	176	53387	2.66	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.41	200	1706	0.41	ng/ul	0.04
71) Anthracene-d10	17.07	188	77969	2.61	ng/ul	0.00
79) Pyrene-d10	19.39	212	86623	2.82	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.24	264	94872	2.39	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM041119\
Data File : BM019901.D
Acq On : 11 Apr 2019 19:29
Operator : JU/SJ
Sample : K2304-11DL 5X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BF895DL

Manual Integrations
APPROVED

Sohil
4/12/2019 5:21:06 PM

Quant Time: Apr 12 05:37:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 12 04:12:05 2019
Response via : Initial Calibration

