




Quant Time: Apr 14 13:33:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM040825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 09 04:00:55 2025
Response via : Initial Calibration



TIC: BM049909.D\data.ms

(70) Pentachlorophenol (C)




16.815min (-0.000) 254292.86 ng

response **630700**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.30	64.41
264.00	62.30	62.11
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041425\
Data File : BM049909.D
Acq On : 14 Apr 2025 11:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 14 13:33:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM040825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 09 04:00:55 2025
Response via : Initial Calibration



TIC: BM049909.D\data.ms

(77) Benzidine**19.403min (-0.000) 101434.92 ng****response 3159652**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.30	13.80
183.00	11.00	10.89
0.00	0.00	0.00