

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_M\DATA\BM041519\  
 Data File : BM019920.D  
 Acq On : 15 Apr 2019 19:11  
 Operator : JU/SJ  
 Sample : PB118232BL  
 Misc :  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 PB118232BL

Quant Time: Apr 16 03:34:34 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\8270-BM041519.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Apr 15 16:16:29 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.53  | 152  | 115482   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.32 | 136  | 427334   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.20 | 164  | 248544   | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 16.97 | 188  | 523132   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.18 | 240  | 442816   | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 23.38 | 264  | 427194   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.15  | 112  | 886661   | 124.43 | ng    | 0.00     |
| 7) Phenol-d6                | 6.73  | 99   | 1205702  | 112.43 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 8.70  | 82   | 724097   | 75.78  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 15.71 | 330  | 425983   | 106.47 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 12.80 | 172  | 1523144  | 76.37  | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.61 | 244  | 1972095  | 78.55  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_M\DATA\BM041519\  
Data File : BM019920.D  
Acq On : 15 Apr 2019 19:11  
Operator : JU/SJ  
Sample : PB118232BL  
Misc :  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
PB118232BL

Quant Time: Apr 16 03:34:34 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\8270-BM041519.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Apr 15 16:16:29 2019  
Response via : Initial Calibration

