

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM041520\
 Data File : BM025960.D
 Acq On : 15 Apr 2020 11:15
 Operator : MA/SJ
 Sample : SSTD1.678
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD1.678

Quant Time: Apr 15 11:57:54 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 15 11:17:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	1079	0.40	ng/ul	0.00
2) Naphthalene-d8	10.32	136	3881	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.19	164	2394	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.94	188	5332	0.40	ng/ul	0.00
16) Chrysene-d12	21.13	240	5031	0.40	ng/ul	0.00
20) Perylene-d12	23.28	264	4092	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	11.92	152	11191	1.62	ng/ul	0.00
14) Fluoranthene-d10	18.97	212	27043	1.62	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Naphthalene	10.37	128	20750	1.818	ng/ul	97
5) 2-Methylnaphthalene	11.99	142	14202	1.779	ng/ul	98
7) Acenaphthylene	13.91	152	19463	1.901	ng/ul	98
8) Acenaphthene	14.25	153	16030	1.887	ng/ul	99
9) Fluorene	15.25	166	19318	1.833	ng/ul	99
11) Pentachlorophenol	16.60	266	2554	1.951	ng/ul	94
12) Phenanthrene	16.98	178	31608	1.882	ng/ul	98
13) Anthracene	17.07	178	28401	1.914	ng/ul	97
15) Fluoranthene	19.00	202	38664	1.863	ng/ul	98
17) Pyrene	19.37	202	37929	2.105	ng/ul	98
18) Benzo(a)anthracene	21.12	228	34172	1.900	ng/ul	98
19) Chrysene	21.17	228	38074	1.940	ng/ul	99
21) Benzo(b)fluoranthene	22.64	252	33210	2.205	ng/ul	92
22) Benzo(k)fluoranthene	22.68	252	35746	2.214	ng/ul#	91
23) Benzo(a)pyrene	23.19	252	27518	2.045	ng/ul#	89
24) Indeno(1,2,3-cd)pyrene	25.40	276	32089	1.780	ng/ul#	94
25) Dibenzo(a,h)anthracene	25.41	278	25485	1.694	ng/ul#	89
26) Benzo(g,h,i)perylene	26.04	276	27901	1.830	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM041520\
 Data File : BM025960.D
 Acq On : 15 Apr 2020 11:15
 Operator : MA/SJ
 Sample : SSTD1.678
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 Client Sampled :
 SSTD1.678

Quant Time: Apr 15 11:57:54 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 15 11:17:24 2020
 Response via : Initial Calibration

